

Appendix G

2004/05 QA/QC Results: Matrix Spike RPD Success Rates

<i>Event ID</i>	<i>Classification</i>	<i>Method</i>	<i>Total Number</i>	<i>Number Outside DQO</i>	<i>Success Rate</i>
2004/05-1	Metal	EPA 1631E	2	0	100
2004/05-1	Metal	EPA 200.8	10	0	100
2004/05-1	Metal	SM 3500-Cr	1	0	100
2004/05-1	Nutrient	SM 4500-NH3 F	1	0	100
2004/05-1	Nutrient	SM 4500-NO2 B	1	0	100
2004/05-1	Nutrient	SM 4500-NO3 E	1	0	100
2004/05-1	Nutrient	SM 4500-P C	2	1	50
2004/05-1	Organic	EPA 625	40	1	97.5
2004/05-1	PCB	EPA 625	46	0	100
2004/05-1	Pesticide	EPA 625	43	0	100
2004/05-1	Pesticide	EPA 8151A	3	0	100
2004/05-2	Anion	EPA 314.0	1	0	100
2004/05-2	Anion	SM 4500-Cl E	1	0	100
2004/05-2	Metal	EPA 1631E	2	0	100
2004/05-2	Metal	EPA 200.8	11	0	100
2004/05-2	Nutrient	EPA 351.3	1	0	100
2004/05-2	Nutrient	SM 4500-NH3 F	1	0	100
2004/05-2	Nutrient	SM 4500-NO2 B	1	0	100
2004/05-2	Nutrient	SM 4500-NO3 E	1	0	100
2004/05-2	Nutrient	SM 4500-P C	2	0	100
2004/05-2	Organic	EPA 625	39	0	100
2004/05-2	PCB	EPA 625	46	0	100
2004/05-2	Pesticide	EPA 625	43	0	100
2004/05-2	Pesticide	EPA 8151A	3	0	100
2004/05-3	Anion	SM 4500-Cl E	1	0	100
2004/05-3	Conventional	EPA 415.1	1	0	100
2004/05-3	Metal	EPA 1631E	1	0	100
2004/05-3	Metal	EPA 200.8	11	0	100
2004/05-3	Metal	SM 3500-Cr	1	0	100
2004/05-3	Nutrient	EPA 351.3	1	0	100
2004/05-3	Nutrient	SM 4500-NO3 E	1	0	100
2004/05-3	Nutrient	SM 4500-P C	3	0	100
2004/05-3	Organic	EPA 625	39	0	100
2004/05-3	PCB	EPA 625	46	0	100

Appendix G

2004/05 QA/QC Results: Matrix Spike RPD Success Rates

<i>Event ID</i>	<i>Classification</i>	<i>Method</i>	<i>Total Number</i>	<i>Number Outside DQO</i>	<i>Success Rate</i>
2004/05-3	Pesticide	EPA 625	43	0	100
2004/05-3	Pesticide	EPA 8151A	3	0	100
2004/05-4	Anion	SM 4500-CI E	1	0	100
2004/05-4	Metal	EPA 1631E	1	0	100
2004/05-4	Metal	EPA 200.8	11	0	100
2004/05-4	Metal	SM 3500-Cr	2	0	100
2004/05-4	Nutrient	SM 4500-NO2 B	1	0	100
2004/05-4	Nutrient	SM 4500-NO3 E	1	0	100
2004/05-4	Nutrient	SM 4500-P C	3	0	100
2004/05-4	Organic	EPA 625	38	22	42.1
2004/05-4	PCB	EPA 625	46	2	95.7
2004/05-4	Pesticide	EPA 625	42	1	97.6
2004/05-4	Pesticide	EPA 8151A	3	0	100
2004/05-5	Anion	EPA 314.0	1	0	100
2004/05-5	Anion	SM 4500-CI E	1	0	100
2004/05-5	Conventional	EPA 415.1	1	0	100
2004/05-5	Metal	EPA 1631E	1	0	100
2004/05-5	Metal	EPA 200.8	11	2	81.8
2004/05-5	Metal	SM 3500-Cr	1	0	100
2004/05-5	Nutrient	EPA 300.0	3	0	100
2004/05-5	Nutrient	SM 4500-NH3 F	1	0	100
2004/05-5	Nutrient	SM 4500-P C	2	0	100
2004/05-5	Organic	EPA 625	38	0	100
2004/05-5	PCB	EPA 625	46	0	100
2004/05-5	Pesticide	EPA 625	44	0	100
2004/05-5	Pesticide	EPA 8151A	3	0	100
2004/05-6	Anion	SM 4500-CI E	1	0	100
2004/05-6	Metal	EPA 1631E	1	0	100
2004/05-6	Metal	EPA 200.8	11	0	100
2004/05-6	Metal	SM 3500-Cr	1	0	100
2004/05-6	Nutrient	EPA 300.0	3	0	100
2004/05-6	Nutrient	SM 4500-NH3 F	1	0	100
2004/05-6	Organic	EPA 625	39	0	100
2004/05-6	PCB	EPA 625	46	0	100

Appendix G

2004/05 QA/QC Results: Matrix Spike RPD Success Rates

<i>Event ID</i>	<i>Classification</i>	<i>Method</i>	<i>Total Number</i>	<i>Number Outside DQO</i>	<i>Success Rate</i>
2004/05-6	Pesticide	EPA 625	43	0	100