

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Anion	Bromide	n/a	=	11.3	%	SM 4500-Br		0	30	
2005/06-1	Lab	method blank	11/15/2005	Anion	Bromide	n/a	<	0.001	mg/L	SM 4500-Br	0.001		0.001	
2005/06-1	ME-VR2	field blank	11/15/2005	Anion	Bromide	n/a	<	0.001	mg/L	SM 4500-Br	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Anion	Chloride	n/a	=	0.8	%	SM 4500-Cl E		0	30	
2005/06-1	Lab	LCS dup, rec	10/20/2005	Anion	Chloride	n/a	=	101	%	SM 4500-Cl E		70	130	
2005/06-1	Lab	LCS, rec	10/20/2005	Anion	Chloride	n/a	=	98	%	SM 4500-Cl E		70	130	
2005/06-1	Lab	LCS, RPD	10/20/2005	Anion	Chloride	n/a	=	3	%	SM 4500-Cl E		0	30	
2005/06-1	Lab	method blank	10/26/2005	Anion	Chloride	n/a	<	0.01	mg/L	SM 4500-Cl E	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	10/26/2005	Anion	Perchlorate	n/a	=	0	%	EPA 314.0		0	25	
2005/06-1	Lab	LCS dup, rec	10/25/2005	Anion	Perchlorate	n/a	=	103	%	EPA 314.0		85	115	
2005/06-1	Lab	LCS, rec	10/25/2005	Anion	Perchlorate	n/a	=	103	%	EPA 314.0		85	115	
2005/06-1	Lab	LCS, RPD	10/25/2005	Anion	Perchlorate	n/a	=	0	%	EPA 314.0		0	15	
2005/06-1	Lab	method blank	10/25/2005	Anion	Perchlorate	n/a	<	2	µg/L	EPA 314.0	2		2	
2005/06-1	A-1	lab duplicate, RPD	10/17/2005	Bacteriological	E. Coli	n/a	=	6.2	%	MMO-MUG		0	30	
2005/06-1	ME-VR2	field blank	10/17/2005	Bacteriological	E. Coli	n/a	<	10	MPN/100 mL	MMO-MUG	10		10	
2005/06-1	A-1	lab duplicate, RPD	10/17/2005	Bacteriological	Enterococcus	n/a	=	5.4	%	Enterolert		0	30	
2005/06-1	ME-VR2	field blank	10/17/2005	Bacteriological	Enterococcus	n/a	<	10	MPN/100 mL	Enterolert	10		10	
2005/06-1	A-1	lab duplicate, RPD	10/17/2005	Bacteriological	Fecal Coliform	n/a	=	12.7	%	SM 9221 E		0	30	
2005/06-1	ME-VR2	field blank	10/17/2005	Bacteriological	Fecal Coliform	n/a	<	2	MPN/100 mL	SM 9221 E	2		2	
2005/06-1	A-1	lab duplicate, RPD	10/17/2005	Bacteriological	Total Coliform	n/a	=	0.6	%	MMO-MUG		0	30	
2005/06-1	ME-VR2	field blank	10/17/2005	Bacteriological	Total Coliform	n/a	<	10	MPN/100 mL	MMO-MUG	10		10	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Conventional	BOD	n/a	=	22.2	%	EPA 405.1		0	25	
2005/06-1	Lab	method blank	10/20/2005	Conventional	BOD	n/a	<	1	mg/L	EPA 405.1	1		1	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Conventional	Conductivity	n/a	=	0	%	SM 2510		0	30	
2005/06-1	Lab	method blank	10/21/2005	Conventional	Conductivity	n/a	<	1	µmhos/cm	SM 2510	1		1	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Conventional	Hardness as CaCO3	Total	=	1.4	%	SM 2340 B		0	30	
2005/06-1	Lab	method blank	11/3/2005	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1		1	
2005/06-1	ME-VR2	field blank	11/3/2005	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1		1	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Conventional	pH	n/a	=	0.8	%	EPA 150.1		0	30	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Conventional	Total Dissolved Solids	n/a	=	3.4	%	SM 2540 C		0	30	
2005/06-1	Lab	LCS dup, rec	10/20/2005	Conventional	Total Dissolved Solids	n/a	=	101	%	SM 2540 C		70	130	
2005/06-1	Lab	LCS, rec	10/20/2005	Conventional	Total Dissolved Solids	n/a	=	101	%	SM 2540 C		70	130	
2005/06-1	Lab	LCS, RPD	10/20/2005	Conventional	Total Dissolved Solids	n/a	=	0	%	SM 2540 C		0	30	
2005/06-1	Lab	method blank	10/24/2005	Conventional	Total Dissolved Solids	n/a	<	0.1	mg/L	SM 2540 C	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	10/27/2005	Conventional	Total Organic Carbon	n/a	=	13.3	%	EPA 415.1		0	25	
2005/06-1	Lab	LCS, rec	10/25/2005	Conventional	Total Organic Carbon	n/a	=	94	%	EPA 415.1		80	120	
2005/06-1	Lab	method blank	10/27/2006	Conventional	Total Organic Carbon	n/a	<	0.5	mg/L	EPA 415.1	0.5		0.5	
2005/06-1	ME-CC	matrix spike dup, rec	10/25/2005	Conventional	Total Organic Carbon	n/a	=	91	%	EPA 415.1		70	130	
2005/06-1	ME-CC	matrix spike, rec	10/25/2005	Conventional	Total Organic Carbon	n/a	=	92	%	EPA 415.1		70	130	
2005/06-1	ME-CC	matrix spike, RPD	10/25/2005	Conventional	Total Organic Carbon	n/a	=	1.1	%	EPA 415.1		0	25	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Conventional	Total Suspended Solids	n/a	=	9.1	%	SM 2540 D		0	30	
2005/06-1	Lab	method blank	10/21/2005	Conventional	Total Suspended Solids	n/a	<	0.5	mg/L	SM 2540 D	0.5		0.5	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Conventional	Turbidity	n/a	=	5.2	%	EPA 180.1		0	30	
2005/06-1	Lab	method blank	10/21/2005	Conventional	Turbidity	n/a	<	1	NTU	EPA 180.1	1		1	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Hydrocarbon	Oil and Grease	n/a	=	0	%	EPA 1664A		0	30	
2005/06-1	Lab	LCS dup, rec	10/20/2005	Hydrocarbon	Oil and Grease	n/a	=	94	%	EPA 1664A		70	130	
2005/06-1	Lab	LCS, rec	10/20/2005	Hydrocarbon	Oil and Grease	n/a	=	96	%	EPA 1664A		70	130	
2005/06-1	Lab	LCS, RPD	10/20/2005	Hydrocarbon	Oil and Grease	n/a	=	2	%	EPA 1664A		0	30	
2005/06-1	Lab	method blank	10/27/2005	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1		1	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Hydrocarbon	TRPH	n/a	=	0	%	EPA 418.1		0	30	
2005/06-1	Lab	LCS dup, rec	10/20/2005	Hydrocarbon	TRPH	n/a	=	88	%	EPA 418.1		70	130	
2005/06-1	Lab	LCS, rec	10/20/2005	Hydrocarbon	TRPH	n/a	=	91	%	EPA 418.1		70	130	
2005/06-1	Lab	LCS, RPD	10/20/2005	Hydrocarbon	TRPH	n/a	=	3	%	EPA 418.1		0	30	
2005/06-1	Lab	method blank	11/1/2005	Hydrocarbon	TRPH	n/a	<	0.01	mg/L	EPA 418.1	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Aluminum	Dissolved	=	35.3	%	EPA 200.8m		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Aluminum	Total	=	119.7	%	EPA 200.8m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	Lab	LCS dup, rec	11/7/2005	Metal	Aluminum	Total	=	102	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, rec	11/7/2005	Metal	Aluminum	Total	=	100	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, RPD	11/7/2005	Metal	Aluminum	Total	=	2	%	EPA 200.8m		0	30	
2005/06-1	Lab	method blank	11/7/2005	Metal	Aluminum	Total	<	1	µg/L	EPA 200.8m	1		1	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Aluminum	Total	=	94	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Aluminum	Total	=	97	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Aluminum	Total	=	3	%	EPA 200.8m		0	30	
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Aluminum	Total	<	1	µg/L	EPA 200.8m	1		1	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Arsenic	Dissolved	=	5	%	EPA 200.8m		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Arsenic	Total	=	5.6	%	EPA 200.8m		0	30	
2005/06-1	Lab	LCS dup, rec	11/7/2005	Metal	Arsenic	Total	=	102	%	EPA 200.8m		65	135	
2005/06-1	Lab	LCS, rec	11/7/2005	Metal	Arsenic	Total	=	101	%	EPA 200.8m		65	135	
2005/06-1	Lab	LCS, RPD	11/7/2005	Metal	Arsenic	Total	=	1	%	EPA 200.8m		0	30	
2005/06-1	Lab	method blank	11/7/2005	Metal	Arsenic	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Arsenic	Total	=	104	%	EPA 200.8m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Arsenic	Total	=	105	%	EPA 200.8m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Arsenic	Total	=	1	%	EPA 200.8m		0	30	
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Arsenic	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Cadmium	Dissolved	=	17.9	%	EPA 200.8m		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Cadmium	Total	=	32.1	%	EPA 200.8m		0	30	
2005/06-1	Lab	LCS dup, rec	11/7/2005	Metal	Cadmium	Total	=	101	%	EPA 200.8m		60	140	
2005/06-1	Lab	LCS, rec	11/7/2005	Metal	Cadmium	Total	=	101	%	EPA 200.8m		60	140	
2005/06-1	Lab	LCS, RPD	11/7/2005	Metal	Cadmium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	Lab	method blank	11/7/2005	Metal	Cadmium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Cadmium	Total	=	111	%	EPA 200.8m		60	140	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Cadmium	Total	=	111	%	EPA 200.8m		60	140	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Cadmium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Cadmium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Chromium	Dissolved	=	0.2	%	EPA 200.8m		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Chromium	Total	=	80.6	%	EPA 200.8m		0	30	
2005/06-1	Lab	LCS dup, rec	11/7/2005	Metal	Chromium	Total	=	101	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, rec	11/7/2005	Metal	Chromium	Total	=	100	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, RPD	11/7/2005	Metal	Chromium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-1	Lab	method blank	11/7/2005	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Chromium	Total	=	89	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Chromium	Total	=	89	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Chromium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Metal	Chromium VI	Total	=	18.2	%	SM 3500-Cr		0	30	
2005/06-1	Lab	LCS dup, rec	10/20/2005	Metal	Chromium VI	Total	=	101	%	SM 3500-Cr		70	130	
2005/06-1	Lab	LCS, rec	10/20/2005	Metal	Chromium VI	Total	=	103	%	SM 3500-Cr		70	130	
2005/06-1	Lab	LCS, RPD	10/20/2005	Metal	Chromium VI	Total	=	2	%	SM 3500-Cr		0	30	
2005/06-1	Lab	method blank	10/20/2005	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr	5		5	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Copper	Dissolved	=	1.6	%	EPA 200.8m		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Copper	Total	=	33.9	%	EPA 200.8m		0	30	
2005/06-1	Lab	LCS dup, rec	11/7/2005	Metal	Copper	Total	=	101	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, rec	11/7/2005	Metal	Copper	Total	=	100	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, RPD	11/7/2005	Metal	Copper	Total	=	1	%	EPA 200.8m		0	30	
2005/06-1	Lab	method blank	11/7/2005	Metal	Copper	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Copper	Total	=	89	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Copper	Total	=	88	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Copper	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Copper	Total	=	0.21	µg/L	EPA 200.8m	0.1		0.1	EST
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Lead	Dissolved	=	3.8	%	EPA 200.8m		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Lead	Total	=	33.5	%	EPA 200.8m		0	30	
2005/06-1	Lab	LCS dup, rec	11/7/2005	Metal	Lead	Total	=	101	%	EPA 200.8m		75	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	Lab	LCS, rec	11/7/2005	Metal	Lead	Total	=	101	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, RPD	11/7/2005	Metal	Lead	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	Lab	method blank	11/7/2005	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Lead	Total	=	112	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Lead	Total	=	112	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Lead	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Lead	Total	=	0.15	µg/L	EPA 200.8m	0.05		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Mercury	Dissolved	=	13	%	EPA 1631E		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Mercury	Total	=	0.7	%	EPA 1631E		0	30	
2005/06-1	Lab	method blank	10/19/2005	Metal	Mercury	Total	<	0.5	ng/L	EPA 1631E	0.5		500	
2005/06-1	Lab	LCS dup, rec	11/4/2005	Metal	Mercury	Total	=	102	%	EPA 1631E		75	125	
2005/06-1	Lab	LCS, rec	11/4/2005	Metal	Mercury	Total	=	97	%	EPA 1631E		75	125	
2005/06-1	Lab	LCS, RPD	11/4/2005	Metal	Mercury	Total	=	5	%	EPA 1631E		0	30	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Mercury	Total	=	119	%	EPA 1631E		75	125	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Mercury	Total	=	106	%	EPA 1631E		75	125	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Mercury	Total	=	11.6	%	EPA 1631E		0	30	
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Mercury	Total	=	1.6	ng/L	EPA 1631E	0.5		0.5	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Nickel	Dissolved	=	0.4	%	EPA 200.8m		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Nickel	Total	=	14.4	%	EPA 200.8m		0	30	
2005/06-1	Lab	LCS dup, rec	11/7/2005	Metal	Nickel	Total	=	101	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, rec	11/7/2005	Metal	Nickel	Total	=	100	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, RPD	11/7/2005	Metal	Nickel	Total	=	1	%	EPA 200.8m		0	30	
2005/06-1	Lab	method blank	11/7/2005	Metal	Nickel	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Nickel	Total	=	83	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Nickel	Total	=	83	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Nickel	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Nickel	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Selenium	Dissolved	=	0.6	%	EPA 200.8m		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Selenium	Total	=	5.9	%	EPA 200.8m		0	30	
2005/06-1	Lab	LCS dup, rec	11/7/2005	Metal	Selenium	Total	=	101	%	EPA 200.8m		40	160	
2005/06-1	Lab	LCS, rec	11/7/2005	Metal	Selenium	Total	=	100	%	EPA 200.8m		40	160	
2005/06-1	Lab	LCS, RPD	11/7/2005	Metal	Selenium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-1	Lab	method blank	11/7/2005	Metal	Selenium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Selenium	Total	=	114	%	EPA 200.8m		40	160	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Selenium	Total	=	114	%	EPA 200.8m		40	160	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Selenium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Selenium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Silver	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Silver	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	Lab	LCS dup, rec	11/7/2005	Metal	Silver	Total	=	103	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, rec	11/7/2005	Metal	Silver	Total	=	100	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, RPD	11/7/2005	Metal	Silver	Total	=	3	%	EPA 200.8m		0	30	
2005/06-1	Lab	method blank	11/7/2005	Metal	Silver	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Silver	Total	=	109	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Silver	Total	=	100	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Silver	Total	=	9	%	EPA 200.8m		0	30	
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Silver	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Thallium	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Thallium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	Lab	LCS dup, rec	11/7/2005	Metal	Thallium	Total	=	100	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, rec	11/7/2005	Metal	Thallium	Total	=	100	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, RPD	11/7/2005	Metal	Thallium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	Lab	method blank	11/7/2005	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Thallium	Total	=	109	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Thallium	Total	=	108	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Thallium	Total	=	1	%	EPA 200.8m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Zinc	Dissolved	=	0.5	%	EPA 200.8m		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/7/2005	Metal	Zinc	Total	=	39.5	%	EPA 200.8m		0	30	
2005/06-1	Lab	LCS dup, rec	11/7/2005	Metal	Zinc	Total	=	100	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, rec	11/7/2005	Metal	Zinc	Total	=	100	%	EPA 200.8m		75	125	
2005/06-1	Lab	LCS, RPD	11/7/2005	Metal	Zinc	Total	=	0	%	EPA 200.8m		0	30	
2005/06-1	Lab	method blank	11/7/2005	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/7/2005	Metal	Zinc	Total	=	98	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, rec	11/7/2005	Metal	Zinc	Total	=	97	%	EPA 200.8m		75	125	
2005/06-1	ME-CC	matrix spike, RPD	11/7/2005	Metal	Zinc	Total	=	1	%	EPA 200.8m		0	30	
2005/06-1	ME-VR2	field blank	11/7/2005	Metal	Zinc	Total	=	0.21	µg/L	EPA 200.8m	0.1		0.1	EST
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Nutrient	Ammonia as N	n/a	=	0	%	SM 4500-NH3 F		0	30	
2005/06-1	A-1	matrix spike dup, rec	10/20/2005	Nutrient	Ammonia as N	n/a	=	122	%	SM 4500-NH3 F		70	130	
2005/06-1	A-1	matrix spike, rec	10/20/2005	Nutrient	Ammonia as N	n/a	=	120	%	SM 4500-NH3 F		70	130	
2005/06-1	A-1	matrix spike, RPD	10/20/2005	Nutrient	Ammonia as N	n/a	=	1.7	%	SM 4500-NH3 F		0	30	
2005/06-1	Lab	LCS dup, rec	10/20/2005	Nutrient	Ammonia as N	n/a	=	108	%	SM 4500-NH3 F		70	130	
2005/06-1	Lab	LCS, rec	10/20/2005	Nutrient	Ammonia as N	n/a	=	100	%	SM 4500-NH3 F		70	130	
2005/06-1	Lab	LCS, RPD	10/20/2005	Nutrient	Ammonia as N	n/a	=	8	%	SM 4500-NH3 F		0	30	
2005/06-1	Lab	method blank	10/20/2005	Nutrient	Ammonia as N	n/a	<	0.01	mg/L	SM 4500-NH3 F	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Nutrient	Nitrate as N	n/a	=	0	%	EPA 300.0		0	30	
2005/06-1	Lab	LCS dup, rec	10/20/2005	Nutrient	Nitrate as N	n/a	=	97	%	EPA 300.0		70	130	
2005/06-1	Lab	LCS, rec	10/20/2005	Nutrient	Nitrate as N	n/a	=	98	%	EPA 300.0		70	130	
2005/06-1	Lab	LCS, RPD	10/20/2005	Nutrient	Nitrate as N	n/a	=	1	%	EPA 300.0		0	30	
2005/06-1	Lab	method blank	10/20/2005	Nutrient	Nitrate as N	n/a	<	0.02	mg/L	EPA 300.0	0.02		0.02	
2005/06-1	ME-CC	matrix spike dup, rec	10/20/2005	Nutrient	Nitrate as N	n/a	=	98	%	EPA 300.0		70	130	
2005/06-1	ME-CC	matrix spike, rec	10/20/2005	Nutrient	Nitrate as N	n/a	=	98	%	EPA 300.0		70	130	
2005/06-1	ME-CC	matrix spike, RPD	10/20/2005	Nutrient	Nitrate as N	n/a	=	2	%	EPA 300.0		0	30	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Nutrient	Nitrite as N	n/a	=	1.9	%	EPA 300.0		0	30	
2005/06-1	Lab	LCS dup, rec	10/20/2005	Nutrient	Nitrite as N	n/a	=	91	%	EPA 300.0		70	130	
2005/06-1	Lab	LCS, rec	10/20/2005	Nutrient	Nitrite as N	n/a	=	98	%	EPA 300.0		70	130	
2005/06-1	Lab	LCS, RPD	10/20/2005	Nutrient	Nitrite as N	n/a	=	7	%	EPA 300.0		0	30	
2005/06-1	Lab	method blank	10/20/2005	Nutrient	Nitrite as N	n/a	<	0.02	mg/L	EPA 300.0	0.02		0.02	
2005/06-1	ME-CC	matrix spike dup, rec	10/20/2005	Nutrient	Nitrite as N	n/a	=	94	%	EPA 300.0		70	130	
2005/06-1	ME-CC	matrix spike, rec	10/20/2005	Nutrient	Nitrite as N	n/a	=	94	%	EPA 300.0		70	130	
2005/06-1	ME-CC	matrix spike, RPD	10/20/2005	Nutrient	Nitrite as N	n/a	=	0	%	EPA 300.0		0	30	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	0.4	%	EPA 300.0		0	30	
2005/06-1	Lab	LCS dup, rec	10/20/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	94	%	EPA 300.0		70	130	
2005/06-1	Lab	LCS, rec	10/20/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	94	%	EPA 300.0		70	130	
2005/06-1	Lab	LCS, RPD	10/20/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	0	%	EPA 300.0		0	30	
2005/06-1	Lab	method blank	10/20/2005	Nutrient	Orthophosphate as P (Diss)	n/a	<	0.0075	mg/L	EPA 300.0	0.0075		0.0075	
2005/06-1	ME-CC	matrix spike dup, rec	10/20/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	93	%	EPA 300.0		70	130	
2005/06-1	ME-CC	matrix spike, rec	10/20/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	94	%	EPA 300.0		70	130	
2005/06-1	ME-CC	matrix spike, RPD	10/20/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	1.1	%	EPA 300.0		0	30	
2005/06-1	A-1	lab duplicate, RPD	10/27/2005	Nutrient	TKN	n/a	=	15.4	%	EPA 351.3		0	25	
2005/06-1	Lab	method blank	10/27/2005	Nutrient	TKN	n/a	<	0.5	mg/L	EPA 351.3	0.5		0.5	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Nutrient	Total Phosphorus	Dissolved	=	20.3	%	SM 4500-P C		0	30	
2005/06-1	Lab	LCS dup, rec	10/20/2005	Nutrient	Total Phosphorus	Dissolved	=	98	%	SM 4500-P C		70	130	
2005/06-1	Lab	LCS, rec	10/20/2005	Nutrient	Total Phosphorus	Dissolved	=	98	%	SM 4500-P C		70	130	
2005/06-1	Lab	LCS, RPD	10/20/2005	Nutrient	Total Phosphorus	Dissolved	=	0	%	SM 4500-P C		0	30	
2005/06-1	Lab	method blank	10/21/2005	Nutrient	Total Phosphorus	Dissolved	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2005/06-1	A-1	lab duplicate, RPD	10/20/2005	Nutrient	Total Phosphorus	Total	=	15.7	%	SM 4500-P C		0	30	
2005/06-1	Lab	LCS dup, rec	10/20/2005	Nutrient	Total Phosphorus	Total	=	98	%	SM 4500-P C		70	130	
2005/06-1	Lab	LCS, rec	10/20/2005	Nutrient	Total Phosphorus	Total	=	98	%	SM 4500-P C		70	130	
2005/06-1	Lab	LCS, RPD	10/20/2005	Nutrient	Total Phosphorus	Total	=	0	%	SM 4500-P C		0	30	
2005/06-1	Lab	method blank	10/21/2005	Nutrient	Total Phosphorus	Total	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	1,2,4-Trichlorobenzene	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	Lab	method blank	11/3/2005	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	1,2,4-Trichlorobenzene	n/a	=	109	%	EPA 625m		44	142	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	1,2,4-Trichlorobenzene	n/a	=	109	%	EPA 625m		44	142	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	1,2,4-Trichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	1,2-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	srgt environ, rec	10/25/2005	Organic	1,2-Dichloroethane-d4	n/a	=	120	%	EPA 8260B		74	146	
2005/06-1	Lab	srgt method blank, rec	10/25/2005	Organic	1,2-Dichloroethane-d4	n/a	=	117	%	EPA 8260B		74	146	
2005/06-1	W-3	srgt environ, rec	10/25/2005	Organic	1,2-Dichloroethane-d4	n/a	=	118	%	EPA 8260B		74	146	
2005/06-1	W-4	srgt environ, rec	10/25/2005	Organic	1,2-Dichloroethane-d4	n/a	=	119	%	EPA 8260B		74	146	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	1,3-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	srgt environ, rec	10/25/2005	Organic	1,4-Bromofluorobenzene	n/a	=	101	%	EPA 8260B		74	110	
2005/06-1	Lab	srgt method blank, rec	10/25/2005	Organic	1,4-Bromofluorobenzene	n/a	=	103	%	EPA 8260B		74	110	
2005/06-1	W-3	srgt environ, rec	10/25/2005	Organic	1,4-Bromofluorobenzene	n/a	=	102	%	EPA 8260B		74	110	
2005/06-1	W-4	srgt environ, rec	10/25/2005	Organic	1,4-Bromofluorobenzene	n/a	=	102	%	EPA 8260B		74	110	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	1,4-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	1,4-Dichlorobenzene	n/a	=	99	%	EPA 625m		20	124	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	1,4-Dichlorobenzene	n/a	=	87	%	EPA 625m		20	124	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	1,4-Dichlorobenzene	n/a	=	13	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	1-Methylnaphthalene	n/a	=	75	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	1-Methylnaphthalene	n/a	=	93	%	EPA 625m		50	120	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	1-Methylnaphthalene	n/a	=	82	%	EPA 625m		50	120	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	1-Methylnaphthalene	n/a	=	13	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	1-Methylphenanthrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	1-Methylphenanthrene	n/a	=	121	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	1-Methylphenanthrene	n/a	=	109	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	1-Methylphenanthrene	n/a	=	10	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	=	93	%	EPA 625m		60	120	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	=	79	%	EPA 625m		60	120	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	=	16	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	srgt environ, rec	11/3/2005	Organic	2,4,6-Tribromophenol	n/a	=	90	%	EPA 625m		11	162	
2005/06-1	Lab	srgt method blank, rec	11/3/2005	Organic	2,4,6-Tribromophenol	n/a	=	86	%	EPA 625m		11	162	
2005/06-1	ME-CC	srgt environ, rec	11/3/2005	Organic	2,4,6-Tribromophenol	n/a	=	82	%	EPA 625m		11	162	
2005/06-1	ME-CC	srgt matrix spike dup, rec	11/3/2005	Organic	2,4,6-Tribromophenol	n/a	=	90	%	EPA 625m		11	162	
2005/06-1	ME-CC	srgt matrix spike, rec	11/3/2005	Organic	2,4,6-Tribromophenol	n/a	=	89	%	EPA 625m		11	162	
2005/06-1	ME-SCR	srgt environ, rec	11/3/2005	Organic	2,4,6-Tribromophenol	n/a	=	88	%	EPA 625m		11	162	
2005/06-1	ME-VR2	srgt environ, rec	11/3/2005	Organic	2,4,6-Tribromophenol	n/a	=	53	%	EPA 625m		11	162	
2005/06-1	ME-VR2	srgt field blank, rec	11/3/2005	Organic	2,4,6-Tribromophenol	n/a	=	40	%	EPA 625m		11	162	
2005/06-1	W-3	srgt environ, rec	11/3/2005	Organic	2,4,6-Tribromophenol	n/a	=	86	%	EPA 625m		11	162	
2005/06-1	W-4	srgt environ, rec	11/3/2005	Organic	2,4,6-Tribromophenol	n/a	=	80	%	EPA 625m		11	162	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2,4,6-Trichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2,4-Dichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	srgt environ, rec	10/28/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	82	%	EPA 8151A		0	123	
2005/06-1	Lab	srgt method blank, rec	10/28/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	102	%	EPA 8151A		0	123	
2005/06-1	ME-CC	srgt environ, rec	10/28/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	21	%	EPA 8151A		0	123	
2005/06-1	ME-SCR	srgt environ, rec	10/28/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	77	%	EPA 8151A		0	123	
2005/06-1	ME-VR2	srgt environ, rec	10/28/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	79	%	EPA 8151A		0	123	
2005/06-1	W-3	srgt environ, rec	10/28/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	67	%	EPA 8151A		0	123	
2005/06-1	W-4	srgt environ, rec	10/28/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	79	%	EPA 8151A		0	123	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2,4-Dimethylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2,4-Dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2,4-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	2,4-Dinitrotoluene	n/a	=	113	%	EPA 625m		39	139	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	2,4-Dinitrotoluene	n/a	=	107	%	EPA 625m		39	139	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	2,4-Dinitrotoluene	n/a	=	5	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2,6-Dimethylnaphthalene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	2,6-Dimethylnaphthalene	n/a	=	89	%	EPA 625m		60	120	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	2,6-Dimethylnaphthalene	n/a	=	77	%	EPA 625m		60	120	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	2,6-Dimethylnaphthalene	n/a	=	14	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2,6-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2-Chloronaphthalene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2-Chlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	2-Chlorophenol	n/a	=	72	%	EPA 625m		23	134	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	2-Chlorophenol	n/a	=	63	%	EPA 625m		23	134	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	2-Chlorophenol	n/a	=	13	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2-Methyl-4,6-dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2-Methylnaphthalene	n/a	=	12.3	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	2-Methylnaphthalene	n/a	=	95	%	EPA 625m		50	120	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	2-Methylnaphthalene	n/a	=	90	%	EPA 625m		50	120	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	2-Methylnaphthalene	n/a	=	5	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	2-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	3,3'-Dichlorobenzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	4-Bromophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	Lab	method blank	11/3/2005	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	4-Chloro-3-methylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	4-Chloro-3-methylphenol	n/a	=	101	%	EPA 625m		22	147	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	4-Chloro-3-methylphenol	n/a	=	94	%	EPA 625m		22	147	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	4-Chloro-3-methylphenol	n/a	=	7	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	4-Chlorophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	4-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	4-Nitrophenol	n/a	=	44	%	EPA 625m		0.1	132	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	4-Nitrophenol	n/a	=	48	%	EPA 625m		0.1	132	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	4-Nitrophenol	n/a	=	9	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Acenaphthene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Acenaphthene	n/a	=	103	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Acenaphthene	n/a	=	97	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Acenaphthene	n/a	=	6	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	srgt environ, rec	11/3/2005	Organic	Acenaphthene-d10	n/a	=	98	%	EPA 625m		52	125	
2005/06-1	Lab	srgt method blank, rec	11/3/2005	Organic	Acenaphthene-d10	n/a	=	82	%	EPA 625m		52	125	
2005/06-1	ME-CC	srgt environ, rec	11/3/2005	Organic	Acenaphthene-d10	n/a	=	89	%	EPA 625m		52	125	
2005/06-1	ME-CC	srgt matrix spike dup, rec	11/3/2005	Organic	Acenaphthene-d10	n/a	=	93	%	EPA 625m		18	133	
2005/06-1	ME-CC	srgt matrix spike, rec	11/3/2005	Organic	Acenaphthene-d10	n/a	=	91	%	EPA 625m		18	133	
2005/06-1	ME-SCR	srgt environ, rec	11/3/2005	Organic	Acenaphthene-d10	n/a	=	97	%	EPA 625m		52	125	
2005/06-1	ME-VR2	srgt environ, rec	11/3/2005	Organic	Acenaphthene-d10	n/a	=	97	%	EPA 625m		52	125	
2005/06-1	ME-VR2	srgt field blank, rec	11/3/2005	Organic	Acenaphthene-d10	n/a	=	96	%	EPA 625m		52	125	
2005/06-1	W-3	srgt environ, rec	11/3/2005	Organic	Acenaphthene-d10	n/a	=	95	%	EPA 625m		52	125	
2005/06-1	W-4	srgt environ, rec	11/3/2005	Organic	Acenaphthene-d10	n/a	=	89	%	EPA 625m		52	125	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Acenaphthylene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Acenaphthylene	n/a	=	92	%	EPA 625m		60	120	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Acenaphthylene	n/a	=	91	%	EPA 625m		60	120	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Acenaphthylene	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Anthracene	n/a	=	115	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Anthracene	n/a	=	114	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Anthracene	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Azobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Benzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Benzo(a)anthracene	n/a	=	53.7	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Benzo(a)anthracene	n/a	=	130	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Benzo(a)anthracene	n/a	=	130	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Benzo(a)anthracene	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Benzo(a)pyrene	n/a	=	27.5	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Benzo(a)pyrene	n/a	=	106	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Benzo(a)pyrene	n/a	=	105	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Benzo(a)pyrene	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Benzo(b)fluoranthene	n/a	=	31.3	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Benzo(b)fluoranthene	n/a	=	145	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Benzo(b)fluoranthene	n/a	=	139	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Benzo(b)fluoranthene	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Benzo(e)pyrene	n/a	=	14.5	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Benzo(e)pyrene	n/a	=	116	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Benzo(e)pyrene	n/a	=	106	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Benzo(e)pyrene	n/a	=	9	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Benzo(g,h,i)perylene	n/a	=	15.9	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Benzo(g,h,i)perylene	n/a	=	99	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Benzo(g,h,i)perylene	n/a	=	95	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Benzo(g,h,i)perylene	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Benzo(k)fluoranthene	n/a	=	5.7	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Benzo(k)fluoranthene	n/a	=	113	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Benzo(k)fluoranthene	n/a	=	97	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Benzo(k)fluoranthene	n/a	=	15	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Biphenyl	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Biphenyl	n/a	=	90	%	EPA 625m		50	120	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Biphenyl	n/a	=	93	%	EPA 625m		50	120	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Biphenyl	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Bis(2-chloroethoxy)methane	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Bis(2-chloroethyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Bis(2-chloroisopropyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	102.2	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.485	µg/L	EPA 625m	0.005		0.005	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	182	%	EPA 625m		8	158	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	189	%	EPA 625m		8	158	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.64	µg/L	EPA 625m	0.005		0.005	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Butyl benzyl phthalate	n/a	=	6.8	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Butyl benzyl phthalate	n/a	=	0.0294	µg/L	EPA 625m	0.005		0.005	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Butyl benzyl phthalate	n/a	=	106	%	EPA 625m		0.005	152	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Butyl benzyl phthalate	n/a	=	105	%	EPA 625m		0.005	152	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Butyl benzyl phthalate	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Butyl benzyl phthalate	n/a	=	0.0515	µg/L	EPA 625m	0.005		0.005	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Chrysene	n/a	=	1.8	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Chrysene	n/a	=	121	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Chrysene	n/a	=	124	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Chrysene	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	srgt environ, rec	11/3/2005	Organic	Chrysene-d12	n/a	=	101	%	EPA 625m		61	126	
2005/06-1	Lab	srgt method blank, rec	11/3/2005	Organic	Chrysene-d12	n/a	=	99	%	EPA 625m		61	126	
2005/06-1	ME-CC	srgt environ, rec	11/3/2005	Organic	Chrysene-d12	n/a	=	95	%	EPA 625m		61	126	
2005/06-1	ME-CC	srgt matrix spike dup, rec	11/3/2005	Organic	Chrysene-d12	n/a	=	103	%	EPA 625m		41	131	
2005/06-1	ME-CC	srgt matrix spike, rec	11/3/2005	Organic	Chrysene-d12	n/a	=	97	%	EPA 625m		41	131	
2005/06-1	ME-SCR	srgt environ, rec	11/3/2005	Organic	Chrysene-d12	n/a	=	102	%	EPA 625m		61	126	
2005/06-1	ME-VR2	srgt environ, rec	11/3/2005	Organic	Chrysene-d12	n/a	=	98	%	EPA 625m		61	126	
2005/06-1	ME-VR2	srgt field blank, rec	11/3/2005	Organic	Chrysene-d12	n/a	=	99	%	EPA 625m		61	126	
2005/06-1	W-3	srgt environ, rec	11/3/2005	Organic	Chrysene-d12	n/a	=	94	%	EPA 625m		61	126	
2005/06-1	W-4	srgt environ, rec	11/3/2005	Organic	Chrysene-d12	n/a	=	96	%	EPA 625m		61	126	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Dibenz(a,h)anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Dibenz(a,h)anthracene	n/a	=	106	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Dibenz(a,h)anthracene	n/a	=	104	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Dibenz(a,h)anthracene	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Dibenzothiophene	n/a	=	14.9	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	srgt environ, rec	10/25/2005	Organic	Dibromofluoromethane	n/a	=	111	%	EPA 8260B		74	140	
2005/06-1	Lab	srgt method blank, rec	10/25/2005	Organic	Dibromofluoromethane	n/a	=	108	%	EPA 8260B		74	140	
2005/06-1	W-3	srgt environ, rec	10/25/2005	Organic	Dibromofluoromethane	n/a	=	108	%	EPA 8260B		74	140	
2005/06-1	W-4	srgt environ, rec	10/25/2005	Organic	Dibromofluoromethane	n/a	=	106	%	EPA 8260B		74	140	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Diethyl phthalate	n/a	=	5.6	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Diethyl phthalate	n/a	=	0.0776	µg/L	EPA 625m	0.005		0.005	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Diethyl phthalate	n/a	=	95	%	EPA 625m		0.005	134	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Diethyl phthalate	n/a	=	82	%	EPA 625m		0.005	134	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Diethyl phthalate	n/a	=	15	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Diethyl phthalate	n/a	=	0.0881	µg/L	EPA 625m	0.005		0.005	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Dimethyl phthalate	n/a	=	8.8	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Dimethyl phthalate	n/a	=	0.0956	µg/L	EPA 625m	0.005		0.005	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Dimethyl phthalate	n/a	=	73	%	EPA 625m		0.005	112	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Dimethyl phthalate	n/a	=	67	%	EPA 625m		0.005	112	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Dimethyl phthalate	n/a	=	9	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Dimethyl phthalate	n/a	=	0.0688	µg/L	EPA 625m	0.005		0.005	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Di-n-butylphthalate	n/a	=	9	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Di-n-butylphthalate	n/a	=	0.0498	µg/L	EPA 625m	0.005		0.005	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Di-n-butylphthalate	n/a	=	97	%	EPA 625m		1	118	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Di-n-butylphthalate	n/a	=	97	%	EPA 625m		1	118	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Di-n-butylphthalate	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Di-n-butylphthalate	n/a	=	0.19	µg/L	EPA 625m	0.005		0.005	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Di-n-octylphthalate	n/a	=	22.7	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Di-n-octylphthalate	n/a	=	103	%	EPA 625m		4	146	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Di-n-octylphthalate	n/a	=	102	%	EPA 625m		4	146	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Di-n-octylphthalate	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Fluoranthene	n/a	=	12.4	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	Lab	method blank	11/3/2005	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Fluoranthene	n/a	=	128	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Fluoranthene	n/a	=	119	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Fluoranthene	n/a	=	7	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Fluoranthene	n/a	=	0.029	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Fluorene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Fluorene	n/a	=	110	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Fluorene	n/a	=	111	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Fluorene	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Hexachlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Hexachlorobutadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Hexachlorocyclopentadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Hexachloroethane	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	150.6	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	111	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	105	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	6	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Isophorone	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	10/25/2005	Organic	Methyl tert-butyl ether (MTBE)	n/a	=	14.3	%	EPA 8260B		0	25	
2005/06-1	Lab	LCS dup, rec	10/25/2005	Organic	Methyl tert-butyl ether (MTBE)	n/a	=	106	%	EPA 8260B		82	118	
2005/06-1	Lab	LCS, rec	10/25/2005	Organic	Methyl tert-butyl ether (MTBE)	n/a	=	106	%	EPA 8260B		82	118	
2005/06-1	Lab	LCS, RPD	10/25/2005	Organic	Methyl tert-butyl ether (MTBE)	n/a	=	0	%	EPA 8260B		0	13	
2005/06-1	Lab	method blank	10/25/2006	Organic	Methyl tert-butyl ether (MTBE)	n/a	<	1	µg/L	EPA 8260B	1		1	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Naphthalene	n/a	=	15.8	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Naphthalene	n/a	=	89	%	EPA 625m		50	120	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Naphthalene	n/a	=	80	%	EPA 625m		50	120	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Naphthalene	n/a	=	11	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	srgt environ, rec	11/3/2005	Organic	Naphthalene-d8	n/a	=	86	%	EPA 625m		47	110	
2005/06-1	Lab	srgt method blank, rec	11/3/2005	Organic	Naphthalene-d8	n/a	=	83	%	EPA 625m		47	110	
2005/06-1	ME-CC	srgt environ, rec	11/3/2005	Organic	Naphthalene-d8	n/a	=	82	%	EPA 625m		47	110	
2005/06-1	ME-CC	srgt matrix spike dup, rec	11/3/2005	Organic	Naphthalene-d8	n/a	=	89	%	EPA 625m		6	136	
2005/06-1	ME-CC	srgt matrix spike, rec	11/3/2005	Organic	Naphthalene-d8	n/a	=	81	%	EPA 625m		6	136	
2005/06-1	ME-SCR	srgt environ, rec	11/3/2005	Organic	Naphthalene-d8	n/a	=	89	%	EPA 625m		47	110	
2005/06-1	ME-VR2	srgt environ, rec	11/3/2005	Organic	Naphthalene-d8	n/a	=	99	%	EPA 625m		47	110	
2005/06-1	ME-VR2	srgt field blank, rec	11/3/2005	Organic	Naphthalene-d8	n/a	=	84	%	EPA 625m		47	110	
2005/06-1	W-3	srgt environ, rec	11/3/2005	Organic	Naphthalene-d8	n/a	=	95	%	EPA 625m		47	110	
2005/06-1	W-4	srgt environ, rec	11/3/2005	Organic	Naphthalene-d8	n/a	=	78	%	EPA 625m		47	110	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Nitrobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	N-Nitrosodimethylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	N-Nitrosodi-N-propylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	N-Nitrosodi-N-propylamine	n/a	=	96	%	EPA 625m		60	140	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	N-Nitrosodi-N-propylamine	n/a	=	91	%	EPA 625m		60	140	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	N-Nitrosodi-N-propylamine	n/a	=	5	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	N-Nitrosodiphenylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Pentachlorophenol	n/a	=	20.1	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Pentachlorophenol	n/a	=	100	%	EPA 625m		14	176	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Pentachlorophenol	n/a	=	101	%	EPA 625m		14	176	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Pentachlorophenol	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Perylene	n/a	=	4.1	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Perylene	n/a	=	90	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Perylene	n/a	=	77	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Perylene	n/a	=	16	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	srgt environ, rec	11/3/2005	Organic	Perylene-d12	n/a	=	103	%	EPA 625m		53	122	
2005/06-1	Lab	srgt method blank, rec	11/3/2005	Organic	Perylene-d12	n/a	=	102	%	EPA 625m		53	122	
2005/06-1	ME-CC	srgt environ, rec	11/3/2005	Organic	Perylene-d12	n/a	=	87	%	EPA 625m		53	122	
2005/06-1	ME-CC	srgt matrix spike dup, rec	11/3/2005	Organic	Perylene-d12	n/a	=	98	%	EPA 625m		34	134	
2005/06-1	ME-CC	srgt matrix spike, rec	11/3/2005	Organic	Perylene-d12	n/a	=	97	%	EPA 625m		34	134	
2005/06-1	ME-SCR	srgt environ, rec	11/3/2005	Organic	Perylene-d12	n/a	=	104	%	EPA 625m		53	122	
2005/06-1	ME-VR2	srgt environ, rec	11/3/2005	Organic	Perylene-d12	n/a	=	96	%	EPA 625m		53	122	
2005/06-1	ME-VR2	srgt field blank, rec	11/3/2005	Organic	Perylene-d12	n/a	=	103	%	EPA 625m		53	122	
2005/06-1	W-3	srgt environ, rec	11/3/2005	Organic	Perylene-d12	n/a	=	93	%	EPA 625m		53	122	
2005/06-1	W-4	srgt environ, rec	11/3/2005	Organic	Perylene-d12	n/a	=	94	%	EPA 625m		53	122	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Phenanthrene	n/a	=	8	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Phenanthrene	n/a	=	115	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Phenanthrene	n/a	=	114	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Phenanthrene	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Phenanthrene	n/a	=	0.007	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	srgt environ, rec	11/3/2005	Organic	Phenanthrene-d10	n/a	=	100	%	EPA 625m		57	128	
2005/06-1	Lab	srgt method blank, rec	11/3/2005	Organic	Phenanthrene-d10	n/a	=	93	%	EPA 625m		57	128	
2005/06-1	ME-CC	srgt environ, rec	11/3/2005	Organic	Phenanthrene-d10	n/a	=	95	%	EPA 625m		57	128	
2005/06-1	ME-CC	srgt matrix spike dup, rec	11/3/2005	Organic	Phenanthrene-d10	n/a	=	101	%	EPA 625m		43	124	
2005/06-1	ME-CC	srgt matrix spike, rec	11/3/2005	Organic	Phenanthrene-d10	n/a	=	99	%	EPA 625m		43	124	
2005/06-1	ME-SCR	srgt environ, rec	11/3/2005	Organic	Phenanthrene-d10	n/a	=	100	%	EPA 625m		57	128	
2005/06-1	ME-VR2	srgt environ, rec	11/3/2005	Organic	Phenanthrene-d10	n/a	=	96	%	EPA 625m		57	128	
2005/06-1	ME-VR2	srgt field blank, rec	11/3/2005	Organic	Phenanthrene-d10	n/a	=	100	%	EPA 625m		57	128	
2005/06-1	W-3	srgt environ, rec	11/3/2005	Organic	Phenanthrene-d10	n/a	=	97	%	EPA 625m		57	128	
2005/06-1	W-4	srgt environ, rec	11/3/2005	Organic	Phenanthrene-d10	n/a	=	95	%	EPA 625m		57	128	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Phenol	n/a	=	14.8	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Phenol	n/a	=	32	%	EPA 625m		5	112	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Phenol	n/a	=	33	%	EPA 625m		5	112	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Phenol	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	A-1	srgt environ, rec	11/3/2005	Organic	Phenol-d5	n/a	=	27	%	EPA 625m		20	100	
2005/06-1	Lab	srgt method blank, rec	11/3/2005	Organic	Phenol-d5	n/a	=	37	%	EPA 625m		20	100	
2005/06-1	ME-CC	srgt environ, rec	11/3/2005	Organic	Phenol-d5	n/a	=	28	%	EPA 625m		20	100	
2005/06-1	ME-CC	srgt matrix spike dup, rec	11/3/2005	Organic	Phenol-d5	n/a	=	32	%	EPA 625m		20	100	
2005/06-1	ME-CC	srgt matrix spike, rec	11/3/2005	Organic	Phenol-d5	n/a	=	34	%	EPA 625m		20	100	
2005/06-1	ME-SCR	srgt environ, rec	11/3/2005	Organic	Phenol-d5	n/a	=	26	%	EPA 625m		20	100	
2005/06-1	ME-VR2	srgt environ, rec	11/3/2005	Organic	Phenol-d5	n/a	=	25	%	EPA 625m		20	100	
2005/06-1	ME-VR2	srgt field blank, rec	11/3/2005	Organic	Phenol-d5	n/a	=	11	%	EPA 625m		20	100	
2005/06-1	W-3	srgt environ, rec	11/3/2005	Organic	Phenol-d5	n/a	=	30	%	EPA 625m		20	100	
2005/06-1	W-4	srgt environ, rec	11/3/2005	Organic	Phenol-d5	n/a	=	22	%	EPA 625m		20	100	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Pyrene	n/a	=	22.7	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Organic	Pyrene	n/a	=	123	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Organic	Pyrene	n/a	=	120	%	EPA 625m		70	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Organic	Pyrene	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Pyrene	n/a	=	0.132	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	srgt environ, rec	11/3/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	102	%	EPA 625m		40	110	
2005/06-1	Lab	srgt method blank, rec	11/3/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	79	%	EPA 625m		40	110	
2005/06-1	ME-CC	srgt environ, rec	11/3/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	83	%	EPA 625m		40	110	
2005/06-1	ME-CC	srgt matrix spike dup, rec	11/3/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	98	%	EPA 625m		20	120	
2005/06-1	ME-CC	srgt matrix spike, rec	11/3/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	95	%	EPA 625m		20	120	
2005/06-1	ME-SCR	srgt environ, rec	11/3/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	87	%	EPA 625m		40	110	
2005/06-1	ME-VR2	srgt environ, rec	11/3/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	85	%	EPA 625m		40	110	
2005/06-1	ME-VR2	srgt field blank, rec	11/3/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	95	%	EPA 625m		40	110	
2005/06-1	W-3	srgt environ, rec	11/3/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	84	%	EPA 625m		40	110	
2005/06-1	W-4	srgt environ, rec	11/3/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	80	%	EPA 625m		40	110	
2005/06-1	A-1	srgt environ, rec	10/25/2005	Organic	Toluene-d8	n/a	=	99	%	EPA 8260B		88	112	
2005/06-1	Lab	srgt method blank, rec	10/25/2005	Organic	Toluene-d8	n/a	=	102	%	EPA 8260B		88	112	
2005/06-1	W-3	srgt environ, rec	10/25/2005	Organic	Toluene-d8	n/a	=	102	%	EPA 8260B		88	112	
2005/06-1	W-4	srgt environ, rec	10/25/2005	Organic	Toluene-d8	n/a	=	99	%	EPA 8260B		88	112	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Organic	Total Detectable PAHs	n/a	=	13.9	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Organic	Total Detectable PAHs	n/a	=	0	µg/L	EPA 625m				
2005/06-1	ME-VR2	field blank	11/3/2005	Organic	Total Detectable PAHs	n/a	=	0.168	µg/L	EPA 625m				
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	Aroclor 1016	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	Aroclor 1221	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	Aroclor 1232	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	Aroclor 1242	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	Aroclor 1248	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	Aroclor 1254	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	Aroclor 1260	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 018	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 018	n/a	=	94	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 018	n/a	=	91	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 018	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 028	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 028	n/a	=	95	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 028	n/a	=	87	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 028	n/a	=	9	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	srgt environ, rec	11/3/2005	PCB	PCB 030	n/a	=	110	%	EPA 625m		46	119	
2005/06-1	Lab	srgt method blank, rec	11/3/2005	PCB	PCB 030	n/a	=	83	%	EPA 625m		46	119	
2005/06-1	ME-CC	srgt environ, rec	11/3/2005	PCB	PCB 030	n/a	=	82	%	EPA 625m		46	119	
2005/06-1	ME-CC	srgt matrix spike dup, rec	11/3/2005	PCB	PCB 030	n/a	=	101	%	EPA 625m		20	120	
2005/06-1	ME-CC	srgt matrix spike, rec	11/3/2005	PCB	PCB 030	n/a	=	99	%	EPA 625m		20	120	
2005/06-1	ME-SCR	srgt environ, rec	11/3/2005	PCB	PCB 030	n/a	=	90	%	EPA 625m		46	119	
2005/06-1	ME-VR2	srgt environ, rec	11/3/2005	PCB	PCB 030	n/a	=	91	%	EPA 625m		46	119	
2005/06-1	ME-VR2	srgt field blank, rec	11/3/2005	PCB	PCB 030	n/a	=	100	%	EPA 625m		46	119	
2005/06-1	W-3	srgt environ, rec	11/3/2005	PCB	PCB 030	n/a	=	89	%	EPA 625m		46	119	
2005/06-1	W-4	srgt environ, rec	11/3/2005	PCB	PCB 030	n/a	=	86	%	EPA 625m		46	119	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 031	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 031	n/a	=	94	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 031	n/a	=	87	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 031	n/a	=	7	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 033	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 033	n/a	=	89	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 033	n/a	=	91	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 033	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 037	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 037	n/a	=	92	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 037	n/a	=	87	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 037	n/a	=	6	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 044	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 044	n/a	=	88	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 044	n/a	=	89	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 044	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 049	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 049	n/a	=	88	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 049	n/a	=	81	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 049	n/a	=	8	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 052	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 052	n/a	=	96	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 052	n/a	=	83	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 052	n/a	=	15	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 066	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 066	n/a	=	95	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 066	n/a	=	81	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 066	n/a	=	16	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 070	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 070	n/a	=	92	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 070	n/a	=	88	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 070	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 074	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 074	n/a	=	100	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 074	n/a	=	88	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 074	n/a	=	13	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 077	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 077	n/a	=	85	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 077	n/a	=	78	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 077	n/a	=	9	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 081	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 081	n/a	=	92	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 081	n/a	=	79	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 081	n/a	=	15	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 087	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 087	n/a	=	90	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 087	n/a	=	85	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 087	n/a	=	6	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 095	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 095	n/a	=	97	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 095	n/a	=	93	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 095	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 097	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 097	n/a	=	89	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 097	n/a	=	83	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 097	n/a	=	7	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 099	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 099	n/a	=	87	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 099	n/a	=	83	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 099	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 101	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 101	n/a	=	91	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 101	n/a	=	89	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 101	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 105	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 105	n/a	=	81	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 105	n/a	=	84	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 105	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 110	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 110	n/a	=	91	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 110	n/a	=	96	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 110	n/a	=	5	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	srgt environ, rec	11/3/2005	PCB	PCB 112	n/a	=	98	%	EPA 625m		52	123	
2005/06-1	Lab	srgt method blank, rec	11/3/2005	PCB	PCB 112	n/a	=	94	%	EPA 625m		52	123	
2005/06-1	ME-CC	srgt environ, rec	11/3/2005	PCB	PCB 112	n/a	=	82	%	EPA 625m		52	123	
2005/06-1	ME-CC	srgt matrix spike dup, rec	11/3/2005	PCB	PCB 112	n/a	=	94	%	EPA 625m		31	120	
2005/06-1	ME-CC	srgt matrix spike, rec	11/3/2005	PCB	PCB 112	n/a	=	97	%	EPA 625m		31	120	
2005/06-1	ME-SCR	srgt environ, rec	11/3/2005	PCB	PCB 112	n/a	=	84	%	EPA 625m		52	123	
2005/06-1	ME-VR2	srgt environ, rec	11/3/2005	PCB	PCB 112	n/a	=	91	%	EPA 625m		52	123	
2005/06-1	ME-VR2	srgt field blank, rec	11/3/2005	PCB	PCB 112	n/a	=	97	%	EPA 625m		52	123	
2005/06-1	W-3	srgt environ, rec	11/3/2005	PCB	PCB 112	n/a	=	80	%	EPA 625m		52	123	
2005/06-1	W-4	srgt environ, rec	11/3/2005	PCB	PCB 112	n/a	=	80	%	EPA 625m		52	123	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 114	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 114	n/a	=	92	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 114	n/a	=	87	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 114	n/a	=	6	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 118	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 118	n/a	=	90	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 118	n/a	=	79	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 118	n/a	=	13	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 119	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 119	n/a	=	89	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 119	n/a	=	84	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 119	n/a	=	6	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 123	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 123	n/a	=	85	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 123	n/a	=	81	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 123	n/a	=	5	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 126	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 126	n/a	=	80	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 126	n/a	=	81	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 126	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 128 + 167	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 128 + 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 128 + 167	n/a	=	95	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 128 + 167	n/a	=	88	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 128 + 167	n/a	=	8	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 128 + 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 138	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 138	n/a	=	86	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 138	n/a	=	84	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 138	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 141	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 141	n/a	=	87	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 141	n/a	=	84	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 141	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 149	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 149	n/a	=	97	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 149	n/a	=	84	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 149	n/a	=	14	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 151	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 151	n/a	=	94	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 151	n/a	=	97	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 151	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 153	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 153	n/a	=	94	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 153	n/a	=	103	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 153	n/a	=	9	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 156	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 156	n/a	=	92	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 156	n/a	=	83	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 156	n/a	=	10	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 157	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 157	n/a	=	84	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 157	n/a	=	76	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 157	n/a	=	10	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 158	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 158	n/a	=	92	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 158	n/a	=	88	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 158	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 168 + 132	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 168 + 132	n/a	=	94	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 168 + 132	n/a	=	94	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 168 + 132	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 169	n/a	<	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 169	n/a	=	82	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 169	n/a	=	90	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 169	n/a	=	9	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 170	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 170	n/a	=	73	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 170	n/a	=	76	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 170	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 177	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 177	n/a	=	87	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 177	n/a	=	90	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 177	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 180	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 180	n/a	=	84	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 180	n/a	=	73	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 180	n/a	=	14	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 183	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 183	n/a	=	89	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 183	n/a	=	86	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 183	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 187	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 187	n/a	=	82	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 187	n/a	=	87	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 187	n/a	=	6	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 189	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 189	n/a	=	83	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 189	n/a	=	91	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 189	n/a	=	9	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 194	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 194	n/a	=	77	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 194	n/a	=	79	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 194	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	srgt environ, rec	11/3/2005	PCB	PCB 198	n/a	=	94	%	EPA 625m		59	123	
2005/06-1	Lab	srgt method blank, rec	11/3/2005	PCB	PCB 198	n/a	=	94	%	EPA 625m		59	123	
2005/06-1	ME-CC	srgt environ, rec	11/3/2005	PCB	PCB 198	n/a	=	76	%	EPA 625m		59	123	
2005/06-1	ME-CC	srgt matrix spike dup, rec	11/3/2005	PCB	PCB 198	n/a	=	93	%	EPA 625m		30	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-CC	srgt matrix spike, rec	11/3/2005	PCB	PCB 198	n/a	=	91	%	EPA 625m		30	125	
2005/06-1	ME-SCR	srgt environ, rec	11/3/2005	PCB	PCB 198	n/a	=	83	%	EPA 625m		59	123	
2005/06-1	ME-VR2	srgt environ, rec	11/3/2005	PCB	PCB 198	n/a	=	89	%	EPA 625m		59	123	
2005/06-1	ME-VR2	srgt field blank, rec	11/3/2005	PCB	PCB 198	n/a	=	93	%	EPA 625m		59	123	
2005/06-1	W-3	srgt environ, rec	11/3/2005	PCB	PCB 198	n/a	=	78	%	EPA 625m		59	123	
2005/06-1	W-4	srgt environ, rec	11/3/2005	PCB	PCB 198	n/a	=	79	%	EPA 625m		59	123	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 200	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 200	n/a	=	85	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 200	n/a	=	75	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 200	n/a	=	12	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 201	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 201	n/a	=	97	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 201	n/a	=	97	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 201	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	PCB 206	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	PCB	PCB 206	n/a	=	89	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	PCB	PCB 206	n/a	=	75	%	EPA 625m		65	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	PCB	PCB 206	n/a	=	17	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	PCB	Total Detectable PCBs	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-1	ME-VR2	field blank	11/3/2005	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-1	A-1	lab duplicate, RPD	10/28/2005	Pesticide	2,4,5-T	n/a	=	0	%	EPA 8151A		0	25	
2005/06-1	Lab	LCS dup, rec	10/27/2005	Pesticide	2,4,5-T	n/a	=	89	%	EPA 8151A		30	130	
2005/06-1	Lab	LCS, rec	10/27/2005	Pesticide	2,4,5-T	n/a	=	88	%	EPA 8151A		30	130	
2005/06-1	Lab	LCS, RPD	10/27/2005	Pesticide	2,4,5-T	n/a	=	1	%	EPA 8151A		0	30	
2005/06-1	Lab	method blank	10/28/2005	Pesticide	2,4,5-T	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-1	ME-CC	matrix spike dup, rec	10/28/2005	Pesticide	2,4,5-T	n/a	=	82	%	EPA 8151A		30	130	
2005/06-1	ME-CC	matrix spike, rec	10/28/2005	Pesticide	2,4,5-T	n/a	=	76	%	EPA 8151A		30	130	
2005/06-1	ME-CC	matrix spike, RPD	10/28/2005	Pesticide	2,4,5-T	n/a	=	8	%	EPA 8151A		0	30	
2005/06-1	A-1	lab duplicate, RPD	10/28/2005	Pesticide	2,4,5-TP (Silvex)	n/a	=	0	%	EPA 8151A		0	25	
2005/06-1	Lab	method blank	10/28/2005	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-1	A-1	lab duplicate, RPD	10/28/2005	Pesticide	2,4-D	n/a	=	0	%	EPA 8151A		0	25	
2005/06-1	Lab	LCS dup, rec	10/27/2005	Pesticide	2,4-D	n/a	=	99	%	EPA 8151A		30	130	
2005/06-1	Lab	LCS, rec	10/27/2005	Pesticide	2,4-D	n/a	=	96	%	EPA 8151A		30	130	
2005/06-1	Lab	LCS, RPD	10/27/2005	Pesticide	2,4-D	n/a	=	4	%	EPA 8151A		0	30	
2005/06-1	Lab	method blank	10/28/2005	Pesticide	2,4-D	n/a	<	5	µg/L	EPA 8151A	5		5	
2005/06-1	ME-CC	matrix spike dup, rec	10/28/2005	Pesticide	2,4-D	n/a	=	83	%	EPA 8151A		30	130	
2005/06-1	ME-CC	matrix spike, rec	10/28/2005	Pesticide	2,4-D	n/a	=	85	%	EPA 8151A		30	130	
2005/06-1	ME-CC	matrix spike, RPD	10/28/2005	Pesticide	2,4-D	n/a	=	3	%	EPA 8151A		0	30	
2005/06-1	A-1	lab duplicate, RPD	10/28/2005	Pesticide	2,4-DB	n/a	=	0	%	EPA 8151A		0	25	
2005/06-1	Lab	LCS dup, rec	10/27/2005	Pesticide	2,4-DB	n/a	=	45	%	EPA 8151A		30	130	
2005/06-1	Lab	LCS, rec	10/27/2005	Pesticide	2,4-DB	n/a	=	43	%	EPA 8151A		30	130	
2005/06-1	Lab	LCS, RPD	10/27/2005	Pesticide	2,4-DB	n/a	=	3	%	EPA 8151A		0	30	
2005/06-1	Lab	method blank	10/28/2005	Pesticide	2,4-DB	n/a	<	5	µg/L	EPA 8151A	5		0.5	
2005/06-1	ME-CC	matrix spike dup, rec	10/28/2005	Pesticide	2,4-DB	n/a	=	36	%	EPA 8151A		30	130	
2005/06-1	ME-CC	matrix spike, rec	10/28/2005	Pesticide	2,4-DB	n/a	=	36	%	EPA 8151A		30	130	
2005/06-1	ME-CC	matrix spike, RPD	10/28/2005	Pesticide	2,4-DB	n/a	=	1	%	EPA 8151A		0	30	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	2,4'-DDD	n/a	=	14.2	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	2,4'-DDD	n/a	=	104	%	EPA 625m		56	129	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	2,4'-DDD	n/a	=	101	%	EPA 625m		56	129	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	2,4'-DDD	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	2,4'-DDE	n/a	=	25.4	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	2,4'-DDE	n/a	=	92	%	EPA 625m		60	129	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	2,4'-DDE	n/a	=	97	%	EPA 625m		60	129	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	2,4'-DDE	n/a	=	5	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	2,4'-DDT	n/a	=	30.7	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	2,4'-DDT	n/a	=	54	%	EPA 625m		39	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	2,4'-DDT	n/a	=	50	%	EPA 625m		39	130	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	2,4'-DDT	n/a	=	8	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	4,4'-DDD	n/a	=	8.5	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	4,4'-DDD	n/a	=	122	%	EPA 625m		46	138	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	4,4'-DDD	n/a	=	135	%	EPA 625m		46	138	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	4,4'-DDD	n/a	=	10	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	4,4'-DDE	n/a	=	22	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	4,4'-DDE	n/a	=	107	%	EPA 625m		69	116	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	4,4'-DDE	n/a	=	117	%	EPA 625m		69	116	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	4,4'-DDE	n/a	=	9	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	4,4'-DDT	n/a	=	25.1	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	4,4'-DDT	n/a	=	40	%	EPA 625m		34	136	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	4,4'-DDT	n/a	=	39	%	EPA 625m		34	136	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	4,4'-DDT	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Aldrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Aldrin	n/a	=	75	%	EPA 625m		43	128	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Aldrin	n/a	=	77	%	EPA 625m		43	128	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Aldrin	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	BHC-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	BHC-alpha	n/a	=	83	%	EPA 625m		60	123	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	BHC-alpha	n/a	=	82	%	EPA 625m		60	123	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	BHC-alpha	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	BHC-beta	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	BHC-beta	n/a	=	84	%	EPA 625m		45	140	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	BHC-beta	n/a	=	95	%	EPA 625m		45	140	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	BHC-beta	n/a	=	12	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	BHC-delta	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	BHC-delta	n/a	=	97	%	EPA 625m		29	113	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	BHC-delta	n/a	=	86	%	EPA 625m		29	113	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	BHC-delta	n/a	=	12	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	BHC-gamma (Lindane)	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	BHC-gamma (Lindane)	n/a	=	83	%	EPA 625m		59	110	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	BHC-gamma (Lindane)	n/a	=	83	%	EPA 625m		59	110	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	BHC-gamma (Lindane)	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Bolstar	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Bolstar	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Bolstar	n/a	=	105	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Bolstar	n/a	=	106	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Bolstar	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Bolstar	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Chlordane-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Chlordane-alpha	n/a	=	76	%	EPA 625m		64	117	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Chlordane-alpha	n/a	=	89	%	EPA 625m		64	117	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Chlordane-alpha	n/a	=	16	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Chlordane-gamma	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Chlordane-gamma	n/a	=	86	%	EPA 625m		46	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Chlordane-gamma	n/a	=	88	%	EPA 625m		46	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Chlordane-gamma	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Chlorpyrifos	n/a	=	30.9	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Chlorpyrifos	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Chlorpyrifos	n/a	=	106	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Chlorpyrifos	n/a	=	104	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Chlorpyrifos	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Chlorpyrifos	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	cis-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	cis-Nonachlor	n/a	=	85	%	EPA 625m		60	140	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	cis-Nonachlor	n/a	=	84	%	EPA 625m		60	140	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	cis-Nonachlor	n/a	=	1.2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	10/28/2005	Pesticide	Dalapon	n/a	=	0	%	EPA 8151A		0	25	
2005/06-1	Lab	method blank	10/28/2005	Pesticide	Dalapon	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Demeton-O	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Demeton-O	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Demeton-O	n/a	=	76	%	EPA 625m		45	105	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Demeton-O	n/a	=	79	%	EPA 625m		45	105	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Demeton-O	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Demeton-O	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Diazinon	n/a	=	88.3	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Diazinon	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Diazinon	n/a	=	101	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Diazinon	n/a	=	99	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Diazinon	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Diazinon	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-1	A-1	lab duplicate, RPD	10/28/2005	Pesticide	Dicamba	n/a	=	0	%	EPA 8151A		0	25	
2005/06-1	Lab	method blank	10/28/2005	Pesticide	Dicamba	n/a	<	13	µg/L	EPA 8151A	13		13	
2005/06-1	A-1	lab duplicate, RPD	10/28/2005	Pesticide	Dichlorprop	n/a	=	0	%	EPA 8151A		0	25	
2005/06-1	Lab	method blank	10/28/2005	Pesticide	Dichlorprop	n/a	<	5	µg/L	EPA 8151A	5		5	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Dichlorvos	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Dichlorvos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Dichlorvos	n/a	=	103	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Dichlorvos	n/a	=	102	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Dichlorvos	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Dichlorvos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Dieldrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Dieldrin	n/a	=	73	%	EPA 625m		46	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Dieldrin	n/a	=	75	%	EPA 625m		46	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Dieldrin	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Dimethoate	n/a	=	15.2	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Dimethoate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Dimethoate	n/a	=	87	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Dimethoate	n/a	=	94	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Dimethoate	n/a	=	8	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Dimethoate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-1	A-1	lab duplicate, RPD	10/28/2005	Pesticide	Dinoseb	n/a	=	0	%	EPA 8151A		0	25	
2005/06-1	Lab	method blank	10/28/2005	Pesticide	Dinoseb	n/a	<	2.5	µg/L	EPA 8151A	2.5		2.5	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Disulfoton	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Disulfoton	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Disulfoton	n/a	=	74	%	EPA 625m		45	105	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Disulfoton	n/a	=	82	%	EPA 625m		45	105	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Disulfoton	n/a	=	10	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Disulfoton	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Endosulfan sulfate	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Endosulfan sulfate	n/a	=	77	%	EPA 625m		25	107	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Endosulfan sulfate	n/a	=	81	%	EPA 625m		25	107	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Endosulfan sulfate	n/a	=	5	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Endosulfan-I	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Endosulfan-I	n/a	=	89	%	EPA 625m		54	141	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Endosulfan-I	n/a	=	103	%	EPA 625m		54	141	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Endosulfan-I	n/a	=	15	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Endosulfan-II	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Endosulfan-II	n/a	=	78	%	EPA 625m		0.001	135	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Endosulfan-II	n/a	=	68	%	EPA 625m		0.001	135	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Endosulfan-II	n/a	=	14	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Endrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Endrin	n/a	=	98	%	EPA 625m		32	141	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Endrin	n/a	=	99	%	EPA 625m		32	141	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Endrin	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Endrin aldehyde	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Endrin ketone	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Endrin ketone	n/a	=	60	%	EPA 625m		50	130	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Endrin ketone	n/a	=	72	%	EPA 625m		50	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Endrin ketone	n/a	=	18	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Ethoprop	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Ethoprop	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Ethoprop	n/a	=	106	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Ethoprop	n/a	=	104	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Ethoprop	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Ethoprop	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Fenclorophos (Ronnel)	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Fenclorophos (Ronnel)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Fenclorophos (Ronnel)	n/a	=	97	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Fenclorophos (Ronnel)	n/a	=	100	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Fenclorophos (Ronnel)	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Fenclorophos (Ronnel)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Fensulfothion	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Fensulfothion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Fensulfothion	n/a	=	99	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Fensulfothion	n/a	=	99	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Fensulfothion	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Fensulfothion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Fenthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Fenthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Fenthion	n/a	=	99	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Fenthion	n/a	=	99	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Fenthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Fenthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	10/26/2005	Pesticide	Glyphosate	n/a	=	3.1	%	EPA 547		0	25	
2005/06-1	Lab	LCS, rec	10/26/2005	Pesticide	Glyphosate	n/a	=	105	%	EPA 547		82	115	
2005/06-1	Lab	method blank	10/26/2005	Pesticide	Glyphosate	n/a	<	6	µg/L	EPA 547	6	0	6	
2005/06-1	Lab	LCS, rec	10/27/2005	Pesticide	Glyphosate	n/a	=	107	%	EPA 547		82	115	
2005/06-1	Lab	method blank	10/27/2005	Pesticide	Glyphosate	n/a	<	6	µg/L	EPA 547	6		6	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Heptachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Heptachlor	n/a	=	74	%	EPA 625m		43	122	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Heptachlor	n/a	=	77	%	EPA 625m		43	122	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Heptachlor	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Heptachlor epoxide	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Heptachlor epoxide	n/a	=	87	%	EPA 625m		56	122	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Heptachlor epoxide	n/a	=	89	%	EPA 625m		56	122	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Heptachlor epoxide	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Malathion	n/a	=	4.4	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Malathion	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Malathion	n/a	=	106	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Malathion	n/a	=	110	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Malathion	n/a	=	4	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Malathion	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-1	A-1	lab duplicate, RPD	10/28/2005	Pesticide	MCPA	n/a	=	0	%	EPA 8151A		0	25	
2005/06-1	Lab	method blank	10/28/2005	Pesticide	MCPA	n/a	<	500	µg/L	EPA 8151A	500		500	
2005/06-1	A-1	lab duplicate, RPD	10/28/2005	Pesticide	MCPP	n/a	=	0	%	EPA 8151A		0	25	
2005/06-1	Lab	method blank	10/28/2005	Pesticide	MCPP	n/a	<	500	µg/L	EPA 8151A	500		500	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Merphos	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Merphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Merphos	n/a	=	92	%	EPA 625m		65	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Merphos	n/a	=	93	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Merphos	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Merphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Methoxychlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Methoxychlor	n/a	=	18	%	EPA 625m		0.001	157	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Methoxychlor	n/a	=	18	%	EPA 625m		0.001	157	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Methoxychlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Methyl parathion	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Methyl parathion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Methyl parathion	n/a	=	104	%	EPA 625m		60	120	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Methyl parathion	n/a	=	109	%	EPA 625m		60	120	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Methyl parathion	n/a	=	5	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Methyl parathion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Mevinphos	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Mevinphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Mevinphos	n/a	=	97	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Mevinphos	n/a	=	100	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Mevinphos	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Mevinphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Mirex	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Mirex	n/a	=	69	%	EPA 625m		56	123	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Mirex	n/a	=	71	%	EPA 625m		56	123	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Mirex	n/a	=	3	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Oxychlorthane	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Oxychlorthane	n/a	=	89	%	EPA 625m		60	140	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Oxychlorthane	n/a	=	81	%	EPA 625m		60	140	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Oxychlorthane	n/a	=	7.1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Phorate	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Phorate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Phorate	n/a	=	92	%	EPA 625m		45	105	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Phorate	n/a	=	94	%	EPA 625m		45	105	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Phorate	n/a	=	2	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Phorate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	92	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	91	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Tokuthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Tokuthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Tokuthion	n/a	=	107	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Tokuthion	n/a	=	100	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Tokuthion	n/a	=	7	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Tokuthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Total Detectable DDTs	n/a	=	19.5	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Toxaphene	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	trans-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	trans-Nonachlor	n/a	=	84	%	EPA 625m		47	143	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	trans-Nonachlor	n/a	=	91	%	EPA 625m		47	143	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	trans-Nonachlor	n/a	=	8	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-1	A-1	lab duplicate, RPD	11/3/2005	Pesticide	Trichloronate	n/a	=	0	%	EPA 625m		0	30	
2005/06-1	Lab	method blank	11/3/2005	Pesticide	Trichloronate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-1	ME-CC	matrix spike dup, rec	11/3/2005	Pesticide	Trichloronate	n/a	=	103	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, rec	11/3/2005	Pesticide	Trichloronate	n/a	=	102	%	EPA 625m		65	125	
2005/06-1	ME-CC	matrix spike, RPD	11/3/2005	Pesticide	Trichloronate	n/a	=	1	%	EPA 625m		0	30	
2005/06-1	ME-VR2	field blank	11/3/2005	Pesticide	Trichloronate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	12/2/2005	Anion	Bromide	n/a	<	0.001	mg/L	SM 4500-Br	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Anion	Bromide	n/a	=	3	%	SM 4500-Br		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Anion	Bromide	n/a	=	0.71	mg/L	SM 4500-Br	0.001			
2005/06-2	Lab	LCS dup, rec	11/15/2005	Anion	Chloride	n/a	=	103	%	SM 4500-Cl E		70	130	
2005/06-2	Lab	LCS, rec	11/15/2005	Anion	Chloride	n/a	=	99	%	SM 4500-Cl E		70	130	
2005/06-2	Lab	LCS, RPD	11/15/2005	Anion	Chloride	n/a	=	4	%	SM 4500-Cl E		0	30	
2005/06-2	Lab	method blank	11/15/2005	Anion	Chloride	n/a	<	0.1	mg/L	SM 4500-Cl E	0.1		0.1	
2005/06-2	ME-VR2	field duplicate	11/15/2005	Anion	Chloride	n/a	=	48.6	mg/L	SM 4500-Cl E	0.1			
2005/06-2	Lab	LCS dup, rec	11/16/2005	Anion	Perchlorate	n/a	=	96	%	EPA 314.0		85	115	
2005/06-2	Lab	LCS, rec	11/16/2005	Anion	Perchlorate	n/a	=	100	%	EPA 314.0		85	115	
2005/06-2	Lab	LCS, RPD	11/16/2005	Anion	Perchlorate	n/a	=	4	%	EPA 314.0		0	15	
2005/06-2	Lab	method blank	11/16/2005	Anion	Perchlorate	n/a	<	2	µg/L	EPA 314.0	2	0		
2005/06-2	ME-VR2	field duplicate	11/16/2005	Anion	Perchlorate	n/a	<	2	µg/L	EPA 314.0	2			
2005/06-2	ME-VR2	matrix spike dup, rec	11/16/2005	Anion	Perchlorate	n/a	=	100	%	EPA 314.0		80	120	
2005/06-2	ME-VR2	matrix spike, rec	11/16/2005	Anion	Perchlorate	n/a	=	100	%	EPA 314.0		80	120	
2005/06-2	ME-VR2	matrix spike, RPD	11/16/2005	Anion	Perchlorate	n/a	=	0	%	EPA 314.0			15	
2005/06-2	ME-CC	field blank	11/9/2005	Bacteriological	E. Coli	n/a	<	10	MPN/100 mL	MMO-MUG	10			
2005/06-2	ME-VR2	field duplicate	11/9/2005	Bacteriological	E. Coli	n/a	=	282	MPN/100 mL	MMO-MUG	100			
2005/06-2	ME-CC	field blank	11/9/2005	Bacteriological	Enterococcus	n/a	<	10	MPN/100 mL	Enterolert	10			
2005/06-2	ME-VR2	field duplicate	11/9/2005	Bacteriological	Enterococcus	n/a	=	222	MPN/100 mL	Enterolert	100			
2005/06-2	ME-CC	field blank	11/9/2005	Bacteriological	Fecal Coliform	n/a	<	2	MPN/100 mL	SM 9221 E	2			
2005/06-2	ME-VR2	field duplicate	11/9/2005	Bacteriological	Fecal Coliform	n/a	=	130	MPN/100 mL	SM 9221 E	2			
2005/06-2	ME-CC	field blank	11/9/2005	Bacteriological	Total Coliform	n/a	<	10	MPN/100 mL	MMO-MUG	10			
2005/06-2	ME-VR2	field duplicate	11/9/2005	Bacteriological	Total Coliform	n/a	=	5475	MPN/100 mL	MMO-MUG	100			
2005/06-2	Lab	method blank	11/11/2005	Conventional	BOD	n/a	<	1	mg/L	EPA 405.1	1		1	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Conventional	BOD	n/a	=	59	mg/L	EPA 405.1	1			EST-FD
2005/06-2	ME-VR2	lab duplicate, RPD	11/11/2005	Conventional	BOD	n/a	=	3	%	EPA 405.1		0	25	
2005/06-2	Lab	method blank	11/11/2005	Conventional	Conductivity	n/a	<	1	µmhos/cm	SM 2510	1		1	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Conventional	Conductivity	n/a	=	900	µmhos/cm	SM 2510	1			
2005/06-2	Lab	method blank	12/2/2005	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1		1	
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Conventional	Hardness as CaCO3	Total	=	1.1	%	SM 2340 B		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Conventional	Hardness as CaCO3	Total	=	402	mg/L	SM 2340 B	1			
2005/06-2	ME-VR2	field duplicate	11/11/2005	Conventional	pH	n/a	=	8.4	pH Units	EPA 150.1	0.1			
2005/06-2	Lab	LCS dup, rec	12/2/2005	Conventional	Total Dissolved Solids	n/a	=	97	%	SM 2540 C		70	130	
2005/06-2	Lab	LCS, rec	12/2/2005	Conventional	Total Dissolved Solids	n/a	=	95	%	SM 2540 C		70	130	
2005/06-2	Lab	LCS, RPD	12/2/2005	Conventional	Total Dissolved Solids	n/a	=	2	%	SM 2540 C		0	30	
2005/06-2	Lab	method blank	12/2/2005	Conventional	Total Dissolved Solids	n/a	<	0.1	mg/L	SM 2540 C	0.1		0.1	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Conventional	Total Dissolved Solids	n/a	=	810	mg/L	SM 2540 C	0.1			
2005/06-2	Lab	LCS, rec	11/14/2005	Conventional	Total Organic Carbon	n/a	=	96	%	EPA 415.1		80	120	
2005/06-2	Lab	method blank	11/14/2005	Conventional	Total Organic Carbon	n/a	<	0.5	mg/L	EPA 415.1	0.5	0		
2005/06-2	ME-VR2	field duplicate	11/11/2005	Conventional	Total Organic Carbon	n/a	=	41	mg/L	EPA 415.1	0.5			
2005/06-2	Lab	method blank	11/15/2005	Conventional	Total Suspended Solids	n/a	<	0.5	mg/L	SM 2540 D	0.5		0.5	
2005/06-2	ME-VR2	field duplicate	11/15/2005	Conventional	Total Suspended Solids	n/a	=	15.8	mg/L	SM 2540 D	0.5			



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	Lab	method blank	11/11/2005	Conventional	Turbidity	n/a	<	1	NTU	EPA 180.1	1			
2005/06-2	ME-VR2	field duplicate	11/11/2005	Conventional	Turbidity	n/a	=	15	NTU	EPA 180.1	1			EST-FD
2005/06-2	ME-VR2	lab duplicate, RPD	11/11/2005	Conventional	Turbidity	n/a	=	12.2	%	EPA 180.1		0	30	
2005/06-2	Lab	LCS dup, rec	11/17/2005	Hydrocarbon	Oil and Grease	n/a	=	83	%	EPA 1664A		70	130	
2005/06-2	Lab	LCS, rec	11/17/2005	Hydrocarbon	Oil and Grease	n/a	=	85	%	EPA 1664A		70	130	
2005/06-2	Lab	LCS, RPD	11/17/2005	Hydrocarbon	Oil and Grease	n/a	=	2	%	EPA 1664A		0	30	
2005/06-2	Lab	method blank	11/17/2005	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1		1	
2005/06-2	ME-VR2	field duplicate	11/17/2005	Hydrocarbon	Oil and Grease	n/a	=	1	mg/L	EPA 1664A	1			EST
2005/06-2	Lab	LCS dup, rec	11/21/2005	Hydrocarbon	TRPH	n/a	=	100	%	EPA 418.1		70	130	
2005/06-2	Lab	LCS, rec	11/21/2005	Hydrocarbon	TRPH	n/a	=	99	%	EPA 418.1		70	130	
2005/06-2	Lab	LCS, RPD	11/21/2005	Hydrocarbon	TRPH	n/a	=	1	%	EPA 418.1		0	30	
2005/06-2	Lab	method blank	11/21/2005	Hydrocarbon	TRPH	n/a	<	0.01	mg/L	EPA 418.1	0.01		0.01	
2005/06-2	ME-VR2	field duplicate	11/21/2005	Hydrocarbon	TRPH	n/a	=	0.2	mg/L	EPA 418.1	0.01			EST-FD
2005/06-2	Lab	method blank	12/2/2005	Metal	Aluminum	Dissolved	<	1	µg/L	EPA 200.8m	1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Aluminum	Dissolved	=	0.2	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Aluminum	Dissolved	=	11	µg/L	EPA 200.8m	1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Aluminum	Total	<	1	µg/L	EPA 200.8m	1			
2005/06-2	ME-CC	field blank	4/27/2006	Metal	Aluminum	Total	<	1	µg/L	EPA 200.8m	1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Aluminum	Total	=	0.3	%	EPA 200.8m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Aluminum	Total	=	119	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Aluminum	Total	=	121	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Aluminum	Total	=	2	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Aluminum	Total	=	124	µg/L	EPA 200.8m	1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Arsenic	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Arsenic	Dissolved	=	6	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Arsenic	Dissolved	=	3.78	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Arsenic	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-CC	field blank	4/27/2006	Metal	Arsenic	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Arsenic	Total	=	0.7	%	EPA 200.8m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Arsenic	Total	=	117	%	EPA 200.8m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Arsenic	Total	=	120	%	EPA 200.8m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Arsenic	Total	=	3	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Arsenic	Total	=	3.57	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Cadmium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Cadmium	Dissolved	=	3.5	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Cadmium	Dissolved	=	0.56	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Cadmium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-CC	field blank	4/27/2006	Metal	Cadmium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Cadmium	Total	=	0.3	%	EPA 200.8m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Cadmium	Total	=	97	%	EPA 200.8m		60	140	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Cadmium	Total	=	97	%	EPA 200.8m		60	140	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Cadmium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Cadmium	Total	=	0.61	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Chromium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Chromium	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Chromium	Dissolved	=	0.43	µg/L	EPA 200.8m	0.1			EST
2005/06-2	Lab	method blank	12/2/2005	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-CC	field blank	4/27/2006	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Chromium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Chromium	Total	=	112	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Chromium	Total	=	112	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Chromium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Chromium	Total	=	0.63	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	LCS dup, rec	11/11/2005	Metal	Chromium VI	Total	=	103	%	SM 3500-Cr		70	130	
2005/06-2	Lab	LCS, rec	11/11/2005	Metal	Chromium VI	Total	=	102	%	SM 3500-Cr		70	130	
2005/06-2	Lab	LCS, RPD	11/11/2005	Metal	Chromium VI	Total	=	1	%	SM 3500-Cr		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	Lab	method blank	11/11/2005	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr	5			
2005/06-2	ME-VR2	field duplicate	11/11/2005	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr	5			
2005/06-2	ME-VR2	matrix spike dup, rec	11/11/2005	Metal	Chromium VI	Total	=	92	%	SM 3500-Cr		70	130	
2005/06-2	ME-VR2	matrix spike, rec	11/11/2005	Metal	Chromium VI	Total	=	88	%	SM 3500-Cr		70	130	
2005/06-2	ME-VR2	matrix spike, RPD	11/11/2005	Metal	Chromium VI	Total	=	4.4	%	SM 3500-Cr		0	30	
2005/06-2	Lab	method blank	12/2/2005	Metal	Copper	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Copper	Dissolved	=	6.8	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Copper	Dissolved	=	7.29	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Copper	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-CC	field blank	4/27/2006	Metal	Copper	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Copper	Total	=	0.8	%	EPA 200.8m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Copper	Total	=	101	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Copper	Total	=	102	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Copper	Total	=	1	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Copper	Total	=	8.26	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Lead	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/5/2005	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05			EST
2005/06-2	Lab	method blank	12/2/2005	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05			
2005/06-2	ME-CC	field blank	4/27/2006	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Lead	Total	=	0.9	%	EPA 200.8m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Lead	Total	=	94	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Lead	Total	=	94	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Lead	Total	=	0	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Lead	Total	=	0.19	µg/L	EPA 200.8m	0.05			EST-FD
2005/06-2	Lab	method blank	11/11/2005	Metal	Mercury	Dissolved	<	0.5	ng/L	EPA 1631E	0.5			
2005/06-2	ME-VR2	field duplicate	11/11/2005	Metal	Mercury	Dissolved	=	7.34	ng/L	EPA 1631E	0.5			
2005/06-2	Lab	method blank	12/2/2005	Metal	Mercury	Total	<	0.5	ng/L	EPA 1631E	0.5			
2005/06-2	ME-CC	field blank	12/2/2005	Metal	Mercury	Total	=	1.13	ng/L	EPA 1631E	0.5			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Mercury	Total	=	4.4	%	EPA 1631E		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Mercury	Total	=	99	%	EPA 1631E		75	125	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Mercury	Total	=	100	%	EPA 1631E		75	125	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Mercury	Total	=	1	%	EPA 1631E		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Mercury	Total	=	13.2	ng/L	EPA 1631E	0.5			
2005/06-2	Lab	method blank	12/2/2005	Metal	Nickel	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Nickel	Dissolved	=	2	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Nickel	Dissolved	=	5.59	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Nickel	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-CC	field blank	4/27/2006	Metal	Nickel	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Nickel	Total	=	0	%	EPA 200.8m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Nickel	Total	=	104	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Nickel	Total	=	102	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Nickel	Total	=	2	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Nickel	Total	=	5.66	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Selenium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Selenium	Dissolved	=	6.8	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Selenium	Dissolved	=	6.45	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Selenium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-CC	field blank	4/27/2006	Metal	Selenium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Selenium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Selenium	Total	=	119	%	EPA 200.8m		40	160	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Selenium	Total	=	121	%	EPA 200.8m		40	160	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Selenium	Total	=	2	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Selenium	Total	=	5.53	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Silver	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Silver	Dissolved	=	0	%	EPA 200.8m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Silver	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Silver	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-CC	field blank	4/27/2006	Metal	Silver	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Silver	Total	=	0	%	EPA 200.8m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Silver	Total	=	99	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Silver	Total	=	95	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Silver	Total	=	4	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Silver	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Thallium	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-CC	field blank	4/27/2006	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Thallium	Total	=	3.6	%	EPA 200.8m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Thallium	Total	=	93	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Thallium	Total	=	94	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Thallium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Zinc	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Zinc	Dissolved	=	33.9	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Zinc	Dissolved	=	6.71	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	method blank	12/2/2005	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-CC	field blank	4/27/2006	Metal	Zinc	Total	=	3.6	µg/L	EPA 200.8m	0.1			
2005/06-2	ME-SCR	lab duplicate, RPD	12/2/2005	Metal	Zinc	Total	=	0.9	%	EPA 200.8m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	12/2/2005	Metal	Zinc	Total	=	97	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, rec	12/2/2005	Metal	Zinc	Total	=	98	%	EPA 200.8m		75	125	
2005/06-2	ME-SCR	matrix spike, RPD	12/2/2005	Metal	Zinc	Total	=	1	%	EPA 200.8m		0	30	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Metal	Zinc	Total	=	6.89	µg/L	EPA 200.8m	0.1			
2005/06-2	Lab	LCS dup, rec	11/11/2005	Nutrient	Ammonia as N	n/a	=	100	%	SM 4500-NH3 F		70	130	
2005/06-2	Lab	LCS, rec	11/11/2005	Nutrient	Ammonia as N	n/a	=	104	%	SM 4500-NH3 F		70	130	
2005/06-2	Lab	LCS, RPD	11/11/2005	Nutrient	Ammonia as N	n/a	=	4	%	SM 4500-NH3 F		0	30	
2005/06-2	Lab	method blank	11/11/2005	Nutrient	Ammonia as N	n/a	<	0.01	mg/L	SM 4500-NH3 F	0.01		0.01	
2005/06-2	ME-VR2	field duplicate	12/2/2005	Nutrient	Ammonia as N	n/a	=	0.01	mg/L	SM 4500-NH3 F	0.01			EST
2005/06-2	Lab	LCS dup, rec	11/11/2005	Nutrient	Nitrate as N	n/a	=	95	%	EPA 300.0		70	130	
2005/06-2	Lab	LCS, rec	11/11/2005	Nutrient	Nitrate as N	n/a	=	95	%	EPA 300.0		70	130	
2005/06-2	Lab	LCS, RPD	11/11/2005	Nutrient	Nitrate as N	n/a	=	0	%	EPA 300.0		0	30	
2005/06-2	Lab	method blank	11/11/2005	Nutrient	Nitrate as N	n/a	<	0.02	mg/L	EPA 300.0	0.02		0.02	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Nutrient	Nitrate as N	n/a	=	0.84	mg/L	EPA 300.0	0.02			
2005/06-2	ME-VR2	matrix spike dup, rec	11/11/2005	Nutrient	Nitrate as N	n/a	=	106	%	EPA 300.0		70	130	
2005/06-2	ME-VR2	matrix spike, rec	11/11/2005	Nutrient	Nitrate as N	n/a	=	105	%	EPA 300.0		70	130	
2005/06-2	ME-VR2	matrix spike, RPD	11/11/2005	Nutrient	Nitrate as N	n/a	=	1	%	EPA 300.0		0	30	
2005/06-2	Lab	LCS dup, rec	11/11/2005	Nutrient	Nitrite as N	n/a	=	91	%	EPA 300.0		70	130	
2005/06-2	Lab	LCS, rec	11/11/2005	Nutrient	Nitrite as N	n/a	=	99	%	EPA 300.0		70	130	
2005/06-2	Lab	LCS, RPD	11/11/2005	Nutrient	Nitrite as N	n/a	=	8	%	EPA 300.0		0	30	
2005/06-2	Lab	method blank	11/11/2005	Nutrient	Nitrite as N	n/a	<	0.02	mg/L	EPA 300.0	0.02		0.02	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Nutrient	Nitrite as N	n/a	=	0.05	mg/L	EPA 300.0	0.02			EST
2005/06-2	ME-VR2	matrix spike dup, rec	11/11/2005	Nutrient	Nitrite as N	n/a	=	83	%	EPA 300.0		70	130	
2005/06-2	ME-VR2	matrix spike, rec	11/11/2005	Nutrient	Nitrite as N	n/a	=	82	%	EPA 300.0		70	130	
2005/06-2	ME-VR2	matrix spike, RPD	11/11/2005	Nutrient	Nitrite as N	n/a	=	1	%	EPA 300.0		0	30	
2005/06-2	Lab	LCS dup, rec	11/11/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	94	%	EPA 300.0		70	130	
2005/06-2	Lab	LCS, rec	11/11/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	94	%	EPA 300.0		70	130	
2005/06-2	Lab	LCS, RPD	11/11/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	0	%	EPA 300.0		0	30	
2005/06-2	Lab	method blank	11/11/2005	Nutrient	Orthophosphate as P (Diss)	n/a	<	0.0075	mg/L	EPA 300.0	0.0075		0.0075	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	0.178	mg/L	EPA 300.0	0.0075			EST-FD
2005/06-2	ME-VR2	matrix spike dup, rec	11/11/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	82	%	EPA 300.0		70	130	
2005/06-2	ME-VR2	matrix spike, rec	11/11/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	108	%	EPA 300.0		70	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-VR2	matrix spike, RPD	11/11/2005	Nutrient	Orthophosphate as P (Diss)	n/a	=	27	%	EPA 300.0		0	30	
2005/06-2	Lab	method blank	11/17/2005	Nutrient	TKN	n/a	<	0.5	mg/L	EPA 351.3	0.5			
2005/06-2	ME-SCR	lab duplicate, RPD	11/17/2005	Nutrient	TKN	n/a	=	8	%	EPA 351.3		0	25	
2005/06-2	ME-VR2	field duplicate	11/17/2005	Nutrient	TKN	n/a	=	6	mg/L	EPA 351.3	0.5			
2005/06-2	Lab	LCS dup, rec	11/14/2005	Nutrient	Total Phosphorus	Dissolved	=	84	%	SM 4500-P C		70	130	
2005/06-2	Lab	LCS, rec	11/14/2005	Nutrient	Total Phosphorus	Dissolved	=	102	%	SM 4500-P C		70	130	
2005/06-2	Lab	LCS, RPD	11/14/2005	Nutrient	Total Phosphorus	Dissolved	=	19	%	SM 4500-P C		0	30	
2005/06-2	Lab	method blank	11/14/2005	Nutrient	Total Phosphorus	Dissolved	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2005/06-2	ME-VR2	matrix spike dup, rec	11/11/2005	Nutrient	Total Phosphorus	Dissolved	=	70	%	SM 4500-P C		70	130	
2005/06-2	ME-VR2	matrix spike, rec	11/11/2005	Nutrient	Total Phosphorus	Dissolved	=	70	%	SM 4500-P C		70	130	
2005/06-2	ME-VR2	matrix spike, RPD	11/11/2005	Nutrient	Total Phosphorus	Dissolved	=	0	%	SM 4500-P C		0	30	
2005/06-2	ME-VR2	field duplicate	11/14/2005	Nutrient	Total Phosphorus	Dissolved	=	0.756	mg/L	SM 4500-P C	0.016			
2005/06-2	Lab	LCS dup, rec	11/14/2005	Nutrient	Total Phosphorus	Total	=	84	%	SM 4500-P C		70	130	
2005/06-2	Lab	LCS, rec	11/14/2005	Nutrient	Total Phosphorus	Total	=	102	%	SM 4500-P C		70	130	
2005/06-2	Lab	LCS, RPD	11/14/2005	Nutrient	Total Phosphorus	Total	=	19	%	SM 4500-P C		0	30	
2005/06-2	Lab	method blank	11/14/2005	Nutrient	Total Phosphorus	Total	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2005/06-2	ME-VR2	field duplicate	11/14/2005	Nutrient	Total Phosphorus	Total	=	0.809	mg/L	SM 4500-P C	0.016			
2005/06-2	Lab	method blank	11/23/2005	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	1,2,4-Trichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	1,2,4-Trichlorobenzene	n/a	=	105	%	EPA 625m		44	142	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	1,2,4-Trichlorobenzene	n/a	=	106	%	EPA 625m		44	142	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	1,2,4-Trichlorobenzene	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	1,2-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	1,3-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	1,4-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	1,4-Dichlorobenzene	n/a	=	88	%	EPA 625m		20	124	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	1,4-Dichlorobenzene	n/a	=	86	%	EPA 625m		20	124	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	1,4-Dichlorobenzene	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	1-Methylnaphthalene	n/a	=	14.6	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	1-Methylnaphthalene	n/a	=	95	%	EPA 625m		50	120	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	1-Methylnaphthalene	n/a	=	92	%	EPA 625m		50	120	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	1-Methylnaphthalene	n/a	=	3	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	1-Methylnaphthalene	n/a	=	0.0098	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	1-Methylphenanthrene	n/a	=	30.8	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	1-Methylphenanthrene	n/a	=	94	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	1-Methylphenanthrene	n/a	=	91	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	1-Methylphenanthrene	n/a	=	3	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	=	120	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	=	104	%	EPA 625m		60	120	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	=	97	%	EPA 625m		60	120	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	=	7	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	srgt method blank, rec	11/23/2005	Organic	2,4,6-Tribromophenol	n/a	=	86	%	EPA 625m		10	130	
2005/06-2	ME-CC	srgt environ, rec	11/23/2005	Organic	2,4,6-Tribromophenol	n/a	=	94	%	EPA 625m		10	130	
2005/06-2	ME-CC	srgt field blank, rec	11/23/2005	Organic	2,4,6-Tribromophenol	n/a	=	82	%	EPA 625m		10	130	
2005/06-2	ME-SCR	srgt environ, rec	11/23/2005	Organic	2,4,6-Tribromophenol	n/a	=	95	%	EPA 625m		10	130	
2005/06-2	ME-SCR	srgt matrix spike dup, rec	11/23/2005	Organic	2,4,6-Tribromophenol	n/a	=	84	%	EPA 625m		10	130	
2005/06-2	ME-SCR	srgt matrix spike, rec	11/23/2005	Organic	2,4,6-Tribromophenol	n/a	=	91	%	EPA 625m		10	130	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	2,4,6-Tribromophenol	n/a	=	100	%	EPA 625m		10	130	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	2,4,6-Tribromophenol	n/a	=	100	%	EPA 625m		10	130	
2005/06-2	Lab	method blank	11/23/2005	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2,4,6-Trichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2,4-Dichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	srgt method blank, rec	11/16/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	121	%	EPA 8151A		0	123	
2005/06-2	ME-CC	srgt environ, rec	11/16/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	121	%	EPA 8151A		0	123	
2005/06-2	ME-SCR	srgt environ, rec	11/16/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	122	%	EPA 8151A		0	123	
2005/06-2	ME-VR2	srgt environ, rec	11/16/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	116	%	EPA 8151A		0	123	
2005/06-2	ME-VR2	srgt environ, rec	11/16/2005	Organic	2,4-Dichlorophenylacetic acid	n/a	=	122	%	EPA 8151A		0	123	
2005/06-2	Lab	method blank	11/23/2005	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2,4-Dimethylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	Lab	method blank	11/23/2005	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2,4-Dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	Lab	method blank	11/23/2005	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2,4-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	2,4-Dinitrotoluene	n/a	=	105	%	EPA 625m		39	139	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	2,4-Dinitrotoluene	n/a	=	105	%	EPA 625m		39	139	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	2,4-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2,6-Dimethylnaphthalene	n/a	=	109.7	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	2,6-Dimethylnaphthalene	n/a	=	107	%	EPA 625m		60	120	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	2,6-Dimethylnaphthalene	n/a	=	93	%	EPA 625m		60	120	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	2,6-Dimethylnaphthalene	n/a	=	14	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2,6-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2-Chloronaphthalene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2-Chlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	2-Chlorophenol	n/a	=	79	%	EPA 625m		23	134	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	2-Chlorophenol	n/a	=	83	%	EPA 625m		23	134	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	2-Chlorophenol	n/a	=	5	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2-Methyl-4,6-dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	Lab	method blank	11/23/2005	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2-Methylnaphthalene	n/a	=	1.4	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	2-Methylnaphthalene	n/a	=	101	%	EPA 625m		50	120	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	2-Methylnaphthalene	n/a	=	103	%	EPA 625m		50	120	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	2-Methylnaphthalene	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2-Methylnaphthalene	n/a	=	0.0215	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	2-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	Lab	method blank	11/23/2005	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	3,3'-Dichlorobenzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	4-Bromophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	4-Chloro-3-methylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	4-Chloro-3-methylphenol	n/a	=	99	%	EPA 625m		22	147	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	4-Chloro-3-methylphenol	n/a	=	97	%	EPA 625m		22	147	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	4-Chloro-3-methylphenol	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	Lab	method blank	11/23/2005	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	4-Chlorophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	4-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	4-Nitrophenol	n/a	=	21	%	EPA 625m		0.1	132	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	4-Nitrophenol	n/a	=	20	%	EPA 625m		0.1	132	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	4-Nitrophenol	n/a	=	5	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	Lab	method blank	11/23/2005	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Acenaphthene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Acenaphthene	n/a	=	87	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Acenaphthene	n/a	=	86	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Acenaphthene	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	srgt method blank, rec	11/23/2005	Organic	Acenaphthene-d10	n/a	=	92	%	EPA 625m		18	133	
2005/06-2	ME-CC	srgt environ, rec	11/23/2005	Organic	Acenaphthene-d10	n/a	=	99	%	EPA 625m		18	133	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-CC	srgt field blank, rec	11/23/2005	Organic	Acenaphthene-d10	n/a	=	82	%	EPA 625m		18	133	
2005/06-2	ME-SCR	srgt environ, rec	11/23/2005	Organic	Acenaphthene-d10	n/a	=	90	%	EPA 625m		18	133	
2005/06-2	ME-SCR	srgt matrix spike dup, rec	11/23/2005	Organic	Acenaphthene-d10	n/a	=	92	%	EPA 625m		18	133	
2005/06-2	ME-SCR	srgt matrix spike, rec	11/23/2005	Organic	Acenaphthene-d10	n/a	=	91	%	EPA 625m		18	133	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Acenaphthene-d10	n/a	=	84	%	EPA 625m		18	133	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Acenaphthene-d10	n/a	=	84	%	EPA 625m		18	133	
2005/06-2	Lab	method blank	11/23/2005	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Acenaphthylene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Acenaphthylene	n/a	=	94	%	EPA 625m		60	120	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Acenaphthylene	n/a	=	90	%	EPA 625m		60	120	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Acenaphthylene	n/a	=	4	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Anthracene	n/a	=	91	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Anthracene	n/a	=	87	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Anthracene	n/a	=	4	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Azobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Benzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Benzo(a)anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Benzo(a)anthracene	n/a	=	120	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Benzo(a)anthracene	n/a	=	98	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Benzo(a)anthracene	n/a	=	20	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Benzo(a)pyrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Benzo(a)pyrene	n/a	=	91	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Benzo(a)pyrene	n/a	=	86	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Benzo(a)pyrene	n/a	=	6	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Benzo(b)fluoranthene	n/a	=	142	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Benzo(b)fluoranthene	n/a	=	105	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Benzo(b)fluoranthene	n/a	=	97	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Benzo(b)fluoranthene	n/a	=	8	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Benzo(e)pyrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Benzo(e)pyrene	n/a	=	96	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Benzo(e)pyrene	n/a	=	95	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Benzo(e)pyrene	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	Lab	method blank	11/23/2005	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Benzo(g,h,i)perylene	n/a	=	33.3	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Benzo(g,h,i)perylene	n/a	=	91	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Benzo(g,h,i)perylene	n/a	=	89	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Benzo(g,h,i)perylene	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Benzo(k)fluoranthene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Benzo(k)fluoranthene	n/a	=	99	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Benzo(k)fluoranthene	n/a	=	94	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Benzo(k)fluoranthene	n/a	=	5	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Biphenyl	n/a	=	29.2	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Biphenyl	n/a	=	98	%	EPA 625m		50	120	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Biphenyl	n/a	=	92	%	EPA 625m		50	120	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Biphenyl	n/a	=	6	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Biphenyl	n/a	=	0.0013	µg/L	EPA 625m	0.001		0.001	EST
2005/06-2	Lab	method blank	11/23/2005	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Bis(2-chloroethoxy)methane	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Bis(2-chloroethyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Bis(2-chloroisopropyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.0309	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.0062	µg/L	EPA 625m	0.005		0.005	EST
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	47.6	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	124	%	EPA 625m		8	158	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	63	%	EPA 625m		8	158	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	65	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.321	µg/L	EPA 625m	0.005		0.005	EST-FD
2005/06-2	Lab	method blank	11/23/2005	Organic	Butyl benzyl phthalate	n/a	=	0.014	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Butyl benzyl phthalate	n/a	=	0.0134	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Butyl benzyl phthalate	n/a	=	25	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Butyl benzyl phthalate	n/a	=	107	%	EPA 625m		0.005	152	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Butyl benzyl phthalate	n/a	=	103	%	EPA 625m		0.005	152	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Butyl benzyl phthalate	n/a	=	4	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Butyl benzyl phthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	Lab	method blank	11/23/2005	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Chrysene	n/a	=	44.2	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Chrysene	n/a	=	110	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Chrysene	n/a	=	97	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Chrysene	n/a	=	13	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	srgt method blank, rec	11/23/2005	Organic	Chrysene-d12	n/a	=	108	%	EPA 625m		41	131	
2005/06-2	ME-CC	srgt environ, rec	11/23/2005	Organic	Chrysene-d12	n/a	=	104	%	EPA 625m		41	131	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-CC	srgt field blank, rec	11/23/2005	Organic	Chrysene-d12	n/a	=	102	%	EPA 625m		41	131	
2005/06-2	ME-SCR	srgt environ, rec	11/23/2005	Organic	Chrysene-d12	n/a	=	90	%	EPA 625m		41	131	
2005/06-2	ME-SCR	srgt matrix spike dup, rec	11/23/2005	Organic	Chrysene-d12	n/a	=	97	%	EPA 625m		41	131	
2005/06-2	ME-SCR	srgt matrix spike, rec	11/23/2005	Organic	Chrysene-d12	n/a	=	94	%	EPA 625m		41	131	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Chrysene-d12	n/a	=	94	%	EPA 625m		41	131	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Chrysene-d12	n/a	=	92	%	EPA 625m		41	131	
2005/06-2	Lab	method blank	11/23/2005	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Dibenz(a,h)anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Dibenz(a,h)anthracene	n/a	=	94	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Dibenz(a,h)anthracene	n/a	=	91	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Dibenz(a,h)anthracene	n/a	=	3	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Dibenzothiophene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Dibenzothiophene	n/a	=	99	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Dibenzothiophene	n/a	=	98	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Dibenzothiophene	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Organic	Diethyl phthalate	n/a	=	0.0632	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Diethyl phthalate	n/a	=	0.0639	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Diethyl phthalate	n/a	=	6.1	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Diethyl phthalate	n/a	=	102	%	EPA 625m		0.005	134	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Diethyl phthalate	n/a	=	103	%	EPA 625m		0.005	134	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Diethyl phthalate	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Diethyl phthalate	n/a	=	0.474	µg/L	EPA 625m	0.005		0.005	
2005/06-2	Lab	method blank	11/23/2005	Organic	Dimethyl phthalate	n/a	=	0.135	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Dimethyl phthalate	n/a	=	0.0754	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Dimethyl phthalate	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Dimethyl phthalate	n/a	=	102	%	EPA 625m		0.005	112	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Dimethyl phthalate	n/a	=	103	%	EPA 625m		0.005	112	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Dimethyl phthalate	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Dimethyl phthalate	n/a	=	0.115	µg/L	EPA 625m	0.005		0.005	
2005/06-2	Lab	method blank	11/23/2005	Organic	Di-n-butylphthalate	n/a	=	0.0285	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Di-n-butylphthalate	n/a	=	0.044	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Di-n-butylphthalate	n/a	=	1.2	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Di-n-butylphthalate	n/a	=	101	%	EPA 625m		1	118	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Di-n-butylphthalate	n/a	=	99	%	EPA 625m		1	118	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Di-n-butylphthalate	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Di-n-butylphthalate	n/a	=	0.046	µg/L	EPA 625m	0.005		0.005	
2005/06-2	Lab	method blank	11/23/2005	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Di-n-octylphthalate	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Di-n-octylphthalate	n/a	=	102	%	EPA 625m		4	146	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Di-n-octylphthalate	n/a	=	100	%	EPA 625m		4	146	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Di-n-octylphthalate	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	Lab	method blank	11/23/2005	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Fluoranthene	n/a	=	63.9	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Fluoranthene	n/a	=	97	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Fluoranthene	n/a	=	92	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Fluoranthene	n/a	=	5	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Fluoranthene	n/a	=	0.0036	µg/L	EPA 625m	0.001		0.001	EST, EST-FD
2005/06-2	Lab	method blank	11/23/2005	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Fluorene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Fluorene	n/a	=	97	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Fluorene	n/a	=	96	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Fluorene	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-2	Lab	method blank	11/23/2005	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Hexachlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-2	Lab	method blank	11/23/2005	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Hexachlorobutadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	Lab	method blank	11/23/2005	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Hexachlorocyclopentadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	Lab	method blank	11/23/2005	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Hexachloroethane	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	Lab	method blank	11/23/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	93	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	92	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-2	Lab	method blank	11/23/2005	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Isophorone	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	Lab	method blank	11/23/2005	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Naphthalene	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Naphthalene	n/a	=	90	%	EPA 625m		50	120	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Naphthalene	n/a	=	88	%	EPA 625m		50	120	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Naphthalene	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Naphthalene	n/a	=	0.103	µg/L	EPA 625m	0.001			0.001
2005/06-2	Lab	srgt method blank, rec	11/23/2005	Organic	Naphthalene-d8	n/a	=	85	%	EPA 625m		6	136	
2005/06-2	ME-CC	srgt environ, rec	11/23/2005	Organic	Naphthalene-d8	n/a	=	85	%	EPA 625m		6	136	
2005/06-2	ME-CC	srgt field blank, rec	11/23/2005	Organic	Naphthalene-d8	n/a	=	72	%	EPA 625m		6	136	
2005/06-2	ME-SCR	srgt environ, rec	11/23/2005	Organic	Naphthalene-d8	n/a	=	79	%	EPA 625m		6	136	
2005/06-2	ME-SCR	srgt matrix spike dup, rec	11/23/2005	Organic	Naphthalene-d8	n/a	=	82	%	EPA 625m		6	136	
2005/06-2	ME-SCR	srgt matrix spike, rec	11/23/2005	Organic	Naphthalene-d8	n/a	=	83	%	EPA 625m		6	136	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Naphthalene-d8	n/a	=	81	%	EPA 625m		6	136	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Naphthalene-d8	n/a	=	81	%	EPA 625m		6	136	
2005/06-2	Lab	method blank	11/23/2005	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Nitrobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	Lab	method blank	11/23/2005	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-CC	field blank	11/23/2005	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	N-Nitrosodimethylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			0.05

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	Lab	method blank	11/23/2005	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	N-Nitrosodi-N-propylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	N-Nitrosodi-N-propylamine	n/a	=	81	%	EPA 625m		60	140	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	N-Nitrosodi-N-propylamine	n/a	=	77	%	EPA 625m		60	140	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	N-Nitrosodi-N-propylamine	n/a	=	5	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	N-Nitrosodiphenylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Pentachlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Pentachlorophenol	n/a	=	73	%	EPA 625m		14	176	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Pentachlorophenol	n/a	=	67	%	EPA 625m		14	176	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Pentachlorophenol	n/a	=	9	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-2	Lab	method blank	11/23/2005	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Perylene	n/a	=	52.9	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Perylene	n/a	=	130	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Perylene	n/a	=	90	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Perylene	n/a	=	36	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	srgt method blank, rec	11/23/2005	Organic	Perylene-d12	n/a	=	100	%	EPA 625m		34	134	
2005/06-2	ME-CC	srgt environ, rec	11/23/2005	Organic	Perylene-d12	n/a	=	98	%	EPA 625m		34	134	
2005/06-2	ME-CC	srgt field blank, rec	11/23/2005	Organic	Perylene-d12	n/a	=	92	%	EPA 625m		34	134	
2005/06-2	ME-SCR	srgt environ, rec	11/23/2005	Organic	Perylene-d12	n/a	=	78	%	EPA 625m		34	134	
2005/06-2	ME-SCR	srgt matrix spike dup, rec	11/23/2005	Organic	Perylene-d12	n/a	=	90	%	EPA 625m		34	134	
2005/06-2	ME-SCR	srgt matrix spike, rec	11/23/2005	Organic	Perylene-d12	n/a	=	87	%	EPA 625m		34	134	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Perylene-d12	n/a	=	93	%	EPA 625m		34	134	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Perylene-d12	n/a	=	81	%	EPA 625m		34	134	
2005/06-2	Lab	method blank	11/23/2005	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Phenanthrene	n/a	=	39.6	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Phenanthrene	n/a	=	95	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Phenanthrene	n/a	=	87	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Phenanthrene	n/a	=	9	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	srgt method blank, rec	11/23/2005	Organic	Phenanthrene-d10	n/a	=	93	%	EPA 625m		43	124	
2005/06-2	ME-CC	srgt environ, rec	11/23/2005	Organic	Phenanthrene-d10	n/a	=	94	%	EPA 625m		43	124	
2005/06-2	ME-CC	srgt field blank, rec	11/23/2005	Organic	Phenanthrene-d10	n/a	=	91	%	EPA 625m		43	124	
2005/06-2	ME-SCR	srgt environ, rec	11/23/2005	Organic	Phenanthrene-d10	n/a	=	93	%	EPA 625m		43	124	
2005/06-2	ME-SCR	srgt matrix spike dup, rec	11/23/2005	Organic	Phenanthrene-d10	n/a	=	95	%	EPA 625m		43	124	
2005/06-2	ME-SCR	srgt matrix spike, rec	11/23/2005	Organic	Phenanthrene-d10	n/a	=	93	%	EPA 625m		43	124	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Phenanthrene-d10	n/a	=	95	%	EPA 625m		43	124	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Phenanthrene-d10	n/a	=	96	%	EPA 625m		43	124	
2005/06-2	Lab	method blank	11/23/2005	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Phenol	n/a	=	10	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Phenol	n/a	=	47	%	EPA 625m		5	112	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Phenol	n/a	=	50	%	EPA 625m		5	112	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Phenol	n/a	=	6	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Phenol	n/a	=	6.35	µg/L	EPA 625m	0.1		0.1	
2005/06-2	Lab	srgt method blank, rec	11/23/2005	Organic	Phenol-d5	n/a	=	41	%	EPA 625m		10	110	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-CC	srgt environ, rec	11/23/2005	Organic	Phenol-d5	n/a	=	41	%	EPA 625m		10	110	
2005/06-2	ME-CC	srgt field blank, rec	11/23/2005	Organic	Phenol-d5	n/a	=	28	%	EPA 625m		10	110	
2005/06-2	ME-SCR	srgt environ, rec	11/23/2005	Organic	Phenol-d5	n/a	=	42	%	EPA 625m		10	110	
2005/06-2	ME-SCR	srgt matrix spike dup, rec	11/23/2005	Organic	Phenol-d5	n/a	=	52	%	EPA 625m		10	110	
2005/06-2	ME-SCR	srgt matrix spike, rec	11/23/2005	Organic	Phenol-d5	n/a	=	66	%	EPA 625m		10	110	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Phenol-d5	n/a	=	59	%	EPA 625m		10	110	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Phenol-d5	n/a	=	58	%	EPA 625m		10	110	
2005/06-2	Lab	method blank	11/23/2005	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Pyrene	n/a	=	88.6	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Organic	Pyrene	n/a	=	104	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Organic	Pyrene	n/a	=	103	%	EPA 625m		70	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Organic	Pyrene	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Pyrene	n/a	=	0.0026	µg/L	EPA 625m	0.001		0.001	EST, EST-FD
2005/06-2	Lab	srgt method blank, rec	11/23/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	89	%	EPA 625m		40	110	
2005/06-2	ME-CC	srgt environ, rec	11/23/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	91	%	EPA 625m		40	110	
2005/06-2	ME-CC	srgt field blank, rec	11/23/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	83	%	EPA 625m		40	110	
2005/06-2	ME-SCR	srgt environ, rec	11/23/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	92	%	EPA 625m		40	110	
2005/06-2	ME-SCR	srgt matrix spike dup, rec	11/23/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	94	%	EPA 625m		40	110	
2005/06-2	ME-SCR	srgt matrix spike, rec	11/23/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	94	%	EPA 625m		40	110	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	91	%	EPA 625m		40	110	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	91	%	EPA 625m		40	110	
2005/06-2	Lab	method blank	11/23/2005	Organic	Total Detectable PAHs	n/a	=	0	µg/L	EPA 625m				
2005/06-2	ME-CC	field blank	11/23/2005	Organic	Total Detectable PAHs	n/a	=	0	µg/L	EPA 625m				
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Organic	Total Detectable PAHs	n/a	=	51.1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Organic	Total Detectable PAHs	n/a	=	0.1418	µg/L	EPA 625m				
2005/06-2	Lab	method blank	11/23/2005	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	Aroclor 1016	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	Aroclor 1221	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	Aroclor 1232	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	Aroclor 1242	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	Aroclor 1248	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	Aroclor 1254	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	Aroclor 1260	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 018	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 018	n/a	=	106	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 018	n/a	=	105	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 018	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 028	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 028	n/a	=	96	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 028	n/a	=	108	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 028	n/a	=	12	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	srgt method blank, rec	11/23/2005	PCB	PCB 030	n/a	=	89	%	EPA 625m		46	119	
2005/06-2	ME-CC	srgt environ, rec	11/23/2005	PCB	PCB 030	n/a	=	99	%	EPA 625m		46	119	
2005/06-2	ME-CC	srgt field blank, rec	11/23/2005	PCB	PCB 030	n/a	=	86	%	EPA 625m		46	119	
2005/06-2	ME-SCR	srgt environ, rec	11/23/2005	PCB	PCB 030	n/a	=	93	%	EPA 625m		46	119	
2005/06-2	ME-SCR	srgt matrix spike dup, rec	11/23/2005	PCB	PCB 030	n/a	=	98	%	EPA 625m		46	119	
2005/06-2	ME-SCR	srgt matrix spike, rec	11/23/2005	PCB	PCB 030	n/a	=	98	%	EPA 625m		46	119	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	PCB	PCB 030	n/a	=	96	%	EPA 625m		46	119	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	PCB	PCB 030	n/a	=	95	%	EPA 625m		46	119	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 031	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 031	n/a	=	99	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 031	n/a	=	96	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 031	n/a	=	3	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 033	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 033	n/a	=	109	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 033	n/a	=	108	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 033	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 037	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 037	n/a	=	102	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 037	n/a	=	108	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 037	n/a	=	6	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 044	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 044	n/a	=	102	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 044	n/a	=	108	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 044	n/a	=	6	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 049	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 049	n/a	=	107	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 049	n/a	=	101	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 049	n/a	=	6	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 052	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 052	n/a	=	103	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 052	n/a	=	107	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 052	n/a	=	4	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 066	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 066	n/a	=	106	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 066	n/a	=	106	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 066	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 070	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 070	n/a	=	108	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 070	n/a	=	107	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 070	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 074	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 074	n/a	=	106	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 074	n/a	=	104	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 074	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 077	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 077	n/a	=	95	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 077	n/a	=	94	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 077	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 081	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 081	n/a	=	94	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 081	n/a	=	108	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 081	n/a	=	14	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 087	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 087	n/a	=	105	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 087	n/a	=	97	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 087	n/a	=	8	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 095	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 095	n/a	=	109	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 095	n/a	=	101	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 095	n/a	=	8	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 097	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 097	n/a	=	95	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 097	n/a	=	101	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 097	n/a	=	6	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 099	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 099	n/a	=	94	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 099	n/a	=	98	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 099	n/a	=	4	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 101	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 101	n/a	=	107	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 101	n/a	=	105	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 101	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 105	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 105	n/a	=	93	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 105	n/a	=	97	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 105	n/a	=	4	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 110	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 110	n/a	=	103	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 110	n/a	=	102	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 110	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	srgt method blank, rec	11/23/2005	PCB	PCB 112	n/a	=	91	%	EPA 625m		52	123	
2005/06-2	ME-CC	srgt environ, rec	11/23/2005	PCB	PCB 112	n/a	=	88	%	EPA 625m		52	123	
2005/06-2	ME-CC	srgt field blank, rec	11/23/2005	PCB	PCB 112	n/a	=	88	%	EPA 625m		52	123	
2005/06-2	ME-SCR	srgt environ, rec	11/23/2005	PCB	PCB 112	n/a	=	92	%	EPA 625m		52	123	
2005/06-2	ME-SCR	srgt matrix spike dup, rec	11/23/2005	PCB	PCB 112	n/a	=	87	%	EPA 625m		52	123	
2005/06-2	ME-SCR	srgt matrix spike, rec	11/23/2005	PCB	PCB 112	n/a	=	91	%	EPA 625m		52	123	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	PCB	PCB 112	n/a	=	90	%	EPA 625m		52	123	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	PCB	PCB 112	n/a	=	91	%	EPA 625m		52	123	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 114	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 114	n/a	=	100	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 114	n/a	=	104	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 114	n/a	=	4	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 118	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 118	n/a	=	100	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 118	n/a	=	91	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 118	n/a	=	9	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 119	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 119	n/a	=	101	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 119	n/a	=	92	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 119	n/a	=	9	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 123	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 123	n/a	=	109	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 123	n/a	=	97	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 123	n/a	=	12	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 126	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 126	n/a	=	94	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 126	n/a	=	87	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 126	n/a	=	8	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 128 + 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 128 + 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 128 + 167	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 128 + 167	n/a	=	97	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 128 + 167	n/a	=	85	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 128 + 167	n/a	=	13	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 128 + 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 138	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 138	n/a	=	93	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 138	n/a	=	87	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 138	n/a	=	7	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 141	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 141	n/a	=	101	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 141	n/a	=	102	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 141	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 149	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 149	n/a	=	92	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 149	n/a	=	94	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 149	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 151	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 151	n/a	=	105	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 151	n/a	=	87	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 151	n/a	=	19	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 153	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 153	n/a	=	105	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 153	n/a	=	100	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 153	n/a	=	5	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 156	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 156	n/a	=	92	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 156	n/a	=	86	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 156	n/a	=	7	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 157	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 157	n/a	=	97	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 157	n/a	=	89	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 157	n/a	=	9	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 158	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 158	n/a	=	98	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 158	n/a	=	97	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 158	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 168 + 132	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 168 + 132	n/a	=	96	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 168 + 132	n/a	=	85	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 168 + 132	n/a	=	12	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 169	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 169	n/a	=	101	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 169	n/a	=	86	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 169	n/a	=	16	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 170	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 170	n/a	=	85	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 170	n/a	=	84	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 170	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 177	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 177	n/a	=	95	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 177	n/a	=	107	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 177	n/a	=	12	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 180	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 180	n/a	=	107	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 180	n/a	=	98	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 180	n/a	=	9	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 183	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 183	n/a	=	106	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 183	n/a	=	78	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 183	n/a	=	30.4	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 187	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 187	n/a	=	89	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 187	n/a	=	108	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 187	n/a	=	19	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 189	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 189	n/a	=	97	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 189	n/a	=	86	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 189	n/a	=	12	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 194	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 194	n/a	=	95	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 194	n/a	=	103	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 194	n/a	=	8	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	srgt method blank, rec	11/23/2005	PCB	PCB 198	n/a	=	97	%	EPA 625m		59	123	
2005/06-2	ME-CC	srgt environ, rec	11/23/2005	PCB	PCB 198	n/a	=	88	%	EPA 625m		59	123	
2005/06-2	ME-CC	srgt field blank, rec	11/23/2005	PCB	PCB 198	n/a	=	91	%	EPA 625m		59	123	
2005/06-2	ME-SCR	srgt environ, rec	11/23/2005	PCB	PCB 198	n/a	=	85	%	EPA 625m		59	123	
2005/06-2	ME-SCR	srgt matrix spike dup, rec	11/23/2005	PCB	PCB 198	n/a	=	84	%	EPA 625m		59	123	
2005/06-2	ME-SCR	srgt matrix spike, rec	11/23/2005	PCB	PCB 198	n/a	=	84	%	EPA 625m		59	123	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	PCB	PCB 198	n/a	=	92	%	EPA 625m		59	123	
2005/06-2	ME-VR2	srgt environ, rec	11/23/2005	PCB	PCB 198	n/a	=	84	%	EPA 625m		59	123	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 200	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 200	n/a	=	102	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 200	n/a	=	104	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 200	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 201	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 201	n/a	=	103	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 201	n/a	=	90	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 201	n/a	=	13	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	PCB 206	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	PCB	PCB 206	n/a	=	109	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	PCB	PCB 206	n/a	=	87	%	EPA 625m		65	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	PCB	PCB 206	n/a	=	22	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-2	ME-CC	field blank	11/23/2005	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	PCB	Total Detectable PCBs	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-2	Lab	LCS dup, rec	11/16/2005	Pesticide	2,4,5-T	n/a	=	128	%	EPA 8151A		30	130	
2005/06-2	Lab	LCS, rec	11/16/2005	Pesticide	2,4,5-T	n/a	=	116	%	EPA 8151A		30	130	
2005/06-2	Lab	LCS, RPD	11/16/2005	Pesticide	2,4,5-T	n/a	=	10	%	EPA 8151A		0	30	
2005/06-2	Lab	method blank	11/16/2005	Pesticide	2,4,5-T	n/a	<	0.5	µg/L	EPA 8151A	0.5	0	0.5	
2005/06-2	ME-SCR	matrix spike dup, rec	11/16/2005	Pesticide	2,4,5-T	n/a	=	100	%	EPA 8151A		30	130	
2005/06-2	ME-SCR	matrix spike, rec	11/16/2005	Pesticide	2,4,5-T	n/a	=	118	%	EPA 8151A		30	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/16/2005	Pesticide	2,4,5-T	n/a	=	16	%	EPA 8151A		0	30	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Pesticide	2,4,5-T	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-2	Lab	method blank	11/16/2005	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.5	µg/L	EPA 8151A	0.5	0	0.5	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-2	Lab	LCS dup, rec	11/16/2005	Pesticide	2,4-D	n/a	=	117	%	EPA 8151A		30	130	
2005/06-2	Lab	LCS, rec	11/16/2005	Pesticide	2,4-D	n/a	=	129	%	EPA 8151A		30	130	
2005/06-2	Lab	LCS, RPD	11/16/2005	Pesticide	2,4-D	n/a	=	9	%	EPA 8151A		0	30	
2005/06-2	Lab	method blank	11/16/2005	Pesticide	2,4-D	n/a	<	5	µg/L	EPA 8151A	5	0	5	
2005/06-2	ME-SCR	matrix spike dup, rec	11/16/2005	Pesticide	2,4-D	n/a	=	99	%	EPA 8151A		30	130	
2005/06-2	ME-SCR	matrix spike, rec	11/16/2005	Pesticide	2,4-D	n/a	=	117	%	EPA 8151A		30	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/16/2005	Pesticide	2,4-D	n/a	=	17	%	EPA 8151A		0	30	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Pesticide	2,4-D	n/a	<	5	µg/L	EPA 8151A	5		0.5	
2005/06-2	Lab	LCS dup, rec	11/16/2005	Pesticide	2,4-DB	n/a	=	71	%	EPA 8151A		30	130	
2005/06-2	Lab	LCS, rec	11/16/2005	Pesticide	2,4-DB	n/a	=	76	%	EPA 8151A		30	130	
2005/06-2	Lab	LCS, RPD	11/16/2005	Pesticide	2,4-DB	n/a	=	6	%	EPA 8151A		0	30	
2005/06-2	Lab	method blank	11/16/2005	Pesticide	2,4-DB	n/a	<	5	µg/L	EPA 8151A	5	0	0.5	
2005/06-2	ME-SCR	matrix spike dup, rec	11/16/2005	Pesticide	2,4-DB	n/a	=	101	%	EPA 8151A		30	130	
2005/06-2	ME-SCR	matrix spike, rec	11/16/2005	Pesticide	2,4-DB	n/a	=	113	%	EPA 8151A		30	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/16/2005	Pesticide	2,4-DB	n/a	=	11	%	EPA 8151A		0	30	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Pesticide	2,4-DB	n/a	<	5	µg/L	EPA 8151A	5		5	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	2,4'-DDD	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	2,4'-DDD	n/a	=	86	%	EPA 625m		56	129	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	2,4'-DDD	n/a	=	101	%	EPA 625m		56	129	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	2,4'-DDD	n/a	=	16	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	2,4'-DDE	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	2,4'-DDE	n/a	=	86	%	EPA 625m		60	129	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	2,4'-DDE	n/a	=	95	%	EPA 625m		60	129	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	2,4'-DDE	n/a	=	10	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	2,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	2,4'-DDT	n/a	=	67	%	EPA 625m		39	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	2,4'-DDT	n/a	=	99	%	EPA 625m		39	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	2,4'-DDT	n/a	=	39	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	4,4'-DDD	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	4,4'-DDD	n/a	=	102	%	EPA 625m		46	138	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	4,4'-DDD	n/a	=	99	%	EPA 625m		46	138	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	4,4'-DDD	n/a	=	3	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	4,4'-DDE	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	4,4'-DDE	n/a	=	98	%	EPA 625m		69	116	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	4,4'-DDE	n/a	=	105	%	EPA 625m		69	116	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	4,4'-DDE	n/a	=	7	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	4,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	4,4'-DDT	n/a	=	44	%	EPA 625m		34	136	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	4,4'-DDT	n/a	=	63	%	EPA 625m		34	136	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	4,4'-DDT	n/a	=	36	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Aldrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Aldrin	n/a	=	126	%	EPA 625m		43	128	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Aldrin	n/a	=	97	%	EPA 625m		43	128	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Aldrin	n/a	=	26	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	BHC-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	BHC-alpha	n/a	=	104	%	EPA 625m		60	123	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	BHC-alpha	n/a	=	97	%	EPA 625m		60	123	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	BHC-alpha	n/a	=	7	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	BHC-beta	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	BHC-beta	n/a	=	94	%	EPA 625m		45	140	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	BHC-beta	n/a	=	95	%	EPA 625m		45	140	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	BHC-beta	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	BHC-delta	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	BHC-delta	n/a	=	73	%	EPA 625m		29	113	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	BHC-delta	n/a	=	99	%	EPA 625m		29	113	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	BHC-delta	n/a	=	30	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	BHC-gamma (Lindane)	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	BHC-gamma (Lindane)	n/a	=	98	%	EPA 625m		59	110	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	BHC-gamma (Lindane)	n/a	=	104	%	EPA 625m		59	110	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	BHC-gamma (Lindane)	n/a	=	6	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Bolstar	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Bolstar	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Bolstar	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Bolstar	n/a	=	101	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Bolstar	n/a	=	102	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Bolstar	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Bolstar	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Chlordane-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Chlordane-alpha	n/a	=	97	%	EPA 625m		64	117	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Chlordane-alpha	n/a	=	101	%	EPA 625m		64	117	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Chlordane-alpha	n/a	=	4	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Chlordane-gamma	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Chlordane-gamma	n/a	=	104	%	EPA 625m		46	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Chlordane-gamma	n/a	=	96	%	EPA 625m		46	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Chlordane-gamma	n/a	=	8	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Chlorpyrifos	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Chlorpyrifos	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Chlorpyrifos	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Chlorpyrifos	n/a	=	102	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Chlorpyrifos	n/a	=	106	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Chlorpyrifos	n/a	=	4	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Chlorpyrifos	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	cis-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	cis-Nonachlor	n/a	=	101	%	EPA 625m		60	140	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	cis-Nonachlor	n/a	=	70	%	EPA 625m		60	140	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	cis-Nonachlor	n/a	=	36.3	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-2	Lab	method blank	11/16/2005	Pesticide	Dalapon	n/a	<	13	µg/L	EPA 8151A	13	0	13	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Pesticide	Dalapon	n/a	<	13	µg/L	EPA 8151A	13			
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Demeton-O	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Demeton-O	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Demeton-O	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Demeton-O	n/a	=	104	%	EPA 625m		45	105	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Demeton-O	n/a	=	103	%	EPA 625m		45	105	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Demeton-O	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Demeton-O	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Diazinon	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Diazinon	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Diazinon	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Diazinon	n/a	=	101	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Diazinon	n/a	=	103	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Diazinon	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Diazinon	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	Lab	method blank	11/16/2005	Pesticide	Dicamba	n/a	<	0.5	µg/L	EPA 8151A	0.5	0	0.5	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Pesticide	Dicamba	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	Lab	method blank	11/16/2005	Pesticide	Dichlorprop	n/a	<	5	µg/L	EPA 8151A	5	0	5	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Pesticide	Dichlorprop	n/a	<	5	µg/L	EPA 8151A	5		5	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Dichlorvos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Dichlorvos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Dichlorvos	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Dichlorvos	n/a	=	105	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Dichlorvos	n/a	=	106	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Dichlorvos	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Dichlorvos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Dieldrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Dieldrin	n/a	=	93	%	EPA 625m		46	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Dieldrin	n/a	=	99	%	EPA 625m		46	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Dieldrin	n/a	=	6	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Dimethoate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Dimethoate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Dimethoate	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Dimethoate	n/a	=	103	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Dimethoate	n/a	=	101	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Dimethoate	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Dimethoate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	Lab	method blank	11/16/2005	Pesticide	Dinoseb	n/a	<	2.5	µg/L	EPA 8151A	2.5	0	2.5	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Pesticide	Dinoseb	n/a	<	2.5	µg/L	EPA 8151A	2.5		2.5	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Disulfoton	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Disulfoton	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Disulfoton	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Disulfoton	n/a	=	98	%	EPA 625m		45	105	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Disulfoton	n/a	=	91	%	EPA 625m		45	105	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Disulfoton	n/a	=	7	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Disulfoton	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Endosulfan sulfate	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Endosulfan sulfate	n/a	=	81	%	EPA 625m		25	107	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Endosulfan sulfate	n/a	=	88	%	EPA 625m		25	107	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Endosulfan sulfate	n/a	=	8	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Endosulfan-I	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Endosulfan-I	n/a	=	102	%	EPA 625m		54	141	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Endosulfan-I	n/a	=	82	%	EPA 625m		54	141	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Endosulfan-I	n/a	=	22	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Endosulfan-II	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Endosulfan-II	n/a	=	99	%	EPA 625m		0.001	135	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Endosulfan-II	n/a	=	106	%	EPA 625m		0.001	135	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Endosulfan-II	n/a	=	7	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Endrin	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Endrin	n/a	=	97	%	EPA 625m		32	141	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Endrin	n/a	=	82	%	EPA 625m		32	141	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Endrin	n/a	=	17	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Endrin aldehyde	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Endrin ketone	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Endrin ketone	n/a	=	72	%	EPA 625m		50	130	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Endrin ketone	n/a	=	84	%	EPA 625m		50	130	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Endrin ketone	n/a	=	15	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Ethoprop	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Ethoprop	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Ethoprop	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Ethoprop	n/a	=	103	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Ethoprop	n/a	=	103	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Ethoprop	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Ethoprop	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Fenchlorophos (Ronnel)	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Fenchlorophos (Ronnel)	n/a	=	103	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Fenchlorophos (Ronnel)	n/a	=	100	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Fenchlorophos (Ronnel)	n/a	=	3	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Fensulfothion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Fensulfothion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Fensulfothion	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Fensulfothion	n/a	=	107	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Fensulfothion	n/a	=	107	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Fensulfothion	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Fensulfothion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Fenthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Fenthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Fenthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Fenthion	n/a	=	106	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Fenthion	n/a	=	105	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Fenthion	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Fenthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/18/2005	Pesticide	Glyphosate	n/a	<	6	µg/L	EPA 547	6	0	6	
2005/06-2	Lab	LCS, rec	11/22/2005	Pesticide	Glyphosate	n/a	=	105.8	%	EPA 547		82	115	
2005/06-2	Lab	LCS, rec	11/22/2005	Pesticide	Glyphosate	n/a	=	105.8	%	EPA 547		82	115	
2005/06-2	Lab	method blank	11/22/2005	Pesticide	Glyphosate	n/a	<	6	µg/L	EPA 547	6	83	6	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Heptachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Heptachlor	n/a	=	100	%	EPA 625m		43	122	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Heptachlor	n/a	=	89	%	EPA 625m		43	122	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Heptachlor	n/a	=	12	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Heptachlor epoxide	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Heptachlor epoxide	n/a	=	97	%	EPA 625m		56	122	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Heptachlor epoxide	n/a	=	119	%	EPA 625m		56	122	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Heptachlor epoxide	n/a	=	20	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Malathion	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Malathion	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Malathion	n/a	=	24.3	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Malathion	n/a	=	113	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Malathion	n/a	=	120	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Malathion	n/a	=	6	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Malathion	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-2	Lab	method blank	11/16/2005	Pesticide	MCPA	n/a	<	500	µg/L	EPA 8151A	500	0	500	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Pesticide	MCPA	n/a	<	500	µg/L	EPA 8151A	500		500	
2005/06-2	Lab	method blank	11/16/2005	Pesticide	MCPP	n/a	<	500	µg/L	EPA 8151A	500	0	500	
2005/06-2	ME-VR2	field duplicate	11/11/2005	Pesticide	MCPP	n/a	<	500	µg/L	EPA 8151A	500		500	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Merphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Merphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Merphos	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Merphos	n/a	=	100	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Merphos	n/a	=	103	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Merphos	n/a	=	3	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Merphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Methoxychlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Methoxychlor	n/a	=	28	%	EPA 625m		0.001	157	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Methoxychlor	n/a	=	53	%	EPA 625m		0.001	157	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Methoxychlor	n/a	=	62	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Methyl parathion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Methyl parathion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Methyl parathion	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Methyl parathion	n/a	=	108	%	EPA 625m		60	120	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Methyl parathion	n/a	=	107	%	EPA 625m		60	120	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Methyl parathion	n/a	=	1	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Methyl parathion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Mevinphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Mevinphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Mevinphos	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Mevinphos	n/a	=	102	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Mevinphos	n/a	=	105	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Mevinphos	n/a	=	3	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Mevinphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Mirex	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Mirex	n/a	=	102	%	EPA 625m		56	123	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Mirex	n/a	=	107	%	EPA 625m		56	123	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Mirex	n/a	=	5	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Oxychlorane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Oxychlorane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Oxychlorane	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Oxychlorane	n/a	=	99	%	EPA 625m		60	140	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Oxychlorane	n/a	=	102	%	EPA 625m		60	140	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Oxychlorthane	n/a	=	3	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Phorate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Phorate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Phorate	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Phorate	n/a	=	94	%	EPA 625m		45	105	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Phorate	n/a	=	92	%	EPA 625m		45	105	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Phorate	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Phorate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	101	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	106	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	5	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Tokuthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Tokuthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Tokuthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Tokuthion	n/a	=	105	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Tokuthion	n/a	=	99	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Tokuthion	n/a	=	6	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Tokuthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Total Detectable DDTs	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Toxaphene	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	trans-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	trans-Nonachlor	n/a	=	107	%	EPA 625m		47	143	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	trans-Nonachlor	n/a	=	99	%	EPA 625m		47	143	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	trans-Nonachlor	n/a	=	8	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-2	Lab	method blank	11/23/2005	Pesticide	Trichloronate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-CC	field blank	11/23/2005	Pesticide	Trichloronate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-2	ME-SCR	lab duplicate, RPD	11/23/2005	Pesticide	Trichloronate	n/a	=	0	%	EPA 625m		0	30	
2005/06-2	ME-SCR	matrix spike dup, rec	11/23/2005	Pesticide	Trichloronate	n/a	=	101	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, rec	11/23/2005	Pesticide	Trichloronate	n/a	=	99	%	EPA 625m		65	125	
2005/06-2	ME-SCR	matrix spike, RPD	11/23/2005	Pesticide	Trichloronate	n/a	=	2	%	EPA 625m		0	30	
2005/06-2	ME-VR2	field duplicate	11/23/2005	Pesticide	Trichloronate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-3	Lab	method blank	3/2/2006	Anion	Bromide	n/a	<	0.001	mg/L	SM 4500-Br	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Anion	Bromide	n/a	=	0	%	SM 4500-Br		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Anion	Bromide	n/a	<	0.001	mg/L	SM 4500-Br	0.001			
2005/06-3	Lab	LCS dup, rec	2/23/2006	Anion	Chloride	n/a	=	103	%	SM 4500-Cl E		70	130	
2005/06-3	Lab	LCS, rec	2/23/2006	Anion	Chloride	n/a	=	100	%	SM 4500-Cl E		70	130	
2005/06-3	Lab	LCS, RPD	2/23/2006	Anion	Chloride	n/a	=	3	%	SM 4500-Cl E		0	30	
2005/06-3	Lab	method blank	2/26/2006	Anion	Chloride	n/a	<	0.01	mg/L	SM 4500-Cl E	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	2/26/2006	Anion	Chloride	n/a	=	0	%	SM 4500-Cl E		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	2/26/2006	Anion	Chloride	n/a	=	72	%	SM 4500-Cl E		70	130	
2005/06-3	ME-CC	matrix spike, rec	2/26/2006	Anion	Chloride	n/a	=	81	%	SM 4500-Cl E		70	130	
2005/06-3	ME-CC	matrix spike, RPD	2/26/2006	Anion	Chloride	n/a	=	12	%	SM 4500-Cl E		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	Lab	LCS dup, rec	2/24/2006	Anion	Perchlorate	n/a	=	94	%	EPA 314.0		85	115	
2005/06-3	Lab	LCS, rec	2/24/2006	Anion	Perchlorate	n/a	=	92	%	EPA 314.0		85	115	
2005/06-3	Lab	LCS, RPD	2/24/2006	Anion	Perchlorate	n/a	=	2.2	%	EPA 314.0		0	15	
2005/06-3	Lab	method blank	2/24/2006	Anion	Perchlorate	n/a	<	2	µg/L	EPA 314.0	2		2	
2005/06-3	Lab	LCS dup, rec	2/27/2006	Anion	Perchlorate	n/a	=	105	%	EPA 314.0		85	115	
2005/06-3	Lab	LCS, rec	2/27/2006	Anion	Perchlorate	n/a	=	104	%	EPA 314.0		85	115	
2005/06-3	Lab	LCS, RPD	2/27/2006	Anion	Perchlorate	n/a	=	1	%	EPA 314.0		0	15	
2005/06-3	Lab	method blank	2/27/2006	Anion	Perchlorate	n/a	<	2	µg/L	EPA 314.0	2		2	
2005/06-3	ME-SCR	matrix spike dup, rec	2/27/2006	Anion	Perchlorate	n/a	=	112	%	EPA 314.0		80	120	
2005/06-3	ME-SCR	matrix spike, rec	2/27/2006	Anion	Perchlorate	n/a	=	110	%	EPA 314.0		80	120	
2005/06-3	ME-SCR	matrix spike, RPD	2/27/2006	Anion	Perchlorate	n/a	=	2	%	EPA 314.0		0	15	
2005/06-3	ME-VR2	matrix spike dup, rec	2/24/2006	Anion	Perchlorate	n/a	=	95	%	EPA 314.0		80	120	
2005/06-3	ME-VR2	matrix spike, rec	2/24/2006	Anion	Perchlorate	n/a	=	94	%	EPA 314.0		80	120	
2005/06-3	ME-VR2	matrix spike, RPD	2/24/2006	Anion	Perchlorate	n/a	=	1	%	EPA 314.0		0	15	
2005/06-3	ME-SCR	field blank	2/19/2006	Bacteriological	E. Coli	n/a	<	10	MPN/100 mL	MMO-MUG	10			
2005/06-3	ME-SCR	field blank	2/19/2006	Bacteriological	Enterococcus	n/a	<	10	MPN/100 mL	Enterolert	10			
2005/06-3	ME-SCR	field blank	2/19/2006	Bacteriological	Fecal Coliform	n/a	<	2	MPN/100 mL	SM 9221 E	2			
2005/06-3	ME-SCR	field blank	2/19/2006	Bacteriological	Total Coliform	n/a	<	10	MPN/100 mL	MMO-MUG	10			
2005/06-3	Lab	method blank	2/26/2006	Conventional	BOD	n/a	<	1	mg/L	EPA 405.1	1			
2005/06-3	Lab	method blank	2/22/2006	Conventional	Conductivity	n/a	<	1	µmhos/cm	SM 2510	1		1	
2005/06-3	ME-CC	lab duplicate, RPD	2/22/2006	Conventional	Conductivity	n/a	=	0	%	SM 2510		0	30	
2005/06-3	Lab	method blank	3/2/2006	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1			
2005/06-3	ME-SCR	field blank	3/2/2006	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1			
2005/06-3	ME-CC	lab duplicate, RPD	2/22/2006	Conventional	pH	n/a	=	1	%	EPA 150.1		0	30	
2005/06-3	Lab	LCS dup, rec	2/23/2006	Conventional	Total Dissolved Solids	n/a	=	101	%	SM 2540 C		70	130	
2005/06-3	Lab	LCS, rec	2/23/2006	Conventional	Total Dissolved Solids	n/a	=	100	%	SM 2540 C		70	130	
2005/06-3	Lab	LCS, RPD	2/23/2006	Conventional	Total Dissolved Solids	n/a	=	1	%	SM 2540 C		0	30	
2005/06-3	Lab	method blank	2/26/2006	Conventional	Total Dissolved Solids	n/a	<	5	mg/L	SM 2540 C	5			
2005/06-3	ME-CC	lab duplicate, RPD	2/26/2006	Conventional	Total Dissolved Solids	n/a	=	1	%	SM 2540 C		0	30	
2005/06-3	Lab	LCS, rec	2/23/2006	Conventional	Total Organic Carbon	n/a	=	95	%	EPA 415.1		80	120	
2005/06-3	Lab	method blank	2/23/2006	Conventional	Total Organic Carbon	n/a	<	0.5	mg/L	EPA 415.1	0.5		0.5	
2005/06-3	ME-CC	lab duplicate, RPD	2/23/2006	Conventional	Total Organic Carbon	n/a	=	0	%	EPA 415.1		0	25	
2005/06-3	ME-CC	matrix spike dup, rec	2/23/2006	Conventional	Total Organic Carbon	n/a	=	84	%	EPA 415.1		70	130	
2005/06-3	ME-CC	matrix spike, rec	2/23/2006	Conventional	Total Organic Carbon	n/a	=	80	%	EPA 415.1		70	130	
2005/06-3	ME-CC	matrix spike, RPD	2/23/2006	Conventional	Total Organic Carbon	n/a	=	4.9	%	EPA 415.1		0	25	
2005/06-3	Lab	method blank	2/24/2006	Conventional	Total Suspended Solids	n/a	<	4	mg/L	SM 2540 D	4			
2005/06-3	ME-CC	lab duplicate, RPD	2/24/2006	Conventional	Total Suspended Solids	n/a	=	19	%	SM 2540 D		0	30	
2005/06-3	Lab	method blank	2/22/2006	Conventional	Turbidity	n/a	<	1	NTU	EPA 180.1	1			
2005/06-3	ME-SCR	lab duplicate, RPD	2/22/2006	Conventional	Turbidity	n/a	=	1	%	EPA 180.1		0	30	
2005/06-3	Lab	LCS dup, rec	2/23/2006	Hydrocarbon	Oil and Grease	n/a	=	108	%	EPA 1664A		70	130	
2005/06-3	Lab	LCS, rec	2/23/2006	Hydrocarbon	Oil and Grease	n/a	=	103	%	EPA 1664A		70	130	
2005/06-3	Lab	LCS, RPD	2/23/2006	Hydrocarbon	Oil and Grease	n/a	=	4.7	%	EPA 1664A		0	30	
2005/06-3	Lab	method blank	3/7/2006	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1			
2005/06-3	Lab	LCS dup, rec	2/23/2006	Hydrocarbon	TRPH	n/a	=	108	%	EPA 1664		70	130	
2005/06-3	Lab	LCS, rec	2/23/2006	Hydrocarbon	TRPH	n/a	=	79	%	EPA 1664		70	130	
2005/06-3	Lab	LCS, RPD	2/23/2006	Hydrocarbon	TRPH	n/a	=	31	%	EPA 1664		0	30	
2005/06-3	Lab	method blank	3/14/2006	Hydrocarbon	TRPH	n/a	<	0.1	mg/L	EPA 1664	0.1			
2005/06-3	Lab	method blank	3/2/2006	Metal	Aluminum	Dissolved	<	1	µg/L	EPA 200.8m	1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Aluminum	Dissolved	=	39	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Aluminum	Dissolved	=	99	%	EPA 200.8m		50	140	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Aluminum	Dissolved	=	106	%	EPA 200.8m		50	140	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Aluminum	Dissolved	=	6.8	%	EPA 200.8m		0	30	
2005/06-3	Lab	method blank	3/2/2006	Metal	Aluminum	Total	<	1	µg/L	EPA 200.8m	1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Aluminum	Total	=	1	%	EPA 200.8m		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Aluminum	Total	<	1	µg/L	EPA 200.8m	1			
2005/06-3	Lab	method blank	3/2/2006	Metal	Arsenic	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Arsenic	Dissolved	=	2	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Arsenic	Dissolved	=	106	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Arsenic	Dissolved	=	110	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Arsenic	Dissolved	=	3.7	%	EPA 200.8m		0	30	
2005/06-3	Lab	method blank	3/2/2006	Metal	Arsenic	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Arsenic	Total	=	2	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Arsenic	Total	=	110	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Arsenic	Total	=	107	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Arsenic	Total	=	3	%	EPA 200.8m		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Arsenic	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	Lab	method blank	3/2/2006	Metal	Cadmium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Cadmium	Dissolved	=	24	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Cadmium	Dissolved	=	100	%	EPA 200.8m		75	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Cadmium	Dissolved	=	100	%	EPA 200.8m		75	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Cadmium	Dissolved	=	8.6	%	EPA 200.8m		0	30	
2005/06-3	Lab	method blank	3/2/2006	Metal	Cadmium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Cadmium	Total	=	5	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Cadmium	Total	=	113	%	EPA 200.8m		75	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Cadmium	Total	=	113	%	EPA 200.8m		75	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Cadmium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Cadmium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	Lab	method blank	3/2/2006	Metal	Chromium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Chromium	Dissolved	=	9	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Chromium	Dissolved	=	99	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Chromium	Dissolved	=	104	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Chromium	Dissolved	=	4.9	%	EPA 200.8m		0	30	
2005/06-3	Lab	method blank	3/2/2006	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Chromium	Total	=	4	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Chromium	Total	=	109	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Chromium	Total	=	110	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Chromium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	Lab	method blank	2/22/2006	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr	5			
2005/06-3	Lab	LCS dup, rec	2/23/2006	Metal	Chromium VI	Total	=	104	%	SM 3500-Cr		70	130	
2005/06-3	Lab	LCS, rec	2/23/2006	Metal	Chromium VI	Total	=	102	%	SM 3500-Cr		70	130	
2005/06-3	Lab	LCS, RPD	2/23/2006	Metal	Chromium VI	Total	=	1.9	%	SM 3500-Cr		0	30	
2005/06-3	ME-CC	lab duplicate, RPD	2/22/2006	Metal	Chromium VI	Total	=	0	%	SM 3500-Cr		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Chromium VI	Total	=	104	%	SM 3500-Cr		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Chromium VI	Total	=	97	%	SM 3500-Cr		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Chromium VI	Total	=	7	%	SM 3500-Cr		0	30	
2005/06-3	Lab	method blank	3/2/2006	Metal	Copper	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Copper	Dissolved	=	3	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Copper	Dissolved	=	90	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Copper	Dissolved	=	95	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Copper	Dissolved	=	5.4	%	EPA 200.8m		0	30	
2005/06-3	Lab	method blank	3/2/2006	Metal	Copper	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Copper	Total	=	1	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Copper	Total	=	107	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Copper	Total	=	106	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Copper	Total	=	1	%	EPA 200.8m		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Copper	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	Lab	method blank	3/2/2006	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Lead	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Lead	Dissolved	=	87	%	EPA 200.8m		65	135	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Lead	Dissolved	=	94	%	EPA 200.8m		65	135	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Lead	Dissolved	=	7.7	%	EPA 200.8m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	Lab	method blank	3/2/2006	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Lead	Total	=	3	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Lead	Total	=	98	%	EPA 200.8m		65	135	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Lead	Total	=	101	%	EPA 200.8m		65	135	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Lead	Total	=	3	%	EPA 200.8m		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05			
2005/06-3	Lab	method blank	2/22/2006	Metal	Mercury	Dissolved	<	0.5	ng/L	EPA 1631E	0.5			
2005/06-3	ME-CC	lab duplicate, RPD	2/22/2006	Metal	Mercury	Dissolved	=	13	%	EPA 1631E		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	2/22/2006	Metal	Mercury	Dissolved	=	96	%	EPA 1631E		60	140	
2005/06-3	ME-CC	matrix spike, rec	2/22/2006	Metal	Mercury	Dissolved	=	122	%	EPA 1631E		60	140	
2005/06-3	ME-CC	matrix spike, RPD	2/22/2006	Metal	Mercury	Dissolved	=	23.9	%	EPA 1631E		0	30	
2005/06-3	Lab	method blank	3/17/2006	Metal	Mercury	Total	<	0.5	ng/L	EPA 1631E	0.5			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Mercury	Total	=	4	%	EPA 1631E		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Mercury	Total	=	92	%	EPA 1631E		60	140	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Mercury	Total	=	75	%	EPA 1631E		60	140	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Mercury	Total	=	20	%	EPA 1631E		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Mercury	Total	=	2.16	ng/L	EPA 1631E	0.5			
2005/06-3	Lab	method blank	3/2/2006	Metal	Nickel	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Nickel	Dissolved	=	7	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Nickel	Dissolved	=	96	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Nickel	Dissolved	=	100	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Nickel	Dissolved	=	4.1	%	EPA 200.8m		0	30	
2005/06-3	Lab	method blank	3/2/2006	Metal	Nickel	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Nickel	Total	=	3	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Nickel	Total	=	118	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Nickel	Total	=	118	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Nickel	Total	=	0	%	EPA 200.8m		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Nickel	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	Lab	method blank	3/2/2006	Metal	Selenium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Selenium	Dissolved	=	6	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Selenium	Dissolved	=	110	%	EPA 200.8m		60	150	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Selenium	Dissolved	=	118	%	EPA 200.8m		60	150	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Selenium	Dissolved	=	7	%	EPA 200.8m		0	30	
2005/06-3	Lab	method blank	3/2/2006	Metal	Selenium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Selenium	Total	=	9	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Selenium	Total	=	120	%	EPA 200.8m		60	150	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Selenium	Total	=	114	%	EPA 200.8m		60	150	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Selenium	Total	=	5	%	EPA 200.8m		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Selenium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	Lab	method blank	3/2/2006	Metal	Silver	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Silver	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Silver	Dissolved	=	95	%	EPA 200.8m		50	155	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Silver	Dissolved	=	101	%	EPA 200.8m		50	155	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Silver	Dissolved	=	6.1	%	EPA 200.8m		0	30	
2005/06-3	Lab	method blank	3/2/2006	Metal	Silver	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Silver	Total	=	0	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Silver	Total	=	104	%	EPA 200.8m		50	155	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Silver	Total	=	108	%	EPA 200.8m		50	155	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Silver	Total	=	4	%	EPA 200.8m		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Silver	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	Lab	method blank	3/2/2006	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Thallium	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Thallium	Dissolved	=	96	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Thallium	Dissolved	=	94	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Thallium	Dissolved	=	2.1	%	EPA 200.8m		0	30	
2005/06-3	Lab	method blank	3/2/2006	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1			

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Thallium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Thallium	Total	=	96	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Thallium	Total	=	97	%	EPA 200.8m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Thallium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	Lab	method blank	3/2/2006	Metal	Zinc	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Zinc	Dissolved	=	6	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Zinc	Dissolved	=	110	%	EPA 200.8m		50	150	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Zinc	Dissolved	=	120	%	EPA 200.8m		50	150	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Zinc	Dissolved	=	8.7	%	EPA 200.8m		0	30	
2005/06-3	Lab	method blank	3/2/2006	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Metal	Zinc	Total	=	5	%	EPA 200.8m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Zinc	Total	=	97	%	EPA 200.8m		50	150	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Metal	Zinc	Total	=	94	%	EPA 200.8m		50	150	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Metal	Zinc	Total	=	3	%	EPA 200.8m		0	30	
2005/06-3	ME-SCR	field blank	3/2/2006	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-3	Lab	LCS dup, rec	2/23/2006	Nutrient	Ammonia as N	n/a	=	100	%	SM 4500-NH3 F		70	130	
2005/06-3	Lab	LCS, rec	2/23/2006	Nutrient	Ammonia as N	n/a	=	100	%	SM 4500-NH3 F		70	130	
2005/06-3	Lab	LCS, RPD	2/23/2006	Nutrient	Ammonia as N	n/a	=	0	%	SM 4500-NH3 F		0	30	
2005/06-3	Lab	method blank	2/23/2006	Nutrient	Ammonia as N	n/a	<	0.01	mg/L	SM 4500-NH3 F	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	2/25/2006	Nutrient	Ammonia as N	n/a	=	3	%	SM 4500-NH3 F		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	2/25/2006	Nutrient	Ammonia as N	n/a	=	96	%	SM 4500-NH3 F		70	130	
2005/06-3	ME-CC	matrix spike, rec	2/25/2006	Nutrient	Ammonia as N	n/a	=	100	%	SM 4500-NH3 F		70	130	
2005/06-3	ME-CC	matrix spike, RPD	2/25/2006	Nutrient	Ammonia as N	n/a	=	4	%	SM 4500-NH3 F		0	30	
2005/06-3	Lab	method blank	2/22/2006	Nutrient	Nitrate as N	n/a	<	0.02	mg/L	EPA 300.0	0.02			
2005/06-3	Lab	LCS dup, rec	2/23/2006	Nutrient	Nitrate as N	n/a	=	98	%	EPA 300.0		70	130	
2005/06-3	Lab	LCS, rec	2/23/2006	Nutrient	Nitrate as N	n/a	=	99	%	EPA 300.0		70	130	
2005/06-3	Lab	LCS, RPD	2/23/2006	Nutrient	Nitrate as N	n/a	=	1	%	EPA 300.0		0	30	
2005/06-3	ME-CC	lab duplicate, RPD	2/22/2006	Nutrient	Nitrate as N	n/a	=	2	%	EPA 300.0		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	2/22/2006	Nutrient	Nitrate as N	n/a	=	87	%	EPA 300.0		70	130	
2005/06-3	ME-CC	matrix spike, rec	2/22/2006	Nutrient	Nitrate as N	n/a	=	88	%	EPA 300.0		70	130	
2005/06-3	ME-CC	matrix spike, RPD	2/22/2006	Nutrient	Nitrate as N	n/a	=	1	%	EPA 300.0		0	30	
2005/06-3	Lab	method blank	2/22/2006	Nutrient	Nitrite as N	n/a	<	0.01	mg/L	EPA 300.0	0.01			
2005/06-3	Lab	LCS dup, rec	2/23/2006	Nutrient	Nitrite as N	n/a	=	96	%	EPA 300.0		70	130	
2005/06-3	Lab	LCS, rec	2/23/2006	Nutrient	Nitrite as N	n/a	=	97	%	EPA 300.0		70	130	
2005/06-3	Lab	LCS, RPD	2/23/2006	Nutrient	Nitrite as N	n/a	=	1	%	EPA 300.0		0	30	
2005/06-3	ME-CC	lab duplicate, RPD	2/22/2006	Nutrient	Nitrite as N	n/a	=	0	%	EPA 300.0		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	2/22/2006	Nutrient	Nitrite as N	n/a	=	88	%	EPA 300.0		70	130	
2005/06-3	ME-CC	matrix spike, rec	2/22/2006	Nutrient	Nitrite as N	n/a	=	88	%	EPA 300.0		70	130	
2005/06-3	ME-CC	matrix spike, RPD	2/22/2006	Nutrient	Nitrite as N	n/a	=	0	%	EPA 300.0		0	30	
2005/06-3	Lab	LCS dup, rec	2/23/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	106	%	EPA 300.0		70	130	
2005/06-3	Lab	LCS, rec	2/23/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	106	%	EPA 300.0		70	130	
2005/06-3	Lab	LCS, RPD	2/23/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	0	%	EPA 300.0		0	30	
2005/06-3	Lab	method blank	3/9/2006	Nutrient	Orthophosphate as P (Diss)	n/a	<	0.0075	mg/L	EPA 300.0	0.0075			
2005/06-3	ME-CC	lab duplicate, RPD	2/22/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	7.5	%	EPA 300.0		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	2/26/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	75	%	EPA 300.0		70	130	
2005/06-3	ME-CC	matrix spike, rec	2/26/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	71	%	EPA 300.0		70	130	
2005/06-3	ME-CC	matrix spike, RPD	2/26/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	5	%	EPA 300.0		0	30	
2005/06-3	Lab	LCS, rec	3/8/2006	Nutrient	TKN	n/a	=	96.5	%	EPA 351.1		80	120	
2005/06-3	Lab	method blank	3/8/2006	Nutrient	TKN	n/a	<	0.05	mg/L	EPA 351.1	0.05		0.05	
2005/06-3	ME-CC	lab duplicate, RPD	3/8/2006	Nutrient	TKN	n/a	=	2.7	%	EPA 351.1		0	20	
2005/06-3	ME-CC	matrix spike dup, rec	3/8/2006	Nutrient	TKN	n/a	=	89.2	%	EPA 351.1		80	120	
2005/06-3	ME-CC	matrix spike, rec	3/8/2006	Nutrient	TKN	n/a	=	88.6	%	EPA 351.1		80	120	
2005/06-3	ME-CC	matrix spike, RPD	3/8/2006	Nutrient	TKN	n/a	=	0.7	%	EPA 351.1		0	20	
2005/06-3	Lab	LCS dup, rec	2/23/2006	Nutrient	Total Phosphorus	Dissolved	=	100	%	SM 4500-P C		70	130	
2005/06-3	Lab	LCS, rec	2/23/2006	Nutrient	Total Phosphorus	Dissolved	=	104	%	SM 4500-P C		70	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	Lab	LCS, RPD	2/23/2006	Nutrient	Total Phosphorus	Dissolved	=	3.9	%	SM 4500-P C		0	30	
2005/06-3	Lab	method blank	2/24/2006	Nutrient	Total Phosphorus	Dissolved	<	0.016	mg/L	SM 4500-P C	0.016			
2005/06-3	ME-CC	lab duplicate, RPD	2/24/2006	Nutrient	Total Phosphorus	Dissolved	=	2	%	SM 4500-P C		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	2/24/2006	Nutrient	Total Phosphorus	Dissolved	=	70	%	SM 4500-P C		70	130	
2005/06-3	ME-CC	matrix spike, rec	2/24/2006	Nutrient	Total Phosphorus	Dissolved	=	96	%	SM 4500-P C		70	130	
2005/06-3	ME-CC	matrix spike, RPD	2/24/2006	Nutrient	Total Phosphorus	Dissolved	=	31	%	SM 4500-P C		0	30	
2005/06-3	Lab	LCS dup, rec	2/23/2006	Nutrient	Total Phosphorus	Total	=	100	%	SM 4500-P C		70	130	
2005/06-3	Lab	LCS, rec	2/23/2006	Nutrient	Total Phosphorus	Total	=	104	%	SM 4500-P C		70	130	
2005/06-3	Lab	LCS, RPD	2/23/2006	Nutrient	Total Phosphorus	Total	=	3.9	%	SM 4500-P C		0	30	
2005/06-3	Lab	method blank	2/24/2006	Nutrient	Total Phosphorus	Total	<	0.016	mg/L	SM 4500-P C	0.016			
2005/06-3	ME-CC	lab duplicate, RPD	2/24/2006	Nutrient	Total Phosphorus	Total	=	7	%	SM 4500-P C		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	2/24/2006	Nutrient	Total Phosphorus	Total	=	102	%	SM 4500-P C		70	130	
2005/06-3	ME-CC	matrix spike, rec	2/24/2006	Nutrient	Total Phosphorus	Total	=	102	%	SM 4500-P C		70	130	
2005/06-3	ME-CC	matrix spike, RPD	2/24/2006	Nutrient	Total Phosphorus	Total	=	0	%	SM 4500-P C		0	30	
2005/06-3	Lab	method blank	3/9/2006	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	104	%	EPA 625m		65	140	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	103	%	EPA 625m		65	140	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	10	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	1,2-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	1,2-Dichlorobenzene	n/a	=	0.028	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	1,3-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	1,4-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	1,4-Dichlorobenzene	n/a	=	93	%	EPA 625m		50	140	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	1,4-Dichlorobenzene	n/a	=	81	%	EPA 625m		50	140	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	1,4-Dichlorobenzene	n/a	=	14	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	1-Methylnaphthalene	n/a	=	39.3	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	1-Methylnaphthalene	n/a	=	100	%	EPA 625m		50	120	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	1-Methylnaphthalene	n/a	=	103	%	EPA 625m		50	120	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	1-Methylnaphthalene	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	1-Methylphenanthrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	1-Methylphenanthrene	n/a	=	96	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	1-Methylphenanthrene	n/a	=	104	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	1-Methylphenanthrene	n/a	=	8	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	112	%	EPA 625m		45	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	123	%	EPA 625m		45	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	9	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	srgt method blank, rec	3/9/2006	Organic	2,4,6-Tribromophenol	n/a	=	63	%	EPA 625m		10	130	
2005/06-3	ME-CC	srgt environ, rec	3/9/2006	Organic	2,4,6-Tribromophenol	n/a	=	75	%	EPA 625m		10	130	
2005/06-3	ME-CC	srgt matrix spike dup, rec	3/9/2006	Organic	2,4,6-Tribromophenol	n/a	=	93	%	EPA 625m		10	130	
2005/06-3	ME-CC	srgt matrix spike, rec	3/9/2006	Organic	2,4,6-Tribromophenol	n/a	=	84	%	EPA 625m		10	130	
2005/06-3	ME-SCR	srgt environ, rec	3/9/2006	Organic	2,4,6-Tribromophenol	n/a	=	70	%	EPA 625m		10	130	
2005/06-3	ME-SCR	srgt field blank, rec	3/9/2006	Organic	2,4,6-Tribromophenol	n/a	=	73	%	EPA 625m		10	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-VR2	srgt environ, rec	3/9/2006	Organic	2,4,6-Tribromophenol	n/a	=	89	%	EPA 625m		10	130	
2005/06-3	Lab	method blank	3/9/2006	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2,4,6-Trichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2,4-Dichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	srgt method blank, rec	2/27/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	86	%	EPA 8151A		0	123	
2005/06-3	ME-CC	srgt environ, rec	2/27/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	84	%	EPA 8151A		0	123	
2005/06-3	ME-SCR	srgt environ, rec	2/27/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	77	%	EPA 8151A		0	123	
2005/06-3	ME-VR2	srgt environ, rec	2/27/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	79	%	EPA 8151A		0	123	
2005/06-3	Lab	method blank	3/9/2006	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2,4-Dimethylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	Lab	method blank	3/9/2006	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2,4-Dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	Lab	method blank	3/9/2006	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2,4-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	2,4-Dinitrotoluene	n/a	=	102	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	2,4-Dinitrotoluene	n/a	=	92	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	2,4-Dinitrotoluene	n/a	=	10	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	48	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	104	%	EPA 625m		55	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	106	%	EPA 625m		55	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2,6-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2-Chloronaphthalene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2-Chlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	2-Chlorophenol	n/a	=	85	%	EPA 625m		35	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	2-Chlorophenol	n/a	=	77	%	EPA 625m		35	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	2-Chlorophenol	n/a	=	10	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2-Methyl-4,6-dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	Lab	method blank	3/9/2006	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2-Methylnaphthalene	n/a	=	37.3	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	2-Methylnaphthalene	n/a	=	98	%	EPA 625m		50	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	2-Methylnaphthalene	n/a	=	104	%	EPA 625m		50	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	2-Methylnaphthalene	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	2-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	Lab	method blank	3/9/2006	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	3,3'-Dichlorobenzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05			

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	Lab	method blank	3/9/2006	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	4-Bromophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	4-Chloro-3-methylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	4-Chloro-3-methylphenol	n/a	=	93	%	EPA 625m		30	150	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	4-Chloro-3-methylphenol	n/a	=	94	%	EPA 625m		30	150	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	4-Chloro-3-methylphenol	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	Lab	method blank	3/9/2006	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	4-Chlorophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	4-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	4-Nitrophenol	n/a	=	9	%	EPA 625m		0.1	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	4-Nitrophenol	n/a	=	18	%	EPA 625m		0.1	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	4-Nitrophenol	n/a	=	66.7	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	Lab	method blank	3/9/2006	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Acenaphthene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Acenaphthene	n/a	=	104	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Acenaphthene	n/a	=	112	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Acenaphthene	n/a	=	7	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	srgt method blank, rec	3/9/2006	Organic	Acenaphthene-d10	n/a	=	108	%	EPA 625m		52	125	
2005/06-3	ME-CC	srgt environ, rec	3/9/2006	Organic	Acenaphthene-d10	n/a	=	100	%	EPA 625m		52	125	
2005/06-3	ME-CC	srgt matrix spike dup, rec	3/9/2006	Organic	Acenaphthene-d10	n/a	=	104	%	EPA 625m		52	125	
2005/06-3	ME-CC	srgt matrix spike, rec	3/9/2006	Organic	Acenaphthene-d10	n/a	=	107	%	EPA 625m		52	125	
2005/06-3	ME-SCR	srgt environ, rec	3/9/2006	Organic	Acenaphthene-d10	n/a	=	103	%	EPA 625m		52	125	
2005/06-3	ME-SCR	srgt field blank, rec	3/9/2006	Organic	Acenaphthene-d10	n/a	=	99	%	EPA 625m		52	125	
2005/06-3	ME-VR2	srgt environ, rec	3/9/2006	Organic	Acenaphthene-d10	n/a	=	99	%	EPA 625m		52	125	
2005/06-3	Lab	method blank	3/9/2006	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Acenaphthylene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Acenaphthylene	n/a	=	108	%	EPA 625m		60	120	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Acenaphthylene	n/a	=	110	%	EPA 625m		60	120	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Acenaphthylene	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Anthracene	n/a	=	101	%	EPA 625m		60	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Anthracene	n/a	=	104	%	EPA 625m		60	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Anthracene	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Azobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Benzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Benzo(a)anthracene	n/a	=	183.9	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Benzo(a)anthracene	n/a	=	82	%	EPA 625m		70	140	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Benzo(a)anthracene	n/a	=	70	%	EPA 625m		70	140	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Benzo(a)anthracene	n/a	=	17	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Benzo(a)pyrene	n/a	=	186.6	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Benzo(a)pyrene	n/a	=	82	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Benzo(a)pyrene	n/a	=	58	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Benzo(a)pyrene	n/a	=	35	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Benzo(b)fluoranthene	n/a	=	184	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Benzo(b)fluoranthene	n/a	=	81	%	EPA 625m		60	140	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Benzo(b)fluoranthene	n/a	=	65	%	EPA 625m		60	140	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Benzo(b)fluoranthene	n/a	=	22	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Benzo(e)pyrene	n/a	=	114.7	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Benzo(e)pyrene	n/a	=	78	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Benzo(e)pyrene	n/a	=	59	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Benzo(e)pyrene	n/a	=	28	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Benzo(g,h,i)perylene	n/a	=	181.2	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Benzo(g,h,i)perylene	n/a	=	91	%	EPA 625m		50	140	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Benzo(g,h,i)perylene	n/a	=	73	%	EPA 625m		50	140	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Benzo(g,h,i)perylene	n/a	=	22	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Benzo(k)fluoranthene	n/a	=	187.5	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Benzo(k)fluoranthene	n/a	=	78	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Benzo(k)fluoranthene	n/a	=	69	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Benzo(k)fluoranthene	n/a	=	13	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Biphenyl	n/a	=	106.1	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Biphenyl	n/a	=	104	%	EPA 625m		50	120	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Biphenyl	n/a	=	108	%	EPA 625m		50	120	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Biphenyl	n/a	=	4	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Bis(2-chloroethoxy)methane	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Bis(2-chloroethyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Bis(2-chloroisopropyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.108	µg/L	EPA 625m	0.005			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	41.5	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	113	%	EPA 625m		20	190	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	182	%	EPA 625m		20	190	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	47	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.184	µg/L	EPA 625m	0.005			
2005/06-3	Lab	method blank	3/9/2006	Organic	Butyl benzyl phthalate	n/a	=	0.0226	µg/L	EPA 625m	0.005			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Butyl benzyl phthalate	n/a	=	3.7	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Butyl benzyl phthalate	n/a	=	114	%	EPA 625m		65	160	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Butyl benzyl phthalate	n/a	=	97	%	EPA 625m		65	160	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Butyl benzyl phthalate	n/a	=	16	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Butyl benzyl phthalate	n/a	=	0.0178	µg/L	EPA 625m	0.005			

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	Lab	method blank	3/9/2006	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Chrysene	n/a	=	175.4	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Chrysene	n/a	=	70	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Chrysene	n/a	=	66	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Chrysene	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	srgt method blank, rec	3/9/2006	Organic	Chrysene-d12	n/a	=	88	%	EPA 625m		61	126	
2005/06-3	ME-CC	srgt environ, rec	3/9/2006	Organic	Chrysene-d12	n/a	=	72	%	EPA 625m		61	126	
2005/06-3	ME-CC	srgt matrix spike dup, rec	3/9/2006	Organic	Chrysene-d12	n/a	=	89	%	EPA 625m		61	126	
2005/06-3	ME-CC	srgt matrix spike, rec	3/9/2006	Organic	Chrysene-d12	n/a	=	73	%	EPA 625m		61	126	
2005/06-3	ME-SCR	srgt environ, rec	3/9/2006	Organic	Chrysene-d12	n/a	=	75	%	EPA 625m		61	126	
2005/06-3	ME-SCR	srgt field blank, rec	3/9/2006	Organic	Chrysene-d12	n/a	=	87	%	EPA 625m		61	126	
2005/06-3	ME-VR2	srgt environ, rec	3/9/2006	Organic	Chrysene-d12	n/a	=	81	%	EPA 625m		61	126	
2005/06-3	Lab	method blank	3/9/2006	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Dibenz(a,h)anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Dibenz(a,h)anthracene	n/a	=	101	%	EPA 625m		60	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Dibenz(a,h)anthracene	n/a	=	73	%	EPA 625m		60	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Dibenz(a,h)anthracene	n/a	=	32	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Dibenzothiophene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Dibenzothiophene	n/a	=	117	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Dibenzothiophene	n/a	=	113	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Dibenzothiophene	n/a	=	3.5	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Diethyl phthalate	n/a	=	0.248	µg/L	EPA 625m	0.005			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Diethyl phthalate	n/a	=	16	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Diethyl phthalate	n/a	=	123	%	EPA 625m		50	150	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Diethyl phthalate	n/a	=	149	%	EPA 625m		50	150	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Diethyl phthalate	n/a	=	19	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Diethyl phthalate	n/a	=	0.178	µg/L	EPA 625m	0.005			
2005/06-3	Lab	method blank	3/9/2006	Organic	Dimethyl phthalate	n/a	=	0.214	µg/L	EPA 625m	0.005			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Dimethyl phthalate	n/a	=	8	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Dimethyl phthalate	n/a	=	92	%	EPA 625m		40	155	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Dimethyl phthalate	n/a	=	116	%	EPA 625m		40	155	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Dimethyl phthalate	n/a	=	23	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Dimethyl phthalate	n/a	=	0.12	µg/L	EPA 625m	0.005			
2005/06-3	Lab	method blank	3/9/2006	Organic	Di-n-butylphthalate	n/a	=	0.0551	µg/L	EPA 625m	0.005			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Di-n-butylphthalate	n/a	=	20.3	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Di-n-butylphthalate	n/a	=	90	%	EPA 625m		65	145	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Di-n-butylphthalate	n/a	=	96	%	EPA 625m		65	145	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Di-n-butylphthalate	n/a	=	7	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Di-n-butylphthalate	n/a	=	0.03	µg/L	EPA 625m	0.005			
2005/06-3	Lab	method blank	3/9/2006	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Di-n-octylphthalate	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Di-n-octylphthalate	n/a	=	104	%	EPA 625m		50	165	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Di-n-octylphthalate	n/a	=	99	%	EPA 625m		50	165	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Di-n-octylphthalate	n/a	=	5	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-3	Lab	method blank	3/9/2006	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Fluoranthene	n/a	=	170.3	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Fluoranthene	n/a	=	71	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Fluoranthene	n/a	=	66	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Fluoranthene	n/a	=	7	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001			

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Fluorene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Fluorene	n/a	=	105	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Fluorene	n/a	=	108	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Fluorene	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Hexachlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Hexachlorobutadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Hexachlorocyclopentadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Hexachloroethane	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	179.7	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	92	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	74	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	22	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Isophorone	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Naphthalene	n/a	=	22.6	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Naphthalene	n/a	=	93	%	EPA 625m		50	120	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Naphthalene	n/a	=	91	%	EPA 625m		50	120	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Naphthalene	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	srgt method blank, rec	3/9/2006	Organic	Naphthalene-d8	n/a	=	108	%	EPA 625m		47	110	
2005/06-3	ME-CC	srgt environ, rec	3/9/2006	Organic	Naphthalene-d8	n/a	=	99	%	EPA 625m		47	110	
2005/06-3	ME-CC	srgt matrix spike dup, rec	3/9/2006	Organic	Naphthalene-d8	n/a	=	100	%	EPA 625m		47	110	
2005/06-3	ME-CC	srgt matrix spike, rec	3/9/2006	Organic	Naphthalene-d8	n/a	=	98	%	EPA 625m		47	110	
2005/06-3	ME-SCR	srgt environ, rec	3/9/2006	Organic	Naphthalene-d8	n/a	=	97	%	EPA 625m		47	110	
2005/06-3	ME-SCR	srgt field blank, rec	3/9/2006	Organic	Naphthalene-d8	n/a	=	95	%	EPA 625m		47	110	
2005/06-3	ME-VR2	srgt environ, rec	3/9/2006	Organic	Naphthalene-d8	n/a	=	100	%	EPA 625m		47	110	
2005/06-3	Lab	method blank	3/9/2006	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Nitrobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	N-Nitrosodimethylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	102	%	EPA 625m		55	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	102	%	EPA 625m		55	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	N-Nitrosodiphenylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Pentachlorophenol	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Pentachlorophenol	n/a	=	97	%	EPA 625m		10	160	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Pentachlorophenol	n/a	=	55	%	EPA 625m		10	160	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Pentachlorophenol	n/a	=	55	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-3	Lab	method blank	3/9/2006	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Perylene	n/a	=	160.8	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Perylene	n/a	=	86	%	EPA 625m		65	135	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Perylene	n/a	=	74	%	EPA 625m		65	135	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Perylene	n/a	=	15	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	srgt method blank, rec	3/9/2006	Organic	Perylene-d12	n/a	=	77	%	EPA 625m		53	122	
2005/06-3	ME-CC	srgt environ, rec	3/9/2006	Organic	Perylene-d12	n/a	=	67	%	EPA 625m		53	122	
2005/06-3	ME-CC	srgt matrix spike dup, rec	3/9/2006	Organic	Perylene-d12	n/a	=	91	%	EPA 625m		53	122	
2005/06-3	ME-CC	srgt matrix spike, rec	3/9/2006	Organic	Perylene-d12	n/a	=	76	%	EPA 625m		53	122	
2005/06-3	ME-SCR	srgt environ, rec	3/9/2006	Organic	Perylene-d12	n/a	=	66	%	EPA 625m		53	122	
2005/06-3	ME-SCR	srgt field blank, rec	3/9/2006	Organic	Perylene-d12	n/a	=	86	%	EPA 625m		53	122	
2005/06-3	ME-VR2	srgt environ, rec	3/9/2006	Organic	Perylene-d12	n/a	=	79	%	EPA 625m		53	122	
2005/06-3	Lab	method blank	3/9/2006	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Phenanthrene	n/a	=	142.2	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Phenanthrene	n/a	=	83	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Phenanthrene	n/a	=	90	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Phenanthrene	n/a	=	8	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	srgt method blank, rec	3/9/2006	Organic	Phenanthrene-d10	n/a	=	100	%	EPA 625m		57	128	
2005/06-3	ME-CC	srgt environ, rec	3/9/2006	Organic	Phenanthrene-d10	n/a	=	92	%	EPA 625m		57	128	
2005/06-3	ME-CC	srgt matrix spike dup, rec	3/9/2006	Organic	Phenanthrene-d10	n/a	=	98	%	EPA 625m		57	128	
2005/06-3	ME-CC	srgt matrix spike, rec	3/9/2006	Organic	Phenanthrene-d10	n/a	=	102	%	EPA 625m		57	128	
2005/06-3	ME-SCR	srgt environ, rec	3/9/2006	Organic	Phenanthrene-d10	n/a	=	96	%	EPA 625m		57	128	
2005/06-3	ME-SCR	srgt field blank, rec	3/9/2006	Organic	Phenanthrene-d10	n/a	=	94	%	EPA 625m		57	128	
2005/06-3	ME-VR2	srgt environ, rec	3/9/2006	Organic	Phenanthrene-d10	n/a	=	98	%	EPA 625m		57	128	
2005/06-3	Lab	method blank	3/9/2006	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Phenol	n/a	=	32	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Phenol	n/a	=	28	%	EPA 625m		0.1	115	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Phenol	n/a	=	27	%	EPA 625m		0.1	115	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Phenol	n/a	=	4	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-3	Lab	srgt method blank, rec	3/9/2006	Organic	Phenol-d5	n/a	=	30	%	EPA 625m		10	110	
2005/06-3	ME-CC	srgt environ, rec	3/9/2006	Organic	Phenol-d5	n/a	=	14	%	EPA 625m		10	110	
2005/06-3	ME-CC	srgt matrix spike dup, rec	3/9/2006	Organic	Phenol-d5	n/a	=	26	%	EPA 625m		10	110	
2005/06-3	ME-CC	srgt matrix spike, rec	3/9/2006	Organic	Phenol-d5	n/a	=	24	%	EPA 625m		10	110	
2005/06-3	ME-SCR	srgt environ, rec	3/9/2006	Organic	Phenol-d5	n/a	=	21	%	EPA 625m		10	110	
2005/06-3	ME-SCR	srgt field blank, rec	3/9/2006	Organic	Phenol-d5	n/a	=	18	%	EPA 625m		10	110	
2005/06-3	ME-VR2	srgt environ, rec	3/9/2006	Organic	Phenol-d5	n/a	=	27	%	EPA 625m		10	110	
2005/06-3	Lab	method blank	3/9/2006	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Pyrene	n/a	=	182	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Organic	Pyrene	n/a	=	99	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Organic	Pyrene	n/a	=	85	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Organic	Pyrene	n/a	=	15	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	srgt method blank, rec	3/9/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	82	%	EPA 625m		40	130	
2005/06-3	ME-CC	srgt environ, rec	3/9/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	86	%	EPA 625m		40	130	
2005/06-3	ME-CC	srgt matrix spike dup, rec	3/9/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	99	%	EPA 625m		40	130	
2005/06-3	ME-CC	srgt matrix spike, rec	3/9/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	97	%	EPA 625m		40	130	
2005/06-3	ME-SCR	srgt environ, rec	3/9/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	77	%	EPA 625m		40	130	
2005/06-3	ME-SCR	srgt field blank, rec	3/9/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	79	%	EPA 625m		40	130	
2005/06-3	ME-VR2	srgt environ, rec	3/9/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	78	%	EPA 625m		40	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	Lab	method blank	3/9/2006	Organic	Total Detectable PAHs	n/a	=	0	µg/L	EPA 625m				
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Organic	Total Detectable PAHs	n/a	=	140.3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Organic	Total Detectable PAHs	n/a	=	0	µg/L	EPA 625m				
2005/06-3	Lab	method blank	3/9/2006	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	Aroclor 1016	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	Aroclor 1221	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	Aroclor 1232	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	Aroclor 1242	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	Aroclor 1248	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	Aroclor 1254	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	Aroclor 1260	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 018	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 018	n/a	=	116	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 018	n/a	=	113	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 018	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 028	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 028	n/a	=	105	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 028	n/a	=	100	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 028	n/a	=	5	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	srgt method blank, rec	3/9/2006	PCB	PCB 030	n/a	=	87	%	EPA 625m		40	130	
2005/06-3	ME-CC	srgt environ, rec	3/9/2006	PCB	PCB 030	n/a	=	86	%	EPA 625m		40	130	
2005/06-3	ME-CC	srgt matrix spike dup, rec	3/9/2006	PCB	PCB 030	n/a	=	98	%	EPA 625m		40	130	
2005/06-3	ME-CC	srgt matrix spike, rec	3/9/2006	PCB	PCB 030	n/a	=	98	%	EPA 625m		40	130	
2005/06-3	ME-SCR	srgt environ, rec	3/9/2006	PCB	PCB 030	n/a	=	75	%	EPA 625m		40	130	
2005/06-3	ME-SCR	srgt field blank, rec	3/9/2006	PCB	PCB 030	n/a	=	78	%	EPA 625m		40	130	
2005/06-3	ME-VR2	srgt environ, rec	3/9/2006	PCB	PCB 030	n/a	=	79	%	EPA 625m		40	130	
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 031	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 031	n/a	=	116	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 031	n/a	=	111	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 031	n/a	=	4	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 033	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 033	n/a	=	111	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 033	n/a	=	112	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 033	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001			

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 037	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 037	n/a	=	106	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 037	n/a	=	96	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 037	n/a	=	10	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 044	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 044	n/a	=	107	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 044	n/a	=	111	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 044	n/a	=	4	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 049	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 049	n/a	=	111	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 049	n/a	=	110	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 049	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 052	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 052	n/a	=	111	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 052	n/a	=	108	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 052	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 066	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 066	n/a	=	106	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 066	n/a	=	105	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 066	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 070	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 070	n/a	=	105	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 070	n/a	=	102	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 070	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 074	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 074	n/a	=	106	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 074	n/a	=	103	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 074	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 077	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 077	n/a	=	102	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 077	n/a	=	88	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 077	n/a	=	15	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 081	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 081	n/a	=	105	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 081	n/a	=	103	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 081	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 087	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 087	n/a	=	106	%	EPA 625m		60	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 087	n/a	=	113	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 087	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 095	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 095	n/a	=	112	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 095	n/a	=	121	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 095	n/a	=	8	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 097	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 097	n/a	=	103	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 097	n/a	=	103	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 097	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 099	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 099	n/a	=	108	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 099	n/a	=	115	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 099	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 101	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 101	n/a	=	107	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 101	n/a	=	102	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 101	n/a	=	5	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 105	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 105	n/a	=	100	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 105	n/a	=	100	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 105	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 110	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 110	n/a	=	102	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 110	n/a	=	101	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 110	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	srgt method blank, rec	3/9/2006	PCB	PCB 112	n/a	=	89	%	EPA 625m		60	120	
2005/06-3	ME-CC	srgt environ, rec	3/9/2006	PCB	PCB 112	n/a	=	82	%	EPA 625m		60	120	
2005/06-3	ME-CC	srgt matrix spike dup, rec	3/9/2006	PCB	PCB 112	n/a	=	99	%	EPA 625m		60	120	
2005/06-3	ME-CC	srgt matrix spike, rec	3/9/2006	PCB	PCB 112	n/a	=	100	%	EPA 625m		60	120	
2005/06-3	ME-SCR	srgt environ, rec	3/9/2006	PCB	PCB 112	n/a	=	69	%	EPA 625m		60	120	
2005/06-3	ME-SCR	srgt field blank, rec	3/9/2006	PCB	PCB 112	n/a	=	84	%	EPA 625m		60	120	
2005/06-3	ME-VR2	srgt environ, rec	3/9/2006	PCB	PCB 112	n/a	=	86	%	EPA 625m		60	120	
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 114	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 114	n/a	=	101	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 114	n/a	=	99	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 114	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 118	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 118	n/a	=	101	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 118	n/a	=	98	%	EPA 625m		60	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 118	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 119	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 119	n/a	=	104	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 119	n/a	=	106	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 119	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 123	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 123	n/a	=	104	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 123	n/a	=	108	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 123	n/a	=	4	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 126	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 126	n/a	=	96	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 126	n/a	=	82	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 126	n/a	=	16	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 128 + 167	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 128 + 167	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 128 + 167	n/a	=	102	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 128 + 167	n/a	=	96	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 128 + 167	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 128 + 167	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 138	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 138	n/a	=	95	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 138	n/a	=	96	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 138	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 141	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 141	n/a	=	94	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 141	n/a	=	99	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 141	n/a	=	5	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 149	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 149	n/a	=	103	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 149	n/a	=	107	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 149	n/a	=	4	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 151	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 151	n/a	=	104	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 151	n/a	=	106	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 151	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 153	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 153	n/a	=	101	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 153	n/a	=	108	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 153	n/a	=	7	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001			



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 156	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 156	n/a	=	115	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 156	n/a	=	89	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 156	n/a	=	25	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 157	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 157	n/a	=	104	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 157	n/a	=	98	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 157	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 158	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 158	n/a	=	98	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 158	n/a	=	100	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 158	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 168 + 132	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 168 + 132	n/a	=	103	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 168 + 132	n/a	=	103	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 168 + 132	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 169	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 169	n/a	=	101	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 169	n/a	=	90	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 169	n/a	=	11	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 170	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 170	n/a	=	92	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 170	n/a	=	83	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 170	n/a	=	10	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 177	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 177	n/a	=	96	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 177	n/a	=	98	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 177	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 180	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 180	n/a	=	90	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 180	n/a	=	85	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 180	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 183	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 183	n/a	=	95	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 183	n/a	=	104	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 183	n/a	=	9	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 187	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 187	n/a	=	92	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 187	n/a	=	105	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 187	n/a	=	13	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 189	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 189	n/a	=	94	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 189	n/a	=	84	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 189	n/a	=	11	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 194	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 194	n/a	=	94	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 194	n/a	=	92	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 194	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	srgt method blank, rec	3/9/2006	PCB	PCB 198	n/a	=	96	%	EPA 625m		60	120	
2005/06-3	ME-CC	srgt environ, rec	3/9/2006	PCB	PCB 198	n/a	=	82	%	EPA 625m		60	120	
2005/06-3	ME-CC	srgt matrix spike dup, rec	3/9/2006	PCB	PCB 198	n/a	=	100	%	EPA 625m		60	120	
2005/06-3	ME-CC	srgt matrix spike, rec	3/9/2006	PCB	PCB 198	n/a	=	99	%	EPA 625m		60	120	
2005/06-3	ME-SCR	srgt environ, rec	3/9/2006	PCB	PCB 198	n/a	=	77	%	EPA 625m		60	120	
2005/06-3	ME-SCR	srgt field blank, rec	3/9/2006	PCB	PCB 198	n/a	=	94	%	EPA 625m		60	120	
2005/06-3	ME-VR2	srgt environ, rec	3/9/2006	PCB	PCB 198	n/a	=	89	%	EPA 625m		60	120	
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 200	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 200	n/a	=	104	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 200	n/a	=	105	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 200	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 201	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 201	n/a	=	91	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 201	n/a	=	101	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 201	n/a	=	10	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	PCB 206	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	PCB	PCB 206	n/a	=	103	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	PCB	PCB 206	n/a	=	73	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	PCB	PCB 206	n/a	=	34	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	PCB	Total Detectable PCBs	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-3	Lab	LCS dup, rec	2/27/2006	Pesticide	2,4,5-T	n/a	=	90	%	EPA 8151A		30	130	
2005/06-3	Lab	LCS, rec	2/27/2006	Pesticide	2,4,5-T	n/a	=	115	%	EPA 8151A		30	130	
2005/06-3	Lab	LCS, RPD	2/27/2006	Pesticide	2,4,5-T	n/a	=	24	%	EPA 8151A		0	30	
2005/06-3	Lab	method blank	2/27/2006	Pesticide	2,4,5-T	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-3	ME-CC	matrix spike dup, rec	2/28/2006	Pesticide	2,4,5-T	n/a	=	112	%	EPA 8151A		30	130	
2005/06-3	ME-CC	matrix spike, rec	2/28/2006	Pesticide	2,4,5-T	n/a	=	115	%	EPA 8151A		30	130	
2005/06-3	ME-CC	matrix spike, RPD	2/28/2006	Pesticide	2,4,5-T	n/a	=	3	%	EPA 8151A		0	30	
2005/06-3	Lab	method blank	2/27/2006	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-3	Lab	LCS dup, rec	2/27/2006	Pesticide	2,4-D	n/a	=	80	%	EPA 8151A		30	130	
2005/06-3	Lab	LCS, rec	2/27/2006	Pesticide	2,4-D	n/a	=	102	%	EPA 8151A		30	130	
2005/06-3	Lab	LCS, RPD	2/27/2006	Pesticide	2,4-D	n/a	=	24	%	EPA 8151A		0	30	
2005/06-3	Lab	method blank	2/27/2006	Pesticide	2,4-D	n/a	<	5	µg/L	EPA 8151A	5		5	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	matrix spike dup, rec	2/28/2006	Pesticide	2,4-D	n/a	=	44	%	EPA 8151A		30	130	
2005/06-3	ME-CC	matrix spike, rec	2/28/2006	Pesticide	2,4-D	n/a	=	42	%	EPA 8151A		30	130	
2005/06-3	ME-CC	matrix spike, RPD	2/28/2006	Pesticide	2,4-D	n/a	=	5	%	EPA 8151A		0	30	
2005/06-3	Lab	LCS dup, rec	2/27/2006	Pesticide	2,4-DB	n/a	=	81	%	EPA 8151A		30	130	
2005/06-3	Lab	LCS, rec	2/27/2006	Pesticide	2,4-DB	n/a	=	106	%	EPA 8151A		30	130	
2005/06-3	Lab	LCS, RPD	2/27/2006	Pesticide	2,4-DB	n/a	=	28	%	EPA 8151A		0	30	
2005/06-3	Lab	method blank	2/27/2006	Pesticide	2,4-DB	n/a	<	5	µg/L	EPA 8151A	5		5	
2005/06-3	ME-CC	matrix spike dup, rec	2/28/2006	Pesticide	2,4-DB	n/a	=	71	%	EPA 8151A		30	130	
2005/06-3	ME-CC	matrix spike, rec	2/28/2006	Pesticide	2,4-DB	n/a	=	70	%	EPA 8151A		30	130	
2005/06-3	ME-CC	matrix spike, RPD	2/28/2006	Pesticide	2,4-DB	n/a	=	0	%	EPA 8151A		0	30	
2005/06-3	Lab	method blank	3/9/2006	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	2,4'-DDD	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	2,4'-DDD	n/a	=	77	%	EPA 625m		50	140	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	2,4'-DDD	n/a	=	76	%	EPA 625m		50	140	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	2,4'-DDD	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	2,4'-DDE	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	2,4'-DDE	n/a	=	71	%	EPA 625m		60	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	2,4'-DDE	n/a	=	72	%	EPA 625m		60	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	2,4'-DDE	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	2,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	2,4'-DDT	n/a	=	53	%	EPA 625m		40	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	2,4'-DDT	n/a	=	53	%	EPA 625m		40	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	2,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	4,4'-DDD	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	4,4'-DDD	n/a	=	102	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	4,4'-DDD	n/a	=	109	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	4,4'-DDD	n/a	=	7	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	4,4'-DDE	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	4,4'-DDE	n/a	=	94	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	4,4'-DDE	n/a	=	96	%	EPA 625m		70	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	4,4'-DDE	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	4,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	4,4'-DDT	n/a	=	61	%	EPA 625m		0.05	150	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	4,4'-DDT	n/a	=	60	%	EPA 625m		0.05	150	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	4,4'-DDT	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Aldrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Aldrin	n/a	=	99	%	EPA 625m		50	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Aldrin	n/a	=	103	%	EPA 625m		50	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Aldrin	n/a	=	4	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	BHC-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	BHC-alpha	n/a	=	99	%	EPA 625m		60	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	BHC-alpha	n/a	=	100	%	EPA 625m		60	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	BHC-alpha	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	BHC-beta	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	BHC-beta	n/a	=	95	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	BHC-beta	n/a	=	96	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	BHC-beta	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	BHC-delta	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	BHC-delta	n/a	=	94	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	BHC-delta	n/a	=	98	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	BHC-delta	n/a	=	4	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	92	%	EPA 625m		50	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	98	%	EPA 625m		50	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Bolstar	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Bolstar	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Bolstar	n/a	=	99	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Bolstar	n/a	=	108	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Bolstar	n/a	=	9	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Bolstar	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Chlordane-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Chlordane-alpha	n/a	=	97	%	EPA 625m		60	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Chlordane-alpha	n/a	=	104	%	EPA 625m		60	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Chlordane-alpha	n/a	=	7	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Chlordane-gamma	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Chlordane-gamma	n/a	=	97	%	EPA 625m		60	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Chlordane-gamma	n/a	=	106	%	EPA 625m		60	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Chlordane-gamma	n/a	=	9	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Chlorpyrifos	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Chlorpyrifos	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Chlorpyrifos	n/a	=	99	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Chlorpyrifos	n/a	=	105	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Chlorpyrifos	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Chlorpyrifos	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	cis-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	cis-Nonachlor	n/a	=	102	%	EPA 625m		60	120	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	cis-Nonachlor	n/a	=	108	%	EPA 625m		60	120	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	cis-Nonachlor	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	2/27/2006	Pesticide	Dalapon	n/a	<	13	µg/L	EPA 8151A	13		13	
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Demeton-O	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Demeton-O	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Demeton-O	n/a	=	83	%	EPA 625m		45	105	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Demeton-O	n/a	=	79	%	EPA 625m		45	105	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Demeton-O	n/a	=	5	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Demeton-O	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Diazinon	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Diazinon	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Diazinon	n/a	=	103	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Diazinon	n/a	=	104	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Diazinon	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Diazinon	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-3	Lab	method blank	2/27/2006	Pesticide	Dicamba	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-3	Lab	method blank	2/27/2006	Pesticide	Dichlorprop	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Dichlorvos	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Dichlorvos	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Dichlorvos	n/a	=	94	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Dichlorvos	n/a	=	89	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Dichlorvos	n/a	=	5	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Dichlorvos	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Dieldrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Dieldrin	n/a	=	92	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Dieldrin	n/a	=	99	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Dieldrin	n/a	=	7	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Dimethoate	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Dimethoate	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Dimethoate	n/a	=	86	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Dimethoate	n/a	=	81	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Dimethoate	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Dimethoate	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-3	Lab	method blank	2/27/2006	Pesticide	Dinoseb	n/a	<	2.5	µg/L	EPA 8151A	2.5		2.5	
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Disulfoton	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Disulfoton	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Disulfoton	n/a	=	90	%	EPA 625m		45	105	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Disulfoton	n/a	=	86	%	EPA 625m		45	105	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Disulfoton	n/a	=	5	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Disulfoton	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Endosulfan sulfate	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Endosulfan sulfate	n/a	=	88	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Endosulfan sulfate	n/a	=	91	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Endosulfan sulfate	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Endosulfan-I	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Endosulfan-I	n/a	=	99	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Endosulfan-I	n/a	=	98	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Endosulfan-I	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Endosulfan-II	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Endosulfan-II	n/a	=	95	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Endosulfan-II	n/a	=	84	%	EPA 625m		60	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Endosulfan-II	n/a	=	13	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Endrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Endrin	n/a	=	81	%	EPA 625m		65	135	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Endrin	n/a	=	96	%	EPA 625m		65	135	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Endrin	n/a	=	17	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Endrin aldehyde	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Endrin ketone	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Endrin ketone	n/a	=	74	%	EPA 625m		40	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Endrin ketone	n/a	=	67	%	EPA 625m		40	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Endrin ketone	n/a	=	10	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Ethoprop	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Ethoprop	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Ethoprop	n/a	=	99	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Ethoprop	n/a	=	97	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Ethoprop	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Ethoprop	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	104	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	100	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	4	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Fensulfothion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Fensulfothion	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Fensulfothion	n/a	=	101	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Fensulfothion	n/a	=	80	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Fensulfothion	n/a	=	23	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Fensulfothion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Fenthion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Fenthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Fenthion	n/a	=	99	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Fenthion	n/a	=	100	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Fenthion	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Fenthion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	LCS, rec	3/2/2006	Pesticide	Glyphosate	n/a	=	100	%	EPA 547		82	115	
2005/06-3	Lab	method blank	3/2/2006	Pesticide	Glyphosate	n/a	<	6	µg/L	EPA 547	6		6	
2005/06-3	ME-CC	lab duplicate, RPD	3/2/2006	Pesticide	Glyphosate	n/a	=	0	%	EPA 547		0	25	
2005/06-3	ME-CC	matrix spike dup, rec	3/2/2006	Pesticide	Glyphosate	n/a	=	105	%	EPA 547		82	115	
2005/06-3	ME-CC	matrix spike, rec	3/2/2006	Pesticide	Glyphosate	n/a	=	103	%	EPA 547		82	115	
2005/06-3	ME-CC	matrix spike, RPD	3/2/2006	Pesticide	Glyphosate	n/a	=	1.9	%	EPA 547		0	20	
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Heptachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Heptachlor	n/a	=	89	%	EPA 625m		45	135	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Heptachlor	n/a	=	89	%	EPA 625m		45	135	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Heptachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Heptachlor epoxide	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Heptachlor epoxide	n/a	=	101	%	EPA 625m		65	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Heptachlor epoxide	n/a	=	106	%	EPA 625m		65	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Heptachlor epoxide	n/a	=	5	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Malathion	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Malathion	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Malathion	n/a	=	106	%	EPA 625m		65	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Malathion	n/a	=	109	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Malathion	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Malathion	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-3	Lab	method blank	2/27/2006	Pesticide	MCPA	n/a	<	500	µg/L	EPA 8151A	500		500	
2005/06-3	Lab	method blank	2/27/2006	Pesticide	MCPA	n/a	<	500	µg/L	EPA 8151A	500		500	
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Merphos	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Merphos	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Merphos	n/a	=	103	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Merphos	n/a	=	94	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Merphos	n/a	=	9	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Merphos	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Methoxychlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Methoxychlor	n/a	=	57	%	EPA 625m		0.001	155	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Methoxychlor	n/a	=	53	%	EPA 625m		0.001	155	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Methoxychlor	n/a	=	7	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Methyl parathion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Methyl parathion	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Methyl parathion	n/a	=	103	%	EPA 625m		60	120	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Methyl parathion	n/a	=	97	%	EPA 625m		60	120	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Methyl parathion	n/a	=	6	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Methyl parathion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Mevinphos	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Mevinphos	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Mevinphos	n/a	=	102	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Mevinphos	n/a	=	98	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Mevinphos	n/a	=	4	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Mevinphos	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Mirex	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Mirex	n/a	=	97	%	EPA 625m		50	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Mirex	n/a	=	99	%	EPA 625m		50	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Mirex	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Oxychlordane	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Oxychlordane	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Oxychlordane	n/a	=	104	%	EPA 625m		60	120	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Oxychlordane	n/a	=	107	%	EPA 625m		60	120	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Oxychlordane	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Oxychlordane	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Phorate	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Phorate	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Phorate	n/a	=	92	%	EPA 625m		45	105	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Phorate	n/a	=	90	%	EPA 625m		45	105	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Phorate	n/a	=	2	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Phorate	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	104	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	101	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Tokuthion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Tokuthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Tokuthion	n/a	=	98	%	EPA 625m		65	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Tokuthion	n/a	=	101	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Tokuthion	n/a	=	3	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Tokuthion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Total Detectable DDTs	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Toxaphene	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	trans-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	trans-Nonachlor	n/a	=	101	%	EPA 625m		55	130	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	trans-Nonachlor	n/a	=	111	%	EPA 625m		55	130	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	trans-Nonachlor	n/a	=	9	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-3	Lab	method blank	3/9/2006	Pesticide	Trichloronate	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-3	ME-CC	lab duplicate, RPD	3/9/2006	Pesticide	Trichloronate	n/a	=	0	%	EPA 625m		0	30	
2005/06-3	ME-CC	matrix spike dup, rec	3/9/2006	Pesticide	Trichloronate	n/a	=	101	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, rec	3/9/2006	Pesticide	Trichloronate	n/a	=	100	%	EPA 625m		65	125	
2005/06-3	ME-CC	matrix spike, RPD	3/9/2006	Pesticide	Trichloronate	n/a	=	1	%	EPA 625m		0	30	
2005/06-3	ME-SCR	field blank	3/9/2006	Pesticide	Trichloronate	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	Lab	method blank	3/9/2006	Anion	Bromide	n/a	<	0.001	mg/L	SM 4500-Br	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Anion	Bromide	n/a	=	0	%	SM 4500-Br		0	30	
2005/06-4	Lab	LCS dup, rec	3/2/2006	Anion	Chloride	n/a	=	93	%	SM 4500-Cl E		70	130	
2005/06-4	Lab	LCS, rec	3/2/2006	Anion	Chloride	n/a	=	89	%	SM 4500-Cl E		70	130	
2005/06-4	Lab	LCS, RPD	3/2/2006	Anion	Chloride	n/a	=	5	%	SM 4500-Cl E		0	30	
2005/06-4	Lab	method blank	3/2/2006	Anion	Chloride	n/a	<	0.01	mg/L	SM 4500-Cl E	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/2/2006	Anion	Chloride	n/a	=	7.4	%	SM 4500-Cl E		0	30	
2005/06-4	Lab	LCS dup, rec	3/6/2006	Anion	Perchlorate	n/a	=	109	%	EPA 314.0		85	115	
2005/06-4	Lab	LCS, rec	3/6/2006	Anion	Perchlorate	n/a	=	107	%	EPA 314.0		85	115	
2005/06-4	Lab	LCS, RPD	3/6/2006	Anion	Perchlorate	n/a	=	2	%	EPA 314.0		0	15	
2005/06-4	Lab	method blank	3/6/2006	Anion	Perchlorate	n/a	<	2	µg/L	EPA 314.0	2		2	
2005/06-4	ME-SCR	lab duplicate, RPD	3/6/2006	Anion	Perchlorate	n/a	=	0	%	EPA 314.0		0	25	
2005/06-4	ME-SCR	lab duplicate, RPD	2/27/2006	Bacteriological	E. Coli	n/a	=	4.1	%	MMO-MUG		0	30	
2005/06-4	ME-SCR	lab duplicate, RPD	2/27/2006	Bacteriological	Enterococcus	n/a	=	0	%	Enterolert		0	30	
2005/06-4	ME-SCR	lab duplicate, RPD	2/27/2006	Bacteriological	Fecal Coliform	n/a	=	6.5	%	SM 9221 E		0	30	
2005/06-4	ME-SCR	lab duplicate, RPD	2/27/2006	Bacteriological	Total Coliform	n/a	=	4	%	MMO-MUG		0	30	
2005/06-4	Lab	method blank	3/1/2006	Conventional	BOD	n/a	<	1	mg/L	EPA 405.1	1		1	
2005/06-4	ME-SCR	lab duplicate, RPD	3/1/2006	Conventional	BOD	n/a	=	5.4	%	EPA 405.1		0	20	
2005/06-4	Lab	method blank	3/1/2006	Conventional	Conductivity	n/a	<	1	µmhos/cm	SM 2510	1		1	
2005/06-4	ME-SCR	lab duplicate, RPD	3/1/2006	Conventional	Conductivity	n/a	=	0	%	SM 2510		0	30	
2005/06-4	Lab	method blank	3/9/2006	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Conventional	Hardness as CaCO3	Total	=	37.6	%	SM 2340 B		0	30	
2005/06-4	ME-SCR	lab duplicate, RPD	3/1/2006	Conventional	pH	n/a	=	2.6	%	EPA 150.1		0	30	
2005/06-4	Lab	LCS dup, rec	3/5/2006	Conventional	Total Dissolved Solids	n/a	=	99	%	SM 2540 C		70	130	
2005/06-4	Lab	LCS, rec	3/5/2006	Conventional	Total Dissolved Solids	n/a	=	100	%	SM 2540 C		70	130	
2005/06-4	Lab	LCS, RPD	3/5/2006	Conventional	Total Dissolved Solids	n/a	=	1	%	SM 2540 C		0	30	
2005/06-4	Lab	method blank	3/5/2006	Conventional	Total Dissolved Solids	n/a	<	0.1	mg/L	SM 2540 C	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/5/2006	Conventional	Total Dissolved Solids	n/a	=	17.2	%	SM 2540 C		0	30	
2005/06-4	Lab	LCS dup, rec	3/6/2006	Conventional	Total Organic Carbon	n/a	=	106	%	EPA 415.1		80	120	
2005/06-4	Lab	LCS, rec	3/6/2006	Conventional	Total Organic Carbon	n/a	=	88	%	EPA 415.1		80	120	
2005/06-4	Lab	LCS, RPD	3/6/2006	Conventional	Total Organic Carbon	n/a	=	18	%	EPA 415.1		0	20	
2005/06-4	Lab	method blank	3/6/2006	Conventional	Total Organic Carbon	n/a	<	0.5	mg/L	EPA 415.1	0.5		0.5	
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Conventional	Total Organic Carbon	n/a	=	37.3	%	EPA 415.1		0	20	
2005/06-4	Lab	method blank	3/3/2006	Conventional	Total Suspended Solids	n/a	<	0.5	mg/L	SM 2540 D	0.5			
2005/06-4	ME-SCR	lab duplicate, RPD	3/3/2006	Conventional	Total Suspended Solids	n/a	=	60.6	%	SM 2540 D		0	30	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	Lab	method blank	3/3/2006	Conventional	Turbidity	n/a	<	1	NTU	EPA 180.1	1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/3/2006	Conventional	Turbidity	n/a	=	80.3	%	EPA 180.1		0	30	
2005/06-4	Lab	LCS dup, rec	3/16/2006	Hydrocarbon	Oil and Grease	n/a	=	108	%	EPA 1664A		70	130	
2005/06-4	Lab	LCS, rec	3/16/2006	Hydrocarbon	Oil and Grease	n/a	=	118	%	EPA 1664A		70	130	
2005/06-4	Lab	LCS, RPD	3/16/2006	Hydrocarbon	Oil and Grease	n/a	=	9	%	EPA 1664A		0	30	
2005/06-4	Lab	method blank	3/16/2006	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/16/2006	Hydrocarbon	Oil and Grease	n/a	=	40	%	EPA 1664A		0	30	
2005/06-4	Lab	LCS dup, rec	3/14/2006	Hydrocarbon	TRPH	n/a	=	108	%	EPA 1664		70	130	
2005/06-4	Lab	LCS, rec	3/14/2006	Hydrocarbon	TRPH	n/a	=	82	%	EPA 1664		70	130	
2005/06-4	Lab	LCS, RPD	3/14/2006	Hydrocarbon	TRPH	n/a	=	28	%	EPA 1664		0	30	
2005/06-4	Lab	method blank	3/14/2006	Hydrocarbon	TRPH	n/a	<	0.01	mg/L	EPA 1664	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/14/2006	Hydrocarbon	TRPH	n/a	=	141.7	%	EPA 1664		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Aluminum	Dissolved	<	1	µg/L	EPA 200.8m	1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Aluminum	Dissolved	=	37.9	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Aluminum	Total	<	1	µg/L	EPA 200.8m	1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Aluminum	Total	=	157	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Arsenic	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Arsenic	Dissolved	=	45.2	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Arsenic	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Arsenic	Total	=	70	%	EPA 200.8m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/9/2006	Metal	Arsenic	Total	=	113	%	EPA 200.8m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/9/2006	Metal	Arsenic	Total	=	114	%	EPA 200.8m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/9/2006	Metal	Arsenic	Total	=	1	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Cadmium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Cadmium	Dissolved	=	46	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Cadmium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Cadmium	Total	=	65.5	%	EPA 200.8m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/9/2006	Metal	Cadmium	Total	=	99	%	EPA 200.8m		75	130	
2005/06-4	ME-VR2	matrix spike, rec	3/9/2006	Metal	Cadmium	Total	=	101	%	EPA 200.8m		75	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/9/2006	Metal	Cadmium	Total	=	2	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Chromium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Chromium	Dissolved	=	45	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Chromium	Total	=	31.3	%	EPA 200.8m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/9/2006	Metal	Chromium	Total	=	114	%	EPA 200.8m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/9/2006	Metal	Chromium	Total	=	114	%	EPA 200.8m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/9/2006	Metal	Chromium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-4	Lab	LCS dup, rec	3/2/2006	Metal	Chromium VI	Total	=	110	%	SM 3500-Cr		70	130	
2005/06-4	Lab	LCS, rec	3/2/2006	Metal	Chromium VI	Total	=	103	%	SM 3500-Cr		70	130	
2005/06-4	Lab	LCS, RPD	3/2/2006	Metal	Chromium VI	Total	=	7	%	SM 3500-Cr		0	30	
2005/06-4	Lab	method blank	3/2/2006	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr	5			
2005/06-4	ME-CC	matrix spike dup, rec	3/2/2006	Metal	Chromium VI	Total	=	102	%	SM 3500-Cr		70	130	
2005/06-4	ME-CC	matrix spike, rec	3/2/2006	Metal	Chromium VI	Total	=	102	%	SM 3500-Cr		70	130	
2005/06-4	ME-CC	matrix spike, RPD	3/2/2006	Metal	Chromium VI	Total	=	0	%	SM 3500-Cr		0	30	
2005/06-4	ME-SCR	lab duplicate, RPD	3/2/2006	Metal	Chromium VI	Total	=	156	%	SM 3500-Cr		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Copper	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Copper	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Copper	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Copper	Total	=	64.3	%	EPA 200.8m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/9/2006	Metal	Copper	Total	=	106	%	EPA 200.8m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/9/2006	Metal	Copper	Total	=	105	%	EPA 200.8m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/9/2006	Metal	Copper	Total	=	1	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Lead	Dissolved	=	49	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Lead	Total	=	44.9	%	EPA 200.8m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	ME-VR2	matrix spike dup, rec	3/9/2006	Metal	Lead	Total	=	93	%	EPA 200.8m		65	135	
2005/06-4	ME-VR2	matrix spike, rec	3/9/2006	Metal	Lead	Total	=	96	%	EPA 200.8m		65	135	
2005/06-4	ME-VR2	matrix spike, RPD	3/9/2006	Metal	Lead	Total	=	3	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/2/2006	Metal	Mercury	Dissolved	<	0.5	ng/L	EPA 1631E	0.5			
2005/06-4	ME-SCR	lab duplicate, RPD	3/2/2006	Metal	Mercury	Dissolved	=	3	%	EPA 1631E		0	30	
2005/06-4	Lab	method blank	3/17/2006	Metal	Mercury	Total	<	0.5	ng/L	EPA 1631E	0.5			
2005/06-4	ME-SCR	lab duplicate, RPD	3/17/2006	Metal	Mercury	Total	=	5	%	EPA 1631E		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/17/2006	Metal	Mercury	Total	=	80	%	EPA 1631E		60	140	
2005/06-4	ME-VR2	matrix spike, rec	3/17/2006	Metal	Mercury	Total	=	85	%	EPA 1631E		60	140	
2005/06-4	ME-VR2	matrix spike, RPD	3/17/2006	Metal	Mercury	Total	=	6	%	EPA 1631E		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Nickel	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Nickel	Dissolved	=	43	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Nickel	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Nickel	Total	=	63.6	%	EPA 200.8m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/9/2006	Metal	Nickel	Total	=	99	%	EPA 200.8m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/9/2006	Metal	Nickel	Total	=	100	%	EPA 200.8m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/9/2006	Metal	Nickel	Total	=	1	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Selenium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Selenium	Dissolved	=	40	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Selenium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Selenium	Total	=	50	%	EPA 200.8m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/9/2006	Metal	Selenium	Total	=	121	%	EPA 200.8m		60	150	
2005/06-4	ME-VR2	matrix spike, rec	3/9/2006	Metal	Selenium	Total	=	118	%	EPA 200.8m		60	150	
2005/06-4	ME-VR2	matrix spike, RPD	3/9/2006	Metal	Selenium	Total	=	3	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Silver	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Silver	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Silver	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Silver	Total	=	0	%	EPA 200.8m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/9/2006	Metal	Silver	Total	=	99	%	EPA 200.8m		50	155	
2005/06-4	ME-VR2	matrix spike, rec	3/9/2006	Metal	Silver	Total	=	98	%	EPA 200.8m		50	155	
2005/06-4	ME-VR2	matrix spike, RPD	3/9/2006	Metal	Silver	Total	=	1	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Thallium	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Thallium	Total	=	26	%	EPA 200.8m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/9/2006	Metal	Thallium	Total	=	97	%	EPA 200.8m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/9/2006	Metal	Thallium	Total	=	95	%	EPA 200.8m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/9/2006	Metal	Thallium	Total	=	2	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Zinc	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Zinc	Dissolved	=	50	%	EPA 200.8m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Metal	Zinc	Total	=	48	%	EPA 200.8m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/9/2006	Metal	Zinc	Total	=	108	%	EPA 200.8m		50	150	
2005/06-4	ME-VR2	matrix spike, rec	3/9/2006	Metal	Zinc	Total	=	106	%	EPA 200.8m		50	150	
2005/06-4	ME-VR2	matrix spike, RPD	3/9/2006	Metal	Zinc	Total	=	2	%	EPA 200.8m		0	30	
2005/06-4	Lab	LCS dup, rec	3/6/2006	Nutrient	Ammonia as N	n/a	=	104	%	SM 4500-NH3 F		70	130	
2005/06-4	Lab	LCS, rec	3/6/2006	Nutrient	Ammonia as N	n/a	=	100	%	SM 4500-NH3 F		70	130	
2005/06-4	Lab	LCS, RPD	3/6/2006	Nutrient	Ammonia as N	n/a	=	4	%	SM 4500-NH3 F		0	30	
2005/06-4	Lab	method blank	3/6/2006	Nutrient	Ammonia as N	n/a	<	0.01	mg/L	SM 4500-NH3 F	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/6/2006	Nutrient	Ammonia as N	n/a	=	30.5	%	SM 4500-NH3 F		0	30	
2005/06-4	Lab	LCS dup, rec	3/2/2006	Nutrient	Nitrate as N	n/a	=	104	%	EPA 300.0		70	130	
2005/06-4	Lab	LCS, rec	3/2/2006	Nutrient	Nitrate as N	n/a	=	101	%	EPA 300.0		70	130	
2005/06-4	Lab	LCS, RPD	3/2/2006	Nutrient	Nitrate as N	n/a	=	3	%	EPA 300.0		0	30	
2005/06-4	Lab	method blank	3/2/2006	Nutrient	Nitrate as N	n/a	<	0.01	mg/L	EPA 300.0	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/3/2006	Nutrient	Nitrate as N	n/a	=	6.4	%	EPA 300.0		0	30	
2005/06-4	Lab	LCS dup, rec	3/2/2006	Nutrient	Nitrite as N	n/a	=	111	%	EPA 300.0		70	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	Lab	LCS, rec	3/2/2006	Nutrient	Nitrite as N	n/a	=	97	%	EPA 300.0		70	130	
2005/06-4	Lab	LCS, RPD	3/2/2006	Nutrient	Nitrite as N	n/a	=	14	%	EPA 300.0		0	30	
2005/06-4	Lab	method blank	3/2/2006	Nutrient	Nitrite as N	n/a	<	0.01	mg/L	EPA 300.0	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/3/2006	Nutrient	Nitrite as N	n/a	=	6.9	%	EPA 300.0		0	30	
2005/06-4	Lab	LCS dup, rec	3/6/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	88	%	EPA 300.0		70	130	
2005/06-4	Lab	LCS, rec	3/6/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	102	%	EPA 300.0		70	130	
2005/06-4	Lab	LCS, RPD	3/6/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	14.7	%	EPA 300.0		0	30	
2005/06-4	Lab	method blank	3/9/2006	Nutrient	Orthophosphate as P (Diss)	n/a	<	0.0075	mg/L	EPA 300.0	0.0075			
2005/06-4	ME-SCR	lab duplicate, RPD	3/2/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	5.7	%	EPA 300.0		0	30	
2005/06-4	Lab	LCS, rec	3/8/2006	Nutrient	TKN	n/a	=	102.3	%	EPA 351.1		80	120	
2005/06-4	Lab	method blank	3/22/2006	Nutrient	TKN	n/a	<	0.05	mg/L	EPA 351.1	0.05		0.05	
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Nutrient	TKN	n/a	=	175.4	%	EPA 351.1		0	20	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Nutrient	TKN	n/a	=	84.4	%	EPA 351.1		80	120	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Nutrient	TKN	n/a	=	83.7	%	EPA 351.1		80	120	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Nutrient	TKN	n/a	=	0.8	%	EPA 351.1		0	20	
2005/06-4	Lab	LCS dup, rec	3/7/2006	Nutrient	Total Phosphorus	Dissolved	=	92	%	SM 4500-P C		70	130	
2005/06-4	Lab	LCS, rec	3/7/2006	Nutrient	Total Phosphorus	Dissolved	=	96	%	SM 4500-P C		70	130	
2005/06-4	Lab	LCS, RPD	3/7/2006	Nutrient	Total Phosphorus	Dissolved	=	5	%	SM 4500-P C		0	30	
2005/06-4	Lab	method blank	3/7/2006	Nutrient	Total Phosphorus	Dissolved	<	0.016	mg/L	SM 4500-P C	0.016			
2005/06-4	ME-SCR	lab duplicate, RPD	3/7/2006	Nutrient	Total Phosphorus	Dissolved	=	2.5	%	SM 4500-P C		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/7/2006	Nutrient	Total Phosphorus	Dissolved	=	78	%	SM 4500-P C		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/7/2006	Nutrient	Total Phosphorus	Dissolved	=	74	%	SM 4500-P C		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/7/2006	Nutrient	Total Phosphorus	Dissolved	=	5	%	SM 4500-P C		0	30	
2005/06-4	Lab	LCS dup, rec	3/6/2006	Nutrient	Total Phosphorus	Total	=	92	%	SM 4500-P C		70	130	
2005/06-4	Lab	LCS, rec	3/6/2006	Nutrient	Total Phosphorus	Total	=	96	%	SM 4500-P C		70	130	
2005/06-4	Lab	LCS, RPD	3/6/2006	Nutrient	Total Phosphorus	Total	=	4	%	SM 4500-P C		0	30	
2005/06-4	Lab	method blank	3/6/2006	Nutrient	Total Phosphorus	Total	<	0.016	mg/L	SM 4500-P C	0.016			
2005/06-4	ME-SCR	lab duplicate, RPD	3/6/2006	Nutrient	Total Phosphorus	Total	=	5.2	%	SM 4500-P C		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	73	%	EPA 625m		65	140	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	69	%	EPA 625m		65	140	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	1,2-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	1,3-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	1,4-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	1,4-Dichlorobenzene	n/a	=	58	%	EPA 625m		50	140	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	1,4-Dichlorobenzene	n/a	=	50	%	EPA 625m		50	140	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	1,4-Dichlorobenzene	n/a	=	15	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	1-Methylnaphthalene	n/a	=	56.7	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	1-Methylnaphthalene	n/a	=	89	%	EPA 625m		50	120	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	1-Methylnaphthalene	n/a	=	79	%	EPA 625m		50	120	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	1-Methylnaphthalene	n/a	=	12	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	1-Methylphenanthrene	n/a	=	62.6	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	1-Methylphenanthrene	n/a	=	93	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	1-Methylphenanthrene	n/a	=	81	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	1-Methylphenanthrene	n/a	=	14	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	62	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	88	%	EPA 625m		45	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	91	%	EPA 625m		45	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	srgt method blank, rec	3/22/2006	Organic	2,4,6-Tribromophenol	n/a	=	78	%	EPA 625m		40	130	
2005/06-4	ME-CC	srgt environ, rec	3/22/2006	Organic	2,4,6-Tribromophenol	n/a	=	76	%	EPA 625m		40	130	
2005/06-4	ME-SCR	srgt environ, rec	3/22/2006	Organic	2,4,6-Tribromophenol	n/a	=	81	%	EPA 625m		40	130	
2005/06-4	ME-VR2	srgt environ, rec	3/22/2006	Organic	2,4,6-Tribromophenol	n/a	=	94	%	EPA 625m		40	130	
2005/06-4	ME-VR2	srgt matrix spike dup, rec	3/22/2006	Organic	2,4,6-Tribromophenol	n/a	=	88	%	EPA 625m		40	130	
2005/06-4	ME-VR2	srgt matrix spike, rec	3/22/2006	Organic	2,4,6-Tribromophenol	n/a	=	92	%	EPA 625m		40	130	
2005/06-4	Lab	method blank	3/22/2006	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2,4,6-Trichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2,4-Dichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	srgt method blank, rec	3/9/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	98	%	EPA 8151A		0	123	
2005/06-4	ME-CC	srgt environ, rec	3/9/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	72	%	EPA 8151A		0	123	
2005/06-4	ME-SCR	srgt environ, rec	3/9/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	70	%	EPA 8151A		0	123	
2005/06-4	ME-VR2	srgt environ, rec	3/9/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	78	%	EPA 8151A		0	123	
2005/06-4	Lab	method blank	3/22/2006	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2,4-Dimethylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2,4-Dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2,4-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	2,4-Dinitrotoluene	n/a	=	88	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	2,4-Dinitrotoluene	n/a	=	88	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	2,4-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	48.9	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	85	%	EPA 625m		55	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	85	%	EPA 625m		55	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2,6-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2-Chloronaphthalene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2-Chlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	2-Chlorophenol	n/a	=	91	%	EPA 625m		35	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	2-Chlorophenol	n/a	=	82	%	EPA 625m		35	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	2-Chlorophenol	n/a	=	10	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2-Methyl-4,6-dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2-Methylnaphthalene	n/a	=	54.4	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	2-Methylnaphthalene	n/a	=	68	%	EPA 625m		50	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	2-Methylnaphthalene	n/a	=	75	%	EPA 625m		50	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	2-Methylnaphthalene	n/a	=	10	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	2-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	3,3'-Dichlorobenzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	4-Bromophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	4-Chloro-3-methylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	4-Chloro-3-methylphenol	n/a	=	98	%	EPA 625m		30	150	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	4-Chloro-3-methylphenol	n/a	=	96	%	EPA 625m		30	150	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	4-Chloro-3-methylphenol	n/a	=	2	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	Lab	method blank	3/22/2006	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	4-Chlorophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	4-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	4-Nitrophenol	n/a	=	17	%	EPA 625m		0.1	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	4-Nitrophenol	n/a	=	15	%	EPA 625m		0.1	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	4-Nitrophenol	n/a	=	12.5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Acenaphthene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Acenaphthene	n/a	=	109	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Acenaphthene	n/a	=	105	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Acenaphthene	n/a	=	4	%	EPA 625m		0	30	
2005/06-4	Lab	srgt method blank, rec	3/22/2006	Organic	Acenaphthene-d10	n/a	=	89	%	EPA 625m		50	130	
2005/06-4	ME-CC	srgt environ, rec	3/22/2006	Organic	Acenaphthene-d10	n/a	=	79	%	EPA 625m		50	130	
2005/06-4	ME-SCR	srgt environ, rec	3/22/2006	Organic	Acenaphthene-d10	n/a	=	83	%	EPA 625m		50	130	
2005/06-4	ME-VR2	srgt environ, rec	3/22/2006	Organic	Acenaphthene-d10	n/a	=	83	%	EPA 625m		50	130	
2005/06-4	ME-VR2	srgt matrix spike dup, rec	3/22/2006	Organic	Acenaphthene-d10	n/a	=	77	%	EPA 625m		50	130	
2005/06-4	ME-VR2	srgt matrix spike, rec	3/22/2006	Organic	Acenaphthene-d10	n/a	=	80	%	EPA 625m		50	130	
2005/06-4	Lab	method blank	3/22/2006	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Acenaphthylene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Acenaphthylene	n/a	=	90	%	EPA 625m		60	120	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Acenaphthylene	n/a	=	90	%	EPA 625m		60	120	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Acenaphthylene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Anthracene	n/a	=	87	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Anthracene	n/a	=	85	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Anthracene	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Azobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Benzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Benzo(a)anthracene	n/a	=	57.5	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Benzo(a)anthracene	n/a	=	106	%	EPA 625m		70	140	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Benzo(a)anthracene	n/a	=	100	%	EPA 625m		70	140	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Benzo(a)anthracene	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Benzo(a)pyrene	n/a	=	58.1	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Benzo(a)pyrene	n/a	=	95	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Benzo(a)pyrene	n/a	=	100	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Benzo(a)pyrene	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Benzo(b)fluoranthene	n/a	=	38.8	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Benzo(b)fluoranthene	n/a	=	93	%	EPA 625m		60	140	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Benzo(b)fluoranthene	n/a	=	89	%	EPA 625m		60	140	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Benzo(b)fluoranthene	n/a	=	4	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Benzo(e)pyrene	n/a	=	51.7	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Benzo(e)pyrene	n/a	=	84	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Benzo(e)pyrene	n/a	=	87	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Benzo(e)pyrene	n/a	=	4	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Benzo(g,h,i)perylene	n/a	=	45.5	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Benzo(g,h,i)perylene	n/a	=	79	%	EPA 625m		50	140	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Benzo(g,h,i)perylene	n/a	=	80	%	EPA 625m		50	140	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Benzo(g,h,i)perylene	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Benzo(k)fluoranthene	n/a	=	55.2	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Benzo(k)fluoranthene	n/a	=	96	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Benzo(k)fluoranthene	n/a	=	94	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Benzo(k)fluoranthene	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Biphenyl	n/a	=	49.3	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Biphenyl	n/a	=	85	%	EPA 625m		50	120	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Biphenyl	n/a	=	85	%	EPA 625m		50	120	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Biphenyl	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Bis(2-chloroethoxy)methane	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Bis(2-chloroethyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Bis(2-chloroisopropyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.0213	µg/L	EPA 625m	0.005			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	7.8	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	169	%	EPA 625m		20	190	
2005/06-4	Lab	method blank	3/22/2006	Organic	Butyl benzyl phthalate	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Butyl benzyl phthalate	n/a	=	37.1	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Butyl benzyl phthalate	n/a	=	120	%	EPA 625m		65	160	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Butyl benzyl phthalate	n/a	=	111	%	EPA 625m		65	160	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Butyl benzyl phthalate	n/a	=	8	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Chrysene	n/a	=	38.9	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Chrysene	n/a	=	98	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Chrysene	n/a	=	92	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Chrysene	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	srgt method blank, rec	3/22/2006	Organic	Chrysene-d12	n/a	=	86	%	EPA 625m		70	130	
2005/06-4	ME-CC	srgt environ, rec	3/22/2006	Organic	Chrysene-d12	n/a	=	93	%	EPA 625m		70	130	
2005/06-4	ME-SCR	srgt environ, rec	3/22/2006	Organic	Chrysene-d12	n/a	=	97	%	EPA 625m		70	130	
2005/06-4	ME-VR2	srgt environ, rec	3/22/2006	Organic	Chrysene-d12	n/a	=	97	%	EPA 625m		70	130	
2005/06-4	ME-VR2	srgt matrix spike dup, rec	3/22/2006	Organic	Chrysene-d12	n/a	=	92	%	EPA 625m		70	130	
2005/06-4	ME-VR2	srgt matrix spike, rec	3/22/2006	Organic	Chrysene-d12	n/a	=	91	%	EPA 625m		70	130	
2005/06-4	Lab	method blank	3/22/2006	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Dibenz(a,h)anthracene	n/a	=	129.8	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Dibenz(a,h)anthracene	n/a	=	86	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Dibenz(a,h)anthracene	n/a	=	84	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Dibenz(a,h)anthracene	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Dibenzothiophene	n/a	=	53.2	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Dibenzothiophene	n/a	=	92	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Dibenzothiophene	n/a	=	80	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Dibenzothiophene	n/a	=	14	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Diethyl phthalate	n/a	=	0.0766	µg/L	EPA 625m	0.005			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Diethyl phthalate	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Diethyl phthalate	n/a	=	91	%	EPA 625m		50	150	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Diethyl phthalate	n/a	=	91	%	EPA 625m		50	150	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Diethyl phthalate	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Dimethyl phthalate	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Dimethyl phthalate	n/a	=	3.2	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Dimethyl phthalate	n/a	=	84	%	EPA 625m		40	155	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Dimethyl phthalate	n/a	=	85	%	EPA 625m		40	155	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Dimethyl phthalate	n/a	=	1	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	Lab	method blank	3/22/2006	Organic	Di-n-butylphthalate	n/a	=	0.0131	µg/L	EPA 625m	0.005			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Di-n-butylphthalate	n/a	=	29.3	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Di-n-butylphthalate	n/a	=	96	%	EPA 625m		65	145	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Di-n-butylphthalate	n/a	=	97	%	EPA 625m		65	145	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Di-n-butylphthalate	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Di-n-octylphthalate	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Di-n-octylphthalate	n/a	=	106	%	EPA 625m		50	165	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Di-n-octylphthalate	n/a	=	113	%	EPA 625m		50	165	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Di-n-octylphthalate	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Fluoranthene	n/a	=	52.3	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Fluoranthene	n/a	=	93	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Fluoranthene	n/a	=	92	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Fluoranthene	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Fluorene	n/a	=	48.2	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Fluorene	n/a	=	90	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Fluorene	n/a	=	89	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Fluorene	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Hexachlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Hexachlorobutadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Hexachlorocyclopentadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Hexachloroethane	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	164	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	87	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	83	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Isophorone	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Naphthalene	n/a	=	60.7	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Naphthalene	n/a	=	85	%	EPA 625m		50	120	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Naphthalene	n/a	=	74	%	EPA 625m		50	120	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Naphthalene	n/a	=	14	%	EPA 625m		0	30	
2005/06-4	Lab	srgt method blank, rec	3/22/2006	Organic	Naphthalene-d8	n/a	=	79	%	EPA 625m		40	120	
2005/06-4	ME-CC	srgt environ, rec	3/22/2006	Organic	Naphthalene-d8	n/a	=	57	%	EPA 625m		40	120	
2005/06-4	ME-SCR	srgt environ, rec	3/22/2006	Organic	Naphthalene-d8	n/a	=	71	%	EPA 625m		40	120	
2005/06-4	ME-VR2	srgt environ, rec	3/22/2006	Organic	Naphthalene-d8	n/a	=	71	%	EPA 625m		40	120	
2005/06-4	ME-VR2	srgt matrix spike dup, rec	3/22/2006	Organic	Naphthalene-d8	n/a	=	69	%	EPA 625m		40	120	
2005/06-4	ME-VR2	srgt matrix spike, rec	3/22/2006	Organic	Naphthalene-d8	n/a	=	68	%	EPA 625m		40	120	
2005/06-4	Lab	method blank	3/22/2006	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Nitrobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	N-Nitrosodimethylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	90	%	EPA 625m		55	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	80	%	EPA 625m		55	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	12	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	N-Nitrosodiphenylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Pentachlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Pentachlorophenol	n/a	=	98	%	EPA 625m		10	160	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Pentachlorophenol	n/a	=	95	%	EPA 625m		10	160	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Pentachlorophenol	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Perylene	n/a	=	51.2	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Perylene	n/a	=	85	%	EPA 625m		65	135	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Perylene	n/a	=	90	%	EPA 625m		65	135	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Perylene	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	srgt method blank, rec	3/22/2006	Organic	Perylene-d12	n/a	=	68	%	EPA 625m		60	140	
2005/06-4	ME-CC	srgt environ, rec	3/22/2006	Organic	Perylene-d12	n/a	=	87	%	EPA 625m		60	140	
2005/06-4	ME-SCR	srgt environ, rec	3/22/2006	Organic	Perylene-d12	n/a	=	93	%	EPA 625m		60	140	
2005/06-4	ME-VR2	srgt environ, rec	3/22/2006	Organic	Perylene-d12	n/a	=	87	%	EPA 625m		60	140	
2005/06-4	ME-VR2	srgt matrix spike dup, rec	3/22/2006	Organic	Perylene-d12	n/a	=	80	%	EPA 625m		60	140	
2005/06-4	ME-VR2	srgt matrix spike, rec	3/22/2006	Organic	Perylene-d12	n/a	=	82	%	EPA 625m		60	140	
2005/06-4	Lab	method blank	3/22/2006	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Phenanthrene	n/a	=	45.9	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Phenanthrene	n/a	=	80	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Phenanthrene	n/a	=	83	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Phenanthrene	n/a	=	4	%	EPA 625m		0	30	
2005/06-4	Lab	srgt method blank, rec	3/22/2006	Organic	Phenanthrene-d10	n/a	=	92	%	EPA 625m		70	130	
2005/06-4	ME-CC	srgt environ, rec	3/22/2006	Organic	Phenanthrene-d10	n/a	=	80	%	EPA 625m		70	130	
2005/06-4	ME-SCR	srgt environ, rec	3/22/2006	Organic	Phenanthrene-d10	n/a	=	86	%	EPA 625m		70	130	
2005/06-4	ME-VR2	srgt environ, rec	3/22/2006	Organic	Phenanthrene-d10	n/a	=	89	%	EPA 625m		70	130	
2005/06-4	ME-VR2	srgt matrix spike dup, rec	3/22/2006	Organic	Phenanthrene-d10	n/a	=	82	%	EPA 625m		70	130	
2005/06-4	ME-VR2	srgt matrix spike, rec	3/22/2006	Organic	Phenanthrene-d10	n/a	=	84	%	EPA 625m		70	130	
2005/06-4	Lab	method blank	3/22/2006	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Phenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Phenol	n/a	=	35	%	EPA 625m		0.1	115	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Phenol	n/a	=	34	%	EPA 625m		0.1	115	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Phenol	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	srgt method blank, rec	3/22/2006	Organic	Phenol-d5	n/a	=	39	%	EPA 625m		10	110	
2005/06-4	ME-CC	srgt environ, rec	3/22/2006	Organic	Phenol-d5	n/a	=	29	%	EPA 625m		10	110	
2005/06-4	ME-SCR	srgt environ, rec	3/22/2006	Organic	Phenol-d5	n/a	=	23	%	EPA 625m		10	110	
2005/06-4	ME-VR2	srgt environ, rec	3/22/2006	Organic	Phenol-d5	n/a	=	30	%	EPA 625m		10	110	
2005/06-4	ME-VR2	srgt matrix spike dup, rec	3/22/2006	Organic	Phenol-d5	n/a	=	45	%	EPA 625m		10	110	
2005/06-4	ME-VR2	srgt matrix spike, rec	3/22/2006	Organic	Phenol-d5	n/a	=	42	%	EPA 625m		10	110	
2005/06-4	Lab	method blank	3/22/2006	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Pyrene	n/a	=	48.5	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Organic	Pyrene	n/a	=	110	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Organic	Pyrene	n/a	=	94	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Organic	Pyrene	n/a	=	16	%	EPA 625m		0	30	
2005/06-4	Lab	srgt method blank, rec	3/22/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	85	%	EPA 625m		40	130	
2005/06-4	ME-CC	srgt environ, rec	3/22/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	68	%	EPA 625m		40	130	
2005/06-4	ME-SCR	srgt environ, rec	3/22/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	67	%	EPA 625m		40	130	
2005/06-4	ME-VR2	srgt environ, rec	3/22/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	80	%	EPA 625m		40	130	
2005/06-4	ME-VR2	srgt matrix spike dup, rec	3/22/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	82	%	EPA 625m		40	130	
2005/06-4	ME-VR2	srgt matrix spike, rec	3/22/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	78	%	EPA 625m		40	130	
2005/06-4	Lab	method blank	3/22/2006	Organic	Total Detectable PAHs	n/a	=	0	µg/L	EPA 625m				
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Organic	Total Detectable PAHs	n/a	=	52.9	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	Aroclor 1016	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	Aroclor 1221	n/a	=	0	%	EPA 625m		0	30	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	Lab	method blank	3/22/2006	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	Aroclor 1232	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	Aroclor 1242	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	Aroclor 1248	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	Aroclor 1254	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	Aroclor 1260	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 018	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 018	n/a	=	88	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 018	n/a	=	86	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 018	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 028	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 028	n/a	=	100	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 028	n/a	=	92	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 028	n/a	=	8	%	EPA 625m		0	30	
2005/06-4	Lab	srvt method blank, rec	3/22/2006	PCB	PCB 030	n/a	=	88	%	EPA 625m		40	130	
2005/06-4	ME-CC	srvt environ, rec	3/22/2006	PCB	PCB 030	n/a	=	66	%	EPA 625m		40	130	
2005/06-4	ME-SCR	srvt environ, rec	3/22/2006	PCB	PCB 030	n/a	=	67	%	EPA 625m		40	130	
2005/06-4	ME-VR2	srvt environ, rec	3/22/2006	PCB	PCB 030	n/a	=	81	%	EPA 625m		40	130	
2005/06-4	ME-VR2	srvt matrix spike dup, rec	3/22/2006	PCB	PCB 030	n/a	=	83	%	EPA 625m		40	130	
2005/06-4	ME-VR2	srvt matrix spike, rec	3/22/2006	PCB	PCB 030	n/a	=	78	%	EPA 625m		40	130	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 031	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 031	n/a	=	100	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 031	n/a	=	100	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 031	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 033	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 033	n/a	=	87	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 033	n/a	=	92	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 033	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 037	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 037	n/a	=	100	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 037	n/a	=	98	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 037	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 044	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 044	n/a	=	86	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 044	n/a	=	87	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 044	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 049	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 049	n/a	=	93	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 049	n/a	=	91	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 049	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 052	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 052	n/a	=	89	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 052	n/a	=	91	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 052	n/a	=	2	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 066	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 066	n/a	=	98	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 066	n/a	=	106	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 066	n/a	=	8	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 070	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 070	n/a	=	94	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 070	n/a	=	93	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 070	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 074	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 074	n/a	=	96	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 074	n/a	=	106	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 074	n/a	=	10	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 077	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 077	n/a	=	104	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 077	n/a	=	105	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 077	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 081	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 081	n/a	=	105	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 081	n/a	=	105	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 081	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 087	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 087	n/a	=	99	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 087	n/a	=	98	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 087	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 095	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 095	n/a	=	88	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 095	n/a	=	88	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 095	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 097	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 097	n/a	=	91	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 097	n/a	=	97	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 097	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 099	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 099	n/a	=	96	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 099	n/a	=	94	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 099	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 101	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 101	n/a	=	95	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 101	n/a	=	99	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 101	n/a	=	4	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 105	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 105	n/a	=	97	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 105	n/a	=	95	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 105	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001			

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 110	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 110	n/a	=	92	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 110	n/a	=	89	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 110	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	srgt method blank, rec	3/22/2006	PCB	PCB 112	n/a	=	93	%	EPA 625m		60	120	
2005/06-4	ME-CC	srgt environ, rec	3/22/2006	PCB	PCB 112	n/a	=	77	%	EPA 625m		60	120	
2005/06-4	ME-SCR	srgt environ, rec	3/22/2006	PCB	PCB 112	n/a	=	67	%	EPA 625m		60	120	
2005/06-4	ME-VR2	srgt environ, rec	3/22/2006	PCB	PCB 112	n/a	=	77	%	EPA 625m		60	120	
2005/06-4	ME-VR2	srgt matrix spike dup, rec	3/22/2006	PCB	PCB 112	n/a	=	88	%	EPA 625m		60	120	
2005/06-4	ME-VR2	srgt matrix spike, rec	3/22/2006	PCB	PCB 112	n/a	=	85	%	EPA 625m		60	120	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 114	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 114	n/a	=	99	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 114	n/a	=	105	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 114	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 118	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 118	n/a	=	97	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 118	n/a	=	100	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 118	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 119	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 119	n/a	=	96	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 119	n/a	=	94	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 119	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 123	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 123	n/a	=	106	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 123	n/a	=	100	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 123	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 126	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 126	n/a	=	110	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 126	n/a	=	117	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 126	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 128 + 167	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 128 + 167	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 128 + 167	n/a	=	102	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 128 + 167	n/a	=	107	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 128 + 167	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 138	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 138	n/a	=	100	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 138	n/a	=	105	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 138	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 141	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 141	n/a	=	102	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 141	n/a	=	100	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 141	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 149	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 149	n/a	=	86	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 149	n/a	=	87	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 149	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001			

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 151	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 151	n/a	=	94	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 151	n/a	=	96	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 151	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 153	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 153	n/a	=	93	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 153	n/a	=	93	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 153	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 156	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 156	n/a	=	112	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 156	n/a	=	114	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 156	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 157	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 157	n/a	=	103	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 157	n/a	=	106	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 157	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 158	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 158	n/a	=	99	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 158	n/a	=	98	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 158	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 168 + 132	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 168 + 132	n/a	=	94	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 168 + 132	n/a	=	94	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 168 + 132	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 169	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 169	n/a	=	122	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 169	n/a	=	110	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 169	n/a	=	10	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 170	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 170	n/a	=	100	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 170	n/a	=	103	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 170	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 177	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 177	n/a	=	99	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 177	n/a	=	104	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 177	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 180	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 180	n/a	=	108	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 180	n/a	=	109	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 180	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 183	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 183	n/a	=	99	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 183	n/a	=	97	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 183	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 187	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 187	n/a	=	99	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 187	n/a	=	95	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 187	n/a	=	4	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 189	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 189	n/a	=	121	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 189	n/a	=	109	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 189	n/a	=	10	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 194	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 194	n/a	=	76	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 194	n/a	=	102	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 194	n/a	=	29	%	EPA 625m		0	30	
2005/06-4	Lab	srgt method blank, rec	3/22/2006	PCB	PCB 198	n/a	=	88	%	EPA 625m		60	120	
2005/06-4	ME-CC	srgt environ, rec	3/22/2006	PCB	PCB 198	n/a	=	87	%	EPA 625m		60	120	
2005/06-4	ME-SCR	srgt environ, rec	3/22/2006	PCB	PCB 198	n/a	=	76	%	EPA 625m		60	120	
2005/06-4	ME-VR2	srgt environ, rec	3/22/2006	PCB	PCB 198	n/a	=	88	%	EPA 625m		60	120	
2005/06-4	ME-VR2	srgt matrix spike dup, rec	3/22/2006	PCB	PCB 198	n/a	=	90	%	EPA 625m		60	120	
2005/06-4	ME-VR2	srgt matrix spike, rec	3/22/2006	PCB	PCB 198	n/a	=	92	%	EPA 625m		60	120	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 200	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 200	n/a	=	94	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 200	n/a	=	97	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 200	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 201	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 201	n/a	=	111	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 201	n/a	=	112	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 201	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	PCB 206	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	PCB	PCB 206	n/a	=	103	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	PCB	PCB 206	n/a	=	100	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	PCB	PCB 206	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	PCB	Total Detectable PCBs	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	LCS, RPD	3/6/2006	Pesticide	2,4,5-T	n/a	=	0	%	EPA 8151A		0	30	
2005/06-4	Lab	LCS dup, rec	3/9/2006	Pesticide	2,4,5-T	n/a	=	114	%	EPA 8151A		30	130	
2005/06-4	Lab	LCS, rec	3/9/2006	Pesticide	2,4,5-T	n/a	=	114	%	EPA 8151A		30	130	
2005/06-4	Lab	method blank	3/9/2006	Pesticide	2,4,5-T	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Pesticide	2,4,5-T	n/a	=	0	%	EPA 8151A		0	20	
2005/06-4	Lab	method blank	3/9/2006	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Pesticide	2,4,5-TP (Silvex)	n/a	=	0	%	EPA 8151A		0	20	
2005/06-4	Lab	LCS, RPD	3/6/2006	Pesticide	2,4-D	n/a	=	3	%	EPA 8151A		0	30	
2005/06-4	Lab	LCS dup, rec	3/9/2006	Pesticide	2,4-D	n/a	=	103	%	EPA 8151A		30	130	
2005/06-4	Lab	LCS, rec	3/9/2006	Pesticide	2,4-D	n/a	=	101	%	EPA 8151A		30	130	
2005/06-4	Lab	method blank	3/9/2006	Pesticide	2,4-D	n/a	<	5	µg/L	EPA 8151A	5		5	
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Pesticide	2,4-D	n/a	=	0	%	EPA 8151A		0	20	
2005/06-4	Lab	LCS, RPD	3/6/2006	Pesticide	2,4-DB	n/a	=	4	%	EPA 8151A		0	30	
2005/06-4	Lab	LCS dup, rec	3/9/2006	Pesticide	2,4-DB	n/a	=	94	%	EPA 8151A		30	130	
2005/06-4	Lab	LCS, rec	3/9/2006	Pesticide	2,4-DB	n/a	=	97	%	EPA 8151A		30	130	
2005/06-4	Lab	method blank	3/9/2006	Pesticide	2,4-DB	n/a	<	5	µg/L	EPA 8151A	5		5	
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Pesticide	2,4-DB	n/a	=	0	%	EPA 8151A		0	20	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	2,4'-DDD	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	2,4'-DDD	n/a	=	99	%	EPA 625m		50	140	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	2,4'-DDD	n/a	=	98	%	EPA 625m		50	140	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	2,4'-DDD	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	2,4'-DDE	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	2,4'-DDE	n/a	=	84	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	2,4'-DDE	n/a	=	83	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	2,4'-DDE	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	2,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	2,4'-DDT	n/a	=	88	%	EPA 625m		40	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	2,4'-DDT	n/a	=	79	%	EPA 625m		40	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	2,4'-DDT	n/a	=	11	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	4,4'-DDD	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	4,4'-DDD	n/a	=	95	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	4,4'-DDD	n/a	=	83	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	4,4'-DDD	n/a	=	13	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	4,4'-DDE	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	4,4'-DDE	n/a	=	88	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	4,4'-DDE	n/a	=	87	%	EPA 625m		70	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	4,4'-DDE	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	4,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	4,4'-DDT	n/a	=	27	%	EPA 625m		0.05	150	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	4,4'-DDT	n/a	=	29	%	EPA 625m		0.05	150	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	4,4'-DDT	n/a	=	7	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Aldrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Aldrin	n/a	=	76	%	EPA 625m		50	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Aldrin	n/a	=	86	%	EPA 625m		50	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Aldrin	n/a	=	12	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	BHC-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	BHC-alpha	n/a	=	82	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	BHC-alpha	n/a	=	78	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	BHC-alpha	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	BHC-beta	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	BHC-beta	n/a	=	80	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	BHC-beta	n/a	=	78	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	BHC-beta	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	BHC-delta	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	BHC-delta	n/a	=	88	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	BHC-delta	n/a	=	89	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	BHC-delta	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	83	%	EPA 625m		50	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	82	%	EPA 625m		50	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Bolstar	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Bolstar	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Bolstar	n/a	=	106	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Bolstar	n/a	=	102	%	EPA 625m		65	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Bolstar	n/a	=	4	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Chlordane-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Chlordane-alpha	n/a	=	79	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Chlordane-alpha	n/a	=	81	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Chlordane-alpha	n/a	=	2.5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Chlordane-gamma	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Chlordane-gamma	n/a	=	80	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Chlordane-gamma	n/a	=	81	%	EPA 625m		60	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Chlordane-gamma	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Chlorpyrifos	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Chlorpyrifos	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Chlorpyrifos	n/a	=	96	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Chlorpyrifos	n/a	=	96	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Chlorpyrifos	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	cis-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	cis-Nonachlor	n/a	=	79	%	EPA 625m		60	120	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	cis-Nonachlor	n/a	=	81	%	EPA 625m		60	120	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	cis-Nonachlor	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Pesticide	Dalapon	n/a	<	13	µg/L	EPA 8151A	13		13	
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Pesticide	Dalapon	n/a	=	0	%	EPA 8151A		0	20	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Demeton-O	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Demeton-O	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Demeton-O	n/a	=	100	%	EPA 625m		45	105	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Demeton-O	n/a	=	84	%	EPA 625m		45	105	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Demeton-O	n/a	=	17	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Diazinon	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Diazinon	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Diazinon	n/a	=	104	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Diazinon	n/a	=	99	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Diazinon	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Pesticide	Dicamba	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Pesticide	Dicamba	n/a	=	0	%	EPA 8151A		0	20	
2005/06-4	Lab	method blank	3/9/2006	Pesticide	Dichlorprop	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Pesticide	Dichlorprop	n/a	=	0	%	EPA 8151A		0	20	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Dichlorvos	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Dichlorvos	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Dichlorvos	n/a	=	104	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Dichlorvos	n/a	=	93	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Dichlorvos	n/a	=	11	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Dieldrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Dieldrin	n/a	=	79	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Dieldrin	n/a	=	77	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Dieldrin	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Dimethoate	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Dimethoate	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Dimethoate	n/a	=	103	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Dimethoate	n/a	=	96	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Dimethoate	n/a	=	7	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Pesticide	Dinoseb	n/a	<	2.5	µg/L	EPA 8151A	2.5		2.5	
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Pesticide	Dinoseb	n/a	=	0	%	EPA 8151A		0	20	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Disulfoton	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Disulfoton	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Disulfoton	n/a	=	93	%	EPA 625m		45	105	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Disulfoton	n/a	=	89	%	EPA 625m		45	105	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Disulfoton	n/a	=	4	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Endosulfan sulfate	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Endosulfan sulfate	n/a	=	87	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Endosulfan sulfate	n/a	=	84	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Endosulfan sulfate	n/a	=	3	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Endosulfan-I	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Endosulfan-I	n/a	=	90	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Endosulfan-I	n/a	=	88	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Endosulfan-I	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Endosulfan-II	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Endosulfan-II	n/a	=	92	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Endosulfan-II	n/a	=	83	%	EPA 625m		60	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Endosulfan-II	n/a	=	10	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Endrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Endrin	n/a	=	102	%	EPA 625m		65	135	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Endrin	n/a	=	82	%	EPA 625m		65	135	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Endrin	n/a	=	23	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Endrin aldehyde	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Endrin ketone	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Endrin ketone	n/a	=	69	%	EPA 625m		40	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Endrin ketone	n/a	=	54	%	EPA 625m		40	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Endrin ketone	n/a	=	24	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Ethoprop	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Ethoprop	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Ethoprop	n/a	=	102	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Ethoprop	n/a	=	97	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Ethoprop	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	105	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	100	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Fensulfothion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Fensulfothion	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Fensulfothion	n/a	=	104	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Fensulfothion	n/a	=	99	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Fensulfothion	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Fenthion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Fenthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Fenthion	n/a	=	102	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Fenthion	n/a	=	97	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Fenthion	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	LCS, rec	3/15/2006	Pesticide	Glyphosate	n/a	=	100	%	EPA 547		71	137	
2005/06-4	Lab	method blank	3/15/2006	Pesticide	Glyphosate	n/a	<	5	µg/L	EPA 547	5		5	
2005/06-4	ME-SCR	lab duplicate, RPD	3/15/2006	Pesticide	Glyphosate	n/a	=	0	%	EPA 547		0	25	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Heptachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Heptachlor	n/a	=	66	%	EPA 625m		45	135	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Heptachlor	n/a	=	61	%	EPA 625m		45	135	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Heptachlor	n/a	=	8	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Heptachlor epoxide	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Heptachlor epoxide	n/a	=	81	%	EPA 625m		65	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Heptachlor epoxide	n/a	=	83	%	EPA 625m		65	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Heptachlor epoxide	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Malathion	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Malathion	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Malathion	n/a	=	95	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Malathion	n/a	=	97	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Malathion	n/a	=	2	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/9/2006	Pesticide	MCPA	n/a	<	500	µg/L	EPA 8151A	500		500	
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Pesticide	MCPA	n/a	=	0	%	EPA 8151A		0	20	
2005/06-4	Lab	method blank	3/9/2006	Pesticide	MCPP	n/a	<	500	µg/L	EPA 8151A	500		500	
2005/06-4	ME-SCR	lab duplicate, RPD	3/9/2006	Pesticide	MCPP	n/a	=	0	%	EPA 8151A		0	20	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Merphos	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Merphos	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Merphos	n/a	=	104	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Merphos	n/a	=	99	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Merphos	n/a	=	5	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Methoxychlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Methoxychlor	n/a	=	14	%	EPA 625m		0.001	155	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Methoxychlor	n/a	=	16	%	EPA 625m		0.001	155	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Methoxychlor	n/a	=	13	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Methyl parathion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Methyl parathion	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Methyl parathion	n/a	=	102	%	EPA 625m		60	120	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Methyl parathion	n/a	=	96	%	EPA 625m		60	120	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Methyl parathion	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Mevinphos	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Mevinphos	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Mevinphos	n/a	=	94	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Mevinphos	n/a	=	104	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Mevinphos	n/a	=	10	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Mirex	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Mirex	n/a	=	83	%	EPA 625m		50	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Mirex	n/a	=	83	%	EPA 625m		50	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Mirex	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Oxychlorane	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Oxychlorane	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Oxychlorane	n/a	=	88	%	EPA 625m		60	120	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Oxychlorane	n/a	=	93	%	EPA 625m		60	120	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Oxychlorane	n/a	=	6	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Phorate	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Phorate	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Phorate	n/a	=	100	%	EPA 625m		45	105	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Phorate	n/a	=	88	%	EPA 625m		45	105	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Phorate	n/a	=	13	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	100	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	101	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	1	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Tokuthion	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Tokuthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Tokuthion	n/a	=	103	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Tokuthion	n/a	=	99	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Tokuthion	n/a	=	4	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Total Detectable DDTs	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Toxaphene	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	trans-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	trans-Nonachlor	n/a	=	84	%	EPA 625m		55	130	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	trans-Nonachlor	n/a	=	85	%	EPA 625m		55	130	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	trans-Nonachlor	n/a	=	1	%	EPA 625m		0	30	
2005/06-4	Lab	method blank	3/22/2006	Pesticide	Trichloronate	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-4	ME-SCR	lab duplicate, RPD	3/22/2006	Pesticide	Trichloronate	n/a	=	0	%	EPA 625m		0	30	
2005/06-4	ME-VR2	matrix spike dup, rec	3/22/2006	Pesticide	Trichloronate	n/a	=	112	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, rec	3/22/2006	Pesticide	Trichloronate	n/a	=	104	%	EPA 625m		65	125	
2005/06-4	ME-VR2	matrix spike, RPD	3/22/2006	Pesticide	Trichloronate	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Anion	Bromide	n/a	=	100	%	EPA 300.0		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Anion	Bromide	n/a	=	98	%	EPA 300.0		70	130	
2005/06-5	Lab	LCS, RPD	6/2/2006	Anion	Bromide	n/a	=	2	%	EPA 300.0		0	30	
2005/06-5	Lab	method blank	6/2/2006	Anion	Bromide	n/a	<	0.001	mg/L	EPA 300.0	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/2/2006	Anion	Bromide	n/a	=	2.8	%	EPA 300.0		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/2/2006	Anion	Bromide	n/a	=	97	%	EPA 300.0		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/2/2006	Anion	Bromide	n/a	=	95	%	EPA 300.0		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/2/2006	Anion	Bromide	n/a	=	3	%	EPA 300.0		0	30	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Anion	Chloride	n/a	=	111	%	SM 4500-Cl E		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Anion	Chloride	n/a	=	111	%	SM 4500-Cl E		70	130	
2005/06-5	Lab	LCS, RPD	6/2/2006	Anion	Chloride	n/a	=	0	%	SM 4500-Cl E		0	30	
2005/06-5	Lab	method blank	6/20/2006	Anion	Chloride	n/a	<	0.01	mg/L	SM 4500-Cl E	0.01		0.01	
2005/06-5	ME-CC	lab duplicate, RPD	6/20/2006	Anion	Chloride	n/a	=	0	%	SM 4500-Cl E		0	30	
2005/06-5	ME-CC	matrix spike dup, rec	6/20/2006	Anion	Chloride	n/a	=	102	%	SM 4500-Cl E		70	130	
2005/06-5	ME-CC	matrix spike, rec	6/20/2006	Anion	Chloride	n/a	=	103	%	SM 4500-Cl E		70	130	
2005/06-5	ME-CC	matrix spike, RPD	6/20/2006	Anion	Chloride	n/a	=	1	%	SM 4500-Cl E		0	30	
2005/06-5	Lab	LCS dup, rec	6/5/2006	Anion	Perchlorate	n/a	=	97	%	EPA 314.0		85	115	
2005/06-5	Lab	LCS, rec	6/5/2006	Anion	Perchlorate	n/a	=	96	%	EPA 314.0		85	115	
2005/06-5	Lab	LCS, RPD	6/5/2006	Anion	Perchlorate	n/a	=	1	%	EPA 314.0		0	15	
2005/06-5	Lab	method blank	6/5/2006	Anion	Perchlorate	n/a	<	0.43	µg/L	EPA 314.0	0.43		0.43	
2005/06-5	ME-CC	field blank	5/31/2006	Bacteriological	E. Coli	n/a	<	10	MPN/100 mL	Colilert-18	10		10	
2005/06-5	ME-CC	field blank	5/31/2006	Bacteriological	Enterococcus	n/a	<	10	MPN/100 mL	Enterolert	10		10	
2005/06-5	ME-CC	field blank	5/31/2006	Bacteriological	Fecal Coliform	n/a	<	2	MPN/100 mL	Colilert-18	2		2	
2005/06-5	ME-CC	field blank	5/31/2006	Bacteriological	Total Coliform	n/a	<	10	MPN/100 mL	Colilert-18	10		10	
2005/06-5	Lab	method blank	6/1/2006	Conventional	BOD	n/a	<	0.58	mg/L	EPA 405.1	0.58		0.58	
2005/06-5	ME-SCR	lab duplicate, RPD	6/1/2006	Conventional	BOD	n/a	=	0	%	EPA 405.1		0	25	
2005/06-5	Lab	method blank	6/1/2006	Conventional	Conductivity	n/a	<	1	µmhos/cm	SM 2510	1		1	
2005/06-5	ME-VR2	lab duplicate, RPD	6/1/2006	Conventional	Conductivity	n/a	=	0	%	SM 2510		0	30	
2005/06-5	Lab	method blank	6/13/2006	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1		1	
2005/06-5	ME-CC	field blank	6/13/2006	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1		1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Conventional	Hardness as CaCO3	Total	=	4	%	SM 2340 B		0	30	
2005/06-5	ME-VR2	lab duplicate, RPD	6/1/2006	Conventional	pH	n/a	=	0	%	EPA 150.1		0	30	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Conventional	Total Dissolved Solids	n/a	=	98	%	SM 2540 C		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Conventional	Total Dissolved Solids	n/a	=	98	%	SM 2540 C		70	130	
2005/06-5	Lab	LCS, RPD	6/2/2006	Conventional	Total Dissolved Solids	n/a	=	0	%	SM 2540 C		0	30	
2005/06-5	Lab	method blank	6/5/2006	Conventional	Total Dissolved Solids	n/a	<	0.1	mg/L	SM 2540 C	0.1		0.1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/5/2006	Conventional	Total Dissolved Solids	n/a	=	2	%	SM 2540 C		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	Lab	LCS, rec	6/6/2006	Conventional	Total Organic Carbon	n/a	=	106	%	EPA 415.1		80	120	
2005/06-5	Lab	method blank	6/6/2006	Conventional	Total Organic Carbon	n/a	<	0.021	mg/L	EPA 415.1	0.021		0.021	
2005/06-5	ME-SCR	matrix spike dup, rec	6/6/2006	Conventional	Total Organic Carbon	n/a	=	99	%	EPA 415.1		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/6/2006	Conventional	Total Organic Carbon	n/a	=	100	%	EPA 415.1		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/6/2006	Conventional	Total Organic Carbon	n/a	=	1	%	EPA 415.1		0	25	
2005/06-5	Lab	method blank	6/5/2006	Conventional	Total Suspended Solids	n/a	<	0.5	mg/L	SM 2540 D	0.5		0.5	
2005/06-5	ME-SCR	lab duplicate, RPD	6/5/2006	Conventional	Total Suspended Solids	n/a	=	0	%	SM 2540 D		0	30	
2005/06-5	Lab	method blank	6/2/2006	Conventional	Turbidity	n/a	<	1	NTU	EPA 180.1	1		1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/2/2006	Conventional	Turbidity	n/a	=	1	%	EPA 180.1		0	30	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Hydrocarbon	Oil and Grease	n/a	=	98	%	EPA 1664A		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Hydrocarbon	Oil and Grease	n/a	=	95	%	EPA 1664A		70	130	
2005/06-5	Lab	LCS, RPD	6/2/2006	Hydrocarbon	Oil and Grease	n/a	=	3	%	EPA 1664A		0	30	
2005/06-5	Lab	method blank	6/16/2006	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1		1	
2005/06-5	ME-VR2	lab duplicate, RPD	6/16/2006	Hydrocarbon	Oil and Grease	n/a	=	0	%	EPA 1664A		0	30	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Hydrocarbon	TRPH	n/a	=	105	%	EPA 1664		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Hydrocarbon	TRPH	n/a	=	119	%	EPA 1664		70	130	
2005/06-5	Lab	LCS, RPD	6/2/2006	Hydrocarbon	TRPH	n/a	=	13	%	EPA 1664		0	30	
2005/06-5	Lab	method blank	6/6/2006	Hydrocarbon	TRPH	n/a	<	0.01	mg/L	EPA 1664	0.01		0.01	
2005/06-5	Lab	method blank	6/13/2006	Metal	Aluminum	Dissolved	<	5	µg/L	EPA 200.8m	5		5	
2005/06-5	Lab	method blank	6/13/2006	Metal	Aluminum	Total	<	5	µg/L	EPA 200.8m	5		5	
2005/06-5	ME-CC	field blank	6/13/2006	Metal	Aluminum	Total	<	5	µg/L	EPA 200.8m	5		5	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Metal	Aluminum	Total	=	3	%	EPA 200.8m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/13/2006	Metal	Aluminum	Total	=	117	%	EPA 200.8m		50	140	
2005/06-5	ME-SCR	matrix spike, rec	6/13/2006	Metal	Aluminum	Total	=	160	%	EPA 200.8m		50	140	
2005/06-5	ME-SCR	matrix spike, RPD	6/13/2006	Metal	Aluminum	Total	=	31	%	EPA 200.8m		0	30	
2005/06-5	Lab	method blank	6/13/2006	Metal	Arsenic	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	Lab	method blank	6/13/2006	Metal	Arsenic	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	ME-CC	field blank	6/13/2006	Metal	Arsenic	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Metal	Arsenic	Total	=	5	%	EPA 200.8m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/13/2006	Metal	Arsenic	Total	=	106	%	EPA 200.8m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/13/2006	Metal	Arsenic	Total	=	104	%	EPA 200.8m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/13/2006	Metal	Arsenic	Total	=	2	%	EPA 200.8m		0	30	
2005/06-5	Lab	method blank	6/13/2006	Metal	Cadmium	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	Lab	method blank	6/13/2006	Metal	Cadmium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	ME-CC	field blank	6/13/2006	Metal	Cadmium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Metal	Cadmium	Total	=	4	%	EPA 200.8m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/13/2006	Metal	Cadmium	Total	=	100	%	EPA 200.8m		75	130	
2005/06-5	ME-SCR	matrix spike, rec	6/13/2006	Metal	Cadmium	Total	=	99	%	EPA 200.8m		75	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/13/2006	Metal	Cadmium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-5	Lab	method blank	6/13/2006	Metal	Chromium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-5	Lab	method blank	6/13/2006	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-5	ME-CC	field blank	6/13/2006	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Metal	Chromium	Total	=	4	%	EPA 200.8m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/13/2006	Metal	Chromium	Total	=	100	%	EPA 200.8m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/13/2006	Metal	Chromium	Total	=	99	%	EPA 200.8m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/13/2006	Metal	Chromium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Metal	Chromium VI	Total	=	100	%	SM 3500-Cr		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Metal	Chromium VI	Total	=	101	%	SM 3500-Cr		70	130	
2005/06-5	Lab	LCS, RPD	6/2/2006	Metal	Chromium VI	Total	=	1	%	SM 3500-Cr		0	30	
2005/06-5	Lab	method blank	6/2/2006	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr	5		5	
2005/06-5	ME-SCR	lab duplicate, RPD	6/2/2006	Metal	Chromium VI	Total	=	0	%	SM 3500-Cr		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/2/2006	Metal	Chromium VI	Total	=	97	%	SM 3500-Cr		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/2/2006	Metal	Chromium VI	Total	=	99	%	SM 3500-Cr		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/2/2006	Metal	Chromium VI	Total	=	2	%	SM 3500-Cr		0	30	
2005/06-5	Lab	method blank	6/13/2006	Metal	Copper	Dissolved	<	0.4	µg/L	EPA 200.8m	0.4		0.4	
2005/06-5	Lab	method blank	6/13/2006	Metal	Copper	Total	<	0.4	µg/L	EPA 200.8m	0.4		0.4	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-CC	field blank	6/13/2006	Metal	Copper	Total	<	0.4	µg/L	EPA 200.8m	0.4		0.4	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Metal	Copper	Total	=	1	%	EPA 200.8m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/13/2006	Metal	Copper	Total	=	92	%	EPA 200.8m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/13/2006	Metal	Copper	Total	=	91	%	EPA 200.8m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/13/2006	Metal	Copper	Total	=	1	%	EPA 200.8m		0	30	
2005/06-5	Lab	method blank	6/13/2006	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2005/06-5	Lab	method blank	6/13/2006	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2005/06-5	ME-CC	field blank	6/13/2006	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Metal	Lead	Total	=	2	%	EPA 200.8m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/13/2006	Metal	Lead	Total	=	83	%	EPA 200.8m		65	135	
2005/06-5	ME-SCR	matrix spike, rec	6/13/2006	Metal	Lead	Total	=	84	%	EPA 200.8m		65	135	
2005/06-5	ME-SCR	matrix spike, RPD	6/13/2006	Metal	Lead	Total	=	1	%	EPA 200.8m		0	30	
2005/06-5	Lab	method blank	6/22/2006	Metal	Mercury	Dissolved	<	0.5	ng/L	EPA 1631E	0.5		0.5	
2005/06-5	Lab	method blank	6/22/2006	Metal	Mercury	Total	<	0.5	ng/L	EPA 1631E	0.5		0.5	
2005/06-5	ME-CC	field blank	6/22/2006	Metal	Mercury	Total	=	1.1	ng/L	EPA 1631E	0.5		0.5	
2005/06-5	ME-SCR	lab duplicate, RPD	6/22/2006	Metal	Mercury	Total	=	5	%	EPA 1631E		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/22/2006	Metal	Mercury	Total	=	106	%	EPA 1631E		60	140	
2005/06-5	ME-SCR	matrix spike, rec	6/22/2006	Metal	Mercury	Total	=	109	%	EPA 1631E		60	140	
2005/06-5	ME-SCR	matrix spike, RPD	6/22/2006	Metal	Mercury	Total	=	3	%	EPA 1631E		0	30	
2005/06-5	Lab	method blank	6/13/2006	Metal	Nickel	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	Lab	method blank	6/13/2006	Metal	Nickel	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	ME-CC	field blank	6/13/2006	Metal	Nickel	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Metal	Nickel	Total	=	1	%	EPA 200.8m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/13/2006	Metal	Nickel	Total	=	92	%	EPA 200.8m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/13/2006	Metal	Nickel	Total	=	90	%	EPA 200.8m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/13/2006	Metal	Nickel	Total	=	2	%	EPA 200.8m		0	30	
2005/06-5	Lab	method blank	6/13/2006	Metal	Selenium	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	Lab	method blank	6/13/2006	Metal	Selenium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	ME-CC	field blank	6/13/2006	Metal	Selenium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Metal	Selenium	Total	=	8	%	EPA 200.8m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/13/2006	Metal	Selenium	Total	=	113	%	EPA 200.8m		60	150	
2005/06-5	ME-SCR	matrix spike, rec	6/13/2006	Metal	Selenium	Total	=	109	%	EPA 200.8m		60	150	
2005/06-5	ME-SCR	matrix spike, RPD	6/13/2006	Metal	Selenium	Total	=	4	%	EPA 200.8m		0	30	
2005/06-5	Lab	method blank	6/13/2006	Metal	Silver	Dissolved	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2005/06-5	Lab	method blank	6/13/2006	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2005/06-5	ME-CC	field blank	6/13/2006	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Metal	Silver	Total	=	0	%	EPA 200.8m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/13/2006	Metal	Silver	Total	=	94	%	EPA 200.8m		50	155	
2005/06-5	ME-SCR	matrix spike, rec	6/13/2006	Metal	Silver	Total	=	92	%	EPA 200.8m		50	155	
2005/06-5	ME-SCR	matrix spike, RPD	6/13/2006	Metal	Silver	Total	=	2	%	EPA 200.8m		0	30	
2005/06-5	Lab	method blank	6/13/2006	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-5	Lab	method blank	6/13/2006	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-5	ME-CC	field blank	6/13/2006	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Metal	Thallium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/13/2006	Metal	Thallium	Total	=	90	%	EPA 200.8m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/13/2006	Metal	Thallium	Total	=	89	%	EPA 200.8m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/13/2006	Metal	Thallium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-5	Lab	method blank	6/13/2006	Metal	Zinc	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-5	Lab	method blank	6/13/2006	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-5	ME-CC	field blank	6/13/2006	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/13/2006	Metal	Zinc	Total	=	1	%	EPA 200.8m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/13/2006	Metal	Zinc	Total	=	91	%	EPA 200.8m		50	150	
2005/06-5	ME-SCR	matrix spike, rec	6/13/2006	Metal	Zinc	Total	=	88	%	EPA 200.8m		50	150	
2005/06-5	ME-SCR	matrix spike, RPD	6/13/2006	Metal	Zinc	Total	=	3	%	EPA 200.8m		0	30	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Nutrient	Ammonia as N	n/a	=	96	%	SM 4500-NH3 F		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Nutrient	Ammonia as N	n/a	=	92	%	SM 4500-NH3 F		70	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	Lab	LCS, RPD	6/2/2006	Nutrient	Ammonia as N	n/a	=	4	%	SM 4500-NH3 F		0	30	
2005/06-5	Lab	method blank	6/2/2006	Nutrient	Ammonia as N	n/a	<	0.01	mg/L	SM 4500-NH3 F	0.01		0.01	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Nutrient	Nitrate as N	n/a	=	94	%	EPA 300.0		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Nutrient	Nitrate as N	n/a	=	95	%	EPA 300.0		70	130	
2005/06-5	Lab	LCS, RPD	6/2/2006	Nutrient	Nitrate as N	n/a	=	1	%	EPA 300.0		0	30	
2005/06-5	Lab	method blank	6/2/2006	Nutrient	Nitrate as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/2/2006	Nutrient	Nitrate as N	n/a	=	2	%	EPA 300.0		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/2/2006	Nutrient	Nitrate as N	n/a	=	106	%	EPA 300.0		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/2/2006	Nutrient	Nitrate as N	n/a	=	108	%	EPA 300.0		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/2/2006	Nutrient	Nitrate as N	n/a	=	2	%	EPA 300.0		0	30	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Nutrient	Nitrite as N	n/a	=	93	%	EPA 300.0		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Nutrient	Nitrite as N	n/a	=	95	%	EPA 300.0		70	130	
2005/06-5	Lab	LCS, RPD	6/2/2006	Nutrient	Nitrite as N	n/a	=	2	%	EPA 300.0		0	30	
2005/06-5	Lab	method blank	6/2/2006	Nutrient	Nitrite as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/2/2006	Nutrient	Nitrite as N	n/a	=	0	%	EPA 300.0		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/2/2006	Nutrient	Nitrite as N	n/a	=	97	%	EPA 300.0		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/2/2006	Nutrient	Nitrite as N	n/a	=	98	%	EPA 300.0		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/2/2006	Nutrient	Nitrite as N	n/a	=	1	%	EPA 300.0		0	30	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	98	%	EPA 300.0		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	98	%	EPA 300.0		70	130	
2005/06-5	Lab	LCS, RPD	6/2/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	0	%	EPA 300.0		0	30	
2005/06-5	Lab	method blank	6/2/2006	Nutrient	Orthophosphate as P (Diss)	n/a	<	0.0075	mg/L	SM 4500-P C	0.0075		0.0075	
2005/06-5	ME-SCR	lab duplicate, RPD	6/2/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	21	%	EPA 300.0		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/2/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	76	%	EPA 300.0		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/2/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	96	%	EPA 300.0		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/2/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	23	%	EPA 300.0		0	30	
2005/06-5	Lab	LCS, rec	6/21/2006	Nutrient	TKN	n/a	=	90.7	%	EPA 351.1		80	120	
2005/06-5	Lab	method blank	6/21/2006	Nutrient	TKN	n/a	<	0.05	mg/L	EPA 351.1	0.05		0.05	
2005/06-5	ME-CC	lab duplicate, RPD	6/21/2006	Nutrient	TKN	n/a	=	7.3	%	EPA 351.1		0	20	
2005/06-5	ME-VR2	matrix spike dup, rec	6/21/2006	Nutrient	TKN	n/a	=	84.5	%	EPA 351.1		80	120	
2005/06-5	ME-VR2	matrix spike, rec	6/21/2006	Nutrient	TKN	n/a	=	87.3	%	EPA 351.1		80	120	
2005/06-5	ME-VR2	matrix spike, RPD	6/26/2006	Nutrient	TKN	n/a	=	3.3	%	EPA 351.1		0	20	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Nutrient	Total Phosphorus	Dissolved	=	110	%	SM 4500-P C		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Nutrient	Total Phosphorus	Dissolved	=	112	%	SM 4500-P C		70	130	
2005/06-5	Lab	LCS, RPD	6/2/2006	Nutrient	Total Phosphorus	Dissolved	=	2	%	SM 4500-P C		0	30	
2005/06-5	Lab	method blank	6/6/2006	Nutrient	Total Phosphorus	Dissolved	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2005/06-5	ME-SCR	lab duplicate, RPD	6/6/2006	Nutrient	Total Phosphorus	Dissolved	=	0	%	SM 4500-P C		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/6/2006	Nutrient	Total Phosphorus	Dissolved	=	86	%	SM 4500-P C		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/6/2006	Nutrient	Total Phosphorus	Dissolved	=	82	%	SM 4500-P C		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/6/2006	Nutrient	Total Phosphorus	Dissolved	=	5	%	SM 4500-P C		0	30	
2005/06-5	Lab	LCS dup, rec	6/2/2006	Nutrient	Total Phosphorus	Total	=	92	%	SM 4500-P C		70	130	
2005/06-5	Lab	LCS, rec	6/2/2006	Nutrient	Total Phosphorus	Total	=	94	%	SM 4500-P C		70	130	
2005/06-5	Lab	LCS, RPD	6/2/2006	Nutrient	Total Phosphorus	Total	=	2	%	SM 4500-P C		0	30	
2005/06-5	Lab	method blank	6/6/2006	Nutrient	Total Phosphorus	Total	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2005/06-5	ME-SCR	lab duplicate, RPD	6/6/2006	Nutrient	Total Phosphorus	Total	=	3	%	SM 4500-P C		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/6/2006	Nutrient	Total Phosphorus	Total	=	73	%	SM 4500-P C		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/6/2006	Nutrient	Total Phosphorus	Total	=	98	%	SM 4500-P C		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/6/2006	Nutrient	Total Phosphorus	Total	=	29	%	SM 4500-P C		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	65	%	EPA 625m		65	140	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	65	%	EPA 625m		65	140	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	1,2-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	1,3-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	1,4-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	1,4-Dichlorobenzene	n/a	=	57	%	EPA 625m		50	140	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	1,4-Dichlorobenzene	n/a	=	58	%	EPA 625m		50	140	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	1,4-Dichlorobenzene	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	1-Methylnaphthalene	n/a	=	25	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	1-Methylnaphthalene	n/a	=	84	%	EPA 625m		50	120	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	1-Methylnaphthalene	n/a	=	93	%	EPA 625m		50	120	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	1-Methylnaphthalene	n/a	=	10	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	1-Methylphenanthrene	n/a	=	12	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	1-Methylphenanthrene	n/a	=	98	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	1-Methylphenanthrene	n/a	=	98	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	1-Methylphenanthrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	8	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	95	%	EPA 625m		45	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	93	%	EPA 625m		45	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	Lab	srgt method blank, rec	6/21/2006	Organic	2,4,6-Tribromophenol	n/a	=	74	%	EPA 625m		40	130	
2005/06-5	ME-CC	srgt environ, rec	6/21/2006	Organic	2,4,6-Tribromophenol	n/a	=	87	%	EPA 625m		40	130	
2005/06-5	ME-SCR	srgt environ, rec	6/21/2006	Organic	2,4,6-Tribromophenol	n/a	=	80	%	EPA 625m		40	130	
2005/06-5	ME-SCR	srgt matrix spike dup, rec	6/21/2006	Organic	2,4,6-Tribromophenol	n/a	=	92	%	EPA 625m		40	130	
2005/06-5	ME-SCR	srgt matrix spike, rec	6/21/2006	Organic	2,4,6-Tribromophenol	n/a	=	91	%	EPA 625m		40	130	
2005/06-5	ME-VR2	srgt environ, rec	6/21/2006	Organic	2,4,6-Tribromophenol	n/a	=	86	%	EPA 625m		40	130	
2005/06-5	Lab	method blank	6/21/2006	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2,4,6-Trichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2,4-Dichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	srgt method blank, rec	6/9/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	121	%	EPA 8151A		0	123	
2005/06-5	ME-CC	srgt environ, rec	6/9/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	95	%	EPA 8151A		0	123	
2005/06-5	ME-SCR	srgt environ, rec	6/9/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	101	%	EPA 8151A		0	123	
2005/06-5	ME-VR2	srgt environ, rec	6/20/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	98	%	EPA 8151A		0	123	
2005/06-5	Lab	method blank	6/21/2006	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2,4-Dimethylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2,4-Dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2,4-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	2,4-Dinitrotoluene	n/a	=	75	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	2,4-Dinitrotoluene	n/a	=	80	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	2,4-Dinitrotoluene	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	63	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	98	%	EPA 625m		55	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	93	%	EPA 625m		55	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	5	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2,6-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2-Chloronaphthalene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2-Chlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	2-Chlorophenol	n/a	=	65	%	EPA 625m		35	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	2-Chlorophenol	n/a	=	67	%	EPA 625m		35	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	2-Chlorophenol	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2-Methyl-4,6-dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2-Methylnaphthalene	n/a	=	18	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	2-Methylnaphthalene	n/a	=	85	%	EPA 625m		50	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	2-Methylnaphthalene	n/a	=	97	%	EPA 625m		50	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	2-Methylnaphthalene	n/a	=	13	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	2-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	3,3'-Dichlorobenzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	4-Bromophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	4-Chloro-3-methylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	4-Chloro-3-methylphenol	n/a	=	82	%	EPA 625m		30	150	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	4-Chloro-3-methylphenol	n/a	=	82	%	EPA 625m		30	150	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	4-Chloro-3-methylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	4-Chlorophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	4-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	4-Nitrophenol	n/a	=	11	%	EPA 625m		0.1	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	4-Nitrophenol	n/a	=	11	%	EPA 625m		0.1	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	4-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Acenaphthene	n/a	=	19	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Acenaphthene	n/a	=	108	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Acenaphthene	n/a	=	105	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Acenaphthene	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	srgt method blank, rec	6/21/2006	Organic	Acenaphthene-d10	n/a	=	78	%	EPA 625m		50	130	
2005/06-5	ME-CC	srgt environ, rec	6/21/2006	Organic	Acenaphthene-d10	n/a	=	82	%	EPA 625m		50	130	
2005/06-5	ME-SCR	srgt environ, rec	6/21/2006	Organic	Acenaphthene-d10	n/a	=	91	%	EPA 625m		50	130	
2005/06-5	ME-SCR	srgt matrix spike dup, rec	6/21/2006	Organic	Acenaphthene-d10	n/a	=	91	%	EPA 625m		50	130	
2005/06-5	ME-SCR	srgt matrix spike, rec	6/21/2006	Organic	Acenaphthene-d10	n/a	=	85	%	EPA 625m		50	130	
2005/06-5	ME-VR2	srgt environ, rec	6/21/2006	Organic	Acenaphthene-d10	n/a	=	81	%	EPA 625m		50	130	
2005/06-5	Lab	method blank	6/21/2006	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Acenaphthylene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Acenaphthylene	n/a	=	85	%	EPA 625m		60	120	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Acenaphthylene	n/a	=	93	%	EPA 625m		60	120	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Acenaphthylene	n/a	=	9	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Anthracene	n/a	=	90	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Anthracene	n/a	=	94	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Anthracene	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Azobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Benzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Benzo(a)anthracene	n/a	=	35	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Benzo(a)anthracene	n/a	=	104	%	EPA 625m		70	140	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Benzo(a)anthracene	n/a	=	102	%	EPA 625m		70	140	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Benzo(a)anthracene	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Benzo(a)pyrene	n/a	=	164	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Benzo(a)pyrene	n/a	=	103	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Benzo(a)pyrene	n/a	=	107	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Benzo(a)pyrene	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Benzo(b)fluoranthene	n/a	=	13	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Benzo(b)fluoranthene	n/a	=	101	%	EPA 625m		60	140	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Benzo(b)fluoranthene	n/a	=	99	%	EPA 625m		60	140	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Benzo(b)fluoranthene	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Benzo(e)pyrene	n/a	=	151	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Benzo(e)pyrene	n/a	=	101	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Benzo(e)pyrene	n/a	=	95	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Benzo(e)pyrene	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Benzo(g,h,i)perylene	n/a	=	51	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Benzo(g,h,i)perylene	n/a	=	92	%	EPA 625m		50	140	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Benzo(g,h,i)perylene	n/a	=	91	%	EPA 625m		50	140	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Benzo(g,h,i)perylene	n/a	=	1	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Benzo(k)fluoranthene	n/a	=	40	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Benzo(k)fluoranthene	n/a	=	98	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Benzo(k)fluoranthene	n/a	=	101	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Benzo(k)fluoranthene	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Biphenyl	n/a	=	67	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Biphenyl	n/a	=	94	%	EPA 625m		50	120	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Biphenyl	n/a	=	86	%	EPA 625m		50	120	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Biphenyl	n/a	=	9	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Bis(2-chloroethoxy)methane	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Bis(2-chloroethyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Bis(2-chloroisopropyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.0484	µg/L	EPA 625m	0.005		0.005	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	12	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	119	%	EPA 625m		20	190	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	116	%	EPA 625m		20	190	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Butyl benzyl phthalate	n/a	=	0.015	µg/L	EPA 625m	0.005		0.005	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Butyl benzyl phthalate	n/a	=	13	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Butyl benzyl phthalate	n/a	=	122	%	EPA 625m		65	160	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Butyl benzyl phthalate	n/a	=	116	%	EPA 625m		65	160	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Butyl benzyl phthalate	n/a	=	5	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Chrysene	n/a	=	14	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Chrysene	n/a	=	101	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Chrysene	n/a	=	97	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Chrysene	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	srgt method blank, rec	6/21/2006	Organic	Chrysene-d12	n/a	=	87	%	EPA 625m		70	130	
2005/06-5	ME-CC	srgt environ, rec	6/21/2006	Organic	Chrysene-d12	n/a	=	82	%	EPA 625m		70	130	
2005/06-5	ME-SCR	srgt environ, rec	6/21/2006	Organic	Chrysene-d12	n/a	=	85	%	EPA 625m		70	130	
2005/06-5	ME-SCR	srgt matrix spike dup, rec	6/21/2006	Organic	Chrysene-d12	n/a	=	94	%	EPA 625m		70	130	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	srgt matrix spike, rec	6/21/2006	Organic	Chrysene-d12	n/a	=	86	%	EPA 625m		70	130	
2005/06-5	ME-VR2	srgt environ, rec	6/21/2006	Organic	Chrysene-d12	n/a	=	89	%	EPA 625m		70	130	
2005/06-5	Lab	method blank	6/21/2006	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Dibenz(a,h)anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Dibenz(a,h)anthracene	n/a	=	99	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Dibenz(a,h)anthracene	n/a	=	101	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Dibenz(a,h)anthracene	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	matrix spike dup, rec	6/2/2006	Organic	Dibenzothiophene	n/a	=	103	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/2/2006	Organic	Dibenzothiophene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Dibenzothiophene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Dibenzothiophene	n/a	=	103	%	EPA 625m		70	130	
2005/06-5	Lab	method blank	6/21/2006	Organic	Diethyl phthalate	n/a	=	0.0421	µg/L	EPA 625m	0.005		0.005	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Diethyl phthalate	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Diethyl phthalate	n/a	=	124	%	EPA 625m		50	150	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Diethyl phthalate	n/a	=	108	%	EPA 625m		50	150	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Diethyl phthalate	n/a	=	14	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Dimethyl phthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Dimethyl phthalate	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Dimethyl phthalate	n/a	=	110	%	EPA 625m		40	155	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Dimethyl phthalate	n/a	=	103	%	EPA 625m		40	155	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Dimethyl phthalate	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Di-n-butylphthalate	n/a	=	0.0402	µg/L	EPA 625m	0.005		0.005	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Di-n-butylphthalate	n/a	=	31	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Di-n-butylphthalate	n/a	=	129	%	EPA 625m		65	145	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Di-n-butylphthalate	n/a	=	128	%	EPA 625m		65	145	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Di-n-butylphthalate	n/a	=	1	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Di-n-octylphthalate	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Di-n-octylphthalate	n/a	=	112	%	EPA 625m		50	165	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Di-n-octylphthalate	n/a	=	106	%	EPA 625m		50	165	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Di-n-octylphthalate	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Fluoranthene	n/a	=	17	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Fluoranthene	n/a	=	101	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Fluoranthene	n/a	=	98	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Fluoranthene	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Fluorene	n/a	=	58	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Fluorene	n/a	=	93	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Fluorene	n/a	=	90	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Fluorene	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Hexachlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Hexachlorobutadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Hexachlorocyclopentadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Hexachloroethane	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	32	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	96	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	98	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Isophorone	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Naphthalene	n/a	=	18	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Naphthalene	n/a	=	85	%	EPA 625m		50	120	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Naphthalene	n/a	=	80	%	EPA 625m		50	120	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Naphthalene	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	srgt method blank, rec	6/21/2006	Organic	Naphthalene-d8	n/a	=	67	%	EPA 625m		40	120	
2005/06-5	ME-CC	srgt environ, rec	6/21/2006	Organic	Naphthalene-d8	n/a	=	73	%	EPA 625m		40	120	
2005/06-5	ME-SCR	srgt environ, rec	6/21/2006	Organic	Naphthalene-d8	n/a	=	90	%	EPA 625m		40	120	
2005/06-5	ME-SCR	srgt matrix spike dup, rec	6/21/2006	Organic	Naphthalene-d8	n/a	=	84	%	EPA 625m		40	120	
2005/06-5	ME-SCR	srgt matrix spike, rec	6/21/2006	Organic	Naphthalene-d8	n/a	=	76	%	EPA 625m		40	120	
2005/06-5	ME-VR2	srgt environ, rec	6/21/2006	Organic	Naphthalene-d8	n/a	=	55	%	EPA 625m		40	120	
2005/06-5	Lab	method blank	6/21/2006	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Nitrobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	N-Nitrosodimethylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	83	%	EPA 625m		55	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	78	%	EPA 625m		55	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	N-Nitrosodiphenylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Pentachlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Pentachlorophenol	n/a	=	89	%	EPA 625m		10	160	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Pentachlorophenol	n/a	=	94	%	EPA 625m		10	160	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Pentachlorophenol	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Perylene	n/a	=	39	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Perylene	n/a	=	110	%	EPA 625m		65	135	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Perylene	n/a	=	109	%	EPA 625m		65	135	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Perylene	n/a	=	1	%	EPA 625m		0	30	
2005/06-5	Lab	srgt method blank, rec	6/21/2006	Organic	Perylene-d12	n/a	=	81	%	EPA 625m		60	140	
2005/06-5	ME-CC	srgt environ, rec	6/21/2006	Organic	Perylene-d12	n/a	=	74	%	EPA 625m		60	140	
2005/06-5	ME-SCR	srgt environ, rec	6/21/2006	Organic	Perylene-d12	n/a	=	88	%	EPA 625m		60	140	
2005/06-5	ME-SCR	srgt matrix spike dup, rec	6/21/2006	Organic	Perylene-d12	n/a	=	90	%	EPA 625m		60	140	
2005/06-5	ME-SCR	srgt matrix spike, rec	6/21/2006	Organic	Perylene-d12	n/a	=	77	%	EPA 625m		60	140	
2005/06-5	ME-VR2	srgt environ, rec	6/21/2006	Organic	Perylene-d12	n/a	=	80	%	EPA 625m		60	140	
2005/06-5	Lab	method blank	6/21/2006	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Phenanthrene	n/a	=	160	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Phenanthrene	n/a	=	96	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Phenanthrene	n/a	=	92	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Phenanthrene	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	srgt method blank, rec	6/21/2006	Organic	Phenanthrene-d10	n/a	=	86	%	EPA 625m		70	130	
2005/06-5	ME-CC	srgt environ, rec	6/21/2006	Organic	Phenanthrene-d10	n/a	=	86	%	EPA 625m		70	130	
2005/06-5	ME-SCR	srgt environ, rec	6/21/2006	Organic	Phenanthrene-d10	n/a	=	90	%	EPA 625m		70	130	
2005/06-5	ME-SCR	srgt matrix spike dup, rec	6/21/2006	Organic	Phenanthrene-d10	n/a	=	93	%	EPA 625m		70	130	
2005/06-5	ME-SCR	srgt matrix spike, rec	6/21/2006	Organic	Phenanthrene-d10	n/a	=	88	%	EPA 625m		70	130	
2005/06-5	ME-VR2	srgt environ, rec	6/21/2006	Organic	Phenanthrene-d10	n/a	=	92	%	EPA 625m		70	130	
2005/06-5	Lab	method blank	6/21/2006	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Phenol	n/a	=	40.3	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Phenol	n/a	=	29	%	EPA 625m		0.1	115	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Phenol	n/a	=	30	%	EPA 625m		0.1	115	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Phenol	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	srgt method blank, rec	6/21/2006	Organic	Phenol-d5	n/a	=	30	%	EPA 625m		10	110	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-CC	srgt environ, rec	6/21/2006	Organic	Phenol-d5	n/a	=	24	%	EPA 625m		10	110	
2005/06-5	ME-SCR	srgt environ, rec	6/21/2006	Organic	Phenol-d5	n/a	=	27	%	EPA 625m		10	110	
2005/06-5	ME-SCR	srgt matrix spike dup, rec	6/21/2006	Organic	Phenol-d5	n/a	=	33	%	EPA 625m		10	110	
2005/06-5	ME-SCR	srgt matrix spike, rec	6/21/2006	Organic	Phenol-d5	n/a	=	33	%	EPA 625m		10	110	
2005/06-5	ME-VR2	srgt environ, rec	6/21/2006	Organic	Phenol-d5	n/a	=	18	%	EPA 625m		10	110	
2005/06-5	Lab	method blank	6/21/2006	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Pyrene	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Organic	Pyrene	n/a	=	104	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Organic	Pyrene	n/a	=	113	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Organic	Pyrene	n/a	=	8	%	EPA 625m		0	30	
2005/06-5	Lab	srgt method blank, rec	6/21/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	72	%	EPA 625m		40	130	
2005/06-5	ME-CC	srgt environ, rec	6/21/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	84	%	EPA 625m		40	130	
2005/06-5	ME-SCR	srgt environ, rec	6/21/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	97	%	EPA 625m		40	130	
2005/06-5	ME-SCR	srgt matrix spike dup, rec	6/21/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	93	%	EPA 625m		40	130	
2005/06-5	ME-SCR	srgt matrix spike, rec	6/21/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	84	%	EPA 625m		40	130	
2005/06-5	ME-VR2	srgt environ, rec	6/21/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	82	%	EPA 625m		40	130	
2005/06-5	Lab	method blank	6/21/2006	Organic	Total Detectable PAHs	n/a	=	0	µg/L	EPA 625m				
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Organic	Total Detectable PAHs	n/a	=	19	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	Aroclor 1016	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	Aroclor 1221	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	Aroclor 1232	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	Aroclor 1242	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	Aroclor 1248	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	Aroclor 1254	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	Aroclor 1260	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 018	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 018	n/a	=	104	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 018	n/a	=	94	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 018	n/a	=	10	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 028	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 028	n/a	=	103	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 028	n/a	=	97	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 028	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	srgt method blank, rec	6/21/2006	PCB	PCB 030	n/a	=	84	%	EPA 625m		40	130	
2005/06-5	ME-CC	srgt environ, rec	6/21/2006	PCB	PCB 030	n/a	=	91	%	EPA 625m		40	130	
2005/06-5	ME-SCR	srgt environ, rec	6/21/2006	PCB	PCB 030	n/a	=	99	%	EPA 625m		40	130	
2005/06-5	ME-SCR	srgt matrix spike dup, rec	6/21/2006	PCB	PCB 030	n/a	=	96	%	EPA 625m		40	130	
2005/06-5	ME-SCR	srgt matrix spike, rec	6/21/2006	PCB	PCB 030	n/a	=	85	%	EPA 625m		40	130	
2005/06-5	ME-VR2	srgt environ, rec	6/16/2006	PCB	PCB 030	n/a	=	89	%	EPA 625m		40	130	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 031	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 031	n/a	=	105	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 031	n/a	=	97	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 031	n/a	=	8	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 033	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 033	n/a	=	107	%	EPA 625m		60	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 033	n/a	=	98	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 033	n/a	=	9	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 037	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 037	n/a	=	116	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 037	n/a	=	107	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 037	n/a	=	8	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 044	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 044	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 044	n/a	=	102	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 044	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 049	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 049	n/a	=	106	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 049	n/a	=	98	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 049	n/a	=	8	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 052	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 052	n/a	=	104	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 052	n/a	=	98	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 052	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 066	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 066	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 066	n/a	=	104	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 066	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 070	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 070	n/a	=	109	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 070	n/a	=	105	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 070	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 074	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 074	n/a	=	110	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 074	n/a	=	107	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 074	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 077	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 077	n/a	=	116	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 077	n/a	=	106	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 077	n/a	=	9	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 081	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 081	n/a	=	113	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 081	n/a	=	105	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 081	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 087	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 087	n/a	=	113	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 087	n/a	=	105	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 087	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 095	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 095	n/a	=	101	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 095	n/a	=	94	%	EPA 625m		60	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 095	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 097	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 097	n/a	=	109	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 097	n/a	=	99	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 097	n/a	=	10	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 099	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 099	n/a	=	113	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 099	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 099	n/a	=	5	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 101	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 101	n/a	=	109	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 101	n/a	=	99	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 101	n/a	=	10	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 105	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 105	n/a	=	113	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 105	n/a	=	104	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 105	n/a	=	8	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 110	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 110	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 110	n/a	=	102	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 110	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	srgt method blank, rec	6/21/2006	PCB	PCB 112	n/a	=	99	%	EPA 625m		60	120	
2005/06-5	ME-CC	srgt environ, rec	6/21/2006	PCB	PCB 112	n/a	=	89	%	EPA 625m		60	120	
2005/06-5	ME-SCR	srgt environ, rec	6/21/2006	PCB	PCB 112	n/a	=	94	%	EPA 625m		60	120	
2005/06-5	ME-SCR	srgt matrix spike dup, rec	6/21/2006	PCB	PCB 112	n/a	=	96	%	EPA 625m		60	120	
2005/06-5	ME-SCR	srgt matrix spike, rec	6/21/2006	PCB	PCB 112	n/a	=	80	%	EPA 625m		60	120	
2005/06-5	ME-VR2	srgt environ, rec	6/21/2006	PCB	PCB 112	n/a	=	93	%	EPA 625m		60	120	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 114	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 114	n/a	=	115	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 114	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 114	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 118	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 118	n/a	=	110	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 118	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 118	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 119	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 119	n/a	=	110	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 119	n/a	=	102	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 119	n/a	=	8	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 123	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 123	n/a	=	117	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 123	n/a	=	111	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 123	n/a	=	5	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 126	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 126	n/a	=	119	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 126	n/a	=	110	%	EPA 625m		60	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 126	n/a	=	8	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 128 + 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 128 + 167	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 128 + 167	n/a	=	112	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 128 + 167	n/a	=	105	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 128 + 167	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 138	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 138	n/a	=	116	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 138	n/a	=	109	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 138	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 141	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 141	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 141	n/a	=	102	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 141	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 149	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 149	n/a	=	102	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 149	n/a	=	100	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 149	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 151	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 151	n/a	=	111	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 151	n/a	=	103	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 151	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 153	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 153	n/a	=	109	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 153	n/a	=	105	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 153	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 156	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 156	n/a	=	118	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 156	n/a	=	109	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 156	n/a	=	8	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 157	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 157	n/a	=	117	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 157	n/a	=	104	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 157	n/a	=	12	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 158	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 158	n/a	=	117	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 158	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 158	n/a	=	8	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 168 + 132	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 168 + 132	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 168 + 132	n/a	=	100	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 168 + 132	n/a	=	8	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 169	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 169	n/a	=	114	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 169	n/a	=	110	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 169	n/a	=	4	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 170	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 170	n/a	=	113	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 170	n/a	=	107	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 170	n/a	=	5	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 177	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 177	n/a	=	118	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 177	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 177	n/a	=	9	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 180	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 180	n/a	=	115	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 180	n/a	=	107	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 180	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 183	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 183	n/a	=	109	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 183	n/a	=	104	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 183	n/a	=	5	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 187	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 187	n/a	=	113	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 187	n/a	=	105	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 187	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 189	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 189	n/a	=	110	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 189	n/a	=	106	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 189	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 194	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 194	n/a	=	120	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 194	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 194	n/a	=	11	%	EPA 625m		0	30	
2005/06-5	Lab	srvt method blank, rec	6/21/2006	PCB	PCB 198	n/a	=	95	%	EPA 625m		60	120	
2005/06-5	ME-CC	srvt environ, rec	6/21/2006	PCB	PCB 198	n/a	=	89	%	EPA 625m		60	120	
2005/06-5	ME-SCR	srvt environ, rec	6/21/2006	PCB	PCB 198	n/a	=	90	%	EPA 625m		60	120	
2005/06-5	ME-SCR	srvt matrix spike dup, rec	6/21/2006	PCB	PCB 198	n/a	=	94	%	EPA 625m		60	120	
2005/06-5	ME-SCR	srvt matrix spike, rec	6/21/2006	PCB	PCB 198	n/a	=	76	%	EPA 625m		60	120	
2005/06-5	ME-VR2	srvt environ, rec	6/21/2006	PCB	PCB 198	n/a	=	87	%	EPA 625m		60	120	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 200	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 200	n/a	=	106	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 200	n/a	=	103	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 200	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 201	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 201	n/a	=	114	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 201	n/a	=	110	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 201	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	PCB 206	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	PCB	PCB 206	n/a	=	108	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	PCB	PCB 206	n/a	=	102	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	PCB	PCB 206	n/a	=	6	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	Lab	method blank	6/21/2006	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	PCB	Total Detectable PCBs	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	LCS dup, rec	6/9/2006	Pesticide	2,4,5-T	n/a	=	112	%	EPA 8151A		30	130	
2005/06-5	Lab	LCS, rec	6/9/2006	Pesticide	2,4,5-T	n/a	=	112	%	EPA 8151A		30	130	
2005/06-5	Lab	LCS, RPD	6/9/2006	Pesticide	2,4,5-T	n/a	=	0	%	EPA 8151A		0	30	
2005/06-5	Lab	method blank	6/9/2006	Pesticide	2,4,5-T	n/a	<	0.17	µg/L	EPA 8151A	0.17			0.17
2005/06-5	ME-SCR	matrix spike dup, rec	6/9/2006	Pesticide	2,4,5-T	n/a	=	82	%	EPA 8151A		30	130	
2005/06-5	ME-SCR	matrix spike, rec	6/9/2006	Pesticide	2,4,5-T	n/a	=	93	%	EPA 8151A		30	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/9/2006	Pesticide	2,4,5-T	n/a	=	13	%	EPA 8151A		0	30	
2005/06-5	Lab	method blank	6/9/2006	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.12	µg/L	EPA 8151A	0.12			0.12
2005/06-5	Lab	LCS dup, rec	6/9/2006	Pesticide	2,4-D	n/a	=	83	%	EPA 8151A		30	130	
2005/06-5	Lab	LCS, rec	6/9/2006	Pesticide	2,4-D	n/a	=	85	%	EPA 8151A		30	130	
2005/06-5	Lab	LCS, RPD	6/9/2006	Pesticide	2,4-D	n/a	=	3	%	EPA 8151A		0	30	
2005/06-5	Lab	method blank	6/9/2006	Pesticide	2,4-D	n/a	<	1.5	µg/L	EPA 8151A	1.5			1.5
2005/06-5	ME-SCR	matrix spike dup, rec	6/9/2006	Pesticide	2,4-D	n/a	=	76	%	EPA 8151A		30	130	
2005/06-5	ME-SCR	matrix spike, rec	6/9/2006	Pesticide	2,4-D	n/a	=	54	%	EPA 8151A		30	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/9/2006	Pesticide	2,4-D	n/a	=	34	%	EPA 8151A		0	30	
2005/06-5	Lab	LCS dup, rec	6/9/2006	Pesticide	2,4-DB	n/a	=	115	%	EPA 8151A		30	130	
2005/06-5	Lab	LCS, rec	6/9/2006	Pesticide	2,4-DB	n/a	=	121	%	EPA 8151A		30	130	
2005/06-5	Lab	LCS, RPD	6/9/2006	Pesticide	2,4-DB	n/a	=	5	%	EPA 8151A		0	30	
2005/06-5	Lab	method blank	6/9/2006	Pesticide	2,4-DB	n/a	<	4.1	µg/L	EPA 8151A	4.1			4.1
2005/06-5	ME-SCR	matrix spike dup, rec	6/9/2006	Pesticide	2,4-DB	n/a	=	60	%	EPA 8151A		30	130	
2005/06-5	ME-SCR	matrix spike, rec	6/9/2006	Pesticide	2,4-DB	n/a	=	52	%	EPA 8151A		30	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/9/2006	Pesticide	2,4-DB	n/a	=	14	%	EPA 8151A		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	2,4'-DDD	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	2,4'-DDD	n/a	=	103	%	EPA 625m		50	140	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	2,4'-DDD	n/a	=	102	%	EPA 625m		50	140	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	2,4'-DDD	n/a	=	1	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	2,4'-DDE	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	2,4'-DDE	n/a	=	101	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	2,4'-DDE	n/a	=	96	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	2,4'-DDE	n/a	=	5	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	2,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	2,4'-DDT	n/a	=	52	%	EPA 625m		40	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	2,4'-DDT	n/a	=	64	%	EPA 625m		40	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	2,4'-DDT	n/a	=	21	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	4,4'-DDD	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	4,4'-DDD	n/a	=	102	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	4,4'-DDD	n/a	=	101	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	4,4'-DDD	n/a	=	1	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	4,4'-DDE	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	4,4'-DDE	n/a	=	106	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	4,4'-DDE	n/a	=	103	%	EPA 625m		70	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	4,4'-DDE	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	4,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	4,4'-DDT	n/a	=	44	%	EPA 625m		0.001	150	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	4,4'-DDT	n/a	=	48	%	EPA 625m		0.001	150	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	4,4'-DDT	n/a	=	9	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Aldrin	n/a	=	0	%	EPA 625m		0	30	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Aldrin	n/a	=	105	%	EPA 625m		50	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Aldrin	n/a	=	101	%	EPA 625m		50	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Aldrin	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	BHC-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	BHC-alpha	n/a	=	100	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	BHC-alpha	n/a	=	100	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	BHC-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	BHC-beta	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	BHC-beta	n/a	=	100	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	BHC-beta	n/a	=	97	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	BHC-beta	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	BHC-delta	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	BHC-delta	n/a	=	102	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	BHC-delta	n/a	=	102	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	BHC-delta	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	92	%	EPA 625m		50	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	90	%	EPA 625m		50	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Bolstar	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Bolstar	n/a	=	109	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Bolstar	n/a	=	111	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Bolstar	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Chlordane-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Chlordane-alpha	n/a	=	105	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Chlordane-alpha	n/a	=	99	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Chlordane-alpha	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Chlordane-gamma	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Chlordane-gamma	n/a	=	102	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Chlordane-gamma	n/a	=	96	%	EPA 625m		60	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Chlordane-gamma	n/a	=	6	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Chlorpyrifos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Chlorpyrifos	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Chlorpyrifos	n/a	=	109	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Chlorpyrifos	n/a	=	102	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Chlorpyrifos	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	cis-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	cis-Nonachlor	n/a	=	91	%	EPA 625m		60	120	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	cis-Nonachlor	n/a	=	96	%	EPA 625m		60	120	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	cis-Nonachlor	n/a	=	5.3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/9/2006	Pesticide	Dalapon	n/a	<	2.6	µg/L	EPA 8151A	2.6		2.6	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Demeton-O	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Demeton-O	n/a	=	103	%	EPA 625m		45	105	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Demeton-O	n/a	=	100	%	EPA 625m		45	105	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Demeton-O	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Diazinon	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Diazinon	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Diazinon	n/a	=	109	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Diazinon	n/a	=	98	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Diazinon	n/a	=	11	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/9/2006	Pesticide	Dicamba	n/a	<	0.13	µg/L	EPA 8151A	0.13		0.13	
2005/06-5	Lab	method blank	6/9/2006	Pesticide	Dichlorprop	n/a	<	1.5	µg/L	EPA 8151A	1.5		1.5	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Dichlorvos	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Dichlorvos	n/a	=	107	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Dichlorvos	n/a	=	100	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Dichlorvos	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Dieldrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Dieldrin	n/a	=	102	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Dieldrin	n/a	=	105	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Dieldrin	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Dimethoate	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Dimethoate	n/a	=	98	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Dimethoate	n/a	=	97	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Dimethoate	n/a	=	1	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/9/2006	Pesticide	Dinoseb	n/a	<	0.3	µg/L	EPA 8151A	0.3		0.3	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Disulfoton	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Disulfoton	n/a	=	98	%	EPA 625m		45	105	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Disulfoton	n/a	=	85	%	EPA 625m		45	105	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Disulfoton	n/a	=	14	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Endosulfan sulfate	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Endosulfan sulfate	n/a	=	99	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Endosulfan sulfate	n/a	=	103	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Endosulfan sulfate	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Endosulfan-I	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Endosulfan-I	n/a	=	100	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Endosulfan-I	n/a	=	93	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Endosulfan-I	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Endosulfan-II	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Endosulfan-II	n/a	=	101	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Endosulfan-II	n/a	=	86	%	EPA 625m		60	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Endosulfan-II	n/a	=	16	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Endrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Endrin	n/a	=	102	%	EPA 625m		65	135	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Endrin	n/a	=	102	%	EPA 625m		65	135	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Endrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Endrin aldehyde	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Endrin ketone	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Endrin ketone	n/a	=	75	%	EPA 625m		40	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Endrin ketone	n/a	=	75	%	EPA 625m		40	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Endrin ketone	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Ethoprop	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Ethoprop	n/a	=	103	%	EPA 625m		65	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Ethoprop	n/a	=	103	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Ethoprop	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	102	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	105	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Fenchlorophos (Ronnel)	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Fensulfothion	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Fensulfothion	n/a	=	115	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Fensulfothion	n/a	=	109	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Fensulfothion	n/a	=	5	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Fenthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Fenthion	n/a	=	102	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Fenthion	n/a	=	106	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Fenthion	n/a	=	4	%	EPA 625m		0	30	
2005/06-5	Lab	LCS, rec	6/6/2006	Pesticide	Glyphosate	n/a	=	120	%	EPA 547		71	137	
2005/06-5	Lab	method blank	6/26/2006	Pesticide	Glyphosate	n/a	<	5	µg/L	EPA 547	5		5	
2005/06-5	ME-VR2	matrix spike dup, rec	6/6/2006	Pesticide	Glyphosate	n/a	=	109	%	EPA 547		68	134	
2005/06-5	ME-VR2	matrix spike, rec	6/6/2006	Pesticide	Glyphosate	n/a	=	107	%	EPA 547		68	134	
2005/06-5	ME-VR2	matrix spike, RPD	6/6/2006	Pesticide	Glyphosate	n/a	=	1.7	%	EPA 547		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Heptachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Heptachlor	n/a	=	93	%	EPA 625m		45	135	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Heptachlor	n/a	=	95	%	EPA 625m		45	135	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Heptachlor	n/a	=	2	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Heptachlor epoxide	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Heptachlor epoxide	n/a	=	97	%	EPA 625m		65	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Heptachlor epoxide	n/a	=	102	%	EPA 625m		65	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Heptachlor epoxide	n/a	=	5	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Malathion	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Malathion	n/a	=	109	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Malathion	n/a	=	110	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Malathion	n/a	=	1	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/9/2006	Pesticide	MCPA	n/a	<	110	µg/L	EPA 8151A	110		110	
2005/06-5	Lab	method blank	6/9/2006	Pesticide	MCPP	n/a	<	110	µg/L	EPA 8151A	110		110	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Merphos	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Merphos	n/a	=	106	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Merphos	n/a	=	111	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Merphos	n/a	=	5	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Methoxychlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Methoxychlor	n/a	=	43	%	EPA 625m		0.001	155	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Methoxychlor	n/a	=	51	%	EPA 625m		0.001	155	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Methoxychlor	n/a	=	17	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Methyl parathion	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Methyl parathion	n/a	=	103	%	EPA 625m		60	120	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Methyl parathion	n/a	=	104	%	EPA 625m		60	120	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Methyl parathion	n/a	=	1	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		0.008	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Mevinphos	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Mevinphos	n/a	=	97	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Mevinphos	n/a	=	92	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Mevinphos	n/a	=	5	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Mirex	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Mirex	n/a	=	83	%	EPA 625m		50	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Mirex	n/a	=	83	%	EPA 625m		50	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Mirex	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Oxychlorthane	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Oxychlorthane	n/a	=	96	%	EPA 625m		60	120	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Oxychlorthane	n/a	=	90	%	EPA 625m		60	120	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Oxychlorthane	n/a	=	7	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006		0.006	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Phorate	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Phorate	n/a	=	104	%	EPA 625m		45	105	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Phorate	n/a	=	103	%	EPA 625m		45	105	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Phorate	n/a	=	1	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	109	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	110	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	1	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Tokuthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Tokuthion	n/a	=	113	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Tokuthion	n/a	=	110	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Tokuthion	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Total Detectable DDTs	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Toxaphene	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	trans-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	trans-Nonachlor	n/a	=	103	%	EPA 625m		55	130	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	trans-Nonachlor	n/a	=	100	%	EPA 625m		55	130	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	trans-Nonachlor	n/a	=	3	%	EPA 625m		0	30	
2005/06-5	Lab	method blank	6/21/2006	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-5	ME-SCR	lab duplicate, RPD	6/21/2006	Pesticide	Trichloronate	n/a	=	0	%	EPA 625m		0	30	
2005/06-5	ME-SCR	matrix spike dup, rec	6/21/2006	Pesticide	Trichloronate	n/a	=	109	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, rec	6/21/2006	Pesticide	Trichloronate	n/a	=	104	%	EPA 625m		65	125	
2005/06-5	ME-SCR	matrix spike, RPD	6/21/2006	Pesticide	Trichloronate	n/a	=	5	%	EPA 625m		0	30	
2005/06-6	Lab	LCS dup, rec	6/15/2006	Anion	Bromide	n/a	=	103	%	EPA 300.0		70	130	
2005/06-6	Lab	LCS, rec	6/15/2006	Anion	Bromide	n/a	=	106	%	EPA 300.0		70	130	
2005/06-6	Lab	LCS, RPD	6/15/2006	Anion	Bromide	n/a	=	2.9	%	EPA 300.0		0	30	
2005/06-6	Lab	method blank	6/15/2006	Anion	Bromide	n/a	<	0.001	mg/L	EPA 300.0	0.001		0.001	
2005/06-6	ME-CC	lab duplicate, RPD	6/15/2006	Anion	Bromide	n/a	=	6.4	%	EPA 300.0		0	30	
2005/06-6	ME-CC	matrix spike dup, rec	6/15/2006	Anion	Bromide	n/a	=	94	%	EPA 300.0		70	130	
2005/06-6	ME-CC	matrix spike, rec	6/15/2006	Anion	Bromide	n/a	=	101	%	EPA 300.0		70	130	
2005/06-6	ME-CC	matrix spike, RPD	6/15/2006	Anion	Bromide	n/a	=	7	%	EPA 300.0		0	30	
2005/06-6	ME-SCR	field duplicate	6/15/2006	Anion	Bromide	n/a	=	0.317	mg/L	EPA 300.0	0.001			
2005/06-6	Lab	LCS dup, rec	6/20/2006	Anion	Chloride	n/a	=	111	%	SM 4500-Cl E		70	130	
2005/06-6	Lab	LCS, rec	6/20/2006	Anion	Chloride	n/a	=	111	%	SM 4500-Cl E		70	130	
2005/06-6	Lab	LCS, RPD	6/20/2006	Anion	Chloride	n/a	=	0	%	SM 4500-Cl E		0	30	
2005/06-6	Lab	method blank	6/20/2006	Anion	Chloride	n/a	<	0.01	mg/L	SM 4500-Cl E	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	6/20/2006	Anion	Chloride	n/a	=	55.2	mg/L	SM 4500-Cl E	0.01			

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	Lab	LCS dup, rec	6/20/2006	Anion	Perchlorate	n/a	=	95	%	EPA 314.0		85	115	
2005/06-6	Lab	LCS, rec	6/20/2006	Anion	Perchlorate	n/a	=	96	%	EPA 314.0		85	115	
2005/06-6	Lab	LCS, RPD	6/20/2006	Anion	Perchlorate	n/a	=	1	%	EPA 314.0		0	15	
2005/06-6	Lab	method blank	6/20/2006	Anion	Perchlorate	n/a	<	0.43	µg/L	EPA 314.0	0.43		0.43	
2005/06-6	ME-SCR	field duplicate	6/21/2006	Anion	Perchlorate	n/a	<	0.43	µg/L	EPA 314.0	0.43			
2005/06-6	ME-SCR	field duplicate	6/13/2006	Bacteriological	E. Coli	n/a	=	41	MPN/100 mL	Colilert-18	10			
2005/06-6	ME-SCR	field duplicate	6/13/2006	Bacteriological	Enterococcus	n/a	<	10	MPN/100 mL	Enterolert	10			
2005/06-6	ME-SCR	field duplicate	6/13/2006	Bacteriological	Fecal Coliform	n/a	=	30	MPN/100 mL	Colilert-18	2			
2005/06-6	ME-SCR	field duplicate	6/13/2006	Bacteriological	Total Coliform	n/a	=	1529	MPN/100 mL	Colilert-18	10			
2005/06-6	Lab	method blank	6/19/2006	Conventional	BOD	n/a	<	0.58	mg/L	EPA 405.1	0.58		0.58	
2005/06-6	ME-SCR	field duplicate	6/14/2006	Conventional	BOD	n/a	=	150	mg/L	EPA 405.1	0.58			
2005/06-6	ME-SCR	lab duplicate, RPD	6/14/2006	Conventional	BOD	n/a	=	0	%	EPA 405.1		0	25	
2005/06-6	Lab	method blank	6/15/2006	Conventional	Conductivity	n/a	<	1	µmhos/cm	SM 2510	1		1	
2005/06-6	ME-CC	lab duplicate, RPD	6/15/2006	Conventional	Conductivity	n/a	=	0	%	SM 2510		0	30	
2005/06-6	ME-SCR	field duplicate	6/15/2006	Conventional	Conductivity	n/a	=	1400	µmhos/cm	SM 2510	1			
2005/06-6	Lab	method blank	6/26/2006	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1		1	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Conventional	Hardness as CaCO3	Total	=	431	mg/L	SM 2340 B	1			
2005/06-6	ME-VR2	lab duplicate, RPD	6/21/2006	Conventional	Hardness as CaCO3	Total	=	0.7	%	SM 2340 B		0	30	
2005/06-6	ME-CC	lab duplicate, RPD	6/15/2006	Conventional	pH	n/a	=	1.2	%	EPA 150.1		0	30	
2005/06-6	ME-SCR	field duplicate	6/15/2006	Conventional	pH	n/a	=	8.3	pH Units	EPA 150.1	0.1			
2005/06-6	Lab	LCS, rec	6/19/2006	Conventional	Total Dissolved Solids	n/a	=	100	%	SM 2540 C		70	130	
2005/06-6	Lab	LCS dup, rec	6/21/2006	Conventional	Total Dissolved Solids	n/a	=	99	%	SM 2540 C		70	130	
2005/06-6	Lab	LCS, RPD	6/21/2006	Conventional	Total Dissolved Solids	n/a	=	1	%	SM 2540 C		0	30	
2005/06-6	Lab	method blank	6/21/2006	Conventional	Total Dissolved Solids	n/a	<	0.1	mg/L	SM 2540 C	0.1		0.1	
2005/06-6	ME-CC	lab duplicate, RPD	6/21/2006	Conventional	Total Dissolved Solids	n/a	=	4.1	%	SM 2540 C		0	30	
2005/06-6	ME-SCR	field duplicate	6/21/2006	Conventional	Total Dissolved Solids	n/a	=	1040	mg/L	SM 2540 C	0.1			
2005/06-6	Lab	LCS, rec	6/19/2006	Conventional	Total Organic Carbon	n/a	=	104	%	EPA 415.1		80	120	
2005/06-6	Lab	method blank	6/19/2006	Conventional	Total Organic Carbon	n/a	<	0.021	mg/L	EPA 415.1	0.021		0.021	
2005/06-6	ME-SCR	field duplicate	6/19/2006	Conventional	Total Organic Carbon	n/a	=	88	mg/L	EPA 415.1	0.21			
2005/06-6	Lab	method blank	6/21/2006	Conventional	Total Suspended Solids	n/a	<	0.5	mg/L	SM 2540 D	0.5		0.5	
2005/06-6	ME-SCR	field duplicate	6/21/2006	Conventional	Total Suspended Solids	n/a	=	90	mg/L	SM 2540 D	0.5			
2005/06-6	ME-VR2	lab duplicate, RPD	6/21/2006	Conventional	Total Suspended Solids	n/a	=	14.3	%	SM 2540 D		0	30	
2005/06-6	Lab	method blank	6/15/2006	Conventional	Turbidity	n/a	<	1	NTU	EPA 180.1	1		1	
2005/06-6	ME-CC	lab duplicate, RPD	6/15/2006	Conventional	Turbidity	n/a	=	0	%	EPA 180.1		0	30	
2005/06-6	ME-SCR	field duplicate	6/15/2006	Conventional	Turbidity	n/a	=	2	NTU	EPA 180.1	1			
2005/06-6	Lab	LCS, rec	6/19/2006	Hydrocarbon	Oil and Grease	n/a	=	87	%	EPA 1664A		70	130	
2005/06-6	Lab	LCS dup, rec	6/28/2006	Hydrocarbon	Oil and Grease	n/a	=	99	%	EPA 1664A		70	130	
2005/06-6	Lab	LCS, RPD	6/28/2006	Hydrocarbon	Oil and Grease	n/a	=	12.9	%	EPA 1664A		0	30	
2005/06-6	Lab	method blank	6/28/2006	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1		1	
2005/06-6	ME-SCR	field duplicate	6/28/2006	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1			
2005/06-6	Lab	LCS, rec	6/19/2006	Hydrocarbon	TRPH	n/a	=	83	%	EPA 1664		70	130	
2005/06-6	Lab	LCS dup, rec	6/28/2006	Hydrocarbon	TRPH	n/a	=	98	%	EPA 1664		70	130	
2005/06-6	Lab	LCS, RPD	6/28/2006	Hydrocarbon	TRPH	n/a	=	16.6	%	EPA 1664		0	30	
2005/06-6	Lab	method blank	6/28/2006	Hydrocarbon	TRPH	n/a	<	0.01	mg/L	EPA 1664	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	6/28/2006	Hydrocarbon	TRPH	n/a	=	0.4	mg/L	EPA 1664	0.01			
2005/06-6	Lab	method blank	6/26/2006	Metal	Aluminum	Dissolved	<	5	µg/L	EPA 200.8m	5		5	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Aluminum	Dissolved	<	5	µg/L	EPA 200.8m	5			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Aluminum	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/26/2006	Metal	Aluminum	Total	=	82	%	EPA 200.8m		50	140	
2005/06-6	Lab	LCS, rec	6/26/2006	Metal	Aluminum	Total	=	84	%	EPA 200.8m		50	140	
2005/06-6	Lab	LCS, RPD	6/26/2006	Metal	Aluminum	Total	=	2.4	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Aluminum	Total	<	5	µg/L	EPA 200.8m	5		5	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Aluminum	Total	=	251	µg/L	EPA 200.8m	5			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Aluminum	Total	=	0	%	EPA 200.8m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Metal	Aluminum	Total	=	94	%	EPA 200.8m		50	140	
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Metal	Aluminum	Total	=	93	%	EPA 200.8m		50	140	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Metal	Aluminum	Total	=	1	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Arsenic	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Arsenic	Dissolved	=	1.3	µg/L	EPA 200.8m	0.2			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Arsenic	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/26/2006	Metal	Arsenic	Total	=	96	%	EPA 200.8m		70	130	
2005/06-6	Lab	LCS, rec	6/26/2006	Metal	Arsenic	Total	=	98	%	EPA 200.8m		70	130	
2005/06-6	Lab	LCS, RPD	6/26/2006	Metal	Arsenic	Total	=	2.1	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Arsenic	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Arsenic	Total	=	1.2	µg/L	EPA 200.8m	0.2			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Arsenic	Total	=	0	%	EPA 200.8m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Metal	Arsenic	Total	=	115	%	EPA 200.8m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Metal	Arsenic	Total	=	114	%	EPA 200.8m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Metal	Arsenic	Total	=	1	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Cadmium	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Cadmium	Dissolved	=	1	µg/L	EPA 200.8m	0.2			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Cadmium	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/26/2006	Metal	Cadmium	Total	=	104	%	EPA 200.8m		75	130	
2005/06-6	Lab	LCS, rec	6/26/2006	Metal	Cadmium	Total	=	105	%	EPA 200.8m		75	130	
2005/06-6	Lab	LCS, RPD	6/26/2006	Metal	Cadmium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Cadmium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Cadmium	Total	=	0.9	µg/L	EPA 200.8m	0.2			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Cadmium	Total	=	22.2	%	EPA 200.8m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Metal	Cadmium	Total	=	103	%	EPA 200.8m		75	130	
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Metal	Cadmium	Total	=	103	%	EPA 200.8m		75	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Metal	Cadmium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Chromium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Chromium	Dissolved	=	0.9	µg/L	EPA 200.8m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Chromium	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/26/2006	Metal	Chromium	Total	=	95	%	EPA 200.8m		70	130	
2005/06-6	Lab	LCS, rec	6/26/2006	Metal	Chromium	Total	=	96	%	EPA 200.8m		70	130	
2005/06-6	Lab	LCS, RPD	6/26/2006	Metal	Chromium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Chromium	Total	=	2.3	µg/L	EPA 200.8m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Chromium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Metal	Chromium	Total	=	103	%	EPA 200.8m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Metal	Chromium	Total	=	103	%	EPA 200.8m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Metal	Chromium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/15/2006	Metal	Chromium VI	Total	=	97	%	SM 3500-Cr		70	130	
2005/06-6	Lab	LCS, rec	6/15/2006	Metal	Chromium VI	Total	=	101	%	SM 3500-Cr		70	130	
2005/06-6	Lab	LCS, RPD	6/15/2006	Metal	Chromium VI	Total	=	4	%	SM 3500-Cr		0	30	
2005/06-6	Lab	method blank	6/15/2006	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr	5		5	
2005/06-6	ME-CC	lab duplicate, RPD	6/15/2006	Metal	Chromium VI	Total	=	0	%	SM 3500-Cr		0	30	
2005/06-6	ME-SCR	field duplicate	6/15/2006	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr	5			
2005/06-6	ME-VR2	matrix spike dup, rec	6/15/2006	Metal	Chromium VI	Total	=	98	%	SM 3500-Cr		70	130	
2005/06-6	ME-VR2	matrix spike, rec	6/15/2006	Metal	Chromium VI	Total	=	97	%	SM 3500-Cr		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/15/2006	Metal	Chromium VI	Total	=	1	%	SM 3500-Cr		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Copper	Dissolved	<	0.4	µg/L	EPA 200.8m	0.4		0.4	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Copper	Dissolved	=	2	µg/L	EPA 200.8m	0.4			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Copper	Dissolved	=	14.3	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/26/2006	Metal	Copper	Total	=	95	%	EPA 200.8m		70	130	
2005/06-6	Lab	LCS, rec	6/26/2006	Metal	Copper	Total	=	95	%	EPA 200.8m		70	130	
2005/06-6	Lab	LCS, RPD	6/26/2006	Metal	Copper	Total	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Copper	Total	<	0.4	µg/L	EPA 200.8m	0.4		0.4	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Copper	Total	=	2.6	µg/L	EPA 200.8m	0.4			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Copper	Total	=	8	%	EPA 200.8m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Metal	Copper	Total	=	98	%	EPA 200.8m		70	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Metal	Copper	Total	=	97	%	EPA 200.8m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Metal	Copper	Total	=	1	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Lead	Dissolved	=	0.14	µg/L	EPA 200.8m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Lead	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/26/2006	Metal	Lead	Total	=	108	%	EPA 200.8m		65	135	
2005/06-6	Lab	LCS, rec	6/26/2006	Metal	Lead	Total	=	103	%	EPA 200.8m		65	135	
2005/06-6	Lab	LCS, RPD	6/26/2006	Metal	Lead	Total	=	4.7	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Lead	Total	=	0.42	µg/L	EPA 200.8m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Lead	Total	=	16	%	EPA 200.8m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Metal	Lead	Total	=	111	%	EPA 200.8m		65	135	
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Metal	Lead	Total	=	114	%	EPA 200.8m		65	135	
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Metal	Lead	Total	=	3	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/22/2006	Metal	Mercury	Dissolved	<	0.5	ng/L	EPA 1631E	0.5		0.5	
2005/06-6	ME-SCR	field duplicate	6/22/2006	Metal	Mercury	Dissolved	=	2.4	ng/L	EPA 1631E	0.5			
2005/06-6	Lab	LCS dup, rec	6/22/2006	Metal	Mercury	Total	=	109	%	EPA 1631E		60	140	
2005/06-6	Lab	LCS, rec	6/22/2006	Metal	Mercury	Total	=	107	%	EPA 1631E		60	140	
2005/06-6	Lab	LCS, RPD	6/22/2006	Metal	Mercury	Total	=	1.9	%	EPA 1631E		0	30	
2005/06-6	Lab	method blank	6/22/2006	Metal	Mercury	Total	<	0.5	ng/L	EPA 1631E	0.5		0.5	
2005/06-6	ME-SCR	field duplicate	6/22/2006	Metal	Mercury	Total	=	4.3	ng/L	EPA 1631E	0.5			
2005/06-6	ME-VR2	lab duplicate, RPD	6/22/2006	Metal	Mercury	Total	=	4.3	%	EPA 1631E		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/22/2006	Metal	Mercury	Total	=	109	%	EPA 1631E		60	140	
2005/06-6	ME-VR2	matrix spike, rec	6/22/2006	Metal	Mercury	Total	=	105	%	EPA 1631E		60	140	
2005/06-6	ME-VR2	matrix spike, RPD	6/22/2006	Metal	Mercury	Total	=	4	%	EPA 1631E		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Nickel	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Nickel	Dissolved	=	2.4	µg/L	EPA 200.8m	0.2			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Nickel	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/26/2006	Metal	Nickel	Total	=	93	%	EPA 200.8m		70	130	
2005/06-6	Lab	LCS, rec	6/26/2006	Metal	Nickel	Total	=	94	%	EPA 200.8m		70	130	
2005/06-6	Lab	LCS, RPD	6/26/2006	Metal	Nickel	Total	=	1.1	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Nickel	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Nickel	Total	=	2.7	µg/L	EPA 200.8m	0.2			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Nickel	Total	=	3.9	%	EPA 200.8m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Metal	Nickel	Total	=	95	%	EPA 200.8m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Metal	Nickel	Total	=	95	%	EPA 200.8m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Metal	Nickel	Total	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Selenium	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Selenium	Dissolved	=	6.8	µg/L	EPA 200.8m	0.2			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Selenium	Dissolved	=	3.8	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/26/2006	Metal	Selenium	Total	=	101	%	EPA 200.8m		60	150	
2005/06-6	Lab	LCS, rec	6/26/2006	Metal	Selenium	Total	=	103	%	EPA 200.8m		60	150	
2005/06-6	Lab	LCS, RPD	6/26/2006	Metal	Selenium	Total	=	2	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Selenium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Selenium	Total	=	5.8	µg/L	EPA 200.8m	0.2			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Selenium	Total	=	4.1	%	EPA 200.8m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Metal	Selenium	Total	=	123	%	EPA 200.8m		60	150	
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Metal	Selenium	Total	=	122	%	EPA 200.8m		60	150	
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Metal	Selenium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Silver	Dissolved	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Silver	Dissolved	<	0.5	µg/L	EPA 200.8m	0.5			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Silver	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/26/2006	Metal	Silver	Total	=	94	%	EPA 200.8m		50	155	
2005/06-6	Lab	LCS, rec	6/26/2006	Metal	Silver	Total	=	104	%	EPA 200.8m		50	155	
2005/06-6	Lab	LCS, RPD	6/26/2006	Metal	Silver	Total	=	10.1	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5		0.5	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Silver	Total	=	0	%	EPA 200.8m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Metal	Silver	Total	=	105	%	EPA 200.8m		50	155	
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Metal	Silver	Total	=	100	%	EPA 200.8m		50	155	
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Metal	Silver	Total	=	5	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Thallium	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/26/2006	Metal	Thallium	Total	=	103	%	EPA 200.8m		70	130	
2005/06-6	Lab	LCS, rec	6/26/2006	Metal	Thallium	Total	=	101	%	EPA 200.8m		70	130	
2005/06-6	Lab	LCS, RPD	6/26/2006	Metal	Thallium	Total	=	2	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Thallium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Metal	Thallium	Total	=	112	%	EPA 200.8m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Metal	Thallium	Total	=	111	%	EPA 200.8m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Metal	Thallium	Total	=	1	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Zinc	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Zinc	Dissolved	=	2.9	µg/L	EPA 200.8m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Zinc	Dissolved	=	0	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/26/2006	Metal	Zinc	Total	=	99	%	EPA 200.8m		50	150	
2005/06-6	Lab	LCS, rec	6/26/2006	Metal	Zinc	Total	=	100	%	EPA 200.8m		50	150	
2005/06-6	Lab	LCS, RPD	6/26/2006	Metal	Zinc	Total	=	1	%	EPA 200.8m		0	30	
2005/06-6	Lab	method blank	6/26/2006	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Metal	Zinc	Total	=	4.6	µg/L	EPA 200.8m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	6/26/2006	Metal	Zinc	Total	=	0	%	EPA 200.8m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Metal	Zinc	Total	=	100	%	EPA 200.8m		50	150	
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Metal	Zinc	Total	=	101	%	EPA 200.8m		50	150	
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Metal	Zinc	Total	=	1	%	EPA 200.8m		0	30	
2005/06-6	Lab	LCS dup, rec	6/19/2006	Nutrient	Ammonia as N	n/a	=	92	%	SM 4500-NH3 F		70	130	
2005/06-6	Lab	LCS, rec	6/19/2006	Nutrient	Ammonia as N	n/a	=	92	%	SM 4500-NH3 F		70	130	
2005/06-6	Lab	LCS, RPD	6/19/2006	Nutrient	Ammonia as N	n/a	=	0	%	SM 4500-NH3 F		0	30	
2005/06-6	Lab	method blank	6/19/2006	Nutrient	Ammonia as N	n/a	<	0.01	mg/L	SM 4500-NH3 F	0.01		0.01	
2005/06-6	ME-CC	matrix spike dup, rec	6/19/2006	Nutrient	Ammonia as N	n/a	=	88	%	SM 4500-NH3 F		70	130	
2005/06-6	ME-CC	matrix spike, rec	6/19/2006	Nutrient	Ammonia as N	n/a	=	88	%	SM 4500-NH3 F		70	130	
2005/06-6	ME-CC	matrix spike, RPD	6/19/2006	Nutrient	Ammonia as N	n/a	=	0	%	SM 4500-NH3 F		0	30	
2005/06-6	ME-SCR	field duplicate	6/19/2006	Nutrient	Ammonia as N	n/a	=	0.03	mg/L	SM 4500-NH3 F	0.01			
2005/06-6	Lab	LCS dup, rec	6/15/2006	Nutrient	Nitrate as N	n/a	=	100	%	EPA 300.0		70	130	
2005/06-6	Lab	LCS, RPD	6/15/2006	Nutrient	Nitrate as N	n/a	=	3	%	EPA 300.0		0	30	
2005/06-6	Lab	method blank	6/15/2006	Nutrient	Nitrate as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2005/06-6	Lab	LCS, rec	6/19/2006	Nutrient	Nitrate as N	n/a	=	103	%	EPA 300.0		70	130	
2005/06-6	ME-CC	lab duplicate, RPD	6/15/2006	Nutrient	Nitrate as N	n/a	=	0.4	%	EPA 300.0		0	30	
2005/06-6	ME-CC	matrix spike dup, rec	6/15/2006	Nutrient	Nitrate as N	n/a	=	29	%	EPA 300.0		70	130	
2005/06-6	ME-CC	matrix spike, rec	6/15/2006	Nutrient	Nitrate as N	n/a	=	34	%	EPA 300.0		70	130	
2005/06-6	ME-CC	matrix spike, RPD	6/15/2006	Nutrient	Nitrate as N	n/a	=	16	%	EPA 300.0		0	30	
2005/06-6	ME-SCR	field duplicate	6/15/2006	Nutrient	Nitrate as N	n/a	=	1.99	mg/L	EPA 300.0	0.01			
2005/06-6	Lab	LCS dup, rec	6/15/2006	Nutrient	Nitrite as N	n/a	=	80	%	EPA 300.0		70	130	
2005/06-6	Lab	LCS, RPD	6/15/2006	Nutrient	Nitrite as N	n/a	=	1.2	%	EPA 300.0		0	30	
2005/06-6	Lab	method blank	6/15/2006	Nutrient	Nitrite as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2005/06-6	Lab	LCS, rec	6/19/2006	Nutrient	Nitrite as N	n/a	=	81	%	EPA 300.0		70	130	
2005/06-6	ME-CC	lab duplicate, RPD	6/15/2006	Nutrient	Nitrite as N	n/a	=	6.1	%	EPA 300.0		0	30	
2005/06-6	ME-CC	matrix spike dup, rec	6/15/2006	Nutrient	Nitrite as N	n/a	=	112	%	EPA 300.0		70	130	
2005/06-6	ME-CC	matrix spike, rec	6/15/2006	Nutrient	Nitrite as N	n/a	=	112	%	EPA 300.0		70	130	
2005/06-6	ME-CC	matrix spike, RPD	6/15/2006	Nutrient	Nitrite as N	n/a	=	0	%	EPA 300.0		0	30	
2005/06-6	ME-SCR	field duplicate	6/15/2006	Nutrient	Nitrite as N	n/a	=	0.09	mg/L	EPA 300.0	0.01			
2005/06-6	Lab	LCS dup, rec	6/15/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	101	%	EPA 300.0		70	130	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	Lab	LCS, RPD	6/15/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	4	%	EPA 300.0		0	30	
2005/06-6	Lab	method blank	6/15/2006	Nutrient	Orthophosphate as P (Diss)	n/a	<	0.0075	mg/L	EPA 300.0	0.0075		0.0075	
2005/06-6	Lab	LCS, rec	6/19/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	97	%	EPA 300.0		70	130	
2005/06-6	ME-CC	lab duplicate, RPD	6/15/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	0.6	%	EPA 300.0		0	30	
2005/06-6	ME-CC	matrix spike dup, rec	6/15/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	95	%	EPA 300.0		70	130	
2005/06-6	ME-CC	matrix spike, rec	6/15/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	95	%	EPA 300.0		70	130	
2005/06-6	ME-CC	matrix spike, RPD	6/15/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	0	%	EPA 300.0		0	30	
2005/06-6	ME-SCR	field duplicate	6/15/2006	Nutrient	Orthophosphate as P (Diss)	n/a	=	0.08	mg/L	EPA 300.0	0.0075			
2005/06-6	Lab	method blank	6/26/2006	Nutrient	TKN	n/a	<	0.01	mg/L	EPA 351.1	0.01		0.01	
2005/06-6	Lab	LCS, rec	6/27/2006	Nutrient	TKN	n/a	=	90.7	%	EPA 351.1		80	120	
2005/06-6	ME-CC	lab duplicate, RPD	6/27/2006	Nutrient	TKN	n/a	=	0	%	EPA 351.1		0	20	
2005/06-6	ME-SCR	field duplicate	6/27/2006	Nutrient	TKN	n/a	=	0.17	mg/L	EPA 351.1	0.01			
2005/06-6	ME-VR2	matrix spike dup, rec	6/26/2006	Nutrient	TKN	n/a	=	83.7	%	EPA 351.1		80	120	
2005/06-6	ME-VR2	matrix spike, rec	6/26/2006	Nutrient	TKN	n/a	=	90.6	%	EPA 351.1		80	120	
2005/06-6	ME-VR2	matrix spike, RPD	6/26/2006	Nutrient	TKN	n/a	=	7.9	%	EPA 351.1		0	20	
2005/06-6	Lab	LCS, rec	6/19/2006	Nutrient	Total Phosphorus	Dissolved	=	96	%	SM 4500-P C		70	130	
2005/06-6	Lab	LCS dup, rec	6/23/2006	Nutrient	Total Phosphorus	Dissolved	=	94	%	SM 4500-P C		70	130	
2005/06-6	Lab	LCS, RPD	6/23/2006	Nutrient	Total Phosphorus	Dissolved	=	2.1	%	SM 4500-P C		0	30	
2005/06-6	Lab	method blank	6/23/2006	Nutrient	Total Phosphorus	Dissolved	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2005/06-6	ME-CC	lab duplicate, RPD	6/23/2006	Nutrient	Total Phosphorus	Dissolved	=	0.6	%	SM 4500-P C		0	30	
2005/06-6	ME-SCR	field duplicate	6/23/2006	Nutrient	Total Phosphorus	Dissolved	=	0.096	mg/L	SM 4500-P C	0.016			
2005/06-6	ME-VR2	matrix spike dup, rec	6/23/2006	Nutrient	Total Phosphorus	Dissolved	=	92	%	SM 4500-P C		70	130	
2005/06-6	ME-VR2	matrix spike, rec	6/23/2006	Nutrient	Total Phosphorus	Dissolved	=	76	%	SM 4500-P C		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/23/2006	Nutrient	Total Phosphorus	Dissolved	=	19	%	SM 4500-P C		0	30	
2005/06-6	Lab	LCS, rec	6/19/2006	Nutrient	Total Phosphorus	Total	=	96	%	SM 4500-P C		70	130	
2005/06-6	Lab	LCS dup, rec	6/23/2006	Nutrient	Total Phosphorus	Total	=	94	%	SM 4500-P C		70	130	
2005/06-6	Lab	LCS, RPD	6/23/2006	Nutrient	Total Phosphorus	Total	=	2.1	%	SM 4500-P C		0	30	
2005/06-6	Lab	method blank	6/23/2006	Nutrient	Total Phosphorus	Total	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2005/06-6	ME-CC	lab duplicate, RPD	6/23/2006	Nutrient	Total Phosphorus	Total	=	5.6	%	SM 4500-P C		0	30	
2005/06-6	ME-SCR	field duplicate	6/23/2006	Nutrient	Total Phosphorus	Total	=	0.056	mg/L	SM 4500-P C	0.016			
2005/06-6	ME-VR2	matrix spike dup, rec	6/23/2006	Nutrient	Total Phosphorus	Total	=	82	%	SM 4500-P C		70	130	
2005/06-6	ME-VR2	matrix spike, rec	6/23/2006	Nutrient	Total Phosphorus	Total	=	84	%	SM 4500-P C		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/23/2006	Nutrient	Total Phosphorus	Total	=	2	%	SM 4500-P C		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	73	%	EPA 625m		65	140	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	65	%	EPA 625m		65	140	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	1,2,4-Trichlorobenzene	n/a	=	12	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	1,2-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	1,3-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	1,4-Dichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	1,4-Dichlorobenzene	n/a	=	56	%	EPA 625m		50	140	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	1,4-Dichlorobenzene	n/a	=	55	%	EPA 625m		50	140	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	1,4-Dichlorobenzene	n/a	=	1	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	1-Methylnaphthalene	n/a	=	0.0041	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	1-Methylnaphthalene	n/a	=	2.9	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	1-Methylnaphthalene	n/a	=	82	%	EPA 625m		50	120	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	1-Methylnaphthalene	n/a	=	77	%	EPA 625m		50	120	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	1-Methylnaphthalene	n/a	=	6	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	1-Methylphenanthrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	1-Methylphenanthrene	n/a	=	95	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	1-Methylphenanthrene	n/a	=	97	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	1-Methylphenanthrene	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	96	%	EPA 625m		45	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	91	%	EPA 625m		45	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	2,3,5-Trimethylnaphthalene	n/a	=	5	%	EPA 625m		0	30	
2005/06-6	Lab	srgt method blank, rec	7/11/2006	Organic	2,4,6-Tribromophenol	n/a	=	74	%	EPA 625m		40	130	
2005/06-6	ME-CC	srgt environ, rec	7/11/2006	Organic	2,4,6-Tribromophenol	n/a	=	84	%	EPA 625m		40	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	2,4,6-Tribromophenol	n/a	=	80	%	EPA 625m		40	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	2,4,6-Tribromophenol	n/a	=	85	%	EPA 625m		40	130	
2005/06-6	ME-VR2	srgt environ, rec	7/11/2006	Organic	2,4,6-Tribromophenol	n/a	=	84	%	EPA 625m		40	130	
2005/06-6	ME-VR2	srgt matrix spike dup, rec	7/11/2006	Organic	2,4,6-Tribromophenol	n/a	=	84	%	EPA 625m		40	130	
2005/06-6	ME-VR2	srgt matrix spike, rec	7/11/2006	Organic	2,4,6-Tribromophenol	n/a	=	80	%	EPA 625m		40	130	
2005/06-6	Lab	method blank	7/11/2006	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2,4,6-Trichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2,4-Dichlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	srgt method blank, rec	6/21/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	59	%	EPA 8151A		0	123	
2005/06-6	ME-CC	srgt environ, rec	6/20/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	86	%	EPA 8151A		0	123	
2005/06-6	ME-SCR	srgt environ, rec	6/20/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	82	%	EPA 8151A		0	123	
2005/06-6	ME-SCR	srgt environ, rec	6/20/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	129	%	EPA 8151A		0	123	
2005/06-6	ME-VR2	srgt environ, rec	6/20/2006	Organic	2,4-Dichlorophenylacetic acid	n/a	=	142	%	EPA 8151A		0	123	
2005/06-6	Lab	method blank	7/11/2006	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2,4-Dimethylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2,4-Dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2,4-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	2,4-Dinitrotoluene	n/a	=	99	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	2,4-Dinitrotoluene	n/a	=	96	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	2,4-Dinitrotoluene	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	12.9	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	89	%	EPA 625m		55	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	77	%	EPA 625m		55	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	2,6-Dimethylnaphthalene	n/a	=	14	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2,6-Dinitrotoluene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2-Chloronaphthalene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2-Chlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	2-Chlorophenol	n/a	=	45	%	EPA 625m		35	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	2-Chlorophenol	n/a	=	55	%	EPA 625m		35	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	2-Chlorophenol	n/a	=	20	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2-Methyl-4,6-dinitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2-Methylnaphthalene	n/a	=	0.0056	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2-Methylnaphthalene	n/a	=	19.1	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	2-Methylnaphthalene	n/a	=	84	%	EPA 625m		50	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	2-Methylnaphthalene	n/a	=	78	%	EPA 625m		50	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	2-Methylnaphthalene	n/a	=	7	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	2-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	3,3'-Dichlorobenzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	4-Bromophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	4-Chloro-3-methylphenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	4-Chloro-3-methylphenol	n/a	=	97	%	EPA 625m		30	150	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	4-Chloro-3-methylphenol	n/a	=	90	%	EPA 625m		30	150	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	4-Chloro-3-methylphenol	n/a	=	7	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	4-Chlorophenyl phenyl ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	4-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	4-Nitrophenol	n/a	=	18	%	EPA 625m		0	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	4-Nitrophenol	n/a	=	20	%	EPA 625m		0	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	4-Nitrophenol	n/a	=	11	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Acenaphthene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Acenaphthene	n/a	=	96	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Acenaphthene	n/a	=	98	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Acenaphthene	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	srgt method blank, rec	7/11/2006	Organic	Acenaphthene-d10	n/a	=	70	%	EPA 625m		50	130	
2005/06-6	ME-CC	srgt environ, rec	7/11/2006	Organic	Acenaphthene-d10	n/a	=	83	%	EPA 625m		50	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Acenaphthene-d10	n/a	=	90	%	EPA 625m		50	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Acenaphthene-d10	n/a	=	95	%	EPA 625m		50	130	
2005/06-6	ME-VR2	srgt environ, rec	7/11/2006	Organic	Acenaphthene-d10	n/a	=	97	%	EPA 625m		50	130	
2005/06-6	ME-VR2	srgt matrix spike dup, rec	7/11/2006	Organic	Acenaphthene-d10	n/a	=	93	%	EPA 625m		50	130	
2005/06-6	ME-VR2	srgt matrix spike, rec	7/11/2006	Organic	Acenaphthene-d10	n/a	=	87	%	EPA 625m		50	130	
2005/06-6	Lab	method blank	7/11/2006	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Acenaphthylene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Acenaphthylene	n/a	=	88	%	EPA 625m		60	120	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Acenaphthylene	n/a	=	82	%	EPA 625m		60	120	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Acenaphthylene	n/a	=	7	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Anthracene	n/a	=	100	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Anthracene	n/a	=	96	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Anthracene	n/a	=	4	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Azobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Benzidine	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Benzo(a)anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Benzo(a)anthracene	n/a	=	95	%	EPA 625m		70	140	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Benzo(a)anthracene	n/a	=	98	%	EPA 625m		70	140	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Benzo(a)anthracene	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Benzo(a)pyrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Benzo(a)pyrene	n/a	=	93	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Benzo(a)pyrene	n/a	=	94	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Benzo(a)pyrene	n/a	=	1	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Benzo(b)fluoranthene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Benzo(b)fluoranthene	n/a	=	99	%	EPA 625m		60	140	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Benzo(b)fluoranthene	n/a	=	97	%	EPA 625m		60	140	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Benzo(b)fluoranthene	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Benzo(e)pyrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Benzo(e)pyrene	n/a	=	101	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Benzo(e)pyrene	n/a	=	98	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Benzo(e)pyrene	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Benzo(g,h,i)perylene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Benzo(g,h,i)perylene	n/a	=	97	%	EPA 625m		50	140	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Benzo(g,h,i)perylene	n/a	=	92	%	EPA 625m		50	140	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Benzo(g,h,i)perylene	n/a	=	5	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Benzo(k)fluoranthene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Benzo(k)fluoranthene	n/a	=	98	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Benzo(k)fluoranthene	n/a	=	98	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Benzo(k)fluoranthene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Biphenyl	n/a	=	0.002	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Biphenyl	n/a	=	77.4	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Biphenyl	n/a	=	85	%	EPA 625m		50	120	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Biphenyl	n/a	=	78	%	EPA 625m		50	120	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Biphenyl	n/a	=	9	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Bis(2-chloroethoxy)methane	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Bis(2-chloroethyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Bis(2-chloroisopropyl)ether	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.0274	µg/L	EPA 625m	0.005		0.005	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.19	µg/L	EPA 625m	0.005			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	23.9	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	91	%	EPA 625m		20	190	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	93	%	EPA 625m		20	190	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Butyl benzyl phthalate	n/a	=	0.014	µg/L	EPA 625m	0.005		0.005	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Butyl benzyl phthalate	n/a	=	0.0253	µg/L	EPA 625m	0.005			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Butyl benzyl phthalate	n/a	=	4.5	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Butyl benzyl phthalate	n/a	=	101	%	EPA 625m		65	160	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Butyl benzyl phthalate	n/a	=	103	%	EPA 625m		65	160	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Butyl benzyl phthalate	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Chrysene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Chrysene	n/a	=	100	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Chrysene	n/a	=	103	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Chrysene	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	srgt method blank, rec	7/11/2006	Organic	Chrysene-d12	n/a	=	89	%	EPA 625m		70	130	
2005/06-6	ME-CC	srgt environ, rec	7/11/2006	Organic	Chrysene-d12	n/a	=	98	%	EPA 625m		70	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Chrysene-d12	n/a	=	103	%	EPA 625m		70	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Chrysene-d12	n/a	=	99	%	EPA 625m		70	130	
2005/06-6	ME-VR2	srgt environ, rec	7/11/2006	Organic	Chrysene-d12	n/a	=	98	%	EPA 625m		70	130	
2005/06-6	ME-VR2	srgt matrix spike dup, rec	7/11/2006	Organic	Chrysene-d12	n/a	=	97	%	EPA 625m		70	130	
2005/06-6	ME-VR2	srgt matrix spike, rec	7/11/2006	Organic	Chrysene-d12	n/a	=	104	%	EPA 625m		70	130	
2005/06-6	Lab	method blank	7/11/2006	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Dibenz(a,h)anthracene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Dibenz(a,h)anthracene	n/a	=	84	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Dibenz(a,h)anthracene	n/a	=	96	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Dibenz(a,h)anthracene	n/a	=	13	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Dibenzothiophene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Dibenzothiophene	n/a	=	100	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Dibenzothiophene	n/a	=	104	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Dibenzothiophene	n/a	=	4	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Diethyl phthalate	n/a	=	0.0812	µg/L	EPA 625m	0.005		0.005	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Diethyl phthalate	n/a	=	0.434	µg/L	EPA 625m	0.005			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Diethyl phthalate	n/a	=	4.7	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Diethyl phthalate	n/a	=	101	%	EPA 625m		50	150	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Diethyl phthalate	n/a	=	90	%	EPA 625m		50	150	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Diethyl phthalate	n/a	=	12	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Dimethyl phthalate	n/a	=	0.0552	µg/L	EPA 625m	0.005		0.005	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Dimethyl phthalate	n/a	=	0.0794	µg/L	EPA 625m	0.005			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Dimethyl phthalate	n/a	=	4.7	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Dimethyl phthalate	n/a	=	91	%	EPA 625m		40	155	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Dimethyl phthalate	n/a	=	77	%	EPA 625m		40	155	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Dimethyl phthalate	n/a	=	17	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Di-n-butylphthalate	n/a	=	0.0387	µg/L	EPA 625m	0.005		0.005	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Di-n-butylphthalate	n/a	=	0.0647	µg/L	EPA 625m	0.005			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Di-n-butylphthalate	n/a	=	9.1	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Di-n-butylphthalate	n/a	=	101	%	EPA 625m		65	145	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Di-n-butylphthalate	n/a	=	92	%	EPA 625m		65	145	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Di-n-butylphthalate	n/a	=	9	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Di-n-octylphthalate	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Di-n-octylphthalate	n/a	=	103	%	EPA 625m		50	165	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Di-n-octylphthalate	n/a	=	103	%	EPA 625m		50	165	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Di-n-octylphthalate	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Fluoranthene	n/a	=	0.0027	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Fluoranthene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Fluoranthene	n/a	=	95	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Fluoranthene	n/a	=	98	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Fluoranthene	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Fluorene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Fluorene	n/a	=	88	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Fluorene	n/a	=	90	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Fluorene	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Hexachlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Hexachlorobutadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Hexachlorocyclopentadiene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Hexachloroethane	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	98	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	95	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Isophorone	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Naphthalene	n/a	=	0.0165	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Naphthalene	n/a	=	95.2	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Naphthalene	n/a	=	67	%	EPA 625m		50	120	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Naphthalene	n/a	=	59	%	EPA 625m		50	120	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Naphthalene	n/a	=	13	%	EPA 625m		0	30	
2005/06-6	Lab	srgt method blank, rec	7/11/2006	Organic	Naphthalene-d8	n/a	=	60	%	EPA 625m		40	120	
2005/06-6	ME-CC	srgt environ, rec	7/11/2006	Organic	Naphthalene-d8	n/a	=	79	%	EPA 625m		40	120	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Naphthalene-d8	n/a	=	72	%	EPA 625m		40	120	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Naphthalene-d8	n/a	=	77	%	EPA 625m		40	120	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-VR2	srgt environ, rec	7/11/2006	Organic	Naphthalene-d8	n/a	=	83	%	EPA 625m		40	120	
2005/06-6	ME-VR2	srgt matrix spike dup, rec	7/11/2006	Organic	Naphthalene-d8	n/a	=	73	%	EPA 625m		40	120	
2005/06-6	ME-VR2	srgt matrix spike, rec	7/11/2006	Organic	Naphthalene-d8	n/a	=	68	%	EPA 625m		40	120	
2005/06-6	Lab	method blank	7/11/2006	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Nitrobenzene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	N-Nitrosodimethylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	111	%	EPA 625m		55	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	91	%	EPA 625m		55	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	N-Nitrosodi-N-propylamine	n/a	=	20	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	N-Nitrosodiphenylamine	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Pentachlorophenol	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Pentachlorophenol	n/a	=	66	%	EPA 625m		10	160	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Pentachlorophenol	n/a	=	79	%	EPA 625m		10	160	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Pentachlorophenol	n/a	=	18	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Perylene	n/a	=	0.0216	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Perylene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Perylene	n/a	=	95	%	EPA 625m		65	135	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Perylene	n/a	=	90	%	EPA 625m		65	135	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Perylene	n/a	=	5	%	EPA 625m		0	30	
2005/06-6	Lab	srgt method blank, rec	7/11/2006	Organic	Perylene-d12	n/a	=	83	%	EPA 625m		60	140	
2005/06-6	ME-CC	srgt environ, rec	7/11/2006	Organic	Perylene-d12	n/a	=	88	%	EPA 625m		60	140	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Perylene-d12	n/a	=	93	%	EPA 625m		60	140	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Perylene-d12	n/a	=	97	%	EPA 625m		60	140	
2005/06-6	ME-VR2	srgt environ, rec	7/11/2006	Organic	Perylene-d12	n/a	=	93	%	EPA 625m		60	140	
2005/06-6	ME-VR2	srgt matrix spike dup, rec	7/11/2006	Organic	Perylene-d12	n/a	=	94	%	EPA 625m		60	140	
2005/06-6	ME-VR2	srgt matrix spike, rec	7/11/2006	Organic	Perylene-d12	n/a	=	98	%	EPA 625m		60	140	
2005/06-6	Lab	method blank	7/11/2006	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Phenanthrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Phenanthrene	n/a	=	100	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Phenanthrene	n/a	=	98	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Phenanthrene	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	srgt method blank, rec	7/11/2006	Organic	Phenanthrene-d10	n/a	=	83	%	EPA 625m		70	130	
2005/06-6	ME-CC	srgt environ, rec	7/11/2006	Organic	Phenanthrene-d10	n/a	=	94	%	EPA 625m		70	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Phenanthrene-d10	n/a	=	96	%	EPA 625m		70	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Phenanthrene-d10	n/a	=	99	%	EPA 625m		70	130	
2005/06-6	ME-VR2	srgt environ, rec	7/11/2006	Organic	Phenanthrene-d10	n/a	=	100	%	EPA 625m		70	130	
2005/06-6	ME-VR2	srgt matrix spike dup, rec	7/11/2006	Organic	Phenanthrene-d10	n/a	=	96	%	EPA 625m		70	130	
2005/06-6	ME-VR2	srgt matrix spike, rec	7/11/2006	Organic	Phenanthrene-d10	n/a	=	101	%	EPA 625m		70	130	
2005/06-6	Lab	method blank	7/11/2006	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Phenol	n/a	=	1.3	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Phenol	n/a	=	40	%	EPA 625m		0	115	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Phenol	n/a	=	30	%	EPA 625m		0	115	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Phenol	n/a	=	29	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	Lab	srgt method blank, rec	7/11/2006	Organic	Phenol-d5	n/a	=	29	%	EPA 625m		10	110	
2005/06-6	ME-CC	srgt environ, rec	7/11/2006	Organic	Phenol-d5	n/a	=	23	%	EPA 625m		10	110	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Phenol-d5	n/a	=	21	%	EPA 625m		10	110	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Phenol-d5	n/a	=	20	%	EPA 625m		10	110	
2005/06-6	ME-VR2	srgt environ, rec	7/11/2006	Organic	Phenol-d5	n/a	=	26	%	EPA 625m		10	110	
2005/06-6	ME-VR2	srgt matrix spike dup, rec	7/11/2006	Organic	Phenol-d5	n/a	=	33	%	EPA 625m		10	110	
2005/06-6	ME-VR2	srgt matrix spike, rec	7/11/2006	Organic	Phenol-d5	n/a	=	31	%	EPA 625m		10	110	
2005/06-6	Lab	method blank	7/11/2006	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Pyrene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Organic	Pyrene	n/a	=	92	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Organic	Pyrene	n/a	=	95	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Organic	Pyrene	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	srgt method blank, rec	7/11/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	75	%	EPA 625m		40	130	
2005/06-6	ME-CC	srgt environ, rec	7/11/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	98	%	EPA 625m		40	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	86	%	EPA 625m		40	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	85	%	EPA 625m		40	130	
2005/06-6	ME-VR2	srgt environ, rec	7/11/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	83	%	EPA 625m		40	130	
2005/06-6	ME-VR2	srgt matrix spike dup, rec	7/11/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	80	%	EPA 625m		40	130	
2005/06-6	ME-VR2	srgt matrix spike, rec	7/11/2006	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	87	%	EPA 625m		40	130	
2005/06-6	Lab	method blank	7/11/2006	Organic	Total Detectable PAHs	n/a	=	0	µg/L	EPA 625m				
2005/06-6	ME-SCR	field duplicate	7/11/2006	Organic	Total Detectable PAHs	n/a	=	0.0525	µg/L	EPA 625m				
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Organic	Total Detectable PAHs	n/a	=	43.3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	Aroclor 1016	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	Aroclor 1221	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	Aroclor 1232	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	Aroclor 1242	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	Aroclor 1248	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	Aroclor 1254	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	Aroclor 1260	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 018	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 018	n/a	=	90	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 018	n/a	=	92	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 018	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 028	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 028	n/a	=	86	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 028	n/a	=	96	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 028	n/a	=	11	%	EPA 625m		0	30	



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	Lab	srgt method blank, rec	7/11/2006	PCB	PCB 030	n/a	=	82	%	EPA 625m		40	130	
2005/06-6	ME-CC	srgt environ, rec	7/11/2006	PCB	PCB 030	n/a	=	96	%	EPA 625m		40	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	PCB	PCB 030	n/a	=	88	%	EPA 625m		40	130	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	PCB	PCB 030	n/a	=	94	%	EPA 625m		40	130	
2005/06-6	ME-VR2	srgt environ, rec	7/11/2006	PCB	PCB 030	n/a	=	85	%	EPA 625m		40	130	
2005/06-6	ME-VR2	srgt matrix spike dup, rec	7/11/2006	PCB	PCB 030	n/a	=	85	%	EPA 625m		40	130	
2005/06-6	ME-VR2	srgt matrix spike, rec	7/11/2006	PCB	PCB 030	n/a	=	92	%	EPA 625m		40	130	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 031	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 031	n/a	=	87	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 031	n/a	=	96	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 031	n/a	=	10	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 033	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 033	n/a	=	85	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 033	n/a	=	95	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 033	n/a	=	11	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 037	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 037	n/a	=	91	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 037	n/a	=	103	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 037	n/a	=	12	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 044	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 044	n/a	=	90	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 044	n/a	=	99	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 044	n/a	=	10	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 049	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 049	n/a	=	85	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 049	n/a	=	93	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 049	n/a	=	9	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 052	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 052	n/a	=	80	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 052	n/a	=	89	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 052	n/a	=	11	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 066	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 066	n/a	=	75	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 066	n/a	=	82	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 066	n/a	=	9	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 070	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 070	n/a	=	81	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 070	n/a	=	91	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 070	n/a	=	12	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 074	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 074	n/a	=	88	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 074	n/a	=	100	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 074	n/a	=	13	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 077	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 077	n/a	=	87	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 077	n/a	=	101	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 077	n/a	=	15	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 081	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 081	n/a	=	84	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 081	n/a	=	92	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 081	n/a	=	9	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 087	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 087	n/a	=	89	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 087	n/a	=	102	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 087	n/a	=	14	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 095	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 095	n/a	=	93	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 095	n/a	=	97	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 095	n/a	=	4	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 097	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 097	n/a	=	102	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 097	n/a	=	104	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 097	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 099	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 099	n/a	=	96	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 099	n/a	=	104	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 099	n/a	=	8	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 101	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 101	n/a	=	94	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 101	n/a	=	102	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 101	n/a	=	8	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 105	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 105	n/a	=	82	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 105	n/a	=	87	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 105	n/a	=	6	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 110	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 110	n/a	=	94	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 110	n/a	=	105	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 110	n/a	=	11	%	EPA 625m		0	30	
2005/06-6	Lab	srgt method blank, rec	7/11/2006	PCB	PCB 112	n/a	=	102	%	EPA 625m		60	120	
2005/06-6	ME-CC	srgt environ, rec	7/11/2006	PCB	PCB 112	n/a	=	104	%	EPA 625m		60	120	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	PCB	PCB 112	n/a	=	91	%	EPA 625m		60	120	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	PCB	PCB 112	n/a	=	96	%	EPA 625m		60	120	
2005/06-6	ME-VR2	srgt environ, rec	7/11/2006	PCB	PCB 112	n/a	=	88	%	EPA 625m		60	120	
2005/06-6	ME-VR2	srgt matrix spike dup, rec	7/11/2006	PCB	PCB 112	n/a	=	91	%	EPA 625m		60	120	
2005/06-6	ME-VR2	srgt matrix spike, rec	7/11/2006	PCB	PCB 112	n/a	=	99	%	EPA 625m		60	120	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 114	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 114	n/a	=	95	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 114	n/a	=	101	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 114	n/a	=	6	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 118	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 118	n/a	=	103	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 118	n/a	=	105	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 118	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 119	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 119	n/a	=	97	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 119	n/a	=	104	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 119	n/a	=	7	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 123	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 123	n/a	=	85	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 123	n/a	=	91	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 123	n/a	=	7	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 126	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 126	n/a	=	90	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 126	n/a	=	104	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 126	n/a	=	14	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 128 + 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 128 + 167	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 128 + 167	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 128 + 167	n/a	=	103	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 128 + 167	n/a	=	103	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 128 + 167	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 138	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 138	n/a	=	100	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 138	n/a	=	115	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 138	n/a	=	14	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 141	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 141	n/a	=	109	%	EPA 625m		60	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 141	n/a	=	99	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 141	n/a	=	10	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 149	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 149	n/a	=	92	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 149	n/a	=	93	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 149	n/a	=	1	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 151	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 151	n/a	=	88	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 151	n/a	=	97	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 151	n/a	=	10	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 153	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 153	n/a	=	90	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 153	n/a	=	93	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 153	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 156	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 156	n/a	=	95	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 156	n/a	=	104	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 156	n/a	=	9	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 157	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 157	n/a	=	99	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 157	n/a	=	104	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 157	n/a	=	5	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 158	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 158	n/a	=	101	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 158	n/a	=	107	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 158	n/a	=	6	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 168 + 132	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 168 + 132	n/a	=	95	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 168 + 132	n/a	=	109	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 168 + 132	n/a	=	14	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 169	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 169	n/a	=	89	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 169	n/a	=	102	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 169	n/a	=	14	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 170	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 170	n/a	=	112	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 170	n/a	=	116	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 170	n/a	=	4	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 177	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 177	n/a	=	114	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 177	n/a	=	122	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 177	n/a	=	7	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 180	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 180	n/a	=	104	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 180	n/a	=	118	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 180	n/a	=	13	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 183	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 183	n/a	=	108	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 183	n/a	=	116	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 183	n/a	=	7	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 187	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 187	n/a	=	109	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 187	n/a	=	115	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 187	n/a	=	5	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 189	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 189	n/a	=	98	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 189	n/a	=	107	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 189	n/a	=	9	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 194	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 194	n/a	=	83	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 194	n/a	=	90	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 194	n/a	=	8	%	EPA 625m		0	30	
2005/06-6	Lab	srgt method blank, rec	7/11/2006	PCB	PCB 198	n/a	=	100	%	EPA 625m		60	120	
2005/06-6	ME-CC	srgt environ, rec	7/11/2006	PCB	PCB 198	n/a	=	106	%	EPA 625m		60	120	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	PCB	PCB 198	n/a	=	97	%	EPA 625m		60	120	
2005/06-6	ME-SCR	srgt environ, rec	7/11/2006	PCB	PCB 198	n/a	=	104	%	EPA 625m		60	120	
2005/06-6	ME-VR2	srgt environ, rec	7/11/2006	PCB	PCB 198	n/a	=	91	%	EPA 625m		60	120	
2005/06-6	ME-VR2	srgt matrix spike dup, rec	7/11/2006	PCB	PCB 198	n/a	=	90	%	EPA 625m		60	120	
2005/06-6	ME-VR2	srgt matrix spike, rec	7/11/2006	PCB	PCB 198	n/a	=	99	%	EPA 625m		60	120	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 200	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 200	n/a	=	108	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 200	n/a	=	112	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 200	n/a	=	4	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 201	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 201	n/a	=	113	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 201	n/a	=	110	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 201	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	PCB 206	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	PCB	PCB 206	n/a	=	115	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	PCB	PCB 206	n/a	=	112	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	PCB	PCB 206	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-6	ME-SCR	field duplicate	7/11/2006	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	PCB	Total Detectable PCBs	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	LCS dup, rec	6/20/2006	Pesticide	2,4,5-T	n/a	=	112	%	EPA 8151A		30	130	
2005/06-6	Lab	LCS, rec	6/20/2006	Pesticide	2,4,5-T	n/a	=	113	%	EPA 8151A		30	130	
2005/06-6	Lab	LCS, RPD	6/20/2006	Pesticide	2,4,5-T	n/a	=	0.9	%	EPA 8151A		0	30	
2005/06-6	Lab	method blank	6/21/2006	Pesticide	2,4,5-T	n/a	<	0.17	µg/L	EPA 8151A	0.17		0.17	
2005/06-6	ME-SCR	field duplicate	6/20/2006	Pesticide	2,4,5-T	n/a	<	0.17	µg/L	EPA 8151A	0.17			
2005/06-6	ME-VR2	matrix spike dup, rec	6/20/2006	Pesticide	2,4,5-T	n/a	=	38	%	EPA 8151A		30	130	
2005/06-6	ME-VR2	matrix spike, rec	6/20/2006	Pesticide	2,4,5-T	n/a	=	34	%	EPA 8151A		30	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/20/2006	Pesticide	2,4,5-T	n/a	=	10	%	EPA 8151A		0	30	
2005/06-6	Lab	method blank	6/21/2006	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.12	µg/L	EPA 8151A	0.12		0.12	
2005/06-6	ME-SCR	field duplicate	6/20/2006	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.12	µg/L	EPA 8151A	0.12			
2005/06-6	Lab	LCS dup, rec	6/20/2006	Pesticide	2,4-D	n/a	=	82	%	EPA 8151A		30	130	
2005/06-6	Lab	LCS, rec	6/20/2006	Pesticide	2,4-D	n/a	=	86	%	EPA 8151A		30	130	
2005/06-6	Lab	LCS, RPD	6/20/2006	Pesticide	2,4-D	n/a	=	4.8	%	EPA 8151A		0	30	
2005/06-6	Lab	method blank	6/21/2006	Pesticide	2,4-D	n/a	<	1.5	µg/L	EPA 8151A	1.5		0	
2005/06-6	ME-SCR	field duplicate	6/20/2006	Pesticide	2,4-D	n/a	<	1.5	µg/L	EPA 8151A	1.5			
2005/06-6	ME-VR2	matrix spike dup, rec	6/20/2006	Pesticide	2,4-D	n/a	=	172	%	EPA 8151A		30	130	
2005/06-6	ME-VR2	matrix spike, rec	6/20/2006	Pesticide	2,4-D	n/a	=	153	%	EPA 8151A		30	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/20/2006	Pesticide	2,4-D	n/a	=	12	%	EPA 8151A		0	30	
2005/06-6	Lab	LCS dup, rec	6/20/2006	Pesticide	2,4-DB	n/a	=	110	%	EPA 8151A		30	130	
2005/06-6	Lab	LCS, rec	6/20/2006	Pesticide	2,4-DB	n/a	=	114	%	EPA 8151A		30	130	
2005/06-6	Lab	LCS, RPD	6/20/2006	Pesticide	2,4-DB	n/a	=	3.6	%	EPA 8151A		0	30	
2005/06-6	Lab	method blank	6/21/2006	Pesticide	2,4-DB	n/a	<	4.1	µg/L	EPA 8151A	4.1		4.1	
2005/06-6	ME-SCR	field duplicate	6/20/2006	Pesticide	2,4-DB	n/a	<	4.1	µg/L	EPA 8151A	4.1			
2005/06-6	ME-VR2	matrix spike dup, rec	6/20/2006	Pesticide	2,4-DB	n/a	=	55	%	EPA 8151A		30	130	
2005/06-6	ME-VR2	matrix spike, rec	6/20/2006	Pesticide	2,4-DB	n/a	=	89	%	EPA 8151A		30	130	
2005/06-6	ME-VR2	matrix spike, RPD	6/20/2006	Pesticide	2,4-DB	n/a	=	47	%	EPA 8151A		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	2,4'-DDD	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	2,4'-DDD	n/a	=	108	%	EPA 625m		50	140	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	2,4'-DDD	n/a	=	114	%	EPA 625m		50	140	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	2,4'-DDD	n/a	=	5	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	2,4'-DDE	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	2,4'-DDE	n/a	=	111	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	2,4'-DDE	n/a	=	109	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	2,4'-DDE	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	2,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	2,4'-DDT	n/a	=	83	%	EPA 625m		40	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	2,4'-DDT	n/a	=	100	%	EPA 625m		40	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	2,4'-DDT	n/a	=	19	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	4,4'-DDD	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	4,4'-DDD	n/a	=	101	%	EPA 625m		70	130	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	4,4'-DDD	n/a	=	109	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	4,4'-DDD	n/a	=	8	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	4,4'-DDE	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	4,4'-DDE	n/a	=	104	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	4,4'-DDE	n/a	=	110	%	EPA 625m		70	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	4,4'-DDE	n/a	=	6	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	4,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	4,4'-DDT	n/a	=	80	%	EPA 625m		0	150	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	4,4'-DDT	n/a	=	99	%	EPA 625m		0	150	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	4,4'-DDT	n/a	=	21	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Aldrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Aldrin	n/a	=	104	%	EPA 625m		50	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Aldrin	n/a	=	101	%	EPA 625m		50	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Aldrin	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	BHC-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	BHC-alpha	n/a	=	102	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	BHC-alpha	n/a	=	102	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	BHC-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	BHC-beta	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	BHC-beta	n/a	=	104	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	BHC-beta	n/a	=	106	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	BHC-beta	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	BHC-delta	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	BHC-delta	n/a	=	101	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	BHC-delta	n/a	=	107	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	BHC-delta	n/a	=	6	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	95	%	EPA 625m		50	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	101	%	EPA 625m		50	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	BHC-gamma (Lindane)	n/a	=	6	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Bolstar	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Bolstar	n/a	=	96	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Bolstar	n/a	=	108	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Bolstar	n/a	=	12	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Chlordane-alpha	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Chlordane-alpha	n/a	=	102	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Chlordane-alpha	n/a	=	102	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Chlordane-alpha	n/a	=	0	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Chlordane-gamma	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Chlordane-gamma	n/a	=	101	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Chlordane-gamma	n/a	=	103	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Chlordane-gamma	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Chlorpyrifos	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Chlorpyrifos	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Chlorpyrifos	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Chlorpyrifos	n/a	=	88	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Chlorpyrifos	n/a	=	94	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Chlorpyrifos	n/a	=	7	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	cis-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	cis-Nonachlor	n/a	=	100	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	cis-Nonachlor	n/a	=	106	%	EPA 625m		60	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	cis-Nonachlor	n/a	=	6	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	6/21/2006	Pesticide	Dalapon	n/a	<	2.6	µg/L	EPA 8151A	2.6			2.6
2005/06-6	ME-SCR	field duplicate	6/20/2006	Pesticide	Dalapon	n/a	<	2.6	µg/L	EPA 8151A	2.6			
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Demeton-O	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Demeton-O	n/a	=	83	%	EPA 625m		45	105	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Demeton-O	n/a	=	100	%	EPA 625m		45	105	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Demeton-O	n/a	=	19	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Diazinon	n/a	<	0.002	µg/L	EPA 625m	0.002			0.002
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Diazinon	n/a	<	0.002	µg/L	EPA 625m	0.002			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Diazinon	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Diazinon	n/a	=	88	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Diazinon	n/a	=	99	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Diazinon	n/a	=	12	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	6/21/2006	Pesticide	Dicamba	n/a	<	0.13	µg/L	EPA 8151A	0.13			0.13
2005/06-6	ME-SCR	field duplicate	6/20/2006	Pesticide	Dicamba	n/a	<	0.13	µg/L	EPA 8151A	0.13			
2005/06-6	Lab	method blank	6/21/2006	Pesticide	Dichlorprop	n/a	<	1.5	µg/L	EPA 8151A	1.5			0
2005/06-6	ME-SCR	field duplicate	6/20/2006	Pesticide	Dichlorprop	n/a	<	1.5	µg/L	EPA 8151A	1.5			
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003			0.003
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Dichlorvos	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Dichlorvos	n/a	=	85	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Dichlorvos	n/a	=	83	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Dichlorvos	n/a	=	2	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Dieldrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Dieldrin	n/a	=	108	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Dieldrin	n/a	=	112	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Dieldrin	n/a	=	4	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003			0.003
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Dimethoate	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Dimethoate	n/a	=	78	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Dimethoate	n/a	=	79	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Dimethoate	n/a	=	1	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	6/21/2006	Pesticide	Dinoseb	n/a	<	0.3	µg/L	EPA 8151A	0.3			0.3
2005/06-6	ME-SCR	field duplicate	6/20/2006	Pesticide	Dinoseb	n/a	<	0.3	µg/L	EPA 8151A	0.3			



**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Disulfoton	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Disulfoton	n/a	=	78	%	EPA 625m		45	105	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Disulfoton	n/a	=	89	%	EPA 625m		45	105	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Disulfoton	n/a	=	13	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Endosulfan sulfate	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Endosulfan sulfate	n/a	=	94	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Endosulfan sulfate	n/a	=	105	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Endosulfan sulfate	n/a	=	11	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Endosulfan-I	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Endosulfan-I	n/a	=	102	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Endosulfan-I	n/a	=	102	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Endosulfan-I	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Endosulfan-II	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Endosulfan-II	n/a	=	100	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Endosulfan-II	n/a	=	109	%	EPA 625m		60	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Endosulfan-II	n/a	=	9	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Endrin	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Endrin	n/a	=	92	%	EPA 625m		65	135	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Endrin	n/a	=	112	%	EPA 625m		65	135	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Endrin	n/a	=	20	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Endrin aldehyde	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Endrin ketone	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Endrin ketone	n/a	=	97	%	EPA 625m		40	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Endrin ketone	n/a	=	100	%	EPA 625m		40	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Endrin ketone	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Ethoprop	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Ethoprop	n/a	=	88	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Ethoprop	n/a	=	96	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Ethoprop	n/a	=	9	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Fenchlorophos (Ronnell)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Fenchlorophos (Ronnell)	n/a	<	0.002	µg/L	EPA 625m	0.002			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Fenchlorophos (Ronnell)	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Fenchlorophos (Ronnell)	n/a	=	88	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Fenchlorophos (Ronnell)	n/a	=	97	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Fenchlorophos (Ronnell)	n/a	=	10	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Fensulfothion	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Fensulfothion	n/a	=	98	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Fensulfothion	n/a	=	107	%	EPA 625m		65	125	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Fensulfothion	n/a	=	9	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Fenthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Fenthion	n/a	=	89	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Fenthion	n/a	=	94	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Fenthion	n/a	=	5	%	EPA 625m		0	30	
2005/06-6	Lab	LCS, rec	6/26/2006	Pesticide	Glyphosate	n/a	=	108	%	EPA 547		71	137	
2005/06-6	Lab	method blank	6/26/2006	Pesticide	Glyphosate	n/a	<	5	µg/L	EPA 547	5		5	
2005/06-6	ME-SCR	field duplicate	6/26/2006	Pesticide	Glyphosate	n/a	<	5	µg/L	EPA 547	5			
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Heptachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Heptachlor	n/a	=	84	%	EPA 625m		45	135	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Heptachlor	n/a	=	102	%	EPA 625m		45	135	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Heptachlor	n/a	=	19	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Heptachlor epoxide	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Heptachlor epoxide	n/a	=	104	%	EPA 625m		65	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Heptachlor epoxide	n/a	=	100	%	EPA 625m		65	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Heptachlor epoxide	n/a	=	4	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Malathion	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Malathion	n/a	=	92	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Malathion	n/a	=	97	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Malathion	n/a	=	5	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	6/21/2006	Pesticide	MCPA	n/a	<	110	µg/L	EPA 8151A	110		110	
2005/06-6	ME-SCR	field duplicate	6/20/2006	Pesticide	MCPA	n/a	<	110	µg/L	EPA 8151A	110			
2005/06-6	Lab	method blank	6/21/2006	Pesticide	MCPP	n/a	<	110	µg/L	EPA 8151A	110		110	
2005/06-6	ME-SCR	field duplicate	6/20/2006	Pesticide	MCPP	n/a	<	110	µg/L	EPA 8151A	110			
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Merphos	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Merphos	n/a	=	88	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Merphos	n/a	=	107	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Merphos	n/a	=	19	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Methoxychlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Methoxychlor	n/a	=	76	%	EPA 625m		0	155	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Methoxychlor	n/a	=	93	%	EPA 625m		0	155	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Methoxychlor	n/a	=	20	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Methyl parathion	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Methyl parathion	n/a	=	98	%	EPA 625m		60	120	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Methyl parathion	n/a	=	104	%	EPA 625m		60	120	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Methyl parathion	n/a	=	6	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		0.008	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Mevinphos	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Mevinphos	n/a	=	97	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Mevinphos	n/a	=	99	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Mevinphos	n/a	=	2	%	EPA 625m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Mirex	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Mirex	n/a	=	95	%	EPA 625m		50	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Mirex	n/a	=	104	%	EPA 625m		50	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Mirex	n/a	=	9	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Oxychlorthane	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Oxychlorthane	n/a	=	95	%	EPA 625m		60	120	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Oxychlorthane	n/a	=	106	%	EPA 625m		60	120	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Oxychlorthane	n/a	=	11	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006		0.006	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Phorate	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Phorate	n/a	=	89	%	EPA 625m		45	105	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Phorate	n/a	=	96	%	EPA 625m		45	105	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Phorate	n/a	=	8	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	112	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	113	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	1	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Tokuthion	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Tokuthion	n/a	=	98	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Tokuthion	n/a	=	105	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Tokuthion	n/a	=	7	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Total Detectable DDTs	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Toxaphene	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	trans-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	trans-Nonachlor	n/a	=	103	%	EPA 625m		55	130	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	trans-Nonachlor	n/a	=	106	%	EPA 625m		55	130	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	trans-Nonachlor	n/a	=	3	%	EPA 625m		0	30	
2005/06-6	Lab	method blank	7/11/2006	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-6	ME-SCR	field duplicate	7/11/2006	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001			
2005/06-6	ME-VR2	lab duplicate, RPD	7/11/2006	Pesticide	Trichloronate	n/a	=	0	%	EPA 625m		0	30	
2005/06-6	ME-VR2	matrix spike dup, rec	7/11/2006	Pesticide	Trichloronate	n/a	=	100	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, rec	7/11/2006	Pesticide	Trichloronate	n/a	=	93	%	EPA 625m		65	125	
2005/06-6	ME-VR2	matrix spike, RPD	7/11/2006	Pesticide	Trichloronate	n/a	=	7	%	EPA 625m		0	30	
2005/06-PRE	Carboy Blank	equipment blank	10/28/2005	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1			
2005/06-PRE	Lab	method blank	10/28/2005	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1		1	
2005/06-PRE	Tubing Blank	equipment blank	10/28/2005	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1			
2005/06-PRE	Carboy Blank	equipment blank	10/28/2005	Metal	Aluminum	Total	=	2.9	µg/L	EPA 200.8m	1		1	EST
2005/06-PRE	Carboy Blank	lab duplicate, RPD	10/28/2005	Metal	Aluminum	Total	=	9	%	EPA 200.8m		0	30	
2005/06-PRE	Lab	method blank	10/28/2005	Metal	Aluminum	Total	<	1	µg/L	EPA 200.8m	1		1	
2005/06-PRE	Tubing Blank	equipment blank	10/28/2005	Metal	Aluminum	Total	=	2.89	µg/L	EPA 200.8m	1		1	EST
2005/06-PRE	Tubing Blank	lab duplicate, RPD	10/28/2005	Metal	Aluminum	Total	=	1	%	EPA 200.8m		0	30	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-PRE	Carboy Blank	equipment blank	10/28/2005	Metal	Arsenic	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Carboy Blank	lab duplicate, RPD	10/28/2005	Metal	Arsenic	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Lab	method blank	10/28/2005	Metal	Arsenic	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/28/2005	Metal	Arsenic	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	lab duplicate, RPD	10/28/2005	Metal	Arsenic	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Carboy Blank	equipment blank	10/28/2005	Metal	Cadmium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Carboy Blank	lab duplicate, RPD	10/28/2005	Metal	Cadmium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Lab	method blank	10/28/2005	Metal	Cadmium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/28/2005	Metal	Cadmium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	lab duplicate, RPD	10/28/2005	Metal	Cadmium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Carboy Blank	equipment blank	10/28/2005	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Carboy Blank	lab duplicate, RPD	10/28/2005	Metal	Chromium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Lab	method blank	10/28/2005	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/28/2005	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	lab duplicate, RPD	10/28/2005	Metal	Chromium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Carboy Blank	equipment blank	10/28/2005	Metal	Copper	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Carboy Blank	lab duplicate, RPD	10/28/2005	Metal	Copper	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Lab	method blank	10/28/2005	Metal	Copper	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/28/2005	Metal	Copper	Total	=	0.47	µg/L	EPA 200.8m	0.1		0.1	EST
2005/06-PRE	Tubing Blank	lab duplicate, RPD	10/28/2005	Metal	Copper	Total	=	4	%	EPA 200.8m		0	30	
2005/06-PRE	Carboy Blank	equipment blank	10/28/2005	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05			
2005/06-PRE	Carboy Blank	lab duplicate, RPD	10/28/2005	Metal	Lead	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Lab	method blank	10/28/2005	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05			
2005/06-PRE	Tubing Blank	equipment blank	10/28/2005	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05			
2005/06-PRE	Tubing Blank	lab duplicate, RPD	10/28/2005	Metal	Lead	Total	=	6	%	EPA 200.8m		0	30	
2005/06-PRE	Carboy Blank	equipment blank	10/27/2005	Metal	Mercury	Total	<	0.0005	ng/L	EPA 1631E	0.0005			
2005/06-PRE	Carboy Blank	lab duplicate, RPD	10/27/2005	Metal	Mercury	Total	=	0	%	EPA 1631E		0	30	
2005/06-PRE	Lab	method blank	10/27/2005	Metal	Mercury	Total	<	0.5	ng/L	EPA 1631E	0.5		0.5	
2005/06-PRE	Tubing Blank	equipment blank	10/27/2005	Metal	Mercury	Total	<	0.0005	ng/L	EPA 1631E	0.0005			
2005/06-PRE	Tubing Blank	lab duplicate, RPD	10/27/2005	Metal	Mercury	Total	=	0	%	EPA 1631E		0	30	
2005/06-PRE	Carboy Blank	equipment blank	10/28/2005	Metal	Nickel	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Carboy Blank	lab duplicate, RPD	10/28/2005	Metal	Nickel	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Lab	method blank	10/28/2005	Metal	Nickel	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/28/2005	Metal	Nickel	Total	=	0.73	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	lab duplicate, RPD	10/28/2005	Metal	Nickel	Total	=	1	%	EPA 200.8m		0	30	
2005/06-PRE	Carboy Blank	equipment blank	10/28/2005	Metal	Selenium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Carboy Blank	lab duplicate, RPD	10/28/2005	Metal	Selenium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Lab	method blank	10/28/2005	Metal	Selenium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/28/2005	Metal	Selenium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	lab duplicate, RPD	10/28/2005	Metal	Selenium	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Carboy Blank	equipment blank	10/28/2005	Metal	Silver	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Carboy Blank	lab duplicate, RPD	10/28/2005	Metal	Silver	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Lab	method blank	10/28/2005	Metal	Silver	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/28/2005	Metal	Silver	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	lab duplicate, RPD	10/28/2005	Metal	Silver	Total	=	0	%	EPA 200.8m		0	30	
2005/06-PRE	Lab	method blank	10/28/2005	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2005/06-PRE	Carboy Blank	equipment blank	10/28/2005	Metal	Zinc	Total	=	1.75	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Carboy Blank	lab duplicate, RPD	10/28/2005	Metal	Zinc	Total	=	10	%	EPA 200.8m		0	30	
2005/06-PRE	Lab	method blank	10/28/2005	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/28/2005	Metal	Zinc	Total	=	7.9	µg/L	EPA 200.8m	0.1		0.1	
2005/06-PRE	Tubing Blank	lab duplicate, RPD	10/28/2005	Metal	Zinc	Total	=	2	%	EPA 200.8m		0	30	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	1,2-Dichlorobenzene	n/a	<	0.0117	µg/L	EPA 625m	0.01		0.01	EST
2005/06-PRE	Lab	method blank	10/21/2005	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	1-Methylnaphthalene	n/a	=	0.0017	µg/L	EPA 625m	0.001		0.001	EST
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	srgt equip blank, rec	10/21/2005	Organic	2,4,6-Tribromophenol	n/a	=	75	%	EPA 625m		11	162	
2005/06-PRE	Lab	srgt method blank, rec	10/21/2005	Organic	2,4,6-Tribromophenol	n/a	=	79	%	EPA 625m		11	162	
2005/06-PRE	Tubing Blank	srgt equip blank, rec	10/21/2005	Organic	2,4,6-Tribromophenol	n/a	=	82	%	EPA 625m		11	162	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2-Methylnaphthalene	n/a	=	0.0033	µg/L	EPA 625m	0.001		0.001	EST
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-PRE	Lab	method blank	10/21/2005	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	srgt equip blank, rec	10/21/2005	Organic	Acenaphthene-d10	n/a	=	89	%	EPA 625m		18	133	
2005/06-PRE	Lab	srgt method blank, rec	10/21/2005	Organic	Acenaphthene-d10	n/a	=	93	%	EPA 625m		18	133	
2005/06-PRE	Tubing Blank	srgt equip blank, rec	10/21/2005	Organic	Acenaphthene-d10	n/a	=	74	%	EPA 625m		18	133	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Biphenyl	n/a	=	0.0023	µg/L	EPA 625m	0.001		0.001	EST
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.0235	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.0288	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.0693	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Butyl benzyl phthalate	n/a	=	0.0085	µg/L	EPA 625m	0.005		0.005	EST
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Butyl benzyl phthalate	n/a	=	0.0297	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Butyl benzyl phthalate	n/a	=	0.0174	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	srgt equip blank, rec	10/21/2005	Organic	Chrysene-d12	n/a	=	102	%	EPA 625m		41	131	
2005/06-PRE	Lab	srgt method blank, rec	10/21/2005	Organic	Chrysene-d12	n/a	=	103	%	EPA 625m		41	131	
2005/06-PRE	Tubing Blank	srgt equip blank, rec	10/21/2005	Organic	Chrysene-d12	n/a	=	107	%	EPA 625m		41	131	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Diethyl phthalate	n/a	=	0.0312	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Diethyl phthalate	n/a	=	0.0393	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Diethyl phthalate	n/a	=	0.687	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Dimethyl phthalate	n/a	=	0.0366	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Dimethyl phthalate	n/a	=	0.0545	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Dimethyl phthalate	n/a	=	0.0761	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Di-n-butylphthalate	n/a	=	0.0261	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Di-n-butylphthalate	n/a	=	0.0534	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Di-n-butylphthalate	n/a	=	0.044	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Di-n-octylphthalate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Naphthalene	n/a	=	0.0081	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	srgt equip blank, rec	10/21/2005	Organic	Naphthalene-d8	n/a	=	71	%	EPA 625m		6	136	
2005/06-PRE	Lab	srgt method blank, rec	10/21/2005	Organic	Naphthalene-d8	n/a	=	81	%	EPA 625m		6	136	
2005/06-PRE	Tubing Blank	srgt equip blank, rec	10/21/2005	Organic	Naphthalene-d8	n/a	=	30	%	EPA 625m		6	136	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	srgt equip blank, rec	10/21/2005	Organic	Perylene-d12	n/a	=	101	%	EPA 625m		34	134	
2005/06-PRE	Lab	srgt method blank, rec	10/21/2005	Organic	Perylene-d12	n/a	=	107	%	EPA 625m		34	134	
2005/06-PRE	Tubing Blank	srgt equip blank, rec	10/21/2005	Organic	Perylene-d12	n/a	=	100	%	EPA 625m		34	134	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	srgt equip blank, rec	10/21/2005	Organic	Phenanthrene-d10	n/a	=	95	%	EPA 625m		43	124	
2005/06-PRE	Lab	srgt method blank, rec	10/21/2005	Organic	Phenanthrene-d10	n/a	=	93	%	EPA 625m		43	124	
2005/06-PRE	Tubing Blank	srgt equip blank, rec	10/21/2005	Organic	Phenanthrene-d10	n/a	=	95	%	EPA 625m		43	124	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Phenol	n/a	=	1.21	µg/L	EPA 625m	0.1		0.1	
2005/06-PRE	Carboy Blank	srgt equip blank, rec	10/21/2005	Organic	Phenol-d5	n/a	=	35	%	EPA 625m		20	100	
2005/06-PRE	Lab	srgt method blank, rec	10/21/2005	Organic	Phenol-d5	n/a	=	66	%	EPA 625m		20	100	
2005/06-PRE	Tubing Blank	srgt equip blank, rec	10/21/2005	Organic	Phenol-d5	n/a	=	14	%	EPA 625m		20	100	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	srgt equip blank, rec	10/21/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	80	%	EPA 625m		20	120	
2005/06-PRE	Lab	srgt method blank, rec	10/21/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	81	%	EPA 625m		20	120	
2005/06-PRE	Tubing Blank	srgt equip blank, rec	10/21/2005	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	68	%	EPA 625m		20	120	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Organic	Total Detectable PAHs	n/a	=	0	µg/L	EPA 625m				
2005/06-PRE	Lab	method blank	10/21/2005	Organic	Total Detectable PAHs	n/a	=	0	µg/L	EPA 625m				
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Organic	Total Detectable PAHs	n/a	=	0.0154	µg/L	EPA 625m				
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	







**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	srgt equip blank, rec	10/21/2005	PCB	PCB 198	n/a	=	104	%	EPA 625m		30	125	
2005/06-PRE	Lab	srgt method blank, rec	10/21/2005	PCB	PCB 198	n/a	=	100	%	EPA 625m		30	125	
2005/06-PRE	Tubing Blank	srgt equip blank, rec	10/21/2005	PCB	PCB 198	n/a	=	98	%	EPA 625m		30	125	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-PRE	Lab	method blank	10/21/2005	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Bolstar	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Bolstar	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Bolstar	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Chlorpyrifos	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Chlorpyrifos	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Chlorpyrifos	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Demeton-O	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Demeton-O	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Demeton-O	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Diazinon	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Diazinon	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Diazinon	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Dichlorvos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Dichlorvos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Dichlorvos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Dimethoate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Dimethoate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Dimethoate	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Disulfoton	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Disulfoton	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Disulfoton	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Ethoprop	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Ethoprop	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Ethoprop	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Fensulfothion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Fensulfothion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Fensulfothion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Fenthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Fenthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Fenthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Malathion	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Malathion	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Malathion	n/a	<	0.005	µg/L	EPA 625m	0.005		0.005	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Merphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Merphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Merphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Methyl parathion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Methyl parathion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Methyl parathion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Mevinphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Mevinphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Mevinphos	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

**Appendix G**  
2005/06 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Oxychlorane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Phorate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Phorate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Phorate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Tokuthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Tokuthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Tokuthion	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2005/06-PRE	Carboy Blank	equipment blank	10/21/2005	Pesticide	Trichloronate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Lab	method blank	10/21/2005	Pesticide	Trichloronate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2005/06-PRE	Tubing Blank	equipment blank	10/21/2005	Pesticide	Trichloronate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	