

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	LCS dup, rec	9/25/2007	Anion	Bromide	n/a	=	85	%	EPA 300.0		70	130	
2007/08-1	Lab	LCS, rec	9/25/2007	Anion	Bromide	n/a	=	91	%	EPA 300.0		70	130	
2007/08-1	Lab	LCS, RPD	9/25/2007	Anion	Bromide	n/a	=	7	%	EPA 300.0		0	30	
2007/08-1	Lab	method blank	9/25/2007	Anion	Bromide	n/a	<	0.001	mg/L	EPA 300.0	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	9/25/2007	Anion	Bromide	n/a	<	0.001	mg/L	EPA 300.0	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	9/25/2007	Anion	Bromide	n/a	=	106	%	EPA 300.0		70	130	
2007/08-1	ME-CC	matrix spike, rec	9/25/2007	Anion	Bromide	n/a	=	102	%	EPA 300.0		70	130	
2007/08-1	ME-CC	matrix spike, RPD	9/25/2007	Anion	Bromide	n/a	=	4	%	EPA 300.0		0	30	
2007/08-1	Lab	LCS dup, rec	10/1/2007	Anion	Chloride	n/a	=	97	%	EPA 300.0		70	130	
2007/08-1	Lab	LCS, rec	10/1/2007	Anion	Chloride	n/a	=	97	%	EPA 300.0		70	130	
2007/08-1	Lab	LCS, RPD	10/1/2007	Anion	Chloride	n/a	=	0	%	EPA 300.0		0	30	
2007/08-1	Lab	method blank	10/1/2007	Anion	Chloride	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/1/2007	Anion	Chloride	n/a	=	158.8	mg/L	EPA 300.0	0.01		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/1/2007	Anion	Chloride	n/a	=	81	%	EPA 300.0		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/1/2007	Anion	Chloride	n/a	=	85	%	EPA 300.0		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/1/2007	Anion	Chloride	n/a	=	5	%	EPA 300.0		0	30	
2007/08-1	Lab	LCS dup, rec	9/26/2007	Anion	Perchlorate	n/a	=	87	%	EPA 314.0		85	115	
2007/08-1	Lab	LCS, rec	9/26/2007	Anion	Perchlorate	n/a	=	87	%	EPA 314.0		85	115	
2007/08-1	Lab	LCS, RPD	9/26/2007	Anion	Perchlorate	n/a	=	0	%	EPA 314.0		0	15	
2007/08-1	Lab	method blank	9/26/2007	Anion	Perchlorate	n/a	<	2	µg/L	EPA 314.0	2		2	
2007/08-1	ME-VR2	field blank	9/22/2007	Bacteriological	E. Coli	n/a	<	10	MPN/100 mL	MMO-MUG	10		10	
2007/08-1	ME-VR2	field blank	9/22/2007	Bacteriological	Enterococcus	n/a	<	10	MPN/100 mL	Enterolert	10		10	
2007/08-1	ME-VR2	field blank	9/22/2007	Bacteriological	Fecal Coliform	n/a	<	2	MPN/100 mL	SM 9221 E	2		2	
2007/08-1	ME-VR2	field blank	9/22/2007	Bacteriological	Total Coliform	n/a	<	10	MPN/100 mL	MMO-MUG	10		10	
2007/08-1	Lab	method blank	9/24/2007	Conventional	BOD	n/a	<	1	mg/L	SM 5210 B	1		1	
2007/08-1	ME-CC	lab duplicate	9/24/2007	Conventional	BOD	n/a	=	9.6	mg/L	SM 5210 B	1		25	
2007/08-1	ME-CC	lab duplicate	9/25/2007	Conventional	Conductivity	n/a	=	1570	µmhos/cm	SM 2510	0.001		30	
2007/08-1	Lab	method blank	10/16/2007	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1		1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Conventional	Hardness as CaCO3	Total	=	254.5	mg/L	SM 2340 B	1		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Conventional	Hardness as CaCO3	Total	=	1.2	mg/L	SM 2340 B	1		1	
2007/08-1	ME-CC	lab duplicate	9/25/2007	Conventional	pH	n/a	=	7.8	pH Units	SM 4500 H+	0.1		30	
2007/08-1	Lab	LCS dup, rec	10/1/2007	Conventional	Total Dissolved Solids	n/a	=	102	%	SM 2540 C		70	130	
2007/08-1	Lab	LCS, rec	10/1/2007	Conventional	Total Dissolved Solids	n/a	=	95	%	SM 2540 C		70	130	
2007/08-1	Lab	LCS, RPD	10/1/2007	Conventional	Total Dissolved Solids	n/a	=	7	%	SM 2540 C		0	30	
2007/08-1	Lab	method blank	10/1/2007	Conventional	Total Dissolved Solids	n/a	<	0.1	mg/L	SM 2540 C	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/1/2007	Conventional	Total Dissolved Solids	n/a	=	924	mg/L	SM 2540 C	0.1		30	
2007/08-1	Lab	LCS dup, rec	9/25/2007	Conventional	Total Organic Carbon	n/a	=	97	%	EPA 415.1		50	150	
2007/08-1	Lab	LCS, rec	9/25/2007	Conventional	Total Organic Carbon	n/a	=	97	%	EPA 415.1		50	150	
2007/08-1	Lab	LCS, RPD	9/25/2007	Conventional	Total Organic Carbon	n/a	=	0	%	EPA 415.1		0	30	
2007/08-1	Lab	method blank	9/25/2007	Conventional	Total Organic Carbon	n/a	<	0.1	mg/L	EPA 415.1	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	9/25/2007	Conventional	Total Organic Carbon	n/a	=	7.4	mg/L	EPA 415.1	0.1		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/10/2007	Conventional	Total Organic Carbon	n/a	=	92	%	EPA 415.1		50	150	
2007/08-1	ME-CC	matrix spike, rec	10/10/2007	Conventional	Total Organic Carbon	n/a	=	93	%	EPA 415.1		50	150	
2007/08-1	ME-CC	matrix spike, RPD	10/10/2007	Conventional	Total Organic Carbon	n/a	=	1	%	EPA 415.1		0	30	
2007/08-1	Lab	method blank	9/28/2007	Conventional	Total Suspended Solids	n/a	<	0.5	mg/L	SM 2540 D	0.5		0.5	
2007/08-1	ME-CC	lab duplicate	9/28/2007	Conventional	Total Suspended Solids	n/a	=	440	mg/L	SM 2540 D	0.5		30	
2007/08-1	ME-CC	lab duplicate	9/25/2007	Conventional	Turbidity	n/a	=	317	NTU	EPA 180.1	1		30	
2007/08-1	Lab	LCS dup, rec	9/29/2007	Hydrocarbon	Oil and Grease	n/a	=	98	%	EPA 1664A		70	130	
2007/08-1	Lab	LCS, rec	9/29/2007	Hydrocarbon	Oil and Grease	n/a	=	99	%	EPA 1664A		70	130	
2007/08-1	Lab	LCS, RPD	9/29/2007	Hydrocarbon	Oil and Grease	n/a	=	1	%	EPA 1664A		0	30	
2007/08-1	Lab	method blank	9/29/2007	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1		1	
2007/08-1	Lab	LCS dup, rec	9/29/2007	Hydrocarbon	TRPH	n/a	=	99	%	EPA 1664		70	130	
2007/08-1	Lab	LCS, rec	9/29/2007	Hydrocarbon	TRPH	n/a	=	100	%	EPA 1664		70	130	
2007/08-1	Lab	LCS, RPD	9/29/2007	Hydrocarbon	TRPH	n/a	=	1	%	EPA 1664		0	30	
2007/08-1	Lab	method blank	9/29/2007	Hydrocarbon	TRPH	n/a	<	1	mg/L	EPA 1664	1		1	
2007/08-1	Lab	method blank	10/16/2007	Metal	Aluminum	Dissolved	<	5	µg/L	EPA 200.8m	5		5	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Aluminum	Dissolved	<	5	µg/L	EPA 200.8m	5		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Metal	Aluminum	Dissolved	=	114	%	EPA 200.8m		50	140	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Metal	Aluminum	Dissolved	=	114	%	EPA 200.8m		50	140	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Metal	Aluminum	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Aluminum	Total	<	5	µg/L	EPA 200.8m	5		5	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Aluminum	Total	=	2349	µg/L	EPA 200.8m	5		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Metal	Aluminum	Total	<	5	µg/L	EPA 200.8m	5		5	
2007/08-1	Lab	method blank	10/16/2007	Metal	Arsenic	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Arsenic	Dissolved	=	4.7	µg/L	EPA 200.8m	0.2		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Metal	Arsenic	Dissolved	=	99	%	EPA 200.8m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Metal	Arsenic	Dissolved	=	98	%	EPA 200.8m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Metal	Arsenic	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Arsenic	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Arsenic	Total	=	6.1	µg/L	EPA 200.8m	0.2		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Metal	Arsenic	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	Lab	method blank	10/16/2007	Metal	Cadmium	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Cadmium	Dissolved	=	0.4	µg/L	EPA 200.8m	0.2		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Metal	Cadmium	Dissolved	=	91	%	EPA 200.8m		75	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Metal	Cadmium	Dissolved	=	91	%	EPA 200.8m		75	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Metal	Cadmium	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Cadmium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Cadmium	Total	=	1.1	µg/L	EPA 200.8m	0.2		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Metal	Cadmium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	Lab	method blank	10/16/2007	Metal	Chromium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Chromium	Dissolved	=	0.3	µg/L	EPA 200.8m	0.1		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Metal	Chromium	Dissolved	=	110	%	EPA 200.8m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Metal	Chromium	Dissolved	=	111	%	EPA 200.8m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Metal	Chromium	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Chromium	Total	=	5.3	µg/L	EPA 200.8m	0.1		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-1	Lab	LCS dup, rec	9/24/2007	Metal	Chromium VI	Total	=	103	%	SM 3500-Cr D		70	130	
2007/08-1	Lab	LCS, rec	9/24/2007	Metal	Chromium VI	Total	=	102	%	SM 3500-Cr D		70	130	
2007/08-1	Lab	LCS, RPD	9/24/2007	Metal	Chromium VI	Total	=	1	%	SM 3500-Cr D		0	30	
2007/08-1	Lab	method blank	9/24/2007	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr D	5		5	
2007/08-1	ME-CC	lab duplicate	9/24/2007	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr D	5		30	
2007/08-1	ME-CC	matrix spike dup, rec	9/24/2007	Metal	Chromium VI	Total	=	98	%	SM 3500-Cr D		70	130	
2007/08-1	ME-CC	matrix spike, rec	9/24/2007	Metal	Chromium VI	Total	=	109	%	SM 3500-Cr D		70	130	
2007/08-1	ME-CC	matrix spike, RPD	9/24/2007	Metal	Chromium VI	Total	=	11	%	SM 3500-Cr D		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Copper	Dissolved	<	0.4	µg/L	EPA 200.8m	0.4		0.4	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Copper	Dissolved	=	5.2	µg/L	EPA 200.8m	0.4		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Metal	Copper	Dissolved	=	103	%	EPA 200.8m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Metal	Copper	Dissolved	=	102	%	EPA 200.8m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Metal	Copper	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Copper	Total	<	0.4	µg/L	EPA 200.8m	0.4		0.4	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Copper	Total	=	22.4	µg/L	EPA 200.8m	0.4		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Metal	Copper	Total	<	0.4	µg/L	EPA 200.8m	0.4		0.4	
2007/08-1	Lab	method blank	10/16/2007	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Lead	Dissolved	=	0.06	µg/L	EPA 200.8m	0.05		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Metal	Lead	Dissolved	=	103	%	EPA 200.8m		65	135	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Metal	Lead	Dissolved	=	102	%	EPA 200.8m		65	135	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Metal	Lead	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Lead	Total	=	6.7	µg/L	EPA 200.8m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2007/08-1	Lab	method blank	10/9/2007	Metal	Mercury	Dissolved	<	0.5	ng/L	EPA 1631Em	0.5		0.5	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	lab duplicate	10/9/2007	Metal	Mercury	Dissolved	=	7.9	ng/L	EPA 1631Em	0.5		30	
2007/08-1	ME-VR2	field blank	10/9/2007	Metal	Mercury	Dissolved	=	0.8	ng/L	EPA 1631Em	0.5		0.5	EST-HT
2007/08-1	Lab	method blank	10/9/2007	Metal	Mercury	Total	<	0.5	ng/L	EPA 1631Em	0.5		0.5	
2007/08-1	ME-CC	lab duplicate	10/9/2007	Metal	Mercury	Total	=	29.8	ng/L	EPA 1631Em	0.5		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/9/2007	Metal	Mercury	Total	=	117	%	EPA 1631Em		60	140	
2007/08-1	ME-CC	matrix spike, rec	10/9/2007	Metal	Mercury	Total	=	112	%	EPA 1631Em		60	140	
2007/08-1	ME-CC	matrix spike, RPD	10/9/2007	Metal	Mercury	Total	=	4	%	EPA 1631Em		0	30	
2007/08-1	ME-VR2	field blank	10/9/2007	Metal	Mercury	Total	=	7.8	ng/L	EPA 1631Em	0.5		0.5	
2007/08-1	Lab	method blank	10/16/2007	Metal	Nickel	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Nickel	Dissolved	=	8.6	µg/L	EPA 200.8m	0.2		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Metal	Nickel	Dissolved	=	105	%	EPA 200.8m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Metal	Nickel	Dissolved	=	106	%	EPA 200.8m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Metal	Nickel	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Nickel	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Nickel	Total	=	21.1	µg/L	EPA 200.8m	0.2		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Metal	Nickel	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	Lab	method blank	10/16/2007	Metal	Selenium	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Selenium	Dissolved	=	2.8	µg/L	EPA 200.8m	0.2		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Metal	Selenium	Dissolved	=	107	%	EPA 200.8m		60	150	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Metal	Selenium	Dissolved	=	106	%	EPA 200.8m		60	150	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Metal	Selenium	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Selenium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Selenium	Total	=	2.8	µg/L	EPA 200.8m	0.2		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Metal	Selenium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-1	Lab	method blank	10/16/2007	Metal	Silver	Dissolved	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Silver	Dissolved	<	0.5	µg/L	EPA 200.8m	0.5		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Metal	Silver	Dissolved	=	92	%	EPA 200.8m		50	155	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Metal	Silver	Dissolved	=	92	%	EPA 200.8m		50	155	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Metal	Silver	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2007/08-1	Lab	method blank	10/16/2007	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Metal	Thallium	Dissolved	=	104	%	EPA 200.8m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Metal	Thallium	Dissolved	=	104	%	EPA 200.8m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Metal	Thallium	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-1	Lab	method blank	10/16/2007	Metal	Zinc	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Zinc	Dissolved	=	19	µg/L	EPA 200.8m	0.1		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Metal	Zinc	Dissolved	=	100	%	EPA 200.8m		50	150	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Metal	Zinc	Dissolved	=	100	%	EPA 200.8m		50	150	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Metal	Zinc	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Metal	Zinc	Total	=	73.9	µg/L	EPA 200.8m	0.1		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-1	Lab	LCS dup, rec	9/25/2007	Nutrient	Ammonia as N	n/a	=	108	%	SM 4500-NH3 F		70	130	
2007/08-1	Lab	LCS, rec	9/25/2007	Nutrient	Ammonia as N	n/a	=	108	%	SM 4500-NH3 F		70	130	
2007/08-1	Lab	LCS, RPD	9/25/2007	Nutrient	Ammonia as N	n/a	=	0	%	SM 4500-NH3 F		0	30	
2007/08-1	Lab	method blank	9/25/2007	Nutrient	Ammonia as N	n/a	<	0.01	mg/L	SM 4500-NH3 F	0.01		0.01	
2007/08-1	ME-VR2	lab duplicate	9/25/2007	Nutrient	Ammonia as N	n/a	=	0.06	mg/L	SM 4500-NH3 F	0.01		30	
2007/08-1	ME-VR2	matrix spike dup, rec	9/25/2007	Nutrient	Ammonia as N	n/a	=	104	%	SM 4500-NH3 F		70	130	
2007/08-1	ME-VR2	matrix spike, rec	9/25/2007	Nutrient	Ammonia as N	n/a	=	100	%	SM 4500-NH3 F		70	130	
2007/08-1	ME-VR2	matrix spike, RPD	9/25/2007	Nutrient	Ammonia as N	n/a	=	4	%	SM 4500-NH3 F		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	LCS dup, rec	9/25/2007	Nutrient	Nitrate as N	n/a	=	92	%	EPA 300.0		70	130	
2007/08-1	Lab	LCS, rec	9/25/2007	Nutrient	Nitrate as N	n/a	=	94	%	EPA 300.0		70	130	
2007/08-1	Lab	LCS, RPD	9/25/2007	Nutrient	Nitrate as N	n/a	=	2	%	EPA 300.0		0	30	
2007/08-1	Lab	method blank	9/25/2007	Nutrient	Nitrate as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	9/25/2007	Nutrient	Nitrate as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		30	
2007/08-1	ME-CC	matrix spike dup, rec	9/25/2007	Nutrient	Nitrate as N	n/a	=	86	%	EPA 300.0		70	130	
2007/08-1	ME-CC	matrix spike, rec	9/25/2007	Nutrient	Nitrate as N	n/a	=	84	%	EPA 300.0		70	130	
2007/08-1	ME-CC	matrix spike, RPD	9/25/2007	Nutrient	Nitrate as N	n/a	=	2	%	EPA 300.0		0	30	
2007/08-1	Lab	LCS dup, rec	9/25/2007	Nutrient	Nitrite as N	n/a	=	90	%	EPA 300.0		70	130	
2007/08-1	Lab	LCS, rec	9/25/2007	Nutrient	Nitrite as N	n/a	=	90	%	EPA 300.0		70	130	
2007/08-1	Lab	LCS, RPD	9/25/2007	Nutrient	Nitrite as N	n/a	=	0	%	EPA 300.0		0	30	
2007/08-1	Lab	method blank	9/25/2007	Nutrient	Nitrite as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	9/25/2007	Nutrient	Nitrite as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		30	
2007/08-1	ME-CC	matrix spike dup, rec	9/25/2007	Nutrient	Nitrite as N	n/a	=	97	%	EPA 300.0		70	130	
2007/08-1	ME-CC	matrix spike, rec	9/25/2007	Nutrient	Nitrite as N	n/a	=	97	%	EPA 300.0		70	130	
2007/08-1	ME-CC	matrix spike, RPD	9/25/2007	Nutrient	Nitrite as N	n/a	=	0	%	EPA 300.0		0	30	
2007/08-1	Lab	LCS dup, rec	9/25/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	96	%	EPA 300.0		70	130	
2007/08-1	Lab	LCS, rec	9/25/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	95	%	EPA 300.0		70	130	
2007/08-1	Lab	LCS, RPD	9/25/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	1	%	EPA 300.0		0	30	
2007/08-1	Lab	method blank	9/25/2007	Nutrient	Orthophosphate as P (Diss)	n/a	<	0.0075	mg/L	EPA 300.0	0.0075		0.0075	
2007/08-1	ME-CC	lab duplicate	9/25/2007	Nutrient	Orthophosphate as P (Diss)	n/a	<	0.0075	mg/L	EPA 300.0	0.0075		30	
2007/08-1	ME-CC	matrix spike dup, rec	9/25/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	78	%	EPA 300.0		70	130	
2007/08-1	ME-CC	matrix spike, rec	9/25/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	80	%	EPA 300.0		70	130	
2007/08-1	ME-CC	matrix spike, RPD	9/25/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	3	%	EPA 300.0		0	30	
2007/08-1	Lab	LCS, rec	10/2/2007	Nutrient	TKN	n/a	=	90.3	%	EPA 351.1		80	120	
2007/08-1	Lab	method blank	10/5/2007	Nutrient	TKN	n/a	<	0.05	mg/L	EPA 351.1	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/2/2007	Nutrient	TKN	n/a	=	0.38	mg/L	EPA 351.1	0.05		20	
2007/08-1	ME-VR2	matrix spike dup, rec	10/5/2007	Nutrient	TKN	n/a	=	91.7	%	EPA 351.1		80	120	
2007/08-1	ME-VR2	matrix spike, rec	10/5/2007	Nutrient	TKN	n/a	=	92.3	%	EPA 351.1		80	120	
2007/08-1	ME-VR2	matrix spike, RPD	10/5/2007	Nutrient	TKN	n/a	=	0.7	%	EPA 351.1		0	20	
2007/08-1	Lab	LCS dup, rec	9/25/2007	Nutrient	Total Phosphorus	Dissolved	=	111	%	SM 4500-P C		70	130	
2007/08-1	Lab	LCS, rec	9/25/2007	Nutrient	Total Phosphorus	Dissolved	=	104	%	SM 4500-P C		70	130	
2007/08-1	Lab	LCS, RPD	9/25/2007	Nutrient	Total Phosphorus	Dissolved	=	7	%	SM 4500-P C		0	30	
2007/08-1	Lab	method blank	9/25/2007	Nutrient	Total Phosphorus	Dissolved	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2007/08-1	ME-CC	lab duplicate	9/25/2007	Nutrient	Total Phosphorus	Dissolved	=	0.66	mg/L	SM 4500-P C	0.016		30	
2007/08-1	ME-CC	matrix spike dup, rec	9/25/2007	Nutrient	Total Phosphorus	Dissolved	=	104	%	SM 4500-P C		70	130	
2007/08-1	ME-CC	matrix spike, rec	9/25/2007	Nutrient	Total Phosphorus	Dissolved	=	104	%	SM 4500-P C		70	130	
2007/08-1	ME-CC	matrix spike, RPD	9/25/2007	Nutrient	Total Phosphorus	Dissolved	=	0	%	SM 4500-P C		0	30	
2007/08-1	Lab	LCS dup, rec	9/25/2007	Nutrient	Total Phosphorus	Total	=	108	%	SM 4500-P C		70	130	
2007/08-1	Lab	LCS, rec	9/25/2007	Nutrient	Total Phosphorus	Total	=	104	%	SM 4500-P C		70	130	
2007/08-1	Lab	LCS, RPD	9/25/2007	Nutrient	Total Phosphorus	Total	=	4	%	SM 4500-P C		0	30	
2007/08-1	Lab	method blank	9/25/2007	Nutrient	Total Phosphorus	Total	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2007/08-1	ME-CC	lab duplicate	9/25/2007	Nutrient	Total Phosphorus	Total	=	2.079	mg/L	SM 4500-P C	0.016		30	
2007/08-1	ME-CC	matrix spike dup, rec	9/25/2007	Nutrient	Total Phosphorus	Total	=	101	%	SM 4500-P C		70	130	
2007/08-1	ME-CC	matrix spike, rec	9/25/2007	Nutrient	Total Phosphorus	Total	=	101	%	SM 4500-P C		70	130	
2007/08-1	ME-CC	matrix spike, RPD	9/25/2007	Nutrient	Total Phosphorus	Total	=	0	%	SM 4500-P C		0	30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	1,2,4-Trichlorobenzene	n/a	=	67	%	EPA 625m		45	140	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	1,2,4-Trichlorobenzene	n/a	=	68	%	EPA 625m		45	140	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	1,2,4-Trichlorobenzene	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	1,2,4-Trichlorobenzene	n/a	=	55	%	EPA 625m		45	140	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	1,2,4-Trichlorobenzene	n/a	=	54	%	EPA 625m		45	140	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	1,2,4-Trichlorobenzene	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	method blank	10/16/2007	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	1,2-Dichlorobenzene	n/a	=	0.017	µg/L	EPA 625m	0.01		0.01	EST
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	srgt method blank, rec	9/28/2007	Organic	1,2-Dichloroethane-d4	n/a	=	96	%	EPA 8260B		74	146	
2007/08-1	W-4	srgt environ, rec	9/28/2007	Organic	1,2-Dichloroethane-d4	n/a	=	91	%	EPA 8260B		74	146	
2007/08-1	Lab	method blank	10/16/2007	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	1,3-Dichlorobenzene	n/a	=	0.027	µg/L	EPA 625m	0.01		0.01	EST
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	srgt method blank, rec	9/28/2007	Organic	1,4-Bromofluorobenzene	n/a	=	91	%	EPA 8260B		74	110	
2007/08-1	W-4	srgt environ, rec	9/28/2007	Organic	1,4-Bromofluorobenzene	n/a	=	91	%	EPA 8260B		74	110	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	1,4-Dichlorobenzene	n/a	=	49	%	EPA 625m		45	140	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	1,4-Dichlorobenzene	n/a	=	58	%	EPA 625m		45	140	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	1,4-Dichlorobenzene	n/a	=	17	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	1,4-Dichlorobenzene	n/a	=	48	%	EPA 625m		45	140	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	1,4-Dichlorobenzene	n/a	=	45	%	EPA 625m		45	140	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	1,4-Dichlorobenzene	n/a	=	6	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	1,4-Dichlorobenzene	n/a	=	0.018	µg/L	EPA 625m	0.01		0.01	EST
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	1-Methylnaphthalene	n/a	=	76	%	EPA 625m		50	120	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	1-Methylnaphthalene	n/a	=	84	%	EPA 625m		50	120	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	1-Methylnaphthalene	n/a	=	10	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	1-Methylnaphthalene	n/a	=	0.0043	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	1-Methylnaphthalene	n/a	=	76	%	EPA 625m		50	120	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	1-Methylnaphthalene	n/a	=	73	%	EPA 625m		50	120	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	1-Methylnaphthalene	n/a	=	4	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	1-Methylnaphthalene	n/a	=	0.0026	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	1-Methylnaphthalene	n/a	=	0.008	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	1-Methylphenanthrene	n/a	=	122	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	1-Methylphenanthrene	n/a	=	91	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	1-Methylphenanthrene	n/a	=	29	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	1-Methylphenanthrene	n/a	=	0.0058	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	1-Methylphenanthrene	n/a	=	95	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	1-Methylphenanthrene	n/a	=	95	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	1-Methylphenanthrene	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	2,3,5-Trimethylnaphthalene	n/a	=	93	%	EPA 625m		45	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	2,3,5-Trimethylnaphthalene	n/a	=	80	%	EPA 625m		45	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	2,3,5-Trimethylnaphthalene	n/a	=	15	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2,3,5-Trimethylnaphthalene	n/a	=	0.0046	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	2,3,5-Trimethylnaphthalene	n/a	=	82	%	EPA 625m		45	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	2,3,5-Trimethylnaphthalene	n/a	=	80	%	EPA 625m		45	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	2,3,5-Trimethylnaphthalene	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	srgt LCS dup, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	67	%	EPA 625m		40	130	
2007/08-1	Lab	srgt LCS, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	74	%	EPA 625m		40	130	
2007/08-1	Lab	srgt method blank, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	67	%	EPA 625m		40	130	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	70	%	EPA 625m		40	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	67	%	EPA 625m		40	130	
2007/08-1	ME-CC	srgt matrix spike dup, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	73	%	EPA 625m		40	130	
2007/08-1	ME-CC	srgt matrix spike, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	73	%	EPA 625m		40	130	
2007/08-1	ME-SCR	srgt environ, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	72	%	EPA 625m		40	130	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	74	%	EPA 625m		40	130	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	75	%	EPA 625m		40	130	
2007/08-1	ME-VR2	srgt field blank, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	68	%	EPA 625m		40	130	
2007/08-1	W-4	srgt environ, rec	10/16/2007	Organic	2,4,6-Tribromophenol	n/a	=	84	%	EPA 625m		40	130	
2007/08-1	Lab	method blank	10/16/2007	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	srgt method blank, rec	10/1/2007	Organic	2,4-Dichlorophenylacetic acid	n/a	=	90	%	EPA 8151A		0	123	
2007/08-1	ME-CC	srgt environ, rec	10/1/2007	Organic	2,4-Dichlorophenylacetic acid	n/a	=	0	%	EPA 8151A		0	123	
2007/08-1	ME-SCR	srgt environ, rec	10/1/2007	Organic	2,4-Dichlorophenylacetic acid	n/a	=	104	%	EPA 8151A		0	123	
2007/08-1	ME-VR2	srgt environ, rec	10/1/2007	Organic	2,4-Dichlorophenylacetic acid	n/a	=	121	%	EPA 8151A		0	123	
2007/08-1	Lab	method blank	10/16/2007	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	2,4-Dinitrotoluene	n/a	=	111	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	2,4-Dinitrotoluene	n/a	=	95	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	2,4-Dinitrotoluene	n/a	=	16	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	2,4-Dinitrotoluene	n/a	=	100	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	2,4-Dinitrotoluene	n/a	=	107	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	2,4-Dinitrotoluene	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	2,6-Dimethylnaphthalene	n/a	=	75	%	EPA 625m		55	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	2,6-Dimethylnaphthalene	n/a	=	83	%	EPA 625m		55	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	2,6-Dimethylnaphthalene	n/a	=	10	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2,6-Dimethylnaphthalene	n/a	=	0.0028	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	2,6-Dimethylnaphthalene	n/a	=	81	%	EPA 625m		55	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	2,6-Dimethylnaphthalene	n/a	=	79	%	EPA 625m		55	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	2,6-Dimethylnaphthalene	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	2-Chlorophenol	n/a	=	51	%	EPA 625m		35	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	2-Chlorophenol	n/a	=	69	%	EPA 625m		35	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	2-Chlorophenol	n/a	=	30	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	2-Chlorophenol	n/a	=	63	%	EPA 625m		35	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	2-Chlorophenol	n/a	=	59	%	EPA 625m		35	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	2-Chlorophenol	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	2-Methylnaphthalene	n/a	=	83	%	EPA 625m		50	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	2-Methylnaphthalene	n/a	=	82	%	EPA 625m		50	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	2-Methylnaphthalene	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2-Methylnaphthalene	n/a	=	0.0115	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	2-Methylnaphthalene	n/a	=	78	%	EPA 625m		50	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	2-Methylnaphthalene	n/a	=	72	%	EPA 625m		50	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	2-Methylnaphthalene	n/a	=	8	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2-Methylnaphthalene	n/a	=	0.003	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2-Methylnaphthalene	n/a	=	0.0096	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	4-Chloro-3-methylphenol	n/a	=	74	%	EPA 625m		30	150	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	4-Chloro-3-methylphenol	n/a	=	95	%	EPA 625m		30	150	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	4-Chloro-3-methylphenol	n/a	=	25	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	4-Chloro-3-methylphenol	n/a	=	93	%	EPA 625m		30	150	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	4-Chloro-3-methylphenol	n/a	=	91	%	EPA 625m		30	150	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	4-Chloro-3-methylphenol	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	4-Nitrophenol	n/a	=	89	%	EPA 625m		0.1	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	4-Nitrophenol	n/a	=	85	%	EPA 625m		0.1	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	4-Nitrophenol	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	4-Nitrophenol	n/a	=	27	%	EPA 625m		0.1	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	4-Nitrophenol	n/a	=	26	%	EPA 625m		0.1	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	4-Nitrophenol	n/a	=	4	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Acenaphthene	n/a	=	114	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Acenaphthene	n/a	=	117	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Acenaphthene	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Acenaphthene	n/a	=	0.0048	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Acenaphthene	n/a	=	106	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Acenaphthene	n/a	=	108	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Acenaphthene	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Acenaphthene	n/a	=	0.003	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	srgt LCS dup, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	94	%	EPA 625m		50	130	
2007/08-1	Lab	srgt LCS, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	87	%	EPA 625m		50	130	
2007/08-1	Lab	srgt method blank, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	88	%	EPA 625m		50	130	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	74	%	EPA 625m		50	130	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	75	%	EPA 625m		50	130	
2007/08-1	ME-CC	srgt matrix spike dup, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	80	%	EPA 625m		50	130	
2007/08-1	ME-CC	srgt matrix spike, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	78	%	EPA 625m		50	130	
2007/08-1	ME-SCR	srgt environ, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	82	%	EPA 625m		50	130	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	80	%	EPA 625m		50	130	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	84	%	EPA 625m		50	130	
2007/08-1	ME-VR2	srgt field blank, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	76	%	EPA 625m		50	130	
2007/08-1	W-4	srgt environ, rec	10/16/2007	Organic	Acenaphthene-d10	n/a	=	47	%	EPA 625m		50	130	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Acenaphthylene	n/a	=	87	%	EPA 625m		60	120	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Acenaphthylene	n/a	=	87	%	EPA 625m		60	120	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Acenaphthylene	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Acenaphthylene	n/a	=	86	%	EPA 625m		60	120	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Acenaphthylene	n/a	=	81	%	EPA 625m		60	120	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Acenaphthylene	n/a	=	6	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Anthracene	n/a	=	118	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Anthracene	n/a	=	90	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Anthracene	n/a	=	27	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Anthracene	n/a	=	0.0042	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Anthracene	n/a	=	76	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Anthracene	n/a	=	82	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Anthracene	n/a	=	8	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Benzo(a)anthracene	n/a	=	122	%	EPA 625m		70	140	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Benzo(a)anthracene	n/a	=	109	%	EPA 625m		70	140	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Benzo(a)anthracene	n/a	=	11	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Benzo(a)anthracene	n/a	=	0.009	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Benzo(a)anthracene	n/a	=	107	%	EPA 625m		70	140	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Benzo(a)anthracene	n/a	=	106	%	EPA 625m		70	140	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Benzo(a)anthracene	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Benzo(a)anthracene	n/a	=	0.0017	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Benzo(a)pyrene	n/a	=	108	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Benzo(a)pyrene	n/a	=	95	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Benzo(a)pyrene	n/a	=	13	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Benzo(a)pyrene	n/a	=	0.0124	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Benzo(a)pyrene	n/a	=	87	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Benzo(a)pyrene	n/a	=	83	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Benzo(a)pyrene	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Benzo(b)fluoranthene	n/a	=	116	%	EPA 625m		60	140	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Benzo(b)fluoranthene	n/a	=	97	%	EPA 625m		60	140	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Benzo(b)fluoranthene	n/a	=	18	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Benzo(b)fluoranthene	n/a	=	0.0224	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Benzo(b)fluoranthene	n/a	=	104	%	EPA 625m		60	140	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Benzo(b)fluoranthene	n/a	=	102	%	EPA 625m		60	140	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Benzo(b)fluoranthene	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Benzo(e)pyrene	n/a	=	118	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Benzo(e)pyrene	n/a	=	104	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Benzo(e)pyrene	n/a	=	13	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Benzo(e)pyrene	n/a	=	0.0202	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Benzo(e)pyrene	n/a	=	99	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Benzo(e)pyrene	n/a	=	95	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Benzo(e)pyrene	n/a	=	4	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Benzo(g,h,i)perylene	n/a	=	100	%	EPA 625m		50	140	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Benzo(g,h,i)perylene	n/a	=	89	%	EPA 625m		50	140	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Benzo(g,h,i)perylene	n/a	=	12	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Benzo(g,h,i)perylene	n/a	=	0.018	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Benzo(g,h,i)perylene	n/a	=	81	%	EPA 625m		50	140	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Benzo(g,h,i)perylene	n/a	=	80	%	EPA 625m		50	140	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Benzo(g,h,i)perylene	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Benzo(g,h,i)perylene	n/a	=	0.0012	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Benzo(k)fluoranthene	n/a	=	117	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Benzo(k)fluoranthene	n/a	=	103	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Benzo(k)fluoranthene	n/a	=	13	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Benzo(k)fluoranthene	n/a	=	0.0086	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Benzo(k)fluoranthene	n/a	=	102	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Benzo(k)fluoranthene	n/a	=	97	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Benzo(k)fluoranthene	n/a	=	5	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Biphenyl	n/a	=	80	%	EPA 625m		50	120	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Biphenyl	n/a	=	89	%	EPA 625m		50	120	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Biphenyl	n/a	=	11	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Biphenyl	n/a	=	0.0151	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Biphenyl	n/a	=	79	%	EPA 625m		50	120	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Biphenyl	n/a	=	81	%	EPA 625m		50	120	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Biphenyl	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	105	%	EPA 625m		20	190	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	116	%	EPA 625m		20	190	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	10	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Bis(2-ethylhexyl)phthalate	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	2.4494	µg/L	EPA 625m	0.1		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	285	%	EPA 625m		20	190	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	917	%	EPA 625m		20	190	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	105	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Bis(2-ethylhexyl)phthalate	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	3.4223	µg/L	EPA 625m	0.1		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Butyl benzyl phthalate	n/a	=	106	%	EPA 625m		65	160	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Butyl benzyl phthalate	n/a	=	115	%	EPA 625m		65	160	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Butyl benzyl phthalate	n/a	=	8	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Butyl benzyl phthalate	n/a	<	0.025	µg/L	EPA 625m	0.025		0.025	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Butyl benzyl phthalate	n/a	=	0.0752	µg/L	EPA 625m	0.025		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Butyl benzyl phthalate	n/a	=	124	%	EPA 625m		65	160	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Butyl benzyl phthalate	n/a	=	150	%	EPA 625m		65	160	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Butyl benzyl phthalate	n/a	=	19	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Butyl benzyl phthalate	n/a	<	0.025	µg/L	EPA 625m	0.025		0.025	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Butyl benzyl phthalate	n/a	<	0.025	µg/L	EPA 625m	0.025		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Chrysene	n/a	=	114	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Chrysene	n/a	=	106	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Chrysene	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Chrysene	n/a	=	0.0248	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Chrysene	n/a	=	109	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Chrysene	n/a	=	111	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Chrysene	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	srgt LCS dup, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	116	%	EPA 625m		70	130	
2007/08-1	Lab	srgt LCS, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	91	%	EPA 625m		70	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	srgt method blank, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	80	%	EPA 625m		70	130	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	79	%	EPA 625m		70	130	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	76	%	EPA 625m		70	130	
2007/08-1	ME-CC	srgt matrix spike dup, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	87	%	EPA 625m		70	130	
2007/08-1	ME-CC	srgt matrix spike, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	89	%	EPA 625m		70	130	
2007/08-1	ME-SCR	srgt environ, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	87	%	EPA 625m		70	130	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	91	%	EPA 625m		70	130	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	86	%	EPA 625m		70	130	
2007/08-1	ME-VR2	srgt field blank, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	79	%	EPA 625m		70	130	
2007/08-1	W-4	srgt environ, rec	10/16/2007	Organic	Chrysene-d12	n/a	=	54	%	EPA 625m		70	130	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Dibenz(a,h)anthracene	n/a	=	95	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Dibenz(a,h)anthracene	n/a	=	97	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Dibenz(a,h)anthracene	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Dibenz(a,h)anthracene	n/a	=	0.0031	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Dibenz(a,h)anthracene	n/a	=	89	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Dibenz(a,h)anthracene	n/a	=	78	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Dibenz(a,h)anthracene	n/a	=	13	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Dibenz(a,h)anthracene	n/a	=	0.0031	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Dibenzothiophene	n/a	=	115	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Dibenzothiophene	n/a	=	98	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Dibenzothiophene	n/a	=	16	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Dibenzothiophene	n/a	=	104	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Dibenzothiophene	n/a	=	106	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Dibenzothiophene	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	srgt method blank, rec	9/28/2007	Organic	Dibromofluoromethane	n/a	=	102	%	EPA 8260B		74	140	
2007/08-1	W-4	srgt environ, rec	9/28/2007	Organic	Dibromofluoromethane	n/a	=	99	%	EPA 8260B		74	140	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Diethyl phthalate	n/a	=	104	%	EPA 625m		50	150	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Diethyl phthalate	n/a	=	113	%	EPA 625m		50	150	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Diethyl phthalate	n/a	=	8	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Diethyl phthalate	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Diethyl phthalate	n/a	=	0.7221	µg/L	EPA 625m	0.1		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Diethyl phthalate	n/a	=	121	%	EPA 625m		50	150	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Diethyl phthalate	n/a	=	120	%	EPA 625m		50	150	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Diethyl phthalate	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Diethyl phthalate	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Diethyl phthalate	n/a	=	0.4115	µg/L	EPA 625m	0.1		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Dimethyl phthalate	n/a	=	96	%	EPA 625m		40	155	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Dimethyl phthalate	n/a	=	101	%	EPA 625m		40	155	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Dimethyl phthalate	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Dimethyl phthalate	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Dimethyl phthalate	n/a	=	0.0665	µg/L	EPA 625m	0.05		30	EST
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Dimethyl phthalate	n/a	=	102	%	EPA 625m		40	155	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Dimethyl phthalate	n/a	=	100	%	EPA 625m		40	155	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Dimethyl phthalate	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Dimethyl phthalate	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Dimethyl phthalate	n/a	=	0.0542	µg/L	EPA 625m	0.05		30	EST
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	108	%	EPA 625m		65	145	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	109	%	EPA 625m		65	145	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Di-n-butylphthalate	n/a	<	0.075	µg/L	EPA 625m	0.075		0.075	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Di-n-butylphthalate	n/a	<	0.075	µg/L	EPA 625m	0.075		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	123	%	EPA 625m		65	145	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	129	%	EPA 625m		65	145	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Di-n-butylphthalate	n/a	<	0.075	µg/L	EPA 625m	0.075		0.075	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Di-n-butylphthalate	n/a	<	0.075	µg/L	EPA 625m	0.075		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	98	%	EPA 625m		50	165	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	115	%	EPA 625m		50	165	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	16	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Di-n-butylphthalate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	0.0513	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	108	%	EPA 625m		50	165	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	114	%	EPA 625m		50	165	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Di-n-butylphthalate	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Di-n-butylphthalate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Di-n-butylphthalate	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Fluoranthene	n/a	=	124	%	EPA 625m		65	135	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Fluoranthene	n/a	=	104	%	EPA 625m		65	135	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Fluoranthene	n/a	=	18	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Fluoranthene	n/a	=	0.025	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Fluoranthene	n/a	=	102	%	EPA 625m		65	135	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Fluoranthene	n/a	=	102	%	EPA 625m		65	135	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Fluoranthene	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Fluorene	n/a	=	104	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Fluorene	n/a	=	90	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Fluorene	n/a	=	14	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Fluorene	n/a	=	87	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Fluorene	n/a	=	90	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Fluorene	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	97	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	97	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	0.0169	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	86	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	77	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	11	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	0.0014	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Isophorone	n/a	=	0.052	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	LCS dup, rec	9/28/2007	Organic	Methyl tert-butyl ether (MTBE)	n/a	=	94	%	EPA 8260B		82	118	
2007/08-1	Lab	LCS, rec	9/28/2007	Organic	Methyl tert-butyl ether (MTBE)	n/a	=	96	%	EPA 8260B		82	118	
2007/08-1	Lab	LCS, RPD	9/28/2007	Organic	Methyl tert-butyl ether (MTBE)	n/a	=	2	%	EPA 8260B		0	13	
2007/08-1	Lab	method blank	9/28/2007	Organic	Methyl tert-butyl ether (MTBE)	n/a	<	1	µg/L	EPA 8260B	1		1	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Naphthalene	n/a	=	72	%	EPA 625m		50	120	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Naphthalene	n/a	=	80	%	EPA 625m		50	120	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Naphthalene	n/a	=	11	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Naphthalene	n/a	=	0.0244	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Naphthalene	n/a	=	76	%	EPA 625m		50	120	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Naphthalene	n/a	=	68	%	EPA 625m		50	120	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Naphthalene	n/a	=	11	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Naphthalene	n/a	=	0.0159	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Naphthalene	n/a	=	0.032	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	srgt LCS dup, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	73	%	EPA 625m		40	120	
2007/08-1	Lab	srgt LCS, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	86	%	EPA 625m		40	120	
2007/08-1	Lab	srgt method blank, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	83	%	EPA 625m		40	120	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	68	%	EPA 625m		40	120	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	62	%	EPA 625m		40	120	
2007/08-1	ME-CC	srgt matrix spike dup, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	71	%	EPA 625m		40	120	
2007/08-1	ME-CC	srgt matrix spike, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	69	%	EPA 625m		40	120	
2007/08-1	ME-SCR	srgt environ, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	80	%	EPA 625m		40	120	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	70	%	EPA 625m		40	120	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	80	%	EPA 625m		40	120	
2007/08-1	ME-VR2	srgt field blank, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	67	%	EPA 625m		40	120	
2007/08-1	W-4	srgt environ, rec	10/16/2007	Organic	Naphthalene-d8	n/a	=	51	%	EPA 625m		40	120	
2007/08-1	Lab	method blank	10/16/2007	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	N-Nitrosodi-N-propylamine	n/a	=	73	%	EPA 625m		55	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	N-Nitrosodi-N-propylamine	n/a	=	64	%	EPA 625m		55	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	N-Nitrosodi-N-propylamine	n/a	=	13	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	N-Nitrosodi-N-propylamine	n/a	=	57	%	EPA 625m		55	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	N-Nitrosodi-N-propylamine	n/a	=	74	%	EPA 625m		55	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	N-Nitrosodi-N-propylamine	n/a	=	26	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	method blank	10/16/2007	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Pentachlorophenol	n/a	=	96	%	EPA 625m		10	160	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Pentachlorophenol	n/a	=	106	%	EPA 625m		10	160	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Pentachlorophenol	n/a	=	10	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Pentachlorophenol	n/a	=	109	%	EPA 625m		10	160	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Pentachlorophenol	n/a	=	108	%	EPA 625m		10	160	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Pentachlorophenol	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Perylene	n/a	=	104	%	EPA 625m		65	135	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Perylene	n/a	=	91	%	EPA 625m		65	135	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Perylene	n/a	=	13	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Perylene	n/a	=	0.0085	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Perylene	n/a	=	86	%	EPA 625m		65	135	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Perylene	n/a	=	80	%	EPA 625m		65	135	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Perylene	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	srgt LCS dup, rec	10/16/2007	Organic	Perylene-d12	n/a	=	104	%	EPA 625m		60	140	
2007/08-1	Lab	srgt LCS, rec	10/16/2007	Organic	Perylene-d12	n/a	=	86	%	EPA 625m		60	140	
2007/08-1	Lab	srgt method blank, rec	10/16/2007	Organic	Perylene-d12	n/a	=	76	%	EPA 625m		60	140	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Perylene-d12	n/a	=	64	%	EPA 625m		60	140	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Perylene-d12	n/a	=	60	%	EPA 625m		60	140	
2007/08-1	ME-CC	srgt matrix spike dup, rec	10/16/2007	Organic	Perylene-d12	n/a	=	71	%	EPA 625m		60	140	
2007/08-1	ME-CC	srgt matrix spike, rec	10/16/2007	Organic	Perylene-d12	n/a	=	72	%	EPA 625m		60	140	
2007/08-1	ME-SCR	srgt environ, rec	10/16/2007	Organic	Perylene-d12	n/a	=	76	%	EPA 625m		60	140	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Perylene-d12	n/a	=	73	%	EPA 625m		60	140	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Perylene-d12	n/a	=	74	%	EPA 625m		60	140	
2007/08-1	ME-VR2	srgt field blank, rec	10/16/2007	Organic	Perylene-d12	n/a	=	65	%	EPA 625m		60	140	
2007/08-1	W-4	srgt environ, rec	10/16/2007	Organic	Perylene-d12	n/a	=	51	%	EPA 625m		60	140	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Phenanthrene	n/a	=	119	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Phenanthrene	n/a	=	91	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Phenanthrene	n/a	=	27	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Phenanthrene	n/a	=	0.011	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Phenanthrene	n/a	=	97	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Phenanthrene	n/a	=	92	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Phenanthrene	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Phenanthrene	n/a	=	0.0022	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	srgt LCS dup, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	111	%	EPA 625m		70	130	
2007/08-1	Lab	srgt LCS, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	94	%	EPA 625m		70	130	
2007/08-1	Lab	srgt method blank, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	90	%	EPA 625m		70	130	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	80	%	EPA 625m		70	130	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	78	%	EPA 625m		70	130	
2007/08-1	ME-CC	srgt matrix spike dup, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	86	%	EPA 625m		70	130	
2007/08-1	ME-CC	srgt matrix spike, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	87	%	EPA 625m		70	130	
2007/08-1	ME-SCR	srgt environ, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	89	%	EPA 625m		70	130	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	88	%	EPA 625m		70	130	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	88	%	EPA 625m		70	130	
2007/08-1	ME-VR2	srgt field blank, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	82	%	EPA 625m		70	130	
2007/08-1	W-4	srgt environ, rec	10/16/2007	Organic	Phenanthrene-d10	n/a	=	45	%	EPA 625m		70	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Phenol	n/a	=	53	%	EPA 625m		0.1	115	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Phenol	n/a	=	67	%	EPA 625m		0.1	115	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Phenol	n/a	=	23	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Phenol	n/a	=	33	%	EPA 625m		0.1	115	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Phenol	n/a	=	30	%	EPA 625m		0.1	115	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Phenol	n/a	=	10	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Phenol	n/a	=	0.355	µg/L	EPA 625m	0.1		30	
2007/08-1	Lab	srgt LCS dup, rec	10/16/2007	Organic	Phenol-d5	n/a	=	86	%	EPA 625m		10	110	
2007/08-1	Lab	srgt LCS, rec	10/16/2007	Organic	Phenol-d5	n/a	=	95	%	EPA 625m		10	110	
2007/08-1	Lab	srgt method blank, rec	10/16/2007	Organic	Phenol-d5	n/a	=	92	%	EPA 625m		10	110	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Phenol-d5	n/a	=	28	%	EPA 625m		10	110	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Phenol-d5	n/a	=	25	%	EPA 625m		10	110	
2007/08-1	ME-CC	srgt matrix spike dup, rec	10/16/2007	Organic	Phenol-d5	n/a	=	34	%	EPA 625m		10	110	
2007/08-1	ME-CC	srgt matrix spike, rec	10/16/2007	Organic	Phenol-d5	n/a	=	32	%	EPA 625m		10	110	
2007/08-1	ME-SCR	srgt environ, rec	10/16/2007	Organic	Phenol-d5	n/a	=	34	%	EPA 625m		10	110	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Phenol-d5	n/a	=	36	%	EPA 625m		10	110	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Phenol-d5	n/a	=	30	%	EPA 625m		10	110	
2007/08-1	ME-VR2	srgt field blank, rec	10/16/2007	Organic	Phenol-d5	n/a	=	25	%	EPA 625m		40	130	
2007/08-1	W-4	srgt environ, rec	10/16/2007	Organic	Phenol-d5	n/a	=	16	%	EPA 625m		10	110	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Organic	Pyrene	n/a	=	126	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Organic	Pyrene	n/a	=	127	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Organic	Pyrene	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Pyrene	n/a	=	0.0279	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Organic	Pyrene	n/a	=	119	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Organic	Pyrene	n/a	=	122	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Organic	Pyrene	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	srgt LCS dup, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	77	%	EPA 625m		40	130	
2007/08-1	Lab	srgt LCS, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	85	%	EPA 625m		40	130	
2007/08-1	Lab	srgt method blank, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	115	%	EPA 625m		40	130	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	95	%	EPA 625m		40	130	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	102	%	EPA 625m		40	130	
2007/08-1	ME-CC	srgt matrix spike dup, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	73	%	EPA 625m		40	130	
2007/08-1	ME-CC	srgt matrix spike, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	84	%	EPA 625m		40	130	
2007/08-1	ME-SCR	srgt environ, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	111	%	EPA 625m		40	130	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	105	%	EPA 625m		40	130	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	107	%	EPA 625m		40	130	
2007/08-1	ME-VR2	srgt field blank, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	102	%	EPA 625m		40	130	
2007/08-1	W-4	srgt environ, rec	10/16/2007	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	77	%	EPA 625m		40	130	
2007/08-1	Lab	srgt method blank, rec	9/28/2007	Organic	Toluene-d8	n/a	=	98	%	EPA 8260B		88	112	
2007/08-1	W-4	srgt environ, rec	9/28/2007	Organic	Toluene-d8	n/a	=	96	%	EPA 8260B		88	112	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Organic	Total Detectable PAHs	n/a	=	0.2853	µg/L	EPA 625m			30	
2007/08-1	ME-VR2	field blank	10/16/2007	Organic	Total Detectable PAHs	n/a	=	0.0311	µg/L	EPA 625m				
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Organic	Total Detectable PAHs	n/a	=	0.0526	µg/L	EPA 625m			30	
2007/08-1	Lab	method blank	10/16/2007	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	method blank	10/16/2007	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	method blank	10/16/2007	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	method blank	10/16/2007	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	method blank	10/16/2007	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	method blank	10/16/2007	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	method blank	10/16/2007	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 008	n/a	=	85	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 008	n/a	=	84	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 008	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 008	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 008	n/a	=	89	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 008	n/a	=	89	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 008	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 018	n/a	=	86	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 018	n/a	=	84	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 018	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 018	n/a	=	91	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 018	n/a	=	91	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 018	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 028	n/a	=	89	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 028	n/a	=	90	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 028	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 028	n/a	=	90	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 028	n/a	=	91	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 028	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	srgt LCS dup, rec	10/16/2007	PCB	PCB 030	n/a	=	93	%	EPA 625m		40	130	
2007/08-1	Lab	srgt LCS, rec	10/16/2007	PCB	PCB 030	n/a	=	91	%	EPA 625m		40	130	
2007/08-1	Lab	srgt method blank, rec	10/16/2007	PCB	PCB 030	n/a	=	108	%	EPA 625m		40	130	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	PCB	PCB 030	n/a	=	92	%	EPA 625m		40	130	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	PCB	PCB 030	n/a	=	94	%	EPA 625m		40	130	
2007/08-1	ME-CC	srgt matrix spike dup, rec	10/16/2007	PCB	PCB 030	n/a	=	87	%	EPA 625m		40	130	
2007/08-1	ME-CC	srgt matrix spike, rec	10/16/2007	PCB	PCB 030	n/a	=	92	%	EPA 625m		40	130	
2007/08-1	ME-SCR	srgt environ, rec	10/16/2007	PCB	PCB 030	n/a	=	101	%	EPA 625m		40	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	PCB	PCB 030	n/a	=	101	%	EPA 625m		40	130	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	PCB	PCB 030	n/a	=	101	%	EPA 625m		40	130	
2007/08-1	ME-VR2	srgt field blank, rec	10/16/2007	PCB	PCB 030	n/a	=	97	%	EPA 625m		40	130	
2007/08-1	W-4	srgt environ, rec	10/16/2007	PCB	PCB 030	n/a	=	64	%	EPA 625m		40	130	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 031	n/a	=	92	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 031	n/a	=	91	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 031	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 031	n/a	=	86	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 031	n/a	=	88	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 031	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 033	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 033	n/a	=	93	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 033	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 033	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 033	n/a	=	93	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 033	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 037	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 037	n/a	=	103	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 037	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 037	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 037	n/a	=	86	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 037	n/a	=	10	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 044	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 044	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 044	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 044	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 044	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 044	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 049	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 049	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 049	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 049	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 049	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 049	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 052	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 052	n/a	=	96	%	EPA 625m		60	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 052	n/a	=	4	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 052	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 052	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 052	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 066	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 066	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 066	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 066	n/a	=	91	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 066	n/a	=	90	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 066	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 070	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 070	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 070	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 070	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 070	n/a	=	91	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 070	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 074	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 074	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 074	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 074	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 074	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 074	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 077	n/a	=	104	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 077	n/a	=	104	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 077	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 077	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 077	n/a	=	86	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 077	n/a	=	9	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 081	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 081	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 081	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 081	n/a	=	90	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 081	n/a	=	86	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 081	n/a	=	5	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 087	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 087	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 087	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 087	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 087	n/a	=	105	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 087	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 095	n/a	=	93	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 095	n/a	=	92	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 095	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 095	n/a	=	90	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 095	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 095	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 097	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 097	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 097	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 097	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 097	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 097	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 099	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 099	n/a	=	103	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 099	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 099	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 099	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 099	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 101	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 101	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 101	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 101	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 101	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 101	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 105	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 105	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 105	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 105	n/a	=	89	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 105	n/a	=	87	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 105	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 110	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 110	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 110	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 110	n/a	=	92	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 110	n/a	=	93	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 110	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	srgt LCS dup, rec	10/16/2007	PCB	PCB 112	n/a	=	96	%	EPA 625m		60	120	
2007/08-1	Lab	srgt LCS, rec	10/16/2007	PCB	PCB 112	n/a	=	95	%	EPA 625m		60	120	
2007/08-1	Lab	srgt method blank, rec	10/16/2007	PCB	PCB 112	n/a	=	99	%	EPA 625m		60	120	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	PCB	PCB 112	n/a	=	87	%	EPA 625m		60	120	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	PCB	PCB 112	n/a	=	88	%	EPA 625m		60	120	
2007/08-1	ME-CC	srgt matrix spike dup, rec	10/16/2007	PCB	PCB 112	n/a	=	90	%	EPA 625m		60	120	
2007/08-1	ME-CC	srgt matrix spike, rec	10/16/2007	PCB	PCB 112	n/a	=	100	%	EPA 625m		60	120	
2007/08-1	ME-SCR	srgt environ, rec	10/16/2007	PCB	PCB 112	n/a	=	93	%	EPA 625m		60	120	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	PCB	PCB 112	n/a	=	91	%	EPA 625m		60	120	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	PCB	PCB 112	n/a	=	97	%	EPA 625m		60	120	
2007/08-1	ME-VR2	srgt field blank, rec	10/16/2007	PCB	PCB 112	n/a	=	95	%	EPA 625m		60	120	
2007/08-1	W-4	srgt environ, rec	10/16/2007	PCB	PCB 112	n/a	=	60	%	EPA 625m		60	120	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 114	n/a	=	105	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 114	n/a	=	105	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 114	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 114	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 114	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 114	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 118	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 118	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 118	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 118	n/a	=	91	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 118	n/a	=	93	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 118	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 119	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 119	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 119	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 119	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 119	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 119	n/a	=	0	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 123	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 123	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 123	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 123	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 123	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 123	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 126	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 126	n/a	=	108	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 126	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 126	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 126	n/a	=	93	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 126	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 128	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 128	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 128	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 128	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 128	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 128	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 128	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 128	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 128	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 128	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 138	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 138	n/a	=	104	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 138	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 138	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 138	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 138	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 141	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 141	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 141	n/a	=	4	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 141	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 141	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 141	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 149	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 149	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 149	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 149	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 149	n/a	=	92	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 149	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 151	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 151	n/a	=	103	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 151	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 151	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 151	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 151	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 153	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 153	n/a	=	105	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 153	n/a	=	4	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 153	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 153	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 153	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 156	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 156	n/a	=	104	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 156	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 156	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 156	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 156	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 157	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 157	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 157	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 157	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 157	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 157	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 158	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 158	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 158	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 158	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 158	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 158	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 167	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 167	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 167	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 167	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 167	n/a	=	91	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 167	n/a	=	93	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 167	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 167	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 168 + 132	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 168 + 132	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 168 + 132	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 168 + 132	n/a	=	91	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 168 + 132	n/a	=	93	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 168 + 132	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 169	n/a	=	92	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 169	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 169	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 169	n/a	=	103	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 169	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 169	n/a	=	8	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 170	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 170	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 170	n/a	=	6	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 170	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 170	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 170	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 177	n/a	=	108	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 177	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 177	n/a	=	9	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 177	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 177	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 177	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 180	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 180	n/a	=	104	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 180	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 180	n/a	=	99	%	EPA 625m		60	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 180	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 180	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 183	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 183	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 183	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 183	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 183	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 183	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 187	n/a	=	103	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 187	n/a	=	103	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 187	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 187	n/a	=	101	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 187	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 187	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 189	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 189	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 189	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 189	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 189	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 189	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 194	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 194	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 194	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 194	n/a	=	99	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 194	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 194	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 195	n/a	=	108	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 195	n/a	=	110	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 195	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 195	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 195	n/a	=	110	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 195	n/a	=	111	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 195	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	srgt LCS dup, rec	10/16/2007	PCB	PCB 198	n/a	=	97	%	EPA 625m		60	120	
2007/08-1	Lab	srgt LCS, rec	10/16/2007	PCB	PCB 198	n/a	=	95	%	EPA 625m		60	120	
2007/08-1	Lab	srgt method blank, rec	10/16/2007	PCB	PCB 198	n/a	=	89	%	EPA 625m		60	120	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	PCB	PCB 198	n/a	=	78	%	EPA 625m		60	120	
2007/08-1	ME-CC	srgt environ, rec	10/16/2007	PCB	PCB 198	n/a	=	77	%	EPA 625m		60	120	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	srgt matrix spike dup, rec	10/16/2007	PCB	PCB 198	n/a	=	90	%	EPA 625m		60	120	
2007/08-1	ME-CC	srgt matrix spike, rec	10/16/2007	PCB	PCB 198	n/a	=	67	%	EPA 625m		60	120	
2007/08-1	ME-SCR	srgt environ, rec	10/16/2007	PCB	PCB 198	n/a	=	87	%	EPA 625m		60	120	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	PCB	PCB 198	n/a	=	84	%	EPA 625m		60	120	
2007/08-1	ME-VR2	srgt environ, rec	10/16/2007	PCB	PCB 198	n/a	=	82	%	EPA 625m		60	120	
2007/08-1	ME-VR2	srgt field blank, rec	10/16/2007	PCB	PCB 198	n/a	=	79	%	EPA 625m		60	120	
2007/08-1	W-4	srgt environ, rec	10/16/2007	PCB	PCB 198	n/a	=	53	%	EPA 625m		60	120	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 200	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 200	n/a	=	100	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 200	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 200	n/a	=	93	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 200	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 200	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 201	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 201	n/a	=	91	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 201	n/a	=	11	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 201	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 201	n/a	=	97	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 201	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 206	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 206	n/a	=	92	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 206	n/a	=	4	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 206	n/a	=	94	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 206	n/a	=	96	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 206	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	PCB	PCB 209	n/a	=	87	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	PCB	PCB 209	n/a	=	90	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	PCB	PCB 209	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	PCB	PCB 209	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	PCB	PCB 209	n/a	=	93	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	PCB	PCB 209	n/a	=	93	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	PCB	PCB 209	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-CC	lab duplicate	10/16/2007	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m			30	
2007/08-1	ME-VR2	field blank	10/16/2007	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2007/08-1	ME-VR2	lab duplicate	10/16/2007	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m			30	
2007/08-1	Lab	LCS dup, rec	10/1/2007	Pesticide	2,4,5-T	n/a	=	107	%	EPA 8151A		30	130	
2007/08-1	Lab	LCS, rec	10/1/2007	Pesticide	2,4,5-T	n/a	=	98	%	EPA 8151A		30	130	
2007/08-1	Lab	LCS, RPD	10/1/2007	Pesticide	2,4,5-T	n/a	=	9	%	EPA 8151A		0	30	
2007/08-1	Lab	method blank	10/1/2007	Pesticide	2,4,5-T	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2007/08-1	Lab	method blank	10/1/2007	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2007/08-1	Lab	LCS dup, rec	10/1/2007	Pesticide	2,4-D	n/a	=	92	%	EPA 8151A		30	130	
2007/08-1	Lab	LCS, rec	10/1/2007	Pesticide	2,4-D	n/a	=	84	%	EPA 8151A		30	130	
2007/08-1	Lab	LCS, RPD	10/1/2007	Pesticide	2,4-D	n/a	=	10	%	EPA 8151A		0	30	
2007/08-1	Lab	method blank	10/1/2007	Pesticide	2,4-D	n/a	<	5	µg/L	EPA 8151A	5		5	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	LCS dup, rec	10/1/2007	Pesticide	2,4-DB	n/a	=	99	%	EPA 8151A		30	130	
2007/08-1	Lab	LCS, rec	10/1/2007	Pesticide	2,4-DB	n/a	=	89	%	EPA 8151A		30	130	
2007/08-1	Lab	LCS, RPD	10/1/2007	Pesticide	2,4-DB	n/a	=	10	%	EPA 8151A		0	30	
2007/08-1	Lab	method blank	10/1/2007	Pesticide	2,4-DB	n/a	<	5	µg/L	EPA 8151A	5		5	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	2,4'-DDD	n/a	=	113	%	EPA 625m		50	140	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	2,4'-DDD	n/a	=	110	%	EPA 625m		50	140	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	2,4'-DDD	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	2,4'-DDD	n/a	=	123	%	EPA 625m		50	140	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	2,4'-DDD	n/a	=	120	%	EPA 625m		50	140	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	2,4'-DDD	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	2,4'-DDE	n/a	=	97	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	2,4'-DDE	n/a	=	101	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	2,4'-DDE	n/a	=	4	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	2,4'-DDE	n/a	=	94	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	2,4'-DDE	n/a	=	108	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	2,4'-DDE	n/a	=	14	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	2,4'-DDT	n/a	=	85	%	EPA 625m		40	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	2,4'-DDT	n/a	=	72	%	EPA 625m		40	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	2,4'-DDT	n/a	=	17	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	2,4'-DDT	n/a	=	40	%	EPA 625m		40	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	2,4'-DDT	n/a	=	43	%	EPA 625m		40	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	2,4'-DDT	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	4,4'-DDD	n/a	=	111	%	EPA 625m		60	140	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	4,4'-DDD	n/a	=	111	%	EPA 625m		60	140	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	4,4'-DDD	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	4,4'-DDD	n/a	=	0.0704	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	4,4'-DDD	n/a	=	106	%	EPA 625m		60	140	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	4,4'-DDD	n/a	=	110	%	EPA 625m		60	140	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	4,4'-DDD	n/a	=	4	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	4,4'-DDE	n/a	=	113	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	4,4'-DDE	n/a	=	121	%	EPA 625m		70	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	4,4'-DDE	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	matrix spike dup, rec	10/5/2007	Pesticide	4,4'-DDE	n/a	=	89	%	EPA 625m		70	130	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	4,4'-DDE	n/a	=	0.1552	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	4,4'-DDE	n/a	=	92	%	EPA 625m		70	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	4,4'-DDE	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	4,4'-DDT	n/a	=	98	%	EPA 625m		0.001	150	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	4,4'-DDT	n/a	=	75	%	EPA 625m		0.001	150	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	4,4'-DDT	n/a	=	27	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	4,4'-DDT	n/a	=	0.001	%	EPA 625m		0.001	150	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	4,4'-DDT	n/a	=	0.001	%	EPA 625m		0.001	150	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	4,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Aldrin	n/a	=	107	%	EPA 625m		50	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Aldrin	n/a	=	89	%	EPA 625m		50	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Aldrin	n/a	=	18	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Aldrin	n/a	=	95	%	EPA 625m		50	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Aldrin	n/a	=	95	%	EPA 625m		50	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Aldrin	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	BHC-alpha	n/a	=	93	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	BHC-alpha	n/a	=	102	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	BHC-alpha	n/a	=	9	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	BHC-alpha	n/a	=	96	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	BHC-alpha	n/a	=	108	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	BHC-alpha	n/a	=	12	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	BHC-beta	n/a	=	98	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	BHC-beta	n/a	=	115	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	BHC-beta	n/a	=	16	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	BHC-beta	n/a	=	95	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	BHC-beta	n/a	=	112	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	BHC-beta	n/a	=	16	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	BHC-delta	n/a	=	105	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	BHC-delta	n/a	=	104	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	BHC-delta	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	BHC-delta	n/a	=	114	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	BHC-delta	n/a	=	121	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	BHC-delta	n/a	=	6	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	BHC-gamma (Lindane)	n/a	=	89	%	EPA 625m		50	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	BHC-gamma (Lindane)	n/a	=	92	%	EPA 625m		50	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	BHC-gamma (Lindane)	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	BHC-gamma (Lindane)	n/a	=	86	%	EPA 625m		50	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	BHC-gamma (Lindane)	n/a	=	82	%	EPA 625m		50	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	BHC-gamma (Lindane)	n/a	=	5	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Bolstar	n/a	=	108	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Bolstar	n/a	=	98	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Bolstar	n/a	=	10	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Bolstar	n/a	=	97	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Bolstar	n/a	=	89	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Bolstar	n/a	=	9	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Chlordane-alpha	n/a	=	112	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Chlordane-alpha	n/a	=	95	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Chlordane-alpha	n/a	=	16	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Chlordane-alpha	n/a	=	101	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Chlordane-alpha	n/a	=	111	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Chlordane-alpha	n/a	=	9	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Chlordane-gamma	n/a	=	89	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Chlordane-gamma	n/a	=	95	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Chlordane-gamma	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Chlordane-gamma	n/a	=	88	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Chlordane-gamma	n/a	=	91	%	EPA 625m		60	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Chlordane-gamma	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Chlorpyrifos	n/a	=	110	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Chlorpyrifos	n/a	=	92	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Chlorpyrifos	n/a	=	18	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Chlorpyrifos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Chlorpyrifos	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Chlorpyrifos	n/a	=	72	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Chlorpyrifos	n/a	=	71	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Chlorpyrifos	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Chlorpyrifos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Chlorpyrifos	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	cis-Nonachlor	n/a	=	92	%	EPA 625m		60	120	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	cis-Nonachlor	n/a	=	113	%	EPA 625m		60	120	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	cis-Nonachlor	n/a	=	20	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	cis-Nonachlor	n/a	=	100	%	EPA 625m		60	120	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	cis-Nonachlor	n/a	=	99	%	EPA 625m		60	120	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	cis-Nonachlor	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	method blank	10/1/2007	Pesticide	Dalapon	n/a	<	13	µg/L	EPA 8151A	13		13	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Demeton-O	n/a	=	82	%	EPA 625m		45	105	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Demeton-O	n/a	=	64	%	EPA 625m		45	105	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Demeton-O	n/a	=	25	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Demeton-O	n/a	=	12	%	EPA 625m		45	105	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Demeton-O	n/a	=	13	%	EPA 625m		45	105	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Demeton-O	n/a	=	8	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Diazinon	n/a	=	107	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Diazinon	n/a	=	89	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Diazinon	n/a	=	18	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Diazinon	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Diazinon	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Diazinon	n/a	=	91	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Diazinon	n/a	=	96	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Diazinon	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Diazinon	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Diazinon	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-1	Lab	method blank	10/1/2007	Pesticide	Dicamba	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2007/08-1	Lab	method blank	10/1/2007	Pesticide	Dichlorprop	n/a	<	5	µg/L	EPA 8151A	5		5	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Dichlorvos	n/a	=	103	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Dichlorvos	n/a	=	87	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Dichlorvos	n/a	=	17	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Dichlorvos	n/a	=	67	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Dichlorvos	n/a	=	75	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Dichlorvos	n/a	=	11	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Dieldrin	n/a	=	117	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Dieldrin	n/a	=	109	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Dieldrin	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Dieldrin	n/a	=	107	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Dieldrin	n/a	=	103	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Dieldrin	n/a	=	4	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Dimethoate	n/a	=	100	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Dimethoate	n/a	=	75	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Dimethoate	n/a	=	29	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Dimethoate	n/a	=	77	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Dimethoate	n/a	=	87	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Dimethoate	n/a	=	12	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-1	Lab	method blank	10/1/2007	Pesticide	Dinoseb	n/a	<	2.5	µg/L	EPA 8151A	2.5		2.5	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Disulfoton	n/a	=	83	%	EPA 625m		45	105	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Disulfoton	n/a	=	78	%	EPA 625m		45	105	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Disulfoton	n/a	=	6	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Disulfoton	n/a	=	10	%	EPA 625m		45	105	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Disulfoton	n/a	=	9	%	EPA 625m		45	105	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Disulfoton	n/a	=	11	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Endosulfan sulfate	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Endosulfan sulfate	n/a	=	105	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Endosulfan sulfate	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Endosulfan sulfate	n/a	=	95	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Endosulfan sulfate	n/a	=	109	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Endosulfan sulfate	n/a	=	14	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Endosulfan-I	n/a	=	79	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Endosulfan-I	n/a	=	102	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Endosulfan-I	n/a	=	25	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Endosulfan-I	n/a	=	121	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Endosulfan-I	n/a	=	112	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Endosulfan-I	n/a	=	8	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Endosulfan-II	n/a	=	98	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Endosulfan-II	n/a	=	105	%	EPA 625m		60	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Endosulfan-II	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Endosulfan-II	n/a	=	119	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Endosulfan-II	n/a	=	107	%	EPA 625m		60	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Endosulfan-II	n/a	=	11	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Endrin	n/a	=	126	%	EPA 625m		65	135	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Endrin	n/a	=	108	%	EPA 625m		65	135	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Endrin	n/a	=	15	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Endrin	n/a	=	112	%	EPA 625m		65	135	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Endrin	n/a	=	113	%	EPA 625m		65	135	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Endrin	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Endrin aldehyde	n/a	=	94	%	EPA 625m		0.001	149	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Endrin aldehyde	n/a	=	81	%	EPA 625m		0.001	149	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Endrin aldehyde	n/a	=	15	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Endrin aldehyde	n/a	=	86	%	EPA 625m		0.001	149	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Endrin aldehyde	n/a	=	104	%	EPA 625m		0.001	149	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Endrin aldehyde	n/a	=	19	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Endrin ketone	n/a	=	104	%	EPA 625m		40	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Endrin ketone	n/a	=	104	%	EPA 625m		40	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Endrin ketone	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Endrin ketone	n/a	=	100	%	EPA 625m		40	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Endrin ketone	n/a	=	106	%	EPA 625m		40	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Endrin ketone	n/a	=	6	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Ethoprop	n/a	=	104	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Ethoprop	n/a	=	80	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Ethoprop	n/a	=	26	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Ethoprop	n/a	=	73	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Ethoprop	n/a	=	77	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Ethoprop	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Fenchlorophos (Ronnel)	n/a	=	104	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Fenchlorophos (Ronnel)	n/a	=	83	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Fenchlorophos (Ronnel)	n/a	=	22	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Fenchlorophos (Ronnel)	n/a	=	62	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Fenchlorophos (Ronnel)	n/a	=	61	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Fenchlorophos (Ronnel)	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Fensulfothion	n/a	=	79	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Fensulfothion	n/a	=	69	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Fensulfothion	n/a	=	14	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Fensulfothion	n/a	=	70	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Fensulfothion	n/a	=	78	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Fensulfothion	n/a	=	11	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Fenthion	n/a	=	95	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Fenthion	n/a	=	78	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Fenthion	n/a	=	20	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Fenthion	n/a	=	45	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Fenthion	n/a	=	52	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Fenthion	n/a	=	14	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-1	Lab	LCS, rec	10/4/2007	Pesticide	Glyphosate	n/a	=	77	%	EPA 547		71	137	
2007/08-1	Lab	method blank	10/4/2007	Pesticide	Glyphosate	n/a	<	5	µg/L	EPA 547	5		5	
2007/08-1	W-4	matrix spike dup, rec	10/4/2007	Pesticide	Glyphosate	n/a	=	100	%	EPA 547		68	134	
2007/08-1	W-4	matrix spike, rec	10/4/2007	Pesticide	Glyphosate	n/a	=	101	%	EPA 547		68	134	
2007/08-1	W-4	matrix spike, RPD	10/4/2007	Pesticide	Glyphosate	n/a	=	0.4	%	EPA 547		0	30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Heptachlor	n/a	=	106	%	EPA 625m		45	135	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Heptachlor	n/a	=	89	%	EPA 625m		45	135	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Heptachlor	n/a	=	17	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Heptachlor	n/a	=	97	%	EPA 625m		45	135	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Heptachlor	n/a	=	85	%	EPA 625m		45	135	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Heptachlor	n/a	=	13	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Heptachlor epoxide	n/a	=	93	%	EPA 625m		65	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Heptachlor epoxide	n/a	=	100	%	EPA 625m		60	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Heptachlor epoxide	n/a	=	7	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Heptachlor epoxide	n/a	=	96	%	EPA 625m		65	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Heptachlor epoxide	n/a	=	106	%	EPA 625m		65	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Heptachlor epoxide	n/a	=	10	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Malathion	n/a	=	91	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Malathion	n/a	=	69	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Malathion	n/a	=	28	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Malathion	n/a	=	77	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Malathion	n/a	=	92	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Malathion	n/a	=	18	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-1	Lab	method blank	10/1/2007	Pesticide	MCPA	n/a	<	500	µg/L	EPA 8151A	500		500	
2007/08-1	Lab	method blank	10/1/2007	Pesticide	MCPD	n/a	<	500	µg/L	EPA 8151A	500		500	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Merphos	n/a	=	91	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Merphos	n/a	=	89	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Merphos	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Merphos	n/a	=	78	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Merphos	n/a	=	74	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Merphos	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Methoxychlor	n/a	=	104	%	EPA 625m		0.001	155	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Methoxychlor	n/a	=	85	%	EPA 625m		0.001	155	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Methoxychlor	n/a	=	20	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Methoxychlor	n/a	=	0.001	%	EPA 625m		0.001	155	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Methoxychlor	n/a	=	0.001	%	EPA 625m		0.001	155	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Methoxychlor	n/a	=	0	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Methyl parathion	n/a	=	104	%	EPA 625m		60	120	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Methyl parathion	n/a	=	84	%	EPA 625m		60	120	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Methyl parathion	n/a	=	21	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Methyl parathion	n/a	=	114	%	EPA 625m		60	120	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Methyl parathion	n/a	=	111	%	EPA 625m		60	120	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Methyl parathion	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Mevinphos	n/a	=	100	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Mevinphos	n/a	=	85	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Mevinphos	n/a	=	16	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		0.008	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Mevinphos	n/a	=	71	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Mevinphos	n/a	=	69	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Mevinphos	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		0.008	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Mirex	n/a	=	90	%	EPA 625m		50	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Mirex	n/a	=	99	%	EPA 625m		50	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Mirex	n/a	=	10	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Mirex	n/a	=	83	%	EPA 625m		50	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Mirex	n/a	=	71	%	EPA 625m		50	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Mirex	n/a	=	16	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Oxychlordane	n/a	=	93	%	EPA 625m		50	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Oxychlordane	n/a	=	121	%	EPA 625m		50	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Oxychlordane	n/a	=	26	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Oxychlordane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Oxychlordane	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Oxychlordane	n/a	=	120	%	EPA 625m		50	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Oxychlordane	n/a	=	114	%	EPA 625m		50	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Oxychlordane	n/a	=	5	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Oxychlordane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Oxychlordane	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Phorate	n/a	=	96	%	EPA 625m		45	105	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Phorate	n/a	=	70	%	EPA 625m		45	105	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Phorate	n/a	=	31	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006		0.006	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Phorate	n/a	=	52	%	EPA 625m		45	105	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Phorate	n/a	=	49	%	EPA 625m		45	105	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Phorate	n/a	=	6	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006		0.006	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	84	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	83	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	69	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	67	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Tokuthion	n/a	=	100	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Tokuthion	n/a	=	84	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Tokuthion	n/a	=	17	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Tokuthion	n/a	=	73	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Tokuthion	n/a	=	69	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Tokuthion	n/a	=	6	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Total Detectable DDTs	n/a	=	0.2256	µg/L	EPA 625m			30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m			30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	trans-Nonachlor	n/a	=	106	%	EPA 625m		55	130	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	trans-Nonachlor	n/a	=	103	%	EPA 625m		55	130	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	trans-Nonachlor	n/a	=	3	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	trans-Nonachlor	n/a	=	99	%	EPA 625m		55	130	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	trans-Nonachlor	n/a	=	101	%	EPA 625m		55	130	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	trans-Nonachlor	n/a	=	2	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	Lab	LCS dup, rec	10/16/2007	Pesticide	Trichloronate	n/a	=	94	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, rec	10/16/2007	Pesticide	Trichloronate	n/a	=	78	%	EPA 625m		65	125	
2007/08-1	Lab	LCS, RPD	10/16/2007	Pesticide	Trichloronate	n/a	=	19	%	EPA 625m		0	30	
2007/08-1	Lab	method blank	10/16/2007	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-CC	lab duplicate	10/16/2007	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-1	ME-CC	matrix spike dup, rec	10/16/2007	Pesticide	Trichloronate	n/a	=	73	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, rec	10/16/2007	Pesticide	Trichloronate	n/a	=	74	%	EPA 625m		65	125	
2007/08-1	ME-CC	matrix spike, RPD	10/16/2007	Pesticide	Trichloronate	n/a	=	1	%	EPA 625m		0	30	
2007/08-1	ME-VR2	field blank	10/16/2007	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-1	ME-VR2	lab duplicate	10/16/2007	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	A-1	field duplicate	1/4/2008	Anion	Bromide	n/a	=	7.3	mg/L	EPA 300.0	0.001			
2007/08-2	Lab	LCS dup, rec	1/4/2008	Anion	Bromide	n/a	=	100	%	EPA 300.0		70	130	
2007/08-2	Lab	LCS, rec	1/4/2008	Anion	Bromide	n/a	=	100	%	EPA 300.0		70	130	
2007/08-2	Lab	LCS, RPD	1/4/2008	Anion	Bromide	n/a	=	0	%	EPA 300.0		0	30	
2007/08-2	Lab	method blank	1/4/2008	Anion	Bromide	n/a	<	0.001	mg/L	EPA 300.0	0.001		0.001	
2007/08-2	ME-CC	lab duplicate	1/4/2008	Anion	Bromide	n/a	=	0.2	mg/L	EPA 300.0	0.001		30	
2007/08-2	ME-CC	matrix spike dup, rec	1/4/2008	Anion	Bromide	n/a	=	100	%	EPA 300.0		70	130	
2007/08-2	ME-CC	matrix spike, rec	1/4/2008	Anion	Bromide	n/a	=	96	%	EPA 300.0		70	130	
2007/08-2	ME-CC	matrix spike, RPD	1/4/2008	Anion	Bromide	n/a	=	4	%	EPA 300.0		0	30	
2007/08-2	A-1	field duplicate	1/8/2008	Anion	Chloride	n/a	=	129.89	mg/L	EPA 300.0	0.01			
2007/08-2	Lab	LCS dup, rec	1/8/2008	Anion	Chloride	n/a	=	97	%	EPA 300.0		70	130	
2007/08-2	Lab	LCS, rec	1/8/2008	Anion	Chloride	n/a	=	99	%	EPA 300.0		70	130	
2007/08-2	Lab	LCS, RPD	1/8/2008	Anion	Chloride	n/a	=	2	%	EPA 300.0		0	30	
2007/08-2	Lab	method blank	1/8/2008	Anion	Chloride	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2007/08-2	ME-CC	lab duplicate	1/8/2008	Anion	Chloride	n/a	=	61.69	mg/L	EPA 300.0	0.01		30	
2007/08-2	ME-CC	matrix spike dup, rec	1/8/2008	Anion	Chloride	n/a	=	92	%	EPA 300.0		70	130	
2007/08-2	ME-CC	matrix spike, rec	1/8/2008	Anion	Chloride	n/a	=	90	%	EPA 300.0		70	130	
2007/08-2	ME-CC	matrix spike, RPD	1/8/2008	Anion	Chloride	n/a	=	2	%	EPA 300.0		0	30	
2007/08-2	A-1	field duplicate	12/28/2007	Anion	Perchlorate	n/a	<	2	µg/L	EPA 314.0	2			
2007/08-2	Lab	LCS dup, rec	12/28/2007	Anion	Perchlorate	n/a	=	96	%	EPA 314.0		85	115	
2007/08-2	Lab	LCS, rec	12/28/2007	Anion	Perchlorate	n/a	=	95	%	EPA 314.0		85	115	
2007/08-2	Lab	LCS, RPD	12/28/2007	Anion	Perchlorate	n/a	=	1	%	EPA 314.0		0	15	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	method blank	12/28/2007	Anion	Perchlorate	n/a	<	2	µg/L	EPA 314.0	2		2	
2007/08-2	ME-VR2	field duplicate	12/28/2007	Anion	Perchlorate	n/a	<	2	µg/L	EPA 314.0	2			
2007/08-2	A-1	field duplicate	12/18/2007	Bacteriological	E. Coli	n/a	=	9804	MPN/100 mL	MMO-MUG	10			
2007/08-2	ME-CC	field blank	12/18/2007	Bacteriological	E. Coli	n/a	<	10	MPN/100 mL	MMO-MUG	10		10	
2007/08-2	ME-VR2	field duplicate	12/18/2007	Bacteriological	E. Coli	n/a	=	4611	MPN/100 mL	MMO-MUG	10			
2007/08-2	A-1	field duplicate	12/18/2007	Bacteriological	Enterococcus	n/a	=	5910	MPN/100 mL	Enterolert	10			
2007/08-2	ME-CC	field blank	12/18/2007	Bacteriological	Enterococcus	n/a	<	10	MPN/100 mL	Enterolert	10		10	
2007/08-2	ME-VR2	field duplicate	12/18/2007	Bacteriological	Enterococcus	n/a	=	6240	MPN/100 mL	Enterolert	10			
2007/08-2	A-1	field duplicate	12/18/2007	Bacteriological	Fecal Coliform	n/a	=	7000	MPN/100 mL	SM 9221 E	2			
2007/08-2	ME-CC	field blank	12/18/2007	Bacteriological	Fecal Coliform	n/a	<	2	MPN/100 mL	SM 9221 E	2		2	
2007/08-2	ME-VR2	field duplicate	12/18/2007	Bacteriological	Fecal Coliform	n/a	=	2800	MPN/100 mL	SM 9221 E	2			
2007/08-2	A-1	field duplicate	12/18/2007	Bacteriological	Total Coliform	n/a	=	241920	MPN/100 mL	MMO-MUG	10			
2007/08-2	ME-CC	field blank	12/18/2007	Bacteriological	Total Coliform	n/a	<	10	MPN/100 mL	MMO-MUG	10		10	
2007/08-2	ME-VR2	field duplicate	12/18/2007	Bacteriological	Total Coliform	n/a	=	198630	MPN/100 mL	MMO-MUG	10			
2007/08-2	A-1	field duplicate	12/21/2007	Conventional	BOD	n/a	<	2	mg/L	EPA 405.1	2			
2007/08-2	Lab	method blank	12/21/2007	Conventional	BOD	n/a	<	2	mg/L	EPA 405.1	2		2	
2007/08-2	ME-VR2	lab duplicate	12/21/2007	Conventional	BOD	n/a	=	5	mg/L	EPA 405.1	2		30	
2007/08-2	A-1	field duplicate	1/7/2008	Conventional	Conductivity	n/a	=	3310	µmhos/cm	SM 2510	0.001			
2007/08-2	ME-CC	lab duplicate	12/20/2007	Conventional	Conductivity	n/a	=	795	µmhos/cm	SM 2510	0.001		30	
2007/08-2	ME-VR2	field duplicate	12/20/2007	Conventional	Conductivity	n/a	=	1129	µmhos/cm	SM 2510	0.001			
2007/08-2	A-1	field duplicate	1/7/2008	Conventional	Hardness as CaCO3	Total	=	554.8	mg/L	SM 2340 B	1			
2007/08-2	ME-CC	lab duplicate	1/7/2008	Conventional	Hardness as CaCO3	Total	=	92.4	mg/L	SM 2340 B	1		30	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Conventional	Hardness as CaCO3	Total	=	253.8	mg/L	SM 2340 B	1		30	
2007/08-2	A-1	field duplicate	1/7/2008	Conventional	pH	n/a	=	7.2	pH Units	SM 4500 H+	0.1			
2007/08-2	ME-CC	lab duplicate	1/2/2008	Conventional	pH	n/a	=	7.9	pH Units	SM 4500 H+	0.1		30	
2007/08-2	ME-VR2	field duplicate	12/20/2007	Conventional	pH	n/a	=	7.9	pH Units	SM 4500 H+	0.1			
2007/08-2	A-1	field duplicate	12/26/2007	Conventional	Total Dissolved Solids	n/a	=	2360	mg/L	SM 2540 C	0.1			
2007/08-2	Lab	LCS dup, rec	12/26/2007	Conventional	Total Dissolved Solids	n/a	=	108	%	SM 2540 C		70	130	
2007/08-2	Lab	LCS, rec	12/26/2007	Conventional	Total Dissolved Solids	n/a	=	112	%	SM 2540 C		70	130	
2007/08-2	Lab	method blank	12/26/2007	Conventional	Total Dissolved Solids	n/a	<	0.1	mg/L	SM 2540 C	0.1		0.1	
2007/08-2	Lab	LCS, RPD	1/2/2008	Conventional	Total Dissolved Solids	n/a	=	4	%	SM 2540 C		0	30	
2007/08-2	ME-CC	lab duplicate	12/26/2007	Conventional	Total Dissolved Solids	n/a	=	352	mg/L	SM 2540 C	0.1		30	
2007/08-2	A-1	field duplicate	1/10/2008	Conventional	Total Organic Carbon	n/a	=	9.9	mg/L	EPA 415.1	0.1			
2007/08-2	Lab	LCS dup, rec	1/2/2008	Conventional	Total Organic Carbon	n/a	=	86	%	EPA 415.1		50	150	
2007/08-2	Lab	LCS, rec	1/2/2008	Conventional	Total Organic Carbon	n/a	=	84	%	EPA 415.1		50	150	
2007/08-2	Lab	LCS, RPD	1/2/2008	Conventional	Total Organic Carbon	n/a	=	2	%	EPA 415.1		0	30	
2007/08-2	Lab	method blank	1/2/2008	Conventional	Total Organic Carbon	n/a	<	0.1	mg/L	EPA 415.1	0.1		0.1	
2007/08-2	ME-CC	lab duplicate	1/4/2008	Conventional	Total Organic Carbon	n/a	=	12	mg/L	EPA 415.1	0.1		30	
2007/08-2	ME-CC	matrix spike dup, rec	1/4/2008	Conventional	Total Organic Carbon	n/a	=	86	%	EPA 415.1		50	150	
2007/08-2	ME-CC	matrix spike, rec	1/4/2008	Conventional	Total Organic Carbon	n/a	=	97	%	EPA 415.1		50	150	
2007/08-2	ME-CC	matrix spike, RPD	1/4/2008	Conventional	Total Organic Carbon	n/a	=	12	%	EPA 415.1		0	30	
2007/08-2	A-1	field duplicate	12/26/2007	Conventional	Total Suspended Solids	n/a	=	236.7	mg/L	SM 2540 D	0.5			
2007/08-2	A-1	lab duplicate	12/26/2007	Conventional	Total Suspended Solids	n/a	=	178	mg/L	SM 2540 D	0.5		30	
2007/08-2	Lab	method blank	12/26/2007	Conventional	Total Suspended Solids	n/a	<	0.5	mg/L	SM 2540 D	0.5		0.5	
2007/08-2	A-1	field duplicate	12/20/2007	Conventional	Turbidity	n/a	=	214	NTU	EPA 180.1	1			
2007/08-2	Lab	method blank	12/20/2007	Conventional	Turbidity	n/a	<	1	NTU	EPA 180.1	1		1	
2007/08-2	ME-CC	lab duplicate	12/20/2007	Conventional	Turbidity	n/a	=	746	NTU	EPA 180.1	1		30	
2007/08-2	A-1	field duplicate	1/7/2008	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1			
2007/08-2	Lab	LCS dup, rec	1/11/2008	Hydrocarbon	Oil and Grease	n/a	=	93	%	EPA 1664A		70	130	
2007/08-2	Lab	LCS, rec	1/11/2008	Hydrocarbon	Oil and Grease	n/a	=	99	%	EPA 1664A		70	130	
2007/08-2	Lab	LCS, RPD	1/11/2008	Hydrocarbon	Oil and Grease	n/a	=	6	%	EPA 1664A		0	30	
2007/08-2	Lab	method blank	1/11/2008	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1		1	
2007/08-2	ME-VR2	field duplicate	1/11/2008	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1			
2007/08-2	A-1	field duplicate	1/7/2008	Hydrocarbon	TRPH	n/a	=	1	mg/L	EPA 1664	1			EST
2007/08-2	Lab	LCS dup, rec	1/11/2008	Hydrocarbon	TRPH	n/a	=	96	%	EPA 1664		70	130	
2007/08-2	Lab	LCS, rec	1/11/2008	Hydrocarbon	TRPH	n/a	=	97	%	EPA 1664		70	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	LCS, RPD	1/11/2008	Hydrocarbon	TRPH	n/a	=	1	%	EPA 1664		0	30	
2007/08-2	Lab	method blank	1/11/2008	Hydrocarbon	TRPH	n/a	<	1	mg/L	EPA 1664	1		1	
2007/08-2	ME-VR2	field duplicate	1/11/2008	Hydrocarbon	TRPH	n/a	<	1	mg/L	EPA 1664	1			
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Aluminum	Dissolved	<	5	µg/L	EPA 200.8m	5			
2007/08-2	Lab	method blank	1/7/2008	Metal	Aluminum	Dissolved	<	5	µg/L	EPA 200.8m	5		5	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Aluminum	Dissolved	<	5	µg/L	EPA 200.8m	5		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/7/2008	Metal	Aluminum	Dissolved	=	103	%	EPA 200.8m		50	140	
2007/08-2	ME-SCR	matrix spike, rec	1/7/2008	Metal	Aluminum	Dissolved	=	104	%	EPA 200.8m		50	140	
2007/08-2	ME-SCR	matrix spike, RPD	1/7/2008	Metal	Aluminum	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Aluminum	Total	=	1432	µg/L	EPA 200.8m	5			
2007/08-2	Lab	method blank	1/7/2008	Metal	Aluminum	Total	<	5	µg/L	EPA 200.8m	5		5	
2007/08-2	ME-CC	lab duplicate	1/7/2008	Metal	Aluminum	Total	=	2342	µg/L	EPA 200.8m	5		30	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Aluminum	Total	=	14790	µg/L	EPA 200.8m	5		30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Arsenic	Dissolved	=	6.4	µg/L	EPA 200.8m	0.2			
2007/08-2	Lab	method blank	1/7/2008	Metal	Arsenic	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Arsenic	Dissolved	=	0.9	µg/L	EPA 200.8m	0.2		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/7/2008	Metal	Arsenic	Dissolved	=	107	%	EPA 200.8m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/7/2008	Metal	Arsenic	Dissolved	=	109	%	EPA 200.8m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/7/2008	Metal	Arsenic	Dissolved	=	2	%	EPA 200.8m		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Arsenic	Total	=	8.1	µg/L	EPA 200.8m	0.2			
2007/08-2	Lab	method blank	1/7/2008	Metal	Arsenic	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-2	ME-CC	lab duplicate	1/7/2008	Metal	Arsenic	Total	=	4.7	µg/L	EPA 200.8m	0.2		30	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Arsenic	Total	=	18.8	µg/L	EPA 200.8m	0.2		30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Cadmium	Dissolved	=	0.5	µg/L	EPA 200.8m	0.2			
2007/08-2	Lab	method blank	1/7/2008	Metal	Cadmium	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Cadmium	Dissolved	=	0.2	µg/L	EPA 200.8m	0.2		30	EST
2007/08-2	ME-SCR	matrix spike dup, rec	1/7/2008	Metal	Cadmium	Dissolved	=	106	%	EPA 200.8m		75	130	
2007/08-2	ME-SCR	matrix spike, rec	1/7/2008	Metal	Cadmium	Dissolved	=	105	%	EPA 200.8m		75	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/7/2008	Metal	Cadmium	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Cadmium	Total	=	1.1	µg/L	EPA 200.8m	0.2			
2007/08-2	Lab	method blank	1/7/2008	Metal	Cadmium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-2	ME-CC	lab duplicate	1/7/2008	Metal	Cadmium	Total	=	1.2	µg/L	EPA 200.8m	0.2		30	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Cadmium	Total	=	8.1	µg/L	EPA 200.8m	0.2		30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Chromium	Dissolved	=	2.8	µg/L	EPA 200.8m	0.1			
2007/08-2	Lab	method blank	1/7/2008	Metal	Chromium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Chromium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/7/2008	Metal	Chromium	Dissolved	=	101	%	EPA 200.8m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/7/2008	Metal	Chromium	Dissolved	=	101	%	EPA 200.8m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/7/2008	Metal	Chromium	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Chromium	Total	=	5.7	µg/L	EPA 200.8m	0.1			
2007/08-2	Lab	method blank	1/7/2008	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-2	ME-CC	lab duplicate	1/7/2008	Metal	Chromium	Total	=	4.5	µg/L	EPA 200.8m	0.1		30	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Chromium	Total	=	10.6	µg/L	EPA 200.8m	0.1		30	
2007/08-2	A-1	field duplicate	1/8/2008	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr D	5			
2007/08-2	Lab	LCS, rec	1/2/2008	Metal	Chromium VI	Total	=	102	%	SM 3500-Cr D		70	130	
2007/08-2	Lab	LCS dup, rec	1/8/2008	Metal	Chromium VI	Total	=	104	%	SM 3500-Cr D		70	130	
2007/08-2	Lab	LCS, RPD	1/8/2008	Metal	Chromium VI	Total	=	2	%	SM 3500-Cr D		0	30	
2007/08-2	Lab	method blank	1/8/2008	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr D	5		5	
2007/08-2	ME-CC	lab duplicate	1/8/2008	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr D	5		30	
2007/08-2	ME-CC	matrix spike dup, rec	1/8/2008	Metal	Chromium VI	Total	=	93	%	SM 3500-Cr D		70	130	
2007/08-2	ME-CC	matrix spike, rec	1/8/2008	Metal	Chromium VI	Total	=	93	%	SM 3500-Cr D		70	130	
2007/08-2	ME-CC	matrix spike, RPD	1/8/2008	Metal	Chromium VI	Total	=	0	%	SM 3500-Cr D		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Copper	Dissolved	=	6.3	µg/L	EPA 200.8m	0.4			
2007/08-2	Lab	method blank	1/7/2008	Metal	Copper	Dissolved	<	0.4	µg/L	EPA 200.8m	0.4		0.4	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Copper	Dissolved	=	2.7	µg/L	EPA 200.8m	0.4		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/7/2008	Metal	Copper	Dissolved	=	94	%	EPA 200.8m		70	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	matrix spike, rec	1/7/2008	Metal	Copper	Dissolved	=	95	%	EPA 200.8m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/7/2008	Metal	Copper	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Copper	Total	=	18.2	µg/L	EPA 200.8m	0.4			
2007/08-2	Lab	method blank	1/7/2008	Metal	Copper	Total	<	0.4	µg/L	EPA 200.8m	0.4		0.4	
2007/08-2	ME-CC	lab duplicate	1/7/2008	Metal	Copper	Total	=	26.5	µg/L	EPA 200.8m	0.4		30	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Copper	Total	=	111.8	µg/L	EPA 200.8m	0.4		30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05			
2007/08-2	Lab	method blank	1/7/2008	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/7/2008	Metal	Lead	Dissolved	=	92	%	EPA 200.8m		65	135	
2007/08-2	ME-SCR	matrix spike, rec	1/7/2008	Metal	Lead	Dissolved	=	92	%	EPA 200.8m		65	135	
2007/08-2	ME-SCR	matrix spike, RPD	1/7/2008	Metal	Lead	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Lead	Total	=	4.41	µg/L	EPA 200.8m	0.05			
2007/08-2	Lab	method blank	1/7/2008	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2007/08-2	ME-CC	lab duplicate	1/7/2008	Metal	Lead	Total	=	10.36	µg/L	EPA 200.8m	0.05		30	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Lead	Total	=	43.22	µg/L	EPA 200.8m	0.05		30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Mercury	Dissolved	=	3	ng/L	EPA 1631Em	0.5			
2007/08-2	Lab	method blank	1/7/2008	Metal	Mercury	Dissolved	<	0.5	ng/L	EPA 1631Em	0.5		0.5	
2007/08-2	ME-CC	field blank	1/7/2008	Metal	Mercury	Dissolved	<	0.5	ng/L	EPA 1631Em	0.5		0.5	
2007/08-2	ME-VR2	field duplicate	1/7/2008	Metal	Mercury	Dissolved	=	1.2	ng/L	EPA 1631Em	0.5			
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Mercury	Total	=	19.4	ng/L	EPA 1631Em	0.5			
2007/08-2	Lab	method blank	1/7/2008	Metal	Mercury	Total	<	0.5	ng/L	EPA 1631Em	0.5		0.5	
2007/08-2	ME-CC	field blank	1/7/2008	Metal	Mercury	Total	=	4	ng/L	EPA 1631Em	0.5		0.5	
2007/08-2	ME-CC	matrix spike dup, rec	1/7/2008	Metal	Mercury	Total	=	125	%	EPA 1631Em		60	140	
2007/08-2	ME-CC	matrix spike, rec	1/7/2008	Metal	Mercury	Total	=	113	%	EPA 1631Em		60	140	
2007/08-2	ME-CC	matrix spike, RPD	1/7/2008	Metal	Mercury	Total	=	10	%	EPA 1631Em		0	30	
2007/08-2	ME-VR2	field duplicate	1/7/2008	Metal	Mercury	Total	=	7.1	ng/L	EPA 1631Em	0.5			
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Nickel	Dissolved	=	15.9	µg/L	EPA 200.8m	0.2			
2007/08-2	Lab	method blank	1/7/2008	Metal	Nickel	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Nickel	Dissolved	=	4.2	µg/L	EPA 200.8m	0.2		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/7/2008	Metal	Nickel	Dissolved	=	95	%	EPA 200.8m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/7/2008	Metal	Nickel	Dissolved	=	96	%	EPA 200.8m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/7/2008	Metal	Nickel	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Nickel	Total	=	21.6	µg/L	EPA 200.8m	0.2			
2007/08-2	Lab	method blank	1/7/2008	Metal	Nickel	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-2	ME-CC	lab duplicate	1/7/2008	Metal	Nickel	Total	=	15.8	µg/L	EPA 200.8m	0.2		30	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Nickel	Total	=	92.7	µg/L	EPA 200.8m	0.2		30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Selenium	Dissolved	=	5	µg/L	EPA 200.8m	0.2			
2007/08-2	Lab	method blank	1/7/2008	Metal	Selenium	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Selenium	Dissolved	=	6.8	µg/L	EPA 200.8m	0.2		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/7/2008	Metal	Selenium	Dissolved	=	103	%	EPA 200.8m		60	150	
2007/08-2	ME-SCR	matrix spike, rec	1/7/2008	Metal	Selenium	Dissolved	=	101	%	EPA 200.8m		60	150	
2007/08-2	ME-SCR	matrix spike, RPD	1/7/2008	Metal	Selenium	Dissolved	=	2	%	EPA 200.8m		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Selenium	Total	=	4.7	µg/L	EPA 200.8m	0.2			
2007/08-2	Lab	method blank	1/7/2008	Metal	Selenium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-2	ME-CC	lab duplicate	1/7/2008	Metal	Selenium	Total	=	1.5	µg/L	EPA 200.8m	0.2		30	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Selenium	Total	=	7.8	µg/L	EPA 200.8m	0.2		30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Silver	Dissolved	<	0.5	µg/L	EPA 200.8m	0.5			
2007/08-2	Lab	method blank	1/7/2008	Metal	Silver	Dissolved	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Silver	Dissolved	<	0.5	µg/L	EPA 200.8m	0.5		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/7/2008	Metal	Silver	Dissolved	=	104	%	EPA 200.8m		50	155	
2007/08-2	ME-SCR	matrix spike, rec	1/7/2008	Metal	Silver	Dissolved	=	110	%	EPA 200.8m		50	155	
2007/08-2	ME-SCR	matrix spike, RPD	1/7/2008	Metal	Silver	Dissolved	=	6	%	EPA 200.8m		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5			
2007/08-2	Lab	method blank	1/7/2008	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2007/08-2	ME-CC	lab duplicate	1/7/2008	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5		30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5			
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1			
2007/08-2	Lab	method blank	1/7/2008	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/7/2008	Metal	Thallium	Dissolved	=	94	%	EPA 200.8m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/7/2008	Metal	Thallium	Dissolved	=	94	%	EPA 200.8m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/7/2008	Metal	Thallium	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1			
2007/08-2	Lab	method blank	1/7/2008	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-2	ME-CC	lab duplicate	1/7/2008	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		30	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Thallium	Total	=	0.2	µg/L	EPA 200.8m	0.1		30	EST
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Zinc	Dissolved	=	6.3	µg/L	EPA 200.8m	0.1			
2007/08-2	Lab	method blank	1/7/2008	Metal	Zinc	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Zinc	Dissolved	=	0.7	µg/L	EPA 200.8m	0.1		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/7/2008	Metal	Zinc	Dissolved	=	101	%	EPA 200.8m		50	150	
2007/08-2	ME-SCR	matrix spike, rec	1/7/2008	Metal	Zinc	Dissolved	=	101	%	EPA 200.8m		50	150	
2007/08-2	ME-SCR	matrix spike, RPD	1/7/2008	Metal	Zinc	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-2	A-1	field duplicate	1/7/2008	Metal	Zinc	Total	=	44.2	µg/L	EPA 200.8m	0.1			
2007/08-2	Lab	method blank	1/7/2008	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-2	ME-CC	lab duplicate	1/7/2008	Metal	Zinc	Total	=	76.7	µg/L	EPA 200.8m	0.1		30	
2007/08-2	ME-SCR	lab duplicate	1/7/2008	Metal	Zinc	Total	=	176.8	µg/L	EPA 200.8m	0.1		30	
2007/08-2	A-1	field duplicate	1/7/2008	Nutrient	Ammonia as N	n/a	=	0.2	mg/L	SM 4500-NH3 F	0.01			
2007/08-2	Lab	LCS dup, rec	1/2/2008	Nutrient	Ammonia as N	n/a	=	100	%	SM 4500-NH3 F		70	130	
2007/08-2	Lab	LCS, rec	1/2/2008	Nutrient	Ammonia as N	n/a	=	100	%	SM 4500-NH3 F		70	130	
2007/08-2	Lab	LCS, RPD	1/2/2008	Nutrient	Ammonia as N	n/a	=	0	%	SM 4500-NH3 F		0	30	
2007/08-2	Lab	method blank	1/2/2008	Nutrient	Ammonia as N	n/a	<	0.01	mg/L	SM 4500-NH3 F	0.01		0.01	
2007/08-2	ME-CC	lab duplicate	1/2/2008	Nutrient	Ammonia as N	n/a	=	0.55	mg/L	SM 4500-NH3 F	0.01		30	
2007/08-2	ME-CC	matrix spike dup, rec	1/2/2008	Nutrient	Ammonia as N	n/a	=	94	%	SM 4500-NH3 F		70	130	
2007/08-2	ME-CC	matrix spike, rec	1/2/2008	Nutrient	Ammonia as N	n/a	=	98	%	SM 4500-NH3 F		70	130	
2007/08-2	ME-CC	matrix spike, RPD	1/2/2008	Nutrient	Ammonia as N	n/a	=	4	%	SM 4500-NH3 F		0	30	
2007/08-2	ME-VR2	field duplicate	1/2/2008	Nutrient	Ammonia as N	n/a	=	0.06	mg/L	SM 4500-NH3 F	0.01			
2007/08-2	A-1	field duplicate	12/20/2007	Nutrient	Nitrate as N	n/a	=	16.94	mg/L	EPA 300.0	0.01			
2007/08-2	Lab	LCS dup, rec	12/20/2007	Nutrient	Nitrate as N	n/a	=	106	%	EPA 300.0		70	130	
2007/08-2	Lab	LCS, rec	12/20/2007	Nutrient	Nitrate as N	n/a	=	106	%	EPA 300.0		70	130	
2007/08-2	Lab	LCS, RPD	12/20/2007	Nutrient	Nitrate as N	n/a	=	0	%	EPA 300.0		0	30	
2007/08-2	Lab	method blank	12/20/2007	Nutrient	Nitrate as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2007/08-2	ME-CC	lab duplicate	12/20/2007	Nutrient	Nitrate as N	n/a	=	1.58	mg/L	EPA 300.0	0.01		30	
2007/08-2	ME-CC	matrix spike dup, rec	12/20/2007	Nutrient	Nitrate as N	n/a	=	145	%	EPA 300.0		70	130	
2007/08-2	ME-CC	matrix spike, rec	12/20/2007	Nutrient	Nitrate as N	n/a	=	141	%	EPA 300.0		70	130	
2007/08-2	ME-CC	matrix spike, RPD	12/20/2007	Nutrient	Nitrate as N	n/a	=	3	%	EPA 300.0		0	30	
2007/08-2	A-1	field duplicate	12/19/2007	Nutrient	Nitrite as N	n/a	=	1.42	mg/L	EPA 300.0	0.01			
2007/08-2	Lab	LCS dup, rec	12/19/2007	Nutrient	Nitrite as N	n/a	=	90	%	EPA 300.0		70	130	
2007/08-2	Lab	LCS, RPD	12/19/2007	Nutrient	Nitrite as N	n/a	=	0	%	EPA 300.0		0	30	
2007/08-2	Lab	method blank	12/19/2007	Nutrient	Nitrite as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2007/08-2	Lab	LCS, rec	1/2/2008	Nutrient	Nitrite as N	n/a	=	90	%	EPA 300.0		70	130	
2007/08-2	ME-CC	lab duplicate	12/19/2007	Nutrient	Nitrite as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		30	
2007/08-2	ME-CC	matrix spike dup, rec	12/19/2007	Nutrient	Nitrite as N	n/a	=	79	%	EPA 300.0		70	130	
2007/08-2	ME-CC	matrix spike, rec	12/19/2007	Nutrient	Nitrite as N	n/a	=	88	%	EPA 300.0		70	130	
2007/08-2	ME-CC	matrix spike, RPD	12/19/2007	Nutrient	Nitrite as N	n/a	=	11	%	EPA 300.0		0	30	
2007/08-2	A-1	field duplicate	12/19/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	0.1689	mg/L	EPA 300.0	0.0075			
2007/08-2	Lab	LCS dup, rec	12/19/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	92	%	EPA 300.0		70	130	
2007/08-2	Lab	LCS, RPD	12/19/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	20	%	EPA 300.0		0	30	
2007/08-2	Lab	method blank	12/19/2007	Nutrient	Orthophosphate as P (Diss)	n/a	<	0.0075	mg/L	EPA 300.0	0.0075		0.0075	
2007/08-2	Lab	LCS, rec	1/2/2008	Nutrient	Orthophosphate as P (Diss)	n/a	=	113	%	EPA 300.0		70	130	
2007/08-2	ME-CC	lab duplicate	12/19/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	0.9524	mg/L	EPA 300.0	0.0075		30	
2007/08-2	ME-CC	matrix spike dup, rec	12/19/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	87	%	EPA 300.0		70	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-CC	matrix spike, rec	12/19/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	88	%	EPA 300.0		70	130	
2007/08-2	ME-CC	matrix spike, RPD	12/19/2007	Nutrient	Orthophosphate as P (Diss)	n/a	=	1	%	EPA 300.0		0	30	
2007/08-2	A-1	field duplicate	1/2/2008	Nutrient	TKN	n/a	=	0.17	mg/L	EPA 351.1	0.05			
2007/08-2	Lab	LCS, rec	1/2/2008	Nutrient	TKN	n/a	=	96.8	%	EPA 351.1		80	120	
2007/08-2	Lab	method blank	1/2/2008	Nutrient	TKN	n/a	<	0.05	mg/L	EPA 351.1	0.05		0.05	
2007/08-2	ME-CC	lab duplicate	1/2/2008	Nutrient	TKN	n/a	=	0.2	mg/L	EPA 351.1	0.05		20	
2007/08-2	ME-SCR	matrix spike dup, rec	1/2/2008	Nutrient	TKN	n/a	=	92.2	%	EPA 351.1		80	120	
2007/08-2	ME-SCR	matrix spike, rec	1/2/2008	Nutrient	TKN	n/a	=	88.9	%	EPA 351.1		80	120	
2007/08-2	ME-SCR	matrix spike, RPD	1/2/2008	Nutrient	TKN	n/a	=	3.6	%	EPA 351.1		0	20	
2007/08-2	A-1	field duplicate	1/5/2008	Nutrient	Total Phosphorus	Dissolved	=	0.34	mg/L	SM 4500-P C	0.016			
2007/08-2	Lab	LCS dup, rec	1/2/2008	Nutrient	Total Phosphorus	Dissolved	=	103	%	SM 4500-P C		70	130	
2007/08-2	Lab	LCS, rec	1/2/2008	Nutrient	Total Phosphorus	Dissolved	=	103	%	SM 4500-P C		70	130	
2007/08-2	Lab	LCS, RPD	1/2/2008	Nutrient	Total Phosphorus	Dissolved	=	0	%	SM 4500-P C		0	30	
2007/08-2	Lab	method blank	1/2/2008	Nutrient	Total Phosphorus	Dissolved	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2007/08-2	ME-CC	lab duplicate	1/4/2008	Nutrient	Total Phosphorus	Dissolved	=	1.05	mg/L	SM 4500-P C	0.016		30	
2007/08-2	ME-CC	matrix spike dup, rec	1/4/2008	Nutrient	Total Phosphorus	Dissolved	=	88	%	SM 4500-P C		70	130	
2007/08-2	ME-CC	matrix spike, rec	1/4/2008	Nutrient	Total Phosphorus	Dissolved	=	96	%	SM 4500-P C		70	130	
2007/08-2	ME-CC	matrix spike, RPD	1/4/2008	Nutrient	Total Phosphorus	Dissolved	=	9	%	SM 4500-P C		0	30	
2007/08-2	A-1	field duplicate	1/5/2008	Nutrient	Total Phosphorus	Total	=	1.581	mg/L	SM 4500-P C	0.016			
2007/08-2	Lab	LCS dup, rec	1/2/2008	Nutrient	Total Phosphorus	Total	=	101	%	SM 4500-P C		70	130	
2007/08-2	Lab	LCS, rec	1/2/2008	Nutrient	Total Phosphorus	Total	=	102	%	SM 4500-P C		70	130	
2007/08-2	Lab	LCS, RPD	1/2/2008	Nutrient	Total Phosphorus	Total	=	1	%	SM 4500-P C		0	30	
2007/08-2	Lab	method blank	1/2/2008	Nutrient	Total Phosphorus	Total	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2007/08-2	ME-CC	lab duplicate	1/4/2008	Nutrient	Total Phosphorus	Total	=	3.521	mg/L	SM 4500-P C	0.016		30	
2007/08-2	ME-CC	matrix spike dup, rec	1/4/2008	Nutrient	Total Phosphorus	Total	=	117	%	SM 4500-P C		70	130	
2007/08-2	ME-CC	matrix spike, rec	1/4/2008	Nutrient	Total Phosphorus	Total	=	116	%	SM 4500-P C		70	130	
2007/08-2	ME-CC	matrix spike, RPD	1/4/2008	Nutrient	Total Phosphorus	Total	=	0	%	SM 4500-P C		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	72	%	EPA 625m		45	140	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	72	%	EPA 625m		45	140	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	0.015	µg/L	EPA 625m	0.01		30	EST
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	13	%	EPA 625m		45	140	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	31	%	EPA 625m		45	140	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	82	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2007/08-2	Lab	method blank	1/15/2008	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-2	A-1	srgt environ, rec	1/4/2008	Organic	1,2-Dichloroethane-d4	n/a	=	117	%	EPA 8260B		74	146	
2007/08-2	A-1	srgt environ, rec	1/4/2008	Organic	1,2-Dichloroethane-d4	n/a	=	115	%	EPA 8260B		74	146	
2007/08-2	Lab	srgt method blank, rec	1/4/2008	Organic	1,2-Dichloroethane-d4	n/a	=	114	%	EPA 8260B		74	146	
2007/08-2	W-3	srgt environ, rec	1/4/2008	Organic	1,2-Dichloroethane-d4	n/a	=	116	%	EPA 8260B		74	146	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2007/08-2	Lab	method blank	1/15/2008	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-2	A-1	srgt environ, rec	1/4/2008	Organic	1,4-Bromofluorobenzene	n/a	=	96	%	EPA 8260B		74	110	
2007/08-2	A-1	srgt environ, rec	1/4/2008	Organic	1,4-Bromofluorobenzene	n/a	=	97	%	EPA 8260B		74	110	
2007/08-2	Lab	srgt method blank, rec	1/4/2008	Organic	1,4-Bromofluorobenzene	n/a	=	98	%	EPA 8260B		74	110	
2007/08-2	W-3	srgt environ, rec	1/4/2008	Organic	1,4-Bromofluorobenzene	n/a	=	99	%	EPA 8260B		74	110	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	1,4-Dichlorobenzene	n/a	=	64	%	EPA 625m		45	140	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	1,4-Dichlorobenzene	n/a	=	62	%	EPA 625m		45	140	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	1,4-Dichlorobenzene	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	1,4-Dichlorobenzene	n/a	=	11	%	EPA 625m		45	140	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	1,4-Dichlorobenzene	n/a	=	25	%	EPA 625m		45	140	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	1,4-Dichlorobenzene	n/a	=	78	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	1-Methylnaphthalene	n/a	=	0.0022	µg/L	EPA 625m	0.001			EST
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	1-Methylnaphthalene	n/a	=	82	%	EPA 625m		50	120	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	1-Methylnaphthalene	n/a	=	85	%	EPA 625m		50	120	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	1-Methylnaphthalene	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	1-Methylnaphthalene	n/a	=	0.0017	µg/L	EPA 625m	0.001		0.001	EST
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	1-Methylnaphthalene	n/a	=	0.0821	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	1-Methylnaphthalene	n/a	=	17	%	EPA 625m		50	120	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	1-Methylnaphthalene	n/a	=	55	%	EPA 625m		50	120	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	1-Methylnaphthalene	n/a	=	106	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	1-Methylphenanthrene	n/a	=	0.0131	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	1-Methylphenanthrene	n/a	=	94	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	1-Methylphenanthrene	n/a	=	102	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	1-Methylphenanthrene	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	1-Methylphenanthrene	n/a	=	0.0103	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	1-Methylphenanthrene	n/a	=	0.1302	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	1-Methylphenanthrene	n/a	=	14	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	1-Methylphenanthrene	n/a	=	64	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	1-Methylphenanthrene	n/a	=	128	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	78	%	EPA 625m		45	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	82	%	EPA 625m		45	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	0.0643	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	19	%	EPA 625m		45	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	59	%	EPA 625m		45	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	103	%	EPA 625m		0	30	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	126	%	EPA 625m		40	130	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	125	%	EPA 625m		40	130	
2007/08-2	Lab	srgt LCS dup, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	91	%	EPA 625m		40	130	
2007/08-2	Lab	srgt LCS, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	96	%	EPA 625m		40	130	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	125	%	EPA 625m		40	130	
2007/08-2	ME-CC	srgt environ, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	124	%	EPA 625m		40	130	
2007/08-2	ME-CC	srgt field blank, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	127	%	EPA 625m		40	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	37	%	EPA 625m		40	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	56	%	EPA 625m		40	130	
2007/08-2	ME-SCR	srgt matrix spike dup, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	26	%	EPA 625m		40	130	
2007/08-2	ME-SCR	srgt matrix spike, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	64	%	EPA 625m		40	130	
2007/08-2	ME-VR2	srgt environ, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	125	%	EPA 625m		40	130	
2007/08-2	W-3	srgt environ, rec	1/15/2008	Organic	2,4,6-Tribromophenol	n/a	=	126	%	EPA 625m		40	130	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	srgt environ, rec	12/27/2007	Organic	2,4-Dichlorophenylacetic acid	n/a	=	98	%	EPA 8151A		0	123	
2007/08-2	A-1	srgt environ, rec	12/27/2007	Organic	2,4-Dichlorophenylacetic acid	n/a	=	120	%	EPA 8151A		0	123	
2007/08-2	Lab	srgt method blank, rec	12/27/2007	Organic	2,4-Dichlorophenylacetic acid	n/a	=	107	%	EPA 8151A		0	123	
2007/08-2	ME-CC	srgt environ, rec	12/27/2007	Organic	2,4-Dichlorophenylacetic acid	n/a	=	103	%	EPA 8151A		0	123	
2007/08-2	ME-SCR	srgt environ, rec	12/27/2007	Organic	2,4-Dichlorophenylacetic acid	n/a	=	117	%	EPA 8151A		0	123	
2007/08-2	W-3	srgt environ, rec	12/27/2007	Organic	2,4-Dichlorophenylacetic acid	n/a	=	122	%	EPA 8151A		0	123	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2007/08-2	Lab	method blank	1/15/2008	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2007/08-2	Lab	method blank	1/15/2008	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	2,4-Dinitrotoluene	n/a	=	98	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	2,4-Dinitrotoluene	n/a	=	109	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	2,4-Dinitrotoluene	n/a	=	11	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	2,4-Dinitrotoluene	n/a	=	42	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	2,4-Dinitrotoluene	n/a	=	73	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	2,4-Dinitrotoluene	n/a	=	54	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	0.0129	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	73	%	EPA 625m		55	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	80	%	EPA 625m		55	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	9	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	0.0022	µg/L	EPA 625m	0.001		0.001	EST
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	0.1183	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	9	%	EPA 625m		55	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	53	%	EPA 625m		55	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	142	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	2-Chlorophenol	n/a	=	37	%	EPA 625m		35	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	2-Chlorophenol	n/a	=	39	%	EPA 625m		35	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	2-Chlorophenol	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	2-Chlorophenol	n/a	=	6	%	EPA 625m		35	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	2-Chlorophenol	n/a	=	15	%	EPA 625m		35	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	2-Chlorophenol	n/a	=	86	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2007/08-2	Lab	method blank	1/15/2008	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2-Methylnaphthalene	n/a	=	0.0044	µg/L	EPA 625m	0.001			EST
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	2-Methylnaphthalene	n/a	=	96	%	EPA 625m		50	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	2-Methylnaphthalene	n/a	=	89	%	EPA 625m		50	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	2-Methylnaphthalene	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2-Methylnaphthalene	n/a	=	0.0017	µg/L	EPA 625m	0.001		0.001	EST
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2-Methylnaphthalene	n/a	=	0.0687	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	2-Methylnaphthalene	n/a	=	19	%	EPA 625m		50	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	2-Methylnaphthalene	n/a	=	54	%	EPA 625m		50	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	2-Methylnaphthalene	n/a	=	96	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2007/08-2	Lab	method blank	1/15/2008	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	4-Chloro-3-methylphenol	n/a	=	43	%	EPA 625m		30	150	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	4-Chloro-3-methylphenol	n/a	=	46	%	EPA 625m		30	150	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	4-Chloro-3-methylphenol	n/a	=	6	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	4-Chloro-3-methylphenol	n/a	=	10	%	EPA 625m		30	150	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	4-Chloro-3-methylphenol	n/a	=	23	%	EPA 625m		30	150	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	4-Chloro-3-methylphenol	n/a	=	79	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	4-Nitrophenol	n/a	=	50	%	EPA 625m		0	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	4-Nitrophenol	n/a	=	51	%	EPA 625m		0	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	4-Nitrophenol	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	4-Nitrophenol	n/a	=	0	%	EPA 625m		0	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	4-Nitrophenol	n/a	=	0	%	EPA 625m		0	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	4-Nitrophenol	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Acenaphthene	n/a	=	91	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Acenaphthene	n/a	=	91	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Acenaphthene	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Acenaphthene	n/a	=	0.017	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Acenaphthene	n/a	=	21	%	EPA 625m		70	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Acenaphthene	n/a	=	54	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Acenaphthene	n/a	=	88	%	EPA 625m		0	30	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	64	%	EPA 625m		50	130	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	69	%	EPA 625m		50	130	
2007/08-2	Lab	srgt LCS dup, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	66	%	EPA 625m		50	130	
2007/08-2	Lab	srgt LCS, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	72	%	EPA 625m		50	130	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	60	%	EPA 625m		50	130	
2007/08-2	ME-CC	srgt environ, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	61	%	EPA 625m		50	130	
2007/08-2	ME-CC	srgt field blank, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	55	%	EPA 625m		50	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	31	%	EPA 625m		50	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	14	%	EPA 625m		50	130	
2007/08-2	ME-SCR	srgt matrix spike dup, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	17	%	EPA 625m		50	130	
2007/08-2	ME-SCR	srgt matrix spike, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	38	%	EPA 625m		50	130	
2007/08-2	ME-VR2	srgt environ, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	61	%	EPA 625m		50	130	
2007/08-2	W-3	srgt environ, rec	1/15/2008	Organic	Acenaphthene-d10	n/a	=	62	%	EPA 625m		50	130	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Acenaphthylene	n/a	=	74	%	EPA 625m		60	120	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Acenaphthylene	n/a	=	80	%	EPA 625m		60	120	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Acenaphthylene	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Acenaphthylene	n/a	=	19	%	EPA 625m		60	120	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Acenaphthylene	n/a	=	45	%	EPA 625m		60	120	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Acenaphthylene	n/a	=	81	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Anthracene	n/a	=	0.0044	µg/L	EPA 625m	0.001			EST
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Anthracene	n/a	=	81	%	EPA 625m		60	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Anthracene	n/a	=	84	%	EPA 625m		60	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Anthracene	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Anthracene	n/a	<	0.005	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Anthracene	n/a	=	18	%	EPA 625m		60	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Anthracene	n/a	=	55	%	EPA 625m		60	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Anthracene	n/a	=	101	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Benzo(a)anthracene	n/a	=	0.005	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Benzo(a)anthracene	n/a	=	108	%	EPA 625m		70	140	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Benzo(a)anthracene	n/a	=	116	%	EPA 625m		70	140	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Benzo(a)anthracene	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Benzo(a)anthracene	n/a	=	0.0189	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Benzo(a)anthracene	n/a	=	0.049	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Benzo(a)anthracene	n/a	=	26	%	EPA 625m		70	140	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Benzo(a)anthracene	n/a	=	80	%	EPA 625m		70	140	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Benzo(a)anthracene	n/a	=	102	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Benzo(a)pyrene	n/a	=	102	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Benzo(a)pyrene	n/a	=	101	%	EPA 625m		70	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Benzo(a)pyrene	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Benzo(a)pyrene	n/a	=	0.0325	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Benzo(a)pyrene	n/a	=	0.0268	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Benzo(a)pyrene	n/a	=	32	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Benzo(a)pyrene	n/a	=	64	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Benzo(a)pyrene	n/a	=	67	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Benzo(b)fluoranthene	n/a	=	0.0014	µg/L	EPA 625m	0.001			EST
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Benzo(b)fluoranthene	n/a	=	98	%	EPA 625m		60	140	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Benzo(b)fluoranthene	n/a	=	107	%	EPA 625m		60	140	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Benzo(b)fluoranthene	n/a	=	9	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Benzo(b)fluoranthene	n/a	=	0.0553	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Benzo(b)fluoranthene	n/a	=	0.0422	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Benzo(b)fluoranthene	n/a	=	33	%	EPA 625m		60	140	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Benzo(b)fluoranthene	n/a	=	72	%	EPA 625m		60	140	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Benzo(b)fluoranthene	n/a	=	74	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Benzo(e)pyrene	n/a	=	0.0129	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Benzo(e)pyrene	n/a	=	102	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Benzo(e)pyrene	n/a	=	98	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Benzo(e)pyrene	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Benzo(e)pyrene	n/a	=	0.0427	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Benzo(e)pyrene	n/a	=	0.1128	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Benzo(e)pyrene	n/a	=	14	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Benzo(e)pyrene	n/a	=	50	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Benzo(e)pyrene	n/a	=	112	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Benzo(g,h,i)perylene	n/a	=	0.0086	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Benzo(g,h,i)perylene	n/a	=	79	%	EPA 625m		50	140	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Benzo(g,h,i)perylene	n/a	=	90	%	EPA 625m		50	140	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Benzo(g,h,i)perylene	n/a	=	13	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Benzo(g,h,i)perylene	n/a	=	0.0295	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Benzo(g,h,i)perylene	n/a	=	0.0739	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Benzo(g,h,i)perylene	n/a	=	10	%	EPA 625m		50	140	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Benzo(g,h,i)perylene	n/a	=	47	%	EPA 625m		50	140	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Benzo(g,h,i)perylene	n/a	=	130	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Benzo(k)fluoranthene	n/a	=	0.0019	µg/L	EPA 625m	0.001			EST
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Benzo(k)fluoranthene	n/a	=	93	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Benzo(k)fluoranthene	n/a	=	107	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Benzo(k)fluoranthene	n/a	=	14	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Benzo(k)fluoranthene	n/a	=	0.0311	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Benzo(k)fluoranthene	n/a	=	0.0199	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Benzo(k)fluoranthene	n/a	=	22	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Benzo(k)fluoranthene	n/a	=	58	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Benzo(k)fluoranthene	n/a	=	90	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Biphenyl	n/a	=	0.002	µg/L	EPA 625m	0.001			EST
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Biphenyl	n/a	=	73	%	EPA 625m		50	120	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Biphenyl	n/a	=	81	%	EPA 625m		50	120	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Biphenyl	n/a	=	10	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Biphenyl	n/a	=	0.0665	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Biphenyl	n/a	=	12	%	EPA 625m		50	120	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Biphenyl	n/a	=	50	%	EPA 625m		50	120	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Biphenyl	n/a	=	123	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	6.1263	µg/L	EPA 625m	0.1			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	102	%	EPA 625m		20	190	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	102	%	EPA 625m		20	190	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.9092	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	4.9674	µg/L	EPA 625m	0.1		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Butyl benzyl phthalate	n/a	=	0.0611	µg/L	EPA 625m	0.025			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Butyl benzyl phthalate	n/a	=	100	%	EPA 625m		65	160	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Butyl benzyl phthalate	n/a	=	97	%	EPA 625m		65	160	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Butyl benzyl phthalate	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Butyl benzyl phthalate	n/a	<	0.025	µg/L	EPA 625m	0.025		0.025	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Butyl benzyl phthalate	n/a	=	0.1976	µg/L	EPA 625m	0.025		0.025	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Butyl benzyl phthalate	n/a	=	0.1732	µg/L	EPA 625m	0.025		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Butyl benzyl phthalate	n/a	=	84	%	EPA 625m		65	160	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Butyl benzyl phthalate	n/a	=	89	%	EPA 625m		65	160	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Butyl benzyl phthalate	n/a	=	6	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Chrysene	n/a	=	0.0421	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Chrysene	n/a	=	102	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Chrysene	n/a	=	107	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Chrysene	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Chrysene	n/a	=	0.0607	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Chrysene	n/a	=	0.2068	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Chrysene	n/a	=	5	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Chrysene	n/a	=	67	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Chrysene	n/a	=	172	%	EPA 625m		0	30	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	110	%	EPA 625m		70	130	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	112	%	EPA 625m		70	130	
2007/08-2	Lab	srgt LCS dup, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	105	%	EPA 625m		70	130	
2007/08-2	Lab	srgt LCS, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	103	%	EPA 625m		70	130	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	114	%	EPA 625m		70	130	
2007/08-2	ME-CC	srgt environ, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	115	%	EPA 625m		70	130	
2007/08-2	ME-CC	srgt field blank, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	94	%	EPA 625m		70	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	27	%	EPA 625m		70	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	61	%	EPA 625m		70	130	
2007/08-2	ME-SCR	srgt matrix spike dup, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	21	%	EPA 625m		70	130	
2007/08-2	ME-SCR	srgt matrix spike, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	57	%	EPA 625m		70	130	
2007/08-2	ME-VR2	srgt environ, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	117	%	EPA 625m		70	130	
2007/08-2	W-3	srgt environ, rec	1/15/2008	Organic	Chrysene-d12	n/a	=	108	%	EPA 625m		70	130	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Dibenz(a,h)anthracene	n/a	=	96	%	EPA 625m		60	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Dibenz(a,h)anthracene	n/a	=	103	%	EPA 625m		60	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Dibenz(a,h)anthracene	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Dibenz(a,h)anthracene	n/a	=	0.0054	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Dibenz(a,h)anthracene	n/a	=	21	%	EPA 625m		60	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Dibenz(a,h)anthracene	n/a	=	63	%	EPA 625m		60	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Dibenz(a,h)anthracene	n/a	=	100	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Dibenzothiophene	n/a	=	0.0132	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Dibenzothiophene	n/a	=	87	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Dibenzothiophene	n/a	=	87	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Dibenzothiophene	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Dibenzothiophene	n/a	=	0.0096	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Dibenzothiophene	n/a	=	0.0722	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Dibenzothiophene	n/a	=	16	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Dibenzothiophene	n/a	=	52	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Dibenzothiophene	n/a	=	106	%	EPA 625m		0	30	
2007/08-2	A-1	srgt environ, rec	1/4/2008	Organic	Dibromofluoromethane	n/a	=	107	%	EPA 8260B		74	140	
2007/08-2	A-1	srgt environ, rec	1/4/2008	Organic	Dibromofluoromethane	n/a	=	107	%	EPA 8260B		74	140	
2007/08-2	Lab	srgt method blank, rec	1/4/2008	Organic	Dibromofluoromethane	n/a	=	105	%	EPA 8260B		74	140	
2007/08-2	W-3	srgt environ, rec	1/4/2008	Organic	Dibromofluoromethane	n/a	=	110	%	EPA 8260B		74	140	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Diethyl phthalate	n/a	=	0.4952	µg/L	EPA 625m	0.1			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Diethyl phthalate	n/a	=	75	%	EPA 625m		50	150	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Diethyl phthalate	n/a	=	69	%	EPA 625m		50	150	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Diethyl phthalate	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Diethyl phthalate	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Diethyl phthalate	n/a	=	1.4361	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Diethyl phthalate	n/a	=	0.3159	µg/L	EPA 625m	0.1		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Diethyl phthalate	n/a	=	24	%	EPA 625m		50	150	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Diethyl phthalate	n/a	=	62	%	EPA 625m		50	150	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Diethyl phthalate	n/a	=	88	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Dimethyl phthalate	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Dimethyl phthalate	n/a	=	72	%	EPA 625m		40	155	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Dimethyl phthalate	n/a	=	72	%	EPA 625m		40	155	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Dimethyl phthalate	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Dimethyl phthalate	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Dimethyl phthalate	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Dimethyl phthalate	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Dimethyl phthalate	n/a	=	32	%	EPA 625m		40	155	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Dimethyl phthalate	n/a	=	55	%	EPA 625m		40	155	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Dimethyl phthalate	n/a	=	53	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Di-n-butylphthalate	n/a	<	0.075	µg/L	EPA 625m	0.075			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Di-n-butylphthalate	n/a	=	87	%	EPA 625m		65	145	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Di-n-butylphthalate	n/a	=	87	%	EPA 625m		65	145	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Di-n-butylphthalate	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Di-n-butylphthalate	n/a	<	0.075	µg/L	EPA 625m	0.075		0.075	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Di-n-butylphthalate	n/a	=	0.0969	µg/L	EPA 625m	0.075		0.075	EST
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Di-n-butylphthalate	n/a	=	0.1064	µg/L	EPA 625m	0.075		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Di-n-butylphthalate	n/a	=	55	%	EPA 625m		65	145	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Di-n-butylphthalate	n/a	=	88	%	EPA 625m		65	145	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Di-n-butylphthalate	n/a	=	46	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Di-n-octylphthalate	n/a	=	0.0439	µg/L	EPA 625m	0.01			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Di-n-octylphthalate	n/a	=	102	%	EPA 625m		50	165	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Di-n-octylphthalate	n/a	=	103	%	EPA 625m		50	165	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Di-n-octylphthalate	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Di-n-octylphthalate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Di-n-octylphthalate	n/a	=	0.1273	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Di-n-octylphthalate	n/a	=	0.0793	µg/L	EPA 625m	0.01		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Di-n-octylphthalate	n/a	=	48	%	EPA 625m		50	165	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Di-n-octylphthalate	n/a	=	86	%	EPA 625m		50	165	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Di-n-octylphthalate	n/a	=	57	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Fluoranthene	n/a	=	0.0192	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Fluoranthene	n/a	=	99	%	EPA 625m		65	135	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Fluoranthene	n/a	=	102	%	EPA 625m		65	135	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Fluoranthene	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Fluoranthene	n/a	=	0.061	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Fluoranthene	n/a	=	0.0847	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Fluoranthene	n/a	=	20	%	EPA 625m		65	135	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Fluoranthene	n/a	=	75	%	EPA 625m		65	135	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Fluoranthene	n/a	=	116	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Fluorene	n/a	=	0.0017	µg/L	EPA 625m	0.001			EST
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Fluorene	n/a	=	90	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Fluorene	n/a	=	89	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Fluorene	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Fluorene	n/a	=	0.0204	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Fluorene	n/a	=	17	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Fluorene	n/a	=	56	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Fluorene	n/a	=	107	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Hexachlorobenzene	n/a	=	68	%	EPA 625m		65	135	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Hexachlorobenzene	n/a	=	70	%	EPA 625m		65	135	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Hexachlorobenzene	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Hexachlorobenzene	n/a	=	10	%	EPA 625m		65	135	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Hexachlorobenzene	n/a	=	28	%	EPA 625m		65	135	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Hexachlorobenzene	n/a	=	95	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	0.007	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	83	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	92	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	10	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	0.024	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	20	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	50	%	EPA 625m		70	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	86	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/4/2008	Organic	Methyl tert-butyl ether (MTBE)	n/a	<	1	µg/L	EPA 8260B	1			
2007/08-2	Lab	LCS dup, rec	1/3/2008	Organic	Methyl tert-butyl ether (MTBE)	n/a	=	101	%	EPA 8260B		82	118	
2007/08-2	Lab	LCS, rec	1/3/2008	Organic	Methyl tert-butyl ether (MTBE)	n/a	=	94	%	EPA 8260B		82	118	
2007/08-2	Lab	LCS, RPD	1/3/2008	Organic	Methyl tert-butyl ether (MTBE)	n/a	=	7	%	EPA 8260B		0	13	
2007/08-2	Lab	method blank	1/4/2008	Organic	Methyl tert-butyl ether (MTBE)	n/a	<	1	µg/L	EPA 8260B	1		1	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Naphthalene	n/a	=	0.0042	µg/L	EPA 625m	0.001			EST
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Naphthalene	n/a	=	65	%	EPA 625m		50	120	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Naphthalene	n/a	=	67	%	EPA 625m		50	120	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Naphthalene	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Naphthalene	n/a	=	0.0065	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Naphthalene	n/a	=	0.0591	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Naphthalene	n/a	=	12	%	EPA 625m		50	120	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Naphthalene	n/a	=	38	%	EPA 625m		50	120	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Naphthalene	n/a	=	104	%	EPA 625m		0	30	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	48	%	EPA 625m		40	120	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	50	%	EPA 625m		40	120	
2007/08-2	Lab	srgt LCS dup, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	59	%	EPA 625m		40	120	
2007/08-2	Lab	srgt LCS, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	64	%	EPA 625m		40	120	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	49	%	EPA 625m		40	120	
2007/08-2	ME-CC	srgt environ, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	48	%	EPA 625m		40	120	
2007/08-2	ME-CC	srgt field blank, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	47	%	EPA 625m		40	120	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	9	%	EPA 625m		40	120	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	19	%	EPA 625m		40	120	
2007/08-2	ME-SCR	srgt matrix spike dup, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	12	%	EPA 625m		40	120	
2007/08-2	ME-SCR	srgt matrix spike, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	27	%	EPA 625m		40	120	
2007/08-2	ME-VR2	srgt environ, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	47	%	EPA 625m		40	120	
2007/08-2	W-3	srgt environ, rec	1/15/2008	Organic	Naphthalene-d8	n/a	=	48	%	EPA 625m		40	120	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05			
2007/08-2	Lab	method blank	1/15/2008	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	N-Nitrosodiphenylamine	n/a	=	0.083	µg/L	EPA 625m	0.05		30	EST
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Pentachlorophenol	n/a	=	0.155	µg/L	EPA 625m	0.05			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Pentachlorophenol	n/a	=	36	%	EPA 625m		10	160	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Pentachlorophenol	n/a	=	37	%	EPA 625m		10	160	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Pentachlorophenol	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Pentachlorophenol	n/a	=	2	%	EPA 625m		10	160	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Pentachlorophenol	n/a	=	4	%	EPA 625m		10	160	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Pentachlorophenol	n/a	=	67	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Perylene	n/a	=	0.0171	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Perylene	n/a	=	99	%	EPA 625m		65	135	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Perylene	n/a	=	99	%	EPA 625m		65	135	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Perylene	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Perylene	n/a	=	0.0198	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Perylene	n/a	=	2.5987	µg/L	EPA 625m	0.001		30	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Perylene-d12	n/a	=	104	%	EPA 625m		60	140	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Perylene-d12	n/a	=	100	%	EPA 625m		60	140	
2007/08-2	Lab	srgt LCS dup, rec	1/15/2008	Organic	Perylene-d12	n/a	=	85	%	EPA 625m		60	140	
2007/08-2	Lab	srgt LCS, rec	1/15/2008	Organic	Perylene-d12	n/a	=	98	%	EPA 625m		60	140	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	Organic	Perylene-d12	n/a	=	114	%	EPA 625m		60	140	
2007/08-2	ME-CC	srgt environ, rec	1/15/2008	Organic	Perylene-d12	n/a	=	104	%	EPA 625m		60	140	
2007/08-2	ME-CC	srgt field blank, rec	1/15/2008	Organic	Perylene-d12	n/a	=	92	%	EPA 625m		60	140	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Perylene-d12	n/a	=	29	%	EPA 625m		60	140	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Perylene-d12	n/a	=	62	%	EPA 625m		60	140	
2007/08-2	ME-SCR	srgt matrix spike dup, rec	1/15/2008	Organic	Perylene-d12	n/a	=	16	%	EPA 625m		60	140	
2007/08-2	ME-SCR	srgt matrix spike, rec	1/15/2008	Organic	Perylene-d12	n/a	=	45	%	EPA 625m		60	140	
2007/08-2	ME-VR2	srgt environ, rec	1/15/2008	Organic	Perylene-d12	n/a	=	112	%	EPA 625m		60	140	
2007/08-2	W-3	srgt environ, rec	1/15/2008	Organic	Perylene-d12	n/a	=	106	%	EPA 625m		60	140	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Phenanthrene	n/a	=	0.0149	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Phenanthrene	n/a	=	94	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Phenanthrene	n/a	=	100	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Phenanthrene	n/a	=	6	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Phenanthrene	n/a	=	0.0166	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Phenanthrene	n/a	=	0.1879	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Phenanthrene	n/a	=	13	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Phenanthrene	n/a	=	72	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Phenanthrene	n/a	=	139	%	EPA 625m		0	30	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	83	%	EPA 625m		70	130	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	81	%	EPA 625m		70	130	
2007/08-2	Lab	srgt LCS dup, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	89	%	EPA 625m		70	130	
2007/08-2	Lab	srgt LCS, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	90	%	EPA 625m		70	130	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	79	%	EPA 625m		70	130	
2007/08-2	ME-CC	srgt environ, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	75	%	EPA 625m		70	130	
2007/08-2	ME-CC	srgt field blank, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	68	%	EPA 625m		70	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	16	%	EPA 625m		70	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	36	%	EPA 625m		70	130	
2007/08-2	ME-SCR	srgt matrix spike dup, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	19	%	EPA 625m		70	130	
2007/08-2	ME-SCR	srgt matrix spike, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	50	%	EPA 625m		70	130	
2007/08-2	ME-VR2	srgt environ, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	83	%	EPA 625m		70	130	
2007/08-2	W-3	srgt environ, rec	1/15/2008	Organic	Phenanthrene-d10	n/a	=	72	%	EPA 625m		70	130	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Phenol	n/a	=	0.234	µg/L	EPA 625m	0.1			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Phenol	n/a	=	37	%	EPA 625m		0	115	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Phenol	n/a	=	40	%	EPA 625m		0	115	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Phenol	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Phenol	n/a	=	5	%	EPA 625m		0	115	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Phenol	n/a	=	8	%	EPA 625m		0	115	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Phenol	n/a	=	46	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Phenol-d5	n/a	=	39	%	EPA 625m		10	110	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Phenol-d5	n/a	=	34	%	EPA 625m		10	110	
2007/08-2	Lab	srgt LCS dup, rec	1/15/2008	Organic	Phenol-d5	n/a	=	79	%	EPA 625m		10	110	
2007/08-2	Lab	srgt LCS, rec	1/15/2008	Organic	Phenol-d5	n/a	=	84	%	EPA 625m		10	110	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	Organic	Phenol-d5	n/a	=	129	%	EPA 625m		10	110	
2007/08-2	ME-CC	srgt environ, rec	1/15/2008	Organic	Phenol-d5	n/a	=	37	%	EPA 625m		10	110	
2007/08-2	ME-CC	srgt field blank, rec	1/15/2008	Organic	Phenol-d5	n/a	=	30	%	EPA 625m		10	110	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Phenol-d5	n/a	=	23	%	EPA 625m		10	110	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Phenol-d5	n/a	=	25	%	EPA 625m		10	110	
2007/08-2	ME-SCR	srgt matrix spike dup, rec	1/15/2008	Organic	Phenol-d5	n/a	=	18	%	EPA 625m		10	110	
2007/08-2	ME-SCR	srgt matrix spike, rec	1/15/2008	Organic	Phenol-d5	n/a	=	38	%	EPA 625m		10	110	
2007/08-2	ME-VR2	srgt environ, rec	1/15/2008	Organic	Phenol-d5	n/a	=	40	%	EPA 625m		10	110	
2007/08-2	W-3	srgt environ, rec	1/15/2008	Organic	Phenol-d5	n/a	=	35	%	EPA 625m		10	110	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Pyrene	n/a	=	0.0227	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Organic	Pyrene	n/a	=	105	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Organic	Pyrene	n/a	=	104	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Organic	Pyrene	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Pyrene	n/a	=	0.0686	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Pyrene	n/a	=	0.1385	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Organic	Pyrene	n/a	=	25	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Organic	Pyrene	n/a	=	86	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Organic	Pyrene	n/a	=	110	%	EPA 625m		0	30	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	99	%	EPA 625m		40	130	
2007/08-2	A-1	srgt environ, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	91	%	EPA 625m		40	130	
2007/08-2	Lab	srgt LCS dup, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	84	%	EPA 625m		40	130	
2007/08-2	Lab	srgt LCS, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	84	%	EPA 625m		40	130	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	80	%	EPA 625m		40	130	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	84	%	EPA 625m		40	130	
2007/08-2	ME-CC	srgt environ, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	77	%	EPA 625m		40	130	
2007/08-2	ME-CC	srgt field blank, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	81	%	EPA 625m		40	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	30	%	EPA 625m		40	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	46	%	EPA 625m		40	130	
2007/08-2	ME-SCR	srgt matrix spike dup, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	25	%	EPA 625m		40	130	
2007/08-2	ME-SCR	srgt matrix spike, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	63	%	EPA 625m		40	130	
2007/08-2	ME-VR2	srgt environ, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	87	%	EPA 625m		40	130	
2007/08-2	W-3	srgt environ, rec	1/15/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	89	%	EPA 625m		40	130	
2007/08-2	A-1	srgt environ, rec	1/4/2008	Organic	Toluene-d8	n/a	=	92	%	EPA 8260B		88	112	
2007/08-2	A-1	srgt environ, rec	1/4/2008	Organic	Toluene-d8	n/a	=	100	%	EPA 8260B		88	112	
2007/08-2	Lab	srgt method blank, rec	1/4/2008	Organic	Toluene-d8	n/a	=	94	%	EPA 8260B		88	112	
2007/08-2	W-3	srgt environ, rec	1/4/2008	Organic	Toluene-d8	n/a	=	97	%	EPA 8260B		88	112	
2007/08-2	A-1	field duplicate	1/15/2008	Organic	Total Detectable PAHs	n/a	=	0.2109	µg/L	EPA 625m				
2007/08-2	ME-CC	field blank	1/15/2008	Organic	Total Detectable PAHs	n/a	=	0.5031	µg/L	EPA 625m				
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Organic	Total Detectable PAHs	n/a	=	4.24	µg/L	EPA 625m			30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01			
2007/08-2	Lab	method blank	1/15/2008	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01			
2007/08-2	Lab	method blank	1/15/2008	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	A-1	field duplicate	1/15/2008	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01			
2007/08-2	Lab	method blank	1/15/2008	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01			
2007/08-2	Lab	method blank	1/15/2008	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01			
2007/08-2	Lab	method blank	1/15/2008	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01			
2007/08-2	Lab	method blank	1/15/2008	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 008	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 008	n/a	=	62	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 008	n/a	=	63	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 008	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 008	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 008	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 008	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 008	n/a	=	69	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 008	n/a	=	70	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 008	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 018	n/a	=	66	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 018	n/a	=	65	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 018	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 018	n/a	=	78	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 018	n/a	=	72	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 018	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 028	n/a	=	66	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 028	n/a	=	65	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 028	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 028	n/a	=	78	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 028	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 028	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	A-1	srgt environ, rec	1/15/2008	PCB	PCB 030	n/a	=	96	%	EPA 625m		40	130	
2007/08-2	A-1	srgt environ, rec	1/15/2008	PCB	PCB 030	n/a	=	93	%	EPA 625m		40	130	
2007/08-2	Lab	srgt LCS dup, rec	1/15/2008	PCB	PCB 030	n/a	=	92	%	EPA 625m		40	130	
2007/08-2	Lab	srgt LCS, rec	1/15/2008	PCB	PCB 030	n/a	=	106	%	EPA 625m		40	130	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	PCB	PCB 030	n/a	=	85	%	EPA 625m		40	130	
2007/08-2	ME-CC	srgt environ, rec	1/15/2008	PCB	PCB 030	n/a	=	80	%	EPA 625m		40	130	
2007/08-2	ME-CC	srgt field blank, rec	1/15/2008	PCB	PCB 030	n/a	=	78	%	EPA 625m		40	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	PCB	PCB 030	n/a	=	48	%	EPA 625m		40	130	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	PCB	PCB 030	n/a	=	21	%	EPA 625m		40	130	
2007/08-2	ME-SCR	srgt matrix spike dup, rec	1/15/2008	PCB	PCB 030	n/a	=	23	%	EPA 625m		40	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	srgt matrix spike, rec	1/15/2008	PCB	PCB 030	n/a	=	61	%	EPA 625m		40	130	
2007/08-2	ME-VR2	srgt environ, rec	1/15/2008	PCB	PCB 030	n/a	=	88	%	EPA 625m		40	130	
2007/08-2	W-3	srgt environ, rec	1/15/2008	PCB	PCB 030	n/a	=	95	%	EPA 625m		40	130	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 031	n/a	=	63	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 031	n/a	=	65	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 031	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 031	n/a	=	69	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 031	n/a	=	67	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 031	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 033	n/a	=	66	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 033	n/a	=	69	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 033	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 033	n/a	=	79	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 033	n/a	=	72	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 033	n/a	=	9	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 037	n/a	=	64	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 037	n/a	=	69	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 037	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 037	n/a	=	79	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 037	n/a	=	71	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 037	n/a	=	11	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 044	n/a	=	67	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 044	n/a	=	72	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 044	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 044	n/a	=	66	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 044	n/a	=	78	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 044	n/a	=	17	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 049	n/a	=	69	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 049	n/a	=	69	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 049	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 049	n/a	=	76	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 049	n/a	=	77	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 049	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 052	n/a	=	70	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 052	n/a	=	69	%	EPA 625m		60	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 052	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 052	n/a	=	77	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 052	n/a	=	67	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 052	n/a	=	14	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 066	n/a	=	70	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 066	n/a	=	75	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 066	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 066	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 066	n/a	=	75	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 066	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 070	n/a	=	67	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 070	n/a	=	69	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 070	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 070	n/a	=	72	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 070	n/a	=	71	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 070	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 074	n/a	=	71	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 074	n/a	=	73	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 074	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 074	n/a	=	76	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 074	n/a	=	72	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 074	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 077	n/a	=	72	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 077	n/a	=	74	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 077	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 077	n/a	=	74	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 077	n/a	=	74	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 077	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 081	n/a	=	71	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 081	n/a	=	72	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 081	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 081	n/a	=	71	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 081	n/a	=	70	%	EPA 625m		60	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 081	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 087	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 087	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 087	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 087	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 087	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 087	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 095	n/a	=	71	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 095	n/a	=	70	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 095	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 095	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 095	n/a	=	82	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 095	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 097	n/a	=	72	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 097	n/a	=	75	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 097	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 097	n/a	=	82	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 097	n/a	=	71	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 097	n/a	=	14	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 099	n/a	=	78	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 099	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 099	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 099	n/a	=	88	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 099	n/a	=	83	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 099	n/a	=	6	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 101	n/a	=	74	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 101	n/a	=	76	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 101	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 101	n/a	=	82	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 101	n/a	=	88	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 101	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 105	n/a	=	78	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 105	n/a	=	75	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 105	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 105	n/a	=	71	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 105	n/a	=	72	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 105	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 110	n/a	=	74	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 110	n/a	=	76	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 110	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 110	n/a	=	82	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 110	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 110	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	srgt environ, rec	1/15/2008	PCB	PCB 112	n/a	=	95	%	EPA 625m		60	120	
2007/08-2	A-1	srgt environ, rec	1/15/2008	PCB	PCB 112	n/a	=	90	%	EPA 625m		60	120	
2007/08-2	Lab	srgt LCS dup, rec	1/15/2008	PCB	PCB 112	n/a	=	85	%	EPA 625m		60	120	
2007/08-2	Lab	srgt LCS, rec	1/15/2008	PCB	PCB 112	n/a	=	89	%	EPA 625m		60	120	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	PCB	PCB 112	n/a	=	94	%	EPA 625m		60	120	
2007/08-2	ME-CC	srgt environ, rec	1/15/2008	PCB	PCB 112	n/a	=	83	%	EPA 625m		60	120	
2007/08-2	ME-CC	srgt field blank, rec	1/15/2008	PCB	PCB 112	n/a	=	78	%	EPA 625m		60	120	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	PCB	PCB 112	n/a	=	49	%	EPA 625m		60	120	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	PCB	PCB 112	n/a	=	22	%	EPA 625m		60	120	
2007/08-2	ME-SCR	srgt matrix spike dup, rec	1/15/2008	PCB	PCB 112	n/a	=	23	%	EPA 625m		60	120	
2007/08-2	ME-SCR	srgt matrix spike, rec	1/15/2008	PCB	PCB 112	n/a	=	66	%	EPA 625m		60	120	
2007/08-2	ME-VR2	srgt environ, rec	1/15/2008	PCB	PCB 112	n/a	=	93	%	EPA 625m		60	120	
2007/08-2	W-3	srgt environ, rec	1/15/2008	PCB	PCB 112	n/a	=	92	%	EPA 625m		60	120	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 114	n/a	=	74	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 114	n/a	=	73	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 114	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 114	n/a	=	73	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 114	n/a	=	73	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 114	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 118	n/a	=	75	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 118	n/a	=	75	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 118	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 118	n/a	=	76	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 118	n/a	=	77	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 118	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 119	n/a	=	74	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 119	n/a	=	69	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 119	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 119	n/a	=	78	%	EPA 625m		60	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 119	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 119	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 123	n/a	=	72	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 123	n/a	=	77	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 123	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 123	n/a	=	85	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 123	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 123	n/a	=	6	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 126	n/a	=	75	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 126	n/a	=	79	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 126	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 126	n/a	=	70	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 126	n/a	=	73	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 126	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 128	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 128	n/a	=	84	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 128	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 128	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 128	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 128	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 128	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 128	n/a	=	85	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 128	n/a	=	85	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 128	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 138	n/a	=	77	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 138	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 138	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 138	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 138	n/a	=	88	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 138	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 141	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 141	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 141	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 141	n/a	=	82	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 141	n/a	=	76	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 141	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 149	n/a	=	75	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 149	n/a	=	79	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 149	n/a	=	5	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 149	n/a	=	84	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 149	n/a	=	85	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 149	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 151	n/a	=	77	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 151	n/a	=	78	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 151	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 151	n/a	=	83	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 151	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 151	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 153	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 153	n/a	=	79	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 153	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 153	n/a	=	89	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 153	n/a	=	90	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 153	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 156	n/a	=	76	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 156	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 156	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 156	n/a	=	78	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 156	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 156	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 157	n/a	=	77	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 157	n/a	=	78	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 157	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 157	n/a	=	72	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 157	n/a	=	67	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 157	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 158	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 158	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 158	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 158	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 158	n/a	=	79	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 158	n/a	=	2	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 167	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 167	n/a	=	76	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 167	n/a	=	83	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 167	n/a	=	9	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 167	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 167	n/a	=	67	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 167	n/a	=	68	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 167	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 168 + 132	n/a	=	82	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 168 + 132	n/a	=	83	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 168 + 132	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 168 + 132	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 168 + 132	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 168 + 132	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 169	n/a	=	82	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 169	n/a	=	87	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 169	n/a	=	6	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 169	n/a	=	77	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 169	n/a	=	75	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 169	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 170	n/a	=	82	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 170	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 170	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 170	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 170	n/a	=	93	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 170	n/a	=	15	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 174	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 174	n/a	=	92	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 174	n/a	=	91	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 174	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 174	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 174	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 174	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 174	n/a	=	85	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 174	n/a	=	89	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 174	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 177	n/a	=	85	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 177	n/a	=	81	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 177	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 177	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 177	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 177	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 180	n/a	=	84	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 180	n/a	=	88	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 180	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 180	n/a	=	88	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 180	n/a	=	91	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 180	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 183	n/a	=	86	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 183	n/a	=	87	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 183	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 183	n/a	=	86	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 183	n/a	=	82	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 183	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 187	n/a	=	83	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 187	n/a	=	88	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 187	n/a	=	6	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 187	n/a	=	91	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 187	n/a	=	95	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 187	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 189	n/a	=	82	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 189	n/a	=	85	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 189	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 189	n/a	=	80	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 189	n/a	=	83	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 189	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 194	n/a	=	85	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 194	n/a	=	85	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 194	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 194	n/a	=	112	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 194	n/a	=	109	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 194	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 195	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 195	n/a	=	76	%	EPA 625m		60	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 195	n/a	=	74	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 195	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 195	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 195	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 195	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 195	n/a	=	101	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 195	n/a	=	99	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 195	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	A-1	srgt environ, rec	1/15/2008	PCB	PCB 198	n/a	=	91	%	EPA 625m		60	120	
2007/08-2	A-1	srgt environ, rec	1/15/2008	PCB	PCB 198	n/a	=	95	%	EPA 625m		60	120	
2007/08-2	Lab	srgt LCS dup, rec	1/15/2008	PCB	PCB 198	n/a	=	86	%	EPA 625m		60	120	
2007/08-2	Lab	srgt LCS, rec	1/15/2008	PCB	PCB 198	n/a	=	90	%	EPA 625m		60	120	
2007/08-2	Lab	srgt method blank, rec	1/15/2008	PCB	PCB 198	n/a	=	97	%	EPA 625m		60	120	
2007/08-2	ME-CC	srgt environ, rec	1/15/2008	PCB	PCB 198	n/a	=	100	%	EPA 625m		60	120	
2007/08-2	ME-CC	srgt field blank, rec	1/15/2008	PCB	PCB 198	n/a	=	101	%	EPA 625m		60	120	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	PCB	PCB 198	n/a	=	63	%	EPA 625m		60	120	
2007/08-2	ME-SCR	srgt environ, rec	1/15/2008	PCB	PCB 198	n/a	=	36	%	EPA 625m		60	120	
2007/08-2	ME-SCR	srgt matrix spike dup, rec	1/15/2008	PCB	PCB 198	n/a	=	30	%	EPA 625m		60	120	
2007/08-2	ME-SCR	srgt matrix spike, rec	1/15/2008	PCB	PCB 198	n/a	=	75	%	EPA 625m		60	120	
2007/08-2	ME-VR2	srgt environ, rec	1/15/2008	PCB	PCB 198	n/a	=	92	%	EPA 625m		60	120	
2007/08-2	W-3	srgt environ, rec	1/15/2008	PCB	PCB 198	n/a	=	98	%	EPA 625m		60	120	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 200	n/a	=	89	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 200	n/a	=	91	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 200	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 200	n/a	=	84	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 200	n/a	=	87	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 200	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 201	n/a	=	94	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 201	n/a	=	96	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 201	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 201	n/a	=	95	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 201	n/a	=	103	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 201	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 206	n/a	=	85	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 206	n/a	=	79	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 206	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 206	n/a	=	98	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 206	n/a	=	84	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 206	n/a	=	15	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	PCB 209	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	PCB	PCB 209	n/a	=	93	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	PCB	PCB 209	n/a	=	104	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	PCB	PCB 209	n/a	=	11	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	PCB	PCB 209	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-CC	field blank	1/15/2008	PCB	PCB 209	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	PCB 209	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	PCB	PCB 209	n/a	=	87	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	PCB	PCB 209	n/a	=	94	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	PCB	PCB 209	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2007/08-2	ME-CC	field blank	1/15/2008	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2007/08-2	ME-SCR	lab duplicate	1/15/2008	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m			30	
2007/08-2	A-1	field duplicate	12/27/2007	Pesticide	2,4,5-T	n/a	<	0.5	µg/L	EPA 8151A	0.5			
2007/08-2	Lab	LCS dup, rec	12/27/2007	Pesticide	2,4,5-T	n/a	=	124	%	EPA 8151A		30	130	
2007/08-2	Lab	LCS, rec	12/27/2007	Pesticide	2,4,5-T	n/a	=	125	%	EPA 8151A		30	130	
2007/08-2	Lab	LCS, RPD	12/27/2007	Pesticide	2,4,5-T	n/a	=	1	%	EPA 8151A		0	30	
2007/08-2	Lab	method blank	12/27/2007	Pesticide	2,4,5-T	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2007/08-2	ME-SCR	matrix spike dup, rec	12/27/2007	Pesticide	2,4,5-T	n/a	=	222	%	EPA 8151A		30	130	
2007/08-2	ME-SCR	matrix spike, rec	12/27/2007	Pesticide	2,4,5-T	n/a	=	134	%	EPA 8151A		30	130	
2007/08-2	ME-SCR	matrix spike, RPD	12/27/2007	Pesticide	2,4,5-T	n/a	=	49	%	EPA 8151A		0	30	
2007/08-2	A-1	field duplicate	12/27/2007	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.5	µg/L	EPA 8151A	0.5			
2007/08-2	Lab	method blank	12/27/2007	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2007/08-2	A-1	field duplicate	12/27/2007	Pesticide	2,4-D	n/a	<	5	µg/L	EPA 8151A	5			
2007/08-2	Lab	LCS dup, rec	12/27/2007	Pesticide	2,4-D	n/a	=	108	%	EPA 8151A		30	130	
2007/08-2	Lab	LCS, rec	12/27/2007	Pesticide	2,4-D	n/a	=	109	%	EPA 8151A		30	130	
2007/08-2	Lab	LCS, RPD	12/27/2007	Pesticide	2,4-D	n/a	=	1	%	EPA 8151A		0	30	
2007/08-2	Lab	method blank	12/27/2007	Pesticide	2,4-D	n/a	<	5	µg/L	EPA 8151A	5		5	
2007/08-2	ME-SCR	matrix spike dup, rec	12/27/2007	Pesticide	2,4-D	n/a	=	149	%	EPA 8151A		30	130	
2007/08-2	ME-SCR	matrix spike, rec	12/27/2007	Pesticide	2,4-D	n/a	=	144	%	EPA 8151A		30	130	
2007/08-2	ME-SCR	matrix spike, RPD	12/27/2007	Pesticide	2,4-D	n/a	=	3	%	EPA 8151A		0	30	
2007/08-2	A-1	field duplicate	12/27/2007	Pesticide	2,4-DB	n/a	<	5	µg/L	EPA 8151A	5			
2007/08-2	Lab	LCS dup, rec	12/27/2007	Pesticide	2,4-DB	n/a	=	110	%	EPA 8151A		30	130	
2007/08-2	Lab	LCS, rec	12/27/2007	Pesticide	2,4-DB	n/a	=	116	%	EPA 8151A		30	130	
2007/08-2	Lab	LCS, RPD	12/27/2007	Pesticide	2,4-DB	n/a	=	5	%	EPA 8151A		0	30	
2007/08-2	Lab	method blank	12/27/2007	Pesticide	2,4-DB	n/a	<	5	µg/L	EPA 8151A	5		5	
2007/08-2	ME-SCR	matrix spike dup, rec	12/27/2007	Pesticide	2,4-DB	n/a	=	698	%	EPA 8151A		30	130	
2007/08-2	ME-SCR	matrix spike, rec	12/27/2007	Pesticide	2,4-DB	n/a	=	443	%	EPA 8151A		30	130	
2007/08-2	ME-SCR	matrix spike, RPD	12/27/2007	Pesticide	2,4-DB	n/a	=	45	%	EPA 8151A		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	2,4'-DDD	n/a	=	0.0485	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	2,4'-DDD	n/a	=	108	%	EPA 625m		50	140	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	2,4'-DDD	n/a	=	116	%	EPA 625m		50	140	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	2,4'-DDD	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	2,4'-DDD	n/a	=	56	%	EPA 625m		50	140	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	2,4'-DDD	n/a	=	101	%	EPA 625m		50	140	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	2,4'-DDD	n/a	=	57	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	2,4'-DDE	n/a	=	0.011	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	2,4'-DDE	n/a	=	97	%	EPA 625m		60	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	2,4'-DDE	n/a	=	111	%	EPA 625m		60	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	2,4'-DDE	n/a	=	13	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	2,4'-DDE	n/a	=	0.008	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	2,4'-DDE	n/a	=	36	%	EPA 625m		60	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	2,4'-DDE	n/a	=	89	%	EPA 625m		60	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	2,4'-DDE	n/a	=	85	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	2,4'-DDT	n/a	=	0.0175	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	2,4'-DDT	n/a	=	89	%	EPA 625m		40	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	2,4'-DDT	n/a	=	84	%	EPA 625m		40	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	2,4'-DDT	n/a	=	6	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	2,4'-DDT	n/a	=	0.0068	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	2,4'-DDT	n/a	=	57	%	EPA 625m		40	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	2,4'-DDT	n/a	=	116	%	EPA 625m		40	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	2,4'-DDT	n/a	=	68	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	4,4'-DDD	n/a	=	0.126	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	4,4'-DDD	n/a	=	115	%	EPA 625m		60	140	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	4,4'-DDD	n/a	=	117	%	EPA 625m		60	140	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	4,4'-DDD	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	4,4'-DDD	n/a	=	0.0342	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	4,4'-DDD	n/a	=	39	%	EPA 625m		60	140	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	4,4'-DDD	n/a	=	97	%	EPA 625m		60	140	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	4,4'-DDD	n/a	=	85	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	4,4'-DDE	n/a	=	0.5949	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	4,4'-DDE	n/a	=	123	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	4,4'-DDE	n/a	=	119	%	EPA 625m		70	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	4,4'-DDE	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	4,4'-DDE	n/a	=	0.2797	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	4,4'-DDE	n/a	=	39	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	4,4'-DDE	n/a	=	94	%	EPA 625m		70	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	4,4'-DDE	n/a	=	83	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	4,4'-DDT	n/a	=	0.1175	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	4,4'-DDT	n/a	=	67	%	EPA 625m		0	150	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	4,4'-DDT	n/a	=	57	%	EPA 625m		0	150	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	4,4'-DDT	n/a	=	16	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	4,4'-DDT	n/a	=	0	%	EPA 625m		0	150	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	4,4'-DDT	n/a	=	0	%	EPA 625m		0	150	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	4,4'-DDT	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Aldrin	n/a	=	90	%	EPA 625m		50	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Aldrin	n/a	=	95	%	EPA 625m		50	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Aldrin	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Aldrin	n/a	=	47	%	EPA 625m		50	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Aldrin	n/a	=	85	%	EPA 625m		50	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Aldrin	n/a	=	58	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	BHC-alpha	n/a	=	83	%	EPA 625m		60	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	BHC-alpha	n/a	=	80	%	EPA 625m		60	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	BHC-alpha	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	BHC-alpha	n/a	=	35	%	EPA 625m		60	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	BHC-alpha	n/a	=	71	%	EPA 625m		60	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	BHC-alpha	n/a	=	68	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	BHC-beta	n/a	=	97	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	BHC-beta	n/a	=	85	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	BHC-beta	n/a	=	13	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	BHC-beta	n/a	=	51	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	BHC-beta	n/a	=	83	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	BHC-beta	n/a	=	48	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	BHC-delta	n/a	=	78	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	BHC-delta	n/a	=	96	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	BHC-delta	n/a	=	21	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	BHC-delta	n/a	=	50	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	BHC-delta	n/a	=	65	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	BHC-delta	n/a	=	26	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	74	%	EPA 625m		50	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	73	%	EPA 625m		50	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	51	%	EPA 625m		50	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	75	%	EPA 625m		50	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	38	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Bolstar	n/a	=	101	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Bolstar	n/a	=	102	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Bolstar	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Bolstar	n/a	=	37	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Bolstar	n/a	=	80	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Bolstar	n/a	=	74	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Chlordane-alpha	n/a	=	87	%	EPA 625m		60	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Chlordane-alpha	n/a	=	95	%	EPA 625m		60	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Chlordane-alpha	n/a	=	9	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Chlordane-alpha	n/a	=	0.001	µg/L	EPA 625m	0.001		0.001	EST
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Chlordane-alpha	n/a	=	24	%	EPA 625m		60	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Chlordane-alpha	n/a	=	77	%	EPA 625m		60	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Chlordane-alpha	n/a	=	105	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Chlordane-gamma	n/a	=	100	%	EPA 625m		60	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Chlordane-gamma	n/a	=	101	%	EPA 625m		60	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Chlordane-gamma	n/a	=	1	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Chlordane-gamma	n/a	=	25	%	EPA 625m		60	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Chlordane-gamma	n/a	=	69	%	EPA 625m		60	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Chlordane-gamma	n/a	=	94	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Chlorpyrifos	n/a	=	0.1725	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Chlorpyrifos	n/a	=	103	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Chlorpyrifos	n/a	=	98	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Chlorpyrifos	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Chlorpyrifos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Chlorpyrifos	n/a	=	0.2119	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Chlorpyrifos	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Chlorpyrifos	n/a	=	36	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Chlorpyrifos	n/a	=	68	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Chlorpyrifos	n/a	=	62	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	cis-Nonachlor	n/a	=	95	%	EPA 625m		60	120	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	cis-Nonachlor	n/a	=	97	%	EPA 625m		60	120	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	cis-Nonachlor	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	cis-Nonachlor	n/a	=	20	%	EPA 625m		60	120	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	cis-Nonachlor	n/a	=	62	%	EPA 625m		60	120	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	cis-Nonachlor	n/a	=	102	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	12/27/2007	Pesticide	Dalapon	n/a	<	13	µg/L	EPA 8151A	13			
2007/08-2	Lab	method blank	12/27/2007	Pesticide	Dalapon	n/a	<	13	µg/L	EPA 8151A	13		13	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Demeton-O	n/a	=	101	%	EPA 625m		45	105	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Demeton-O	n/a	=	96	%	EPA 625m		45	105	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Demeton-O	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Demeton-O	n/a	=	17	%	EPA 625m		45	105	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Demeton-O	n/a	=	42	%	EPA 625m		45	105	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Demeton-O	n/a	=	85	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Diazinon	n/a	=	0.0166	µg/L	EPA 625m	0.002			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Diazinon	n/a	=	93	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Diazinon	n/a	=	93	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Diazinon	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Diazinon	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Diazinon	n/a	=	0.0204	µg/L	EPA 625m	0.002		0.002	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Diazinon	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Diazinon	n/a	=	42	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Diazinon	n/a	=	83	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Diazinon	n/a	=	66	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	12/27/2007	Pesticide	Dicamba	n/a	<	0.5	µg/L	EPA 8151A	0.5			
2007/08-2	Lab	method blank	12/27/2007	Pesticide	Dicamba	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2007/08-2	A-1	field duplicate	12/27/2007	Pesticide	Dichlorprop	n/a	<	5	µg/L	EPA 8151A	5			
2007/08-2	Lab	method blank	12/27/2007	Pesticide	Dichlorprop	n/a	<	5	µg/L	EPA 8151A	5		5	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Dichlorvos	n/a	=	107	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Dichlorvos	n/a	=	106	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Dichlorvos	n/a	=	1	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Dichlorvos	n/a	=	28	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Dichlorvos	n/a	=	77	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Dichlorvos	n/a	=	93	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Dieldrin	n/a	=	80	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Dieldrin	n/a	=	107	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Dieldrin	n/a	=	29	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Dieldrin	n/a	=	40	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Dieldrin	n/a	=	76	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Dieldrin	n/a	=	62	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Dimethoate	n/a	=	104	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Dimethoate	n/a	=	103	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Dimethoate	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Dimethoate	n/a	=	34	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Dimethoate	n/a	=	70	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Dimethoate	n/a	=	69	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	12/27/2007	Pesticide	Dinoseb	n/a	<	2.5	µg/L	EPA 8151A	2.5			
2007/08-2	Lab	method blank	12/27/2007	Pesticide	Dinoseb	n/a	<	2.5	µg/L	EPA 8151A	2.5		2.5	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Disulfoton	n/a	=	87	%	EPA 625m		45	105	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Disulfoton	n/a	=	86	%	EPA 625m		45	105	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Disulfoton	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Disulfoton	n/a	=	15	%	EPA 625m		45	105	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Disulfoton	n/a	=	52	%	EPA 625m		45	105	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Disulfoton	n/a	=	110	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Endosulfan sulfate	n/a	=	103	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Endosulfan sulfate	n/a	=	101	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Endosulfan sulfate	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Endosulfan sulfate	n/a	=	26	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Endosulfan sulfate	n/a	=	41	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Endosulfan sulfate	n/a	=	45	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Endosulfan-I	n/a	=	121	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Endosulfan-I	n/a	=	118	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Endosulfan-I	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Endosulfan-I	n/a	=	115	%	EPA 625m		60	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Endosulfan-I	n/a	=	92	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Endosulfan-I	n/a	=	22	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Endosulfan-II	n/a	=	119	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Endosulfan-II	n/a	=	114	%	EPA 625m		60	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Endosulfan-II	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Endosulfan-II	n/a	=	46	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Endosulfan-II	n/a	=	79	%	EPA 625m		60	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Endosulfan-II	n/a	=	53	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Endrin	n/a	=	90	%	EPA 625m		65	135	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Endrin	n/a	=	103	%	EPA 625m		65	135	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Endrin	n/a	=	13	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Endrin	n/a	=	67	%	EPA 625m		65	135	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Endrin	n/a	=	94	%	EPA 625m		65	135	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Endrin	n/a	=	34	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Endrin aldehyde	n/a	=	103	%	EPA 625m		0	149	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Endrin aldehyde	n/a	=	107	%	EPA 625m		0	149	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Endrin aldehyde	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Endrin aldehyde	n/a	=	33	%	EPA 625m		0	149	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Endrin aldehyde	n/a	=	79	%	EPA 625m		0	149	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Endrin aldehyde	n/a	=	82	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Endrin ketone	n/a	=	94	%	EPA 625m		40	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Endrin ketone	n/a	=	91	%	EPA 625m		40	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Endrin ketone	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Endrin ketone	n/a	=	29	%	EPA 625m		40	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Endrin ketone	n/a	=	59	%	EPA 625m		40	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Endrin ketone	n/a	=	68	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Ethoprop	n/a	=	97	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Ethoprop	n/a	=	101	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Ethoprop	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Ethoprop	n/a	=	29	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Ethoprop	n/a	=	70	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Ethoprop	n/a	=	83	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Fenchlorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Fenchlorophos (Ronnel)	n/a	=	103	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Fenchlorophos (Ronnel)	n/a	=	99	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Fenchlorophos (Ronnel)	n/a	=	4	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Fenclorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Fenclorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Fenclorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Fenclorophos (Ronnel)	n/a	=	21	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Fenclorophos (Ronnel)	n/a	=	68	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Fenclorophos (Ronnel)	n/a	=	106	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Fensulfothion	n/a	=	103	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Fensulfothion	n/a	=	98	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Fensulfothion	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Fensulfothion	n/a	=	45	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Fensulfothion	n/a	=	95	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Fensulfothion	n/a	=	71	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Fenthion	n/a	=	93	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Fenthion	n/a	=	110	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Fenthion	n/a	=	17	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Fenthion	n/a	=	26	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Fenthion	n/a	=	72	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Fenthion	n/a	=	94	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/3/2008	Pesticide	Glyphosate	n/a	=	11	µg/L	EPA 547	5			
2007/08-2	Lab	LCS, rec	1/3/2008	Pesticide	Glyphosate	n/a	=	92	%	EPA 547		71	137	
2007/08-2	Lab	method blank	1/3/2008	Pesticide	Glyphosate	n/a	<	5	µg/L	EPA 547	5		5	
2007/08-2	ME-SCR	matrix spike dup, rec	1/3/2008	Pesticide	Glyphosate	n/a	=	73	%	EPA 547		68	134	
2007/08-2	ME-SCR	matrix spike, rec	1/3/2008	Pesticide	Glyphosate	n/a	=	73	%	EPA 547		68	134	
2007/08-2	ME-SCR	matrix spike, RPD	1/3/2008	Pesticide	Glyphosate	n/a	=	0.5	%	EPA 547		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Heptachlor	n/a	=	79	%	EPA 625m		45	135	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Heptachlor	n/a	=	86	%	EPA 625m		45	135	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Heptachlor	n/a	=	8	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Heptachlor	n/a	=	29	%	EPA 625m		45	135	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Heptachlor	n/a	=	70	%	EPA 625m		45	135	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Heptachlor	n/a	=	83	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Heptachlor epoxide	n/a	=	100	%	EPA 625m		65	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Heptachlor epoxide	n/a	=	99	%	EPA 625m		65	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Heptachlor epoxide	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Heptachlor epoxide	n/a	=	35	%	EPA 625m		65	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Heptachlor epoxide	n/a	=	113	%	EPA 625m		65	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Heptachlor epoxide	n/a	=	105	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Malathion	n/a	=	0.2119	µg/L	EPA 625m	0.003			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Malathion	n/a	=	98	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Malathion	n/a	=	99	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Malathion	n/a	=	1	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Malathion	n/a	=	0.1403	µg/L	EPA 625m	0.003		0.003	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Malathion	n/a	=	53	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Malathion	n/a	=	88	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Malathion	n/a	=	50	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	12/27/2007	Pesticide	MCPA	n/a	<	500	µg/L	EPA 8151A	500			
2007/08-2	Lab	method blank	12/27/2007	Pesticide	MCPA	n/a	<	500	µg/L	EPA 8151A	500		500	
2007/08-2	A-1	field duplicate	12/27/2007	Pesticide	MCPP	n/a	<	500	µg/L	EPA 8151A	500			
2007/08-2	Lab	method blank	12/27/2007	Pesticide	MCPP	n/a	<	500	µg/L	EPA 8151A	500		500	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Merphos	n/a	=	107	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Merphos	n/a	=	106	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Merphos	n/a	=	1	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Merphos	n/a	=	33	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Merphos	n/a	=	80	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Merphos	n/a	=	83	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Methoxychlor	n/a	=	76	%	EPA 625m		0	155	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Methoxychlor	n/a	=	71	%	EPA 625m		0	155	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Methoxychlor	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Methoxychlor	n/a	=	0	%	EPA 625m		0	155	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Methoxychlor	n/a	=	0	%	EPA 625m		0	155	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Methoxychlor	n/a	=	0	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Methyl parathion	n/a	=	118	%	EPA 625m		60	120	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Methyl parathion	n/a	=	113	%	EPA 625m		60	120	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Methyl parathion	n/a	=	4	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Methyl parathion	n/a	=	79	%	EPA 625m		60	120	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Methyl parathion	n/a	=	77	%	EPA 625m		60	120	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Methyl parathion	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Mevinphos	n/a	=	99	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Mevinphos	n/a	=	97	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Mevinphos	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		0.008	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		0.008	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Mevinphos	n/a	=	42	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Mevinphos	n/a	=	87	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Mevinphos	n/a	=	70	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Mirex	n/a	=	91	%	EPA 625m		50	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Mirex	n/a	=	96	%	EPA 625m		50	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Mirex	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Mirex	n/a	=	15	%	EPA 625m		50	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Mirex	n/a	=	39	%	EPA 625m		50	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Mirex	n/a	=	89	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Oxychlorthane	n/a	=	90	%	EPA 625m		50	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Oxychlorthane	n/a	=	84	%	EPA 625m		50	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Oxychlorthane	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Oxychlorthane	n/a	=	93	%	EPA 625m		50	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Oxychlorthane	n/a	=	100	%	EPA 625m		50	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Oxychlorthane	n/a	=	7	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Phorate	n/a	=	107	%	EPA 625m		45	105	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Phorate	n/a	=	102	%	EPA 625m		45	105	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Phorate	n/a	=	5	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006			0.006
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006			0.006
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Phorate	n/a	=	21	%	EPA 625m		45	105	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Phorate	n/a	=	56	%	EPA 625m		45	105	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Phorate	n/a	=	91	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	103	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	109	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	6	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002			0.002
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002			0.002
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	25	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	78	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	103	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Tokuthion	n/a	=	102	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Tokuthion	n/a	=	105	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Tokuthion	n/a	=	3	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003			0.003
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003			0.003
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003			30
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Tokuthion	n/a	=	29	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Tokuthion	n/a	=	75	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Tokuthion	n/a	=	88	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Total Detectable DDTs	n/a	=	0.9154	µg/L	EPA 625m				
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Total Detectable DDTs	n/a	=	0.3287	µg/L	EPA 625m				
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01			
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01			0.01
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01			0.01
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01			30
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	trans-Nonachlor	n/a	=	93	%	EPA 625m		55	130	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	trans-Nonachlor	n/a	=	105	%	EPA 625m		55	130	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	trans-Nonachlor	n/a	=	12	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001			0.001

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	trans-Nonachlor	n/a	=	20	%	EPA 625m		55	130	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	trans-Nonachlor	n/a	=	65	%	EPA 625m		55	130	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	trans-Nonachlor	n/a	=	106	%	EPA 625m		0	30	
2007/08-2	A-1	field duplicate	1/15/2008	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001			
2007/08-2	Lab	LCS dup, rec	1/15/2008	Pesticide	Trichloronate	n/a	=	97	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, rec	1/15/2008	Pesticide	Trichloronate	n/a	=	99	%	EPA 625m		65	125	
2007/08-2	Lab	LCS, RPD	1/15/2008	Pesticide	Trichloronate	n/a	=	2	%	EPA 625m		0	30	
2007/08-2	Lab	method blank	1/15/2008	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-CC	field blank	1/15/2008	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-2	ME-SCR	lab duplicate	1/15/2008	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-2	ME-SCR	matrix spike dup, rec	1/15/2008	Pesticide	Trichloronate	n/a	=	22	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, rec	1/15/2008	Pesticide	Trichloronate	n/a	=	70	%	EPA 625m		65	125	
2007/08-2	ME-SCR	matrix spike, RPD	1/15/2008	Pesticide	Trichloronate	n/a	=	104	%	EPA 625m		0	30	
2007/08-3	Lab	LCS dup, rec	2/7/2008	Anion	Bromide	n/a	=	80	%	EPA 300.0		70	130	
2007/08-3	Lab	LCS, rec	2/7/2008	Anion	Bromide	n/a	=	80	%	EPA 300.0		70	130	
2007/08-3	Lab	LCS, RPD	2/7/2008	Anion	Bromide	n/a	=	0	%	EPA 300.0		0	30	
2007/08-3	Lab	method blank	2/7/2008	Anion	Bromide	n/a	<	0.001	mg/L	EPA 300.0	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/7/2008	Anion	Bromide	n/a	=	0.2	mg/L	EPA 300.0	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/7/2008	Anion	Bromide	n/a	=	80	%	EPA 300.0		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/7/2008	Anion	Bromide	n/a	=	80	%	EPA 300.0		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/7/2008	Anion	Bromide	n/a	=	0	%	EPA 300.0		0	30	
2007/08-3	Lab	LCS dup, rec	2/1/2008	Anion	Chloride	n/a	=	99	%	EPA 300.0		70	130	
2007/08-3	Lab	LCS, rec	2/1/2008	Anion	Chloride	n/a	=	96	%	EPA 300.0		70	130	
2007/08-3	Lab	LCS, RPD	2/1/2008	Anion	Chloride	n/a	=	3	%	EPA 300.0		0	30	
2007/08-3	Lab	method blank	2/1/2008	Anion	Chloride	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/1/2008	Anion	Chloride	n/a	=	35.6	mg/L	EPA 300.0	0.01		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/7/2008	Anion	Chloride	n/a	=	92	%	EPA 300.0		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/7/2008	Anion	Chloride	n/a	=	92	%	EPA 300.0		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/7/2008	Anion	Chloride	n/a	=	0	%	EPA 300.0		0	30	
2007/08-3	Lab	LCS dup, rec	2/1/2008	Anion	Perchlorate	n/a	=	102	%	EPA 314.0		85	115	
2007/08-3	Lab	LCS, rec	2/1/2008	Anion	Perchlorate	n/a	=	101	%	EPA 314.0		85	115	
2007/08-3	Lab	LCS, RPD	2/1/2008	Anion	Perchlorate	n/a	=	1	%	EPA 314.0		0	15	
2007/08-3	Lab	method blank	2/1/2008	Anion	Perchlorate	n/a	<	2	µg/L	EPA 314.0	2		2	
2007/08-3	ME-SCR	field blank	1/23/2008	Bacteriological	E. Coli	n/a	<	10	MPN/100 mL	MMO-MUG	10		10	
2007/08-3	ME-SCR	field blank	1/23/2008	Bacteriological	Enterococcus	n/a	<	10	MPN/100 mL	Enterolert	10		10	
2007/08-3	ME-SCR	field blank	1/23/2008	Bacteriological	Fecal Coliform	n/a	<	2	MPN/100 mL	SM 9221 E	2		2	
2007/08-3	ME-SCR	field blank	1/23/2008	Bacteriological	Total Coliform	n/a	<	10	MPN/100 mL	MMO-MUG	10		10	
2007/08-3	Lab	method blank	1/25/2008	Conventional	BOD	n/a	<	2	mg/L	EPA 405.1	2		2	
2007/08-3	ME-VR2	lab duplicate	1/25/2008	Conventional	BOD	n/a	=	5.8	mg/L	EPA 405.1	2		30	
2007/08-3	ME-CC	lab duplicate	1/25/2008	Conventional	Conductivity	n/a	=	397	µmhos/cm	SM 2510	0.001		30	
2007/08-3	Lab	method blank	2/18/2008	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1		1	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Conventional	Hardness as CaCO3	Total	=	64.5	mg/L	SM 2340 B	1		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Conventional	Hardness as CaCO3	Total	<	1	mg/L	SM 2340 B	1		1	
2007/08-3	ME-CC	lab duplicate	1/25/2008	Conventional	pH	n/a	=	7.9	pH Units	SM 4500 H+	0.1		30	
2007/08-3	Lab	LCS dup, rec	1/30/2008	Conventional	Total Dissolved Solids	n/a	=	102	%	SM 2540 C		70	130	
2007/08-3	Lab	LCS, rec	1/30/2008	Conventional	Total Dissolved Solids	n/a	=	90	%	SM 2540 C		70	130	
2007/08-3	Lab	LCS, RPD	1/30/2008	Conventional	Total Dissolved Solids	n/a	=	12	%	SM 2540 C		0	30	
2007/08-3	Lab	method blank	1/30/2008	Conventional	Total Dissolved Solids	n/a	<	0.1	mg/L	SM 2540 C	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	1/30/2008	Conventional	Total Dissolved Solids	n/a	=	273	mg/L	SM 2540 C	0.1		30	
2007/08-3	Lab	LCS dup, rec	2/4/2008	Conventional	Total Organic Carbon	n/a	=	108	%	EPA 415.1		50	150	
2007/08-3	Lab	LCS, rec	2/4/2008	Conventional	Total Organic Carbon	n/a	=	106	%	EPA 415.1		50	150	
2007/08-3	Lab	LCS, RPD	2/4/2008	Conventional	Total Organic Carbon	n/a	=	2	%	EPA 415.1		0	30	
2007/08-3	Lab	method blank	2/4/2008	Conventional	Total Organic Carbon	n/a	<	0.1	mg/L	EPA 415.1	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/7/2008	Conventional	Total Organic Carbon	n/a	=	19.7	mg/L	EPA 415.1	0.1		30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	matrix spike dup, rec	2/7/2008	Conventional	Total Organic Carbon	n/a	=	105	%	EPA 415.1		50	150	
2007/08-3	ME-CC	matrix spike, rec	2/7/2008	Conventional	Total Organic Carbon	n/a	=	109	%	EPA 415.1		50	150	
2007/08-3	ME-CC	matrix spike, RPD	2/7/2008	Conventional	Total Organic Carbon	n/a	=	4	%	EPA 415.1		0	30	
2007/08-3	Lab	method blank	1/31/2008	Conventional	Total Suspended Solids	n/a	<	0.5	mg/L	SM 2540 D	0.5		0.5	
2007/08-3	ME-CC	lab duplicate	1/31/2008	Conventional	Total Suspended Solids	n/a	=	1190	mg/L	SM 2540 D	0.5		30	
2007/08-3	Lab	method blank	1/25/2008	Conventional	Turbidity	n/a	<	1	NTU	EPA 180.1	1		1	
2007/08-3	ME-CC	lab duplicate	1/25/2008	Conventional	Turbidity	n/a	=	1112	NTU	EPA 180.1	1		30	
2007/08-3	Lab	LCS dup, rec	2/8/2008	Hydrocarbon	Oil and Grease	n/a	=	100	%	EPA 1664A		70	130	
2007/08-3	Lab	LCS, rec	2/8/2008	Hydrocarbon	Oil and Grease	n/a	=	97	%	EPA 1664A		70	130	
2007/08-3	Lab	LCS, RPD	2/8/2008	Hydrocarbon	Oil and Grease	n/a	=	3	%	EPA 1664A		0	30	
2007/08-3	Lab	method blank	2/8/2008	Hydrocarbon	Oil and Grease	n/a	<	1	mg/L	EPA 1664A	1		1	
2007/08-3	Lab	LCS dup, rec	2/14/2008	Hydrocarbon	TRPH	n/a	=	97	%	EPA 1664		70	130	
2007/08-3	Lab	LCS, rec	2/14/2008	Hydrocarbon	TRPH	n/a	=	93	%	EPA 1664		70	130	
2007/08-3	Lab	LCS, RPD	2/14/2008	Hydrocarbon	TRPH	n/a	=	4	%	EPA 1664		0	30	
2007/08-3	Lab	method blank	2/14/2008	Hydrocarbon	TRPH	n/a	<	1	mg/L	EPA 1664	1		1	
2007/08-3	Lab	method blank	2/18/2008	Metal	Aluminum	Dissolved	<	5	µg/L	EPA 200.8m	5		5	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Aluminum	Dissolved	=	16	µg/L	EPA 200.8m	5		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/18/2008	Metal	Aluminum	Dissolved	=	106	%	EPA 200.8m		50	140	
2007/08-3	ME-CC	matrix spike, rec	2/18/2008	Metal	Aluminum	Dissolved	=	103	%	EPA 200.8m		50	140	
2007/08-3	ME-CC	matrix spike, RPD	2/18/2008	Metal	Aluminum	Dissolved	=	3	%	EPA 200.8m		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Aluminum	Total	<	5	µg/L	EPA 200.8m	5		5	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Aluminum	Total	=	6647	µg/L	EPA 200.8m	5		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Metal	Aluminum	Total	<	5	µg/L	EPA 200.8m	5		5	
2007/08-3	Lab	method blank	2/18/2008	Metal	Arsenic	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Arsenic	Dissolved	=	2.7	µg/L	EPA 200.8m	0.2		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/18/2008	Metal	Arsenic	Dissolved	=	120	%	EPA 200.8m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/18/2008	Metal	Arsenic	Dissolved	=	120	%	EPA 200.8m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/18/2008	Metal	Arsenic	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Arsenic	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Arsenic	Total	=	4.6	µg/L	EPA 200.8m	0.2		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Metal	Arsenic	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	Lab	method blank	2/18/2008	Metal	Cadmium	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Cadmium	Dissolved	=	0.2	µg/L	EPA 200.8m	0.2		30	EST
2007/08-3	ME-CC	matrix spike dup, rec	2/18/2008	Metal	Cadmium	Dissolved	=	97	%	EPA 200.8m		75	130	
2007/08-3	ME-CC	matrix spike, rec	2/18/2008	Metal	Cadmium	Dissolved	=	97	%	EPA 200.8m		75	130	
2007/08-3	ME-CC	matrix spike, RPD	2/18/2008	Metal	Cadmium	Dissolved	=	0	%	EPA 200.8m		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Cadmium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Cadmium	Total	=	2.7	µg/L	EPA 200.8m	0.2		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Metal	Cadmium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	Lab	method blank	2/18/2008	Metal	Chromium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Chromium	Dissolved	=	0.4	µg/L	EPA 200.8m	0.1		30	EST
2007/08-3	ME-CC	matrix spike dup, rec	2/18/2008	Metal	Chromium	Dissolved	=	103	%	EPA 200.8m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/18/2008	Metal	Chromium	Dissolved	=	102	%	EPA 200.8m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/18/2008	Metal	Chromium	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Chromium	Total	=	14.5	µg/L	EPA 200.8m	0.1		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Metal	Chromium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-3	Lab	LCS dup, rec	2/7/2008	Metal	Chromium VI	Total	=	100	%	SM 3500-Cr D		70	130	
2007/08-3	Lab	LCS, rec	2/7/2008	Metal	Chromium VI	Total	=	97	%	SM 3500-Cr D		70	130	
2007/08-3	Lab	LCS, RPD	2/7/2008	Metal	Chromium VI	Total	=	3	%	SM 3500-Cr D		0	30	
2007/08-3	Lab	method blank	2/7/2008	Metal	Chromium VI	Total	<	5	µg/L	SM 3500-Cr D	5		5	
2007/08-3	ME-CC	lab duplicate	2/7/2008	Metal	Chromium VI	Total	=	33	µg/L	SM 3500-Cr D	5		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/7/2008	Metal	Chromium VI	Total	=	89	%	SM 3500-Cr D		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/7/2008	Metal	Chromium VI	Total	=	96	%	SM 3500-Cr D		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/7/2008	Metal	Chromium VI	Total	=	7.6	%	SM 3500-Cr D		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Copper	Dissolved	<	0.4	µg/L	EPA 200.8m	0.4		0.4	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Copper	Dissolved	=	3.5	µg/L	EPA 200.8m	0.4		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/18/2008	Metal	Copper	Dissolved	=	93	%	EPA 200.8m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/18/2008	Metal	Copper	Dissolved	=	92	%	EPA 200.8m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/18/2008	Metal	Copper	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Copper	Total	<	0.4	µg/L	EPA 200.8m	0.4		0.4	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Copper	Total	=	49.6	µg/L	EPA 200.8m	0.4		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Metal	Copper	Total	<	0.4	µg/L	EPA 200.8m	0.4		0.4	
2007/08-3	Lab	method blank	2/18/2008	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Lead	Dissolved	<	0.05	µg/L	EPA 200.8m	0.05		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/18/2008	Metal	Lead	Dissolved	=	94	%	EPA 200.8m		65	135	
2007/08-3	ME-CC	matrix spike, rec	2/18/2008	Metal	Lead	Dissolved	=	93	%	EPA 200.8m		65	135	
2007/08-3	ME-CC	matrix spike, RPD	2/18/2008	Metal	Lead	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Lead	Total	=	14.58	µg/L	EPA 200.8m	0.05		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Metal	Lead	Total	<	0.05	µg/L	EPA 200.8m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/8/2008	Metal	Mercury	Dissolved	=	116	%	EPA 1631Em		60	140	
2007/08-3	Lab	LCS, rec	2/8/2008	Metal	Mercury	Dissolved	=	110	%	EPA 1631Em		60	140	
2007/08-3	Lab	LCS, RPD	2/8/2008	Metal	Mercury	Dissolved	=	5.3	%	EPA 1631Em		0	30	
2007/08-3	Lab	method blank	2/8/2008	Metal	Mercury	Dissolved	<	0.5	ng/L	EPA 1631Em	0.5		0.5	
2007/08-3	ME-CC	lab duplicate	2/8/2008	Metal	Mercury	Dissolved	=	3	ng/L	EPA 1631Em	0.5		30	
2007/08-3	ME-SCR	field blank	2/8/2008	Metal	Mercury	Dissolved	<	0.5	ng/L	EPA 1631Em	0.5		0.5	
2007/08-3	Lab	method blank	2/8/2008	Metal	Mercury	Total	<	0.5	ng/L	EPA 1631Em	0.5		0.5	
2007/08-3	ME-CC	lab duplicate	2/8/2008	Metal	Mercury	Total	=	174.2	ng/L	EPA 1631Em	0.5		30	
2007/08-3	ME-SCR	field blank	2/8/2008	Metal	Mercury	Total	=	0.6	ng/L	EPA 1631Em	0.5		0.5	EST
2007/08-3	Lab	method blank	2/18/2008	Metal	Nickel	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Nickel	Dissolved	=	2.7	µg/L	EPA 200.8m	0.2		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/18/2008	Metal	Nickel	Dissolved	=	93	%	EPA 200.8m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/18/2008	Metal	Nickel	Dissolved	=	92	%	EPA 200.8m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/18/2008	Metal	Nickel	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Nickel	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Nickel	Total	=	38.8	µg/L	EPA 200.8m	0.2		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Metal	Nickel	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	Lab	method blank	2/18/2008	Metal	Selenium	Dissolved	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Selenium	Dissolved	=	1.3	µg/L	EPA 200.8m	0.2		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/18/2008	Metal	Selenium	Dissolved	=	108	%	EPA 200.8m		60	150	
2007/08-3	ME-CC	matrix spike, rec	2/18/2008	Metal	Selenium	Dissolved	=	107	%	EPA 200.8m		60	150	
2007/08-3	ME-CC	matrix spike, RPD	2/18/2008	Metal	Selenium	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Selenium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Selenium	Total	=	1.4	µg/L	EPA 200.8m	0.2		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Metal	Selenium	Total	<	0.2	µg/L	EPA 200.8m	0.2		0.2	
2007/08-3	Lab	method blank	2/18/2008	Metal	Silver	Dissolved	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Silver	Dissolved	<	0.5	µg/L	EPA 200.8m	0.5		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/18/2008	Metal	Silver	Dissolved	=	96	%	EPA 200.8m		50	155	
2007/08-3	ME-CC	matrix spike, rec	2/18/2008	Metal	Silver	Dissolved	=	114	%	EPA 200.8m		50	155	
2007/08-3	ME-CC	matrix spike, RPD	2/18/2008	Metal	Silver	Dissolved	=	17	%	EPA 200.8m		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Metal	Silver	Total	<	0.5	µg/L	EPA 200.8m	0.5		0.5	
2007/08-3	Lab	method blank	2/18/2008	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Thallium	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/18/2008	Metal	Thallium	Dissolved	=	96	%	EPA 200.8m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/18/2008	Metal	Thallium	Dissolved	=	95	%	EPA 200.8m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/18/2008	Metal	Thallium	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Metal	Thallium	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	method blank	2/18/2008	Metal	Zinc	Dissolved	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Zinc	Dissolved	=	3.3	µg/L	EPA 200.8m	0.1		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/18/2008	Metal	Zinc	Dissolved	=	110	%	EPA 200.8m		50	150	
2007/08-3	ME-CC	matrix spike, rec	2/18/2008	Metal	Zinc	Dissolved	=	109	%	EPA 200.8m		50	150	
2007/08-3	ME-CC	matrix spike, RPD	2/18/2008	Metal	Zinc	Dissolved	=	1	%	EPA 200.8m		0	30	
2007/08-3	Lab	method blank	2/18/2008	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Metal	Zinc	Total	=	128.3	µg/L	EPA 200.8m	0.1		30	
2007/08-3	ME-SCR	field blank	2/18/2008	Metal	Zinc	Total	<	0.1	µg/L	EPA 200.8m	0.1		0.1	
2007/08-3	Lab	LCS dup, rec	2/4/2008	Nutrient	Ammonia as N	n/a	=	84	%	SM 4500-NH3 F		70	130	
2007/08-3	Lab	LCS, rec	2/4/2008	Nutrient	Ammonia as N	n/a	=	80	%	SM 4500-NH3 F		70	130	
2007/08-3	Lab	LCS, RPD	2/4/2008	Nutrient	Ammonia as N	n/a	=	5	%	SM 4500-NH3 F		0	30	
2007/08-3	Lab	method blank	2/4/2008	Nutrient	Ammonia as N	n/a	<	0.03	mg/L	SM 4500-NH3 F	0.03		0.03	
2007/08-3	ME-CC	lab duplicate	2/4/2008	Nutrient	Ammonia as N	n/a	=	0.24	mg/L	SM 4500-NH3 F	0.03		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/4/2008	Nutrient	Ammonia as N	n/a	=	98	%	SM 4500-NH3 F		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/4/2008	Nutrient	Ammonia as N	n/a	=	96	%	SM 4500-NH3 F		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/4/2008	Nutrient	Ammonia as N	n/a	=	2	%	SM 4500-NH3 F		0	30	
2007/08-3	Lab	LCS dup, rec	1/26/2008	Nutrient	Nitrate as N	n/a	=	110	%	EPA 300.0		70	130	
2007/08-3	Lab	LCS, rec	1/26/2008	Nutrient	Nitrate as N	n/a	=	110	%	EPA 300.0		70	130	
2007/08-3	Lab	LCS, RPD	1/26/2008	Nutrient	Nitrate as N	n/a	=	0	%	EPA 300.0		0	30	
2007/08-3	Lab	method blank	1/26/2008	Nutrient	Nitrate as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	1/26/2008	Nutrient	Nitrate as N	n/a	=	2.68	mg/L	EPA 300.0	0.01		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/7/2008	Nutrient	Nitrate as N	n/a	=	123	%	EPA 300.0		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/7/2008	Nutrient	Nitrate as N	n/a	=	127	%	EPA 300.0		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/7/2008	Nutrient	Nitrate as N	n/a	=	3	%	EPA 300.0		0	30	
2007/08-3	Lab	LCS dup, rec	1/26/2008	Nutrient	Nitrite as N	n/a	=	72	%	EPA 300.0		70	130	
2007/08-3	Lab	LCS, rec	1/26/2008	Nutrient	Nitrite as N	n/a	=	72	%	EPA 300.0		70	130	
2007/08-3	Lab	LCS, RPD	1/26/2008	Nutrient	Nitrite as N	n/a	=	0	%	EPA 300.0		0	30	
2007/08-3	Lab	method blank	1/26/2008	Nutrient	Nitrite as N	n/a	<	0.01	mg/L	EPA 300.0	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	1/26/2008	Nutrient	Nitrite as N	n/a	=	0.03	mg/L	EPA 300.0	0.01		30	EST
2007/08-3	ME-CC	matrix spike dup, rec	1/26/2008	Nutrient	Nitrite as N	n/a	=	66	%	EPA 300.0		70	130	
2007/08-3	ME-CC	matrix spike, rec	1/26/2008	Nutrient	Nitrite as N	n/a	=	67	%	EPA 300.0		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/7/2008	Nutrient	Nitrite as N	n/a	=	2	%	EPA 300.0		0	30	
2007/08-3	Lab	LCS dup, rec	1/26/2008	Nutrient	Orthophosphate as P (Diss)	n/a	=	94	%	EPA 300.0		70	130	
2007/08-3	Lab	LCS, rec	1/26/2008	Nutrient	Orthophosphate as P (Diss)	n/a	=	96	%	EPA 300.0		70	130	
2007/08-3	Lab	LCS, RPD	1/26/2008	Nutrient	Orthophosphate as P (Diss)	n/a	=	2	%	EPA 300.0		0	30	
2007/08-3	Lab	method blank	1/26/2008	Nutrient	Orthophosphate as P (Diss)	n/a	<	0.0075	mg/L	EPA 300.0	0.0075		0.0075	
2007/08-3	ME-CC	lab duplicate	1/26/2008	Nutrient	Orthophosphate as P (Diss)	n/a	=	0.2218	mg/L	EPA 300.0	0.0075		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/7/2008	Nutrient	Orthophosphate as P (Diss)	n/a	=	86	%	EPA 300.0		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/7/2008	Nutrient	Orthophosphate as P (Diss)	n/a	=	85	%	EPA 300.0		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/7/2008	Nutrient	Orthophosphate as P (Diss)	n/a	=	1	%	EPA 300.0		0	30	
2007/08-3	Lab	LCS, rec	2/18/2008	Nutrient	TKN	n/a	=	93.5	%	EPA 351.1		80	120	
2007/08-3	Lab	method blank	2/18/2008	Nutrient	TKN	n/a	<	0.05	mg/L	EPA 351.1	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/18/2008	Nutrient	TKN	n/a	=	0.81	mg/L	EPA 351.1	0.05		20	
2007/08-3	ME-VR2	matrix spike dup, rec	2/18/2008	Nutrient	TKN	n/a	=	90.6	%	EPA 351.1		80	120	
2007/08-3	ME-VR2	matrix spike, rec	2/18/2008	Nutrient	TKN	n/a	=	94.2	%	EPA 351.1		80	120	
2007/08-3	ME-VR2	matrix spike, RPD	2/18/2008	Nutrient	TKN	n/a	=	3.9	%	EPA 351.1		0	20	
2007/08-3	Lab	LCS dup, rec	2/4/2008	Nutrient	Total Phosphorus	Dissolved	=	73	%	SM 4500-P C		70	130	
2007/08-3	Lab	LCS, rec	2/4/2008	Nutrient	Total Phosphorus	Dissolved	=	73	%	SM 4500-P C		70	130	
2007/08-3	Lab	LCS, RPD	2/4/2008	Nutrient	Total Phosphorus	Dissolved	=	0	%	SM 4500-P C		0	30	
2007/08-3	Lab	method blank	2/4/2008	Nutrient	Total Phosphorus	Dissolved	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2007/08-3	ME-VR2	lab duplicate	1/25/2008	Nutrient	Total Phosphorus	Dissolved	=	0.14	mg/L	SM 4500-P C	0.016		30	
2007/08-3	ME-VR2	matrix spike dup, rec	2/7/2008	Nutrient	Total Phosphorus	Dissolved	=	100	%	SM 4500-P C		70	130	
2007/08-3	ME-VR2	matrix spike, rec	2/7/2008	Nutrient	Total Phosphorus	Dissolved	=	94	%	SM 4500-P C		70	130	
2007/08-3	ME-VR2	matrix spike, RPD	2/7/2008	Nutrient	Total Phosphorus	Dissolved	=	6	%	SM 4500-P C		0	30	
2007/08-3	Lab	LCS dup, rec	2/4/2008	Nutrient	Total Phosphorus	Total	=	70	%	SM 4500-P C		70	130	
2007/08-3	Lab	LCS, rec	2/4/2008	Nutrient	Total Phosphorus	Total	=	70	%	SM 4500-P C		70	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	LCS, RPD	2/4/2008	Nutrient	Total Phosphorus	Total	=	0	%	SM 4500-P C		0	30	
2007/08-3	Lab	method blank	2/4/2008	Nutrient	Total Phosphorus	Total	<	0.016	mg/L	SM 4500-P C	0.016		0.016	
2007/08-3	ME-CC	lab duplicate	2/7/2008	Nutrient	Total Phosphorus	Total	=	3.051	mg/L	SM 4500-P C	0.016		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/7/2008	Nutrient	Total Phosphorus	Total	=	104	%	SM 4500-P C		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/7/2008	Nutrient	Total Phosphorus	Total	=	96	%	SM 4500-P C		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/7/2008	Nutrient	Total Phosphorus	Total	=	8	%	SM 4500-P C		0	30	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	93	%	EPA 625m		45	140	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	99	%	EPA 625m		45	140	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	1,2,4-Trichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	84	%	EPA 625m		45	140	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	88	%	EPA 625m		45	140	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	1,2,4-Trichlorobenzene	n/a	=	0.023	µg/L	EPA 625m	0.01		0.01	EST
2007/08-3	Lab	method blank	2/23/2008	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	1,2-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	1,2-Dichlorobenzene	n/a	=	0.044	µg/L	EPA 625m	0.01		0.01	EST
2007/08-3	Lab	method blank	2/23/2008	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	1,3-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	1,3-Dichlorobenzene	n/a	=	0.057	µg/L	EPA 625m	0.01		0.01	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	1,4-Dichlorobenzene	n/a	=	97	%	EPA 625m		45	140	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	1,4-Dichlorobenzene	n/a	=	99	%	EPA 625m		45	140	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	1,4-Dichlorobenzene	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	1,4-Dichlorobenzene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	1,4-Dichlorobenzene	n/a	=	74	%	EPA 625m		45	140	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	1,4-Dichlorobenzene	n/a	=	79	%	EPA 625m		45	140	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	1,4-Dichlorobenzene	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	1,4-Dichlorobenzene	n/a	=	0.049	µg/L	EPA 625m	0.01		0.01	EST
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	1-Methylnaphthalene	n/a	=	88	%	EPA 625m		50	120	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	1-Methylnaphthalene	n/a	=	92	%	EPA 625m		50	120	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	1-Methylnaphthalene	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	1-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	1-Methylnaphthalene	n/a	=	0.0116	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	1-Methylnaphthalene	n/a	=	80	%	EPA 625m		50	120	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	1-Methylnaphthalene	n/a	=	79	%	EPA 625m		50	120	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	1-Methylnaphthalene	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	1-Methylnaphthalene	n/a	=	0.0028	µg/L	EPA 625m	0.001		0.001	EST
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	1-Methylphenanthrene	n/a	=	101	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	1-Methylphenanthrene	n/a	=	98	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	1-Methylphenanthrene	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	1-Methylphenanthrene	n/a	=	103	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	1-Methylphenanthrene	n/a	=	98	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	1-Methylphenanthrene	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	1-Methylphenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	97	%	EPA 625m		45	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	100	%	EPA 625m		45	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	0.0075	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	91	%	EPA 625m		45	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	92	%	EPA 625m		45	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	=	1	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2,3,5-Trimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	srgt LCS dup, rec	2/23/2008	Organic	2,4,6-Tribromophenol	n/a	=	87	%	EPA 625m		40	130	
2007/08-3	Lab	srgt LCS, rec	2/23/2008	Organic	2,4,6-Tribromophenol	n/a	=	86	%	EPA 625m		40	130	
2007/08-3	Lab	srgt method blank, rec	2/23/2008	Organic	2,4,6-Tribromophenol	n/a	=	86	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	2,4,6-Tribromophenol	n/a	=	96	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	2,4,6-Tribromophenol	n/a	=	80	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt matrix spike dup, rec	2/23/2008	Organic	2,4,6-Tribromophenol	n/a	=	99	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt matrix spike, rec	2/23/2008	Organic	2,4,6-Tribromophenol	n/a	=	100	%	EPA 625m		40	130	
2007/08-3	ME-SCR	srgt environ, rec	2/23/2008	Organic	2,4,6-Tribromophenol	n/a	=	77	%	EPA 625m		40	130	
2007/08-3	ME-SCR	srgt field blank, rec	2/23/2008	Organic	2,4,6-Tribromophenol	n/a	=	84	%	EPA 625m		40	130	
2007/08-3	ME-VR2	srgt environ, rec	2/23/2008	Organic	2,4,6-Tribromophenol	n/a	=	78	%	EPA 625m		40	130	
2007/08-3	Lab	method blank	2/23/2008	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2,4,6-Trichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	method blank	2/23/2008	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2,4-Dichlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	srgt method blank, rec	1/31/2008	Organic	2,4-Dichlorophenylacetic acid	n/a	=	70	%	EPA 8151A		0	123	
2007/08-3	ME-CC	srgt environ, rec	1/31/2008	Organic	2,4-Dichlorophenylacetic acid	n/a	=	5	%	EPA 8151A		0	123	
2007/08-3	ME-SCR	srgt environ, rec	1/31/2008	Organic	2,4-Dichlorophenylacetic acid	n/a	=	7	%	EPA 8151A		0	123	
2007/08-3	ME-VR2	srgt environ, rec	1/31/2008	Organic	2,4-Dichlorophenylacetic acid	n/a	=	6	%	EPA 8151A		0	123	
2007/08-3	Lab	method blank	2/23/2008	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2,4-Dimethylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	Lab	method blank	2/23/2008	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2,4-Dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	2,4-Dinitrotoluene	n/a	=	89	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	2,4-Dinitrotoluene	n/a	=	91	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	2,4-Dinitrotoluene	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	2,4-Dinitrotoluene	n/a	=	93	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	2,4-Dinitrotoluene	n/a	=	89	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	2,4-Dinitrotoluene	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2,4-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	87	%	EPA 625m		55	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	94	%	EPA 625m		55	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	8	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	2,6-Dimethylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	0.0094	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	85	%	EPA 625m		55	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	85	%	EPA 625m		55	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2,6-Dimethylnaphthalene	n/a	=	0.0036	µg/L	EPA 625m	0.001		0.001	EST
2007/08-3	Lab	method blank	2/23/2008	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2,6-Dinitrotoluene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	method blank	2/23/2008	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2-Chloronaphthalene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	2-Chlorophenol	n/a	=	106	%	EPA 625m		35	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	2-Chlorophenol	n/a	=	110	%	EPA 625m		35	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	2-Chlorophenol	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	2-Chlorophenol	n/a	=	79	%	EPA 625m		35	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	2-Chlorophenol	n/a	=	81	%	EPA 625m		35	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	2-Chlorophenol	n/a	=	4.5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2-Chlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	method blank	2/23/2008	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2-Methyl-4,6-dinitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	2-Methylnaphthalene	n/a	=	97	%	EPA 625m		50	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	2-Methylnaphthalene	n/a	=	89	%	EPA 625m		50	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	2-Methylnaphthalene	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	2-Methylnaphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2-Methylnaphthalene	n/a	=	0.0152	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	2-Methylnaphthalene	n/a	=	93	%	EPA 625m		50	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	2-Methylnaphthalene	n/a	=	97	%	EPA 625m		50	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	2-Methylnaphthalene	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2-Methylnaphthalene	n/a	=	0.0098	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	method blank	2/23/2008	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	2-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	Lab	method blank	2/23/2008	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	3,3'-Dichlorobenzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	method blank	2/23/2008	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	4-Bromophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	4-Chloro-3-methylphenol	n/a	=	112	%	EPA 625m		30	150	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	4-Chloro-3-methylphenol	n/a	=	112	%	EPA 625m		30	150	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	4-Chloro-3-methylphenol	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	4-Chloro-3-methylphenol	n/a	=	97	%	EPA 625m		30	150	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	4-Chloro-3-methylphenol	n/a	=	100	%	EPA 625m		30	150	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	4-Chloro-3-methylphenol	n/a	=	3.8	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	4-Chloro-3-methylphenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	Lab	method blank	2/23/2008	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	4-Chlorophenyl phenyl ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	4-Nitrophenol	n/a	=	76	%	EPA 625m		0	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	4-Nitrophenol	n/a	=	86	%	EPA 625m		0	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	4-Nitrophenol	n/a	=	12	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	4-Nitrophenol	n/a	=	36	%	EPA 625m		0	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	4-Nitrophenol	n/a	=	33	%	EPA 625m		0	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	4-Nitrophenol	n/a	=	9.7	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	4-Nitrophenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Acenaphthene	n/a	=	108	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Acenaphthene	n/a	=	110	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Acenaphthene	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Acenaphthene	n/a	=	105	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Acenaphthene	n/a	=	110	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Acenaphthene	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Acenaphthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	srgt LCS dup, rec	2/23/2008	Organic	Acenaphthene-d10	n/a	=	94	%	EPA 625m		50	130	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	srgt LCS, rec	2/23/2008	Organic	Acenaphthene-d10	n/a	=	75	%	EPA 625m		50	130	
2007/08-3	Lab	srgt method blank, rec	2/23/2008	Organic	Acenaphthene-d10	n/a	=	108	%	EPA 625m		50	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Acenaphthene-d10	n/a	=	71	%	EPA 625m		50	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Acenaphthene-d10	n/a	=	89	%	EPA 625m		50	130	
2007/08-3	ME-CC	srgt matrix spike dup, rec	2/23/2008	Organic	Acenaphthene-d10	n/a	=	84	%	EPA 625m		50	130	
2007/08-3	ME-CC	srgt matrix spike, rec	2/23/2008	Organic	Acenaphthene-d10	n/a	=	85	%	EPA 625m		50	130	
2007/08-3	ME-SCR	srgt environ, rec	2/23/2008	Organic	Acenaphthene-d10	n/a	=	71	%	EPA 625m		50	130	
2007/08-3	ME-SCR	srgt field blank, rec	2/23/2008	Organic	Acenaphthene-d10	n/a	=	86	%	EPA 625m		50	130	
2007/08-3	ME-VR2	srgt environ, rec	2/23/2008	Organic	Acenaphthene-d10	n/a	=	65	%	EPA 625m		50	130	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Acenaphthylene	n/a	=	94	%	EPA 625m		60	120	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Acenaphthylene	n/a	=	91	%	EPA 625m		60	120	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Acenaphthylene	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Acenaphthylene	n/a	=	0.0046	µg/L	EPA 625m	0.001		30	EST
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Acenaphthylene	n/a	=	91	%	EPA 625m		60	120	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Acenaphthylene	n/a	=	92	%	EPA 625m		60	120	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Acenaphthylene	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Acenaphthylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Anthracene	n/a	=	89	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Anthracene	n/a	=	99	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Anthracene	n/a	=	11	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Anthracene	n/a	=	0.0081	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Anthracene	n/a	=	78	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Anthracene	n/a	=	85	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Anthracene	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Anthracene	n/a	=	0.0018	µg/L	EPA 625m	0.001		0.001	EST
2007/08-3	Lab	method blank	2/23/2008	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Azobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	method blank	2/23/2008	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Benzidine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Benzo(a)anthracene	n/a	=	89	%	EPA 625m		70	140	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Benzo(a)anthracene	n/a	=	96	%	EPA 625m		70	140	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Benzo(a)anthracene	n/a	=	8	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Benzo(a)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Benzo(a)anthracene	n/a	=	0.0211	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Benzo(a)anthracene	n/a	=	107	%	EPA 625m		70	140	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Benzo(a)anthracene	n/a	=	108	%	EPA 625m		70	140	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Benzo(a)anthracene	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Benzo(a)anthracene	n/a	=	0.0029	µg/L	EPA 625m	0.001		0.001	EST
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Benzo(a)pyrene	n/a	=	87	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Benzo(a)pyrene	n/a	=	101	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Benzo(a)pyrene	n/a	=	15	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Benzo(a)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Benzo(a)pyrene	n/a	=	0.0354	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Benzo(a)pyrene	n/a	=	107	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Benzo(a)pyrene	n/a	=	96	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Benzo(a)pyrene	n/a	=	11	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Benzo(a)pyrene	n/a	=	0.0048	µg/L	EPA 625m	0.001		0.001	EST
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Benzo(b)fluoranthene	n/a	=	96	%	EPA 625m		60	140	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Benzo(b)fluoranthene	n/a	=	99	%	EPA 625m		60	140	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Benzo(b)fluoranthene	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Benzo(b)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Benzo(b)fluoranthene	n/a	=	0.0548	µg/L	EPA 625m	0.001		30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Benzo(b)fluoranthene	n/a	=	125	%	EPA 625m		60	140	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Benzo(b)fluoranthene	n/a	=	109	%	EPA 625m		60	140	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Benzo(b)fluoranthene	n/a	=	14	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Benzo(b)fluoranthene	n/a	=	0.0043	µg/L	EPA 625m	0.001		0.001	EST
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Benzo(e)pyrene	n/a	=	87	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Benzo(e)pyrene	n/a	=	104	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Benzo(e)pyrene	n/a	=	18	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Benzo(e)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Benzo(e)pyrene	n/a	=	0.0585	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Benzo(e)pyrene	n/a	=	103	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Benzo(e)pyrene	n/a	=	94	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Benzo(e)pyrene	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Benzo(e)pyrene	n/a	=	0.006	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Benzo(g,h,i)perylene	n/a	=	99	%	EPA 625m		50	140	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Benzo(g,h,i)perylene	n/a	=	113	%	EPA 625m		50	140	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Benzo(g,h,i)perylene	n/a	=	13	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Benzo(g,h,i)perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Benzo(g,h,i)perylene	n/a	=	0.1045	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Benzo(g,h,i)perylene	n/a	=	132	%	EPA 625m		50	140	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Benzo(g,h,i)perylene	n/a	=	119	%	EPA 625m		50	140	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Benzo(g,h,i)perylene	n/a	=	10	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Benzo(g,h,i)perylene	n/a	=	0.0139	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Benzo(k)fluoranthene	n/a	=	90	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Benzo(k)fluoranthene	n/a	=	102	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Benzo(k)fluoranthene	n/a	=	12	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Benzo(k)fluoranthene	n/a	=	0.025	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Benzo(k)fluoranthene	n/a	=	111	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Benzo(k)fluoranthene	n/a	=	93	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Benzo(k)fluoranthene	n/a	=	18	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Benzo(k)fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Biphenyl	n/a	=	96	%	EPA 625m		50	120	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Biphenyl	n/a	=	93	%	EPA 625m		50	120	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Biphenyl	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Biphenyl	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Biphenyl	n/a	=	0.0091	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Biphenyl	n/a	=	84	%	EPA 625m		50	120	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Biphenyl	n/a	=	92	%	EPA 625m		50	120	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Biphenyl	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Biphenyl	n/a	=	0.0014	µg/L	EPA 625m	0.001		0.001	EST
2007/08-3	Lab	method blank	2/23/2008	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Bis(2-chloroethoxy)methane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	method blank	2/23/2008	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Bis(2-chloroethyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	method blank	2/23/2008	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Bis(2-chloroisopropyl)ether	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	92	%	EPA 625m		20	190	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	90	%	EPA 625m		20	190	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	11.1612	µg/L	EPA 625m	0.1		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Bis(2-ethylhexyl)phthalate	n/a	=	0.1173	µg/L	EPA 625m	0.1		0.1	EST
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Butyl benzyl phthalate	n/a	=	96	%	EPA 625m		65	160	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Butyl benzyl phthalate	n/a	=	96	%	EPA 625m		65	160	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Butyl benzyl phthalate	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Butyl benzyl phthalate	n/a	<	0.025	µg/L	EPA 625m	0.025		0.025	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Butyl benzyl phthalate	n/a	=	0.2904	µg/L	EPA 625m	0.025		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Butyl benzyl phthalate	n/a	=	133	%	EPA 625m		65	160	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Butyl benzyl phthalate	n/a	=	133	%	EPA 625m		65	160	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Butyl benzyl phthalate	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Butyl benzyl phthalate	n/a	<	0.025	µg/L	EPA 625m	0.025		0.025	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Chrysene	n/a	=	92	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Chrysene	n/a	=	84	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Chrysene	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Chrysene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Chrysene	n/a	=	0.0605	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Chrysene	n/a	=	101	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Chrysene	n/a	=	90	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Chrysene	n/a	=	12	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Chrysene	n/a	=	0.0016	µg/L	EPA 625m	0.001		0.001	EST
2007/08-3	Lab	srgt LCS dup, rec	2/23/2008	Organic	Chrysene-d12	n/a	=	86	%	EPA 625m		70	130	
2007/08-3	Lab	srgt LCS, rec	2/23/2008	Organic	Chrysene-d12	n/a	=	102	%	EPA 625m		70	130	
2007/08-3	Lab	srgt method blank, rec	2/23/2008	Organic	Chrysene-d12	n/a	=	76	%	EPA 625m		70	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Chrysene-d12	n/a	=	103	%	EPA 625m		70	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Chrysene-d12	n/a	=	98	%	EPA 625m		70	130	
2007/08-3	ME-CC	srgt matrix spike dup, rec	2/23/2008	Organic	Chrysene-d12	n/a	=	99	%	EPA 625m		70	130	
2007/08-3	ME-CC	srgt matrix spike, rec	2/23/2008	Organic	Chrysene-d12	n/a	=	100	%	EPA 625m		70	130	
2007/08-3	ME-SCR	srgt environ, rec	2/23/2008	Organic	Chrysene-d12	n/a	=	90	%	EPA 625m		70	130	
2007/08-3	ME-SCR	srgt field blank, rec	2/23/2008	Organic	Chrysene-d12	n/a	=	88	%	EPA 625m		70	130	
2007/08-3	ME-VR2	srgt environ, rec	2/23/2008	Organic	Chrysene-d12	n/a	=	83	%	EPA 625m		70	130	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Dibenz(a,h)anthracene	n/a	=	82	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Dibenz(a,h)anthracene	n/a	=	85	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Dibenz(a,h)anthracene	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Dibenz(a,h)anthracene	n/a	=	0.0232	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Dibenz(a,h)anthracene	n/a	=	123	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Dibenz(a,h)anthracene	n/a	=	115	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Dibenz(a,h)anthracene	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Dibenz(a,h)anthracene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Dibenzothiophene	n/a	=	96	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Dibenzothiophene	n/a	=	97	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Dibenzothiophene	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Dibenzothiophene	n/a	=	107	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Dibenzothiophene	n/a	=	104	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Dibenzothiophene	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Dibenzothiophene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Diethyl phthalate	n/a	=	108	%	EPA 625m		50	150	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Diethyl phthalate	n/a	=	110	%	EPA 625m		50	150	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Diethyl phthalate	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Diethyl phthalate	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Diethyl phthalate	n/a	=	3.0094	µg/L	EPA 625m	0.1		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Diethyl phthalate	n/a	=	114	%	EPA 625m		50	150	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Diethyl phthalate	n/a	=	102	%	EPA 625m		50	150	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Diethyl phthalate	n/a	=	11	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Diethyl phthalate	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Dimethyl phthalate	n/a	=	105	%	EPA 625m		40	155	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Dimethyl phthalate	n/a	=	100	%	EPA 625m		40	155	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Dimethyl phthalate	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Dimethyl phthalate	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Dimethyl phthalate	n/a	=	0.109	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Dimethyl phthalate	n/a	=	96	%	EPA 625m		40	155	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Dimethyl phthalate	n/a	=	93	%	EPA 625m		40	155	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Dimethyl phthalate	n/a	=	33	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Dimethyl phthalate	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Di-n-butylphthalate	n/a	=	119	%	EPA 625m		65	145	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Di-n-butylphthalate	n/a	=	119	%	EPA 625m		65	145	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Di-n-butylphthalate	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Di-n-butylphthalate	n/a	<	0.075	µg/L	EPA 625m	0.075		0.075	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Di-n-butylphthalate	n/a	=	0.1169	µg/L	EPA 625m	0.075		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Di-n-butylphthalate	n/a	=	109	%	EPA 625m		65	145	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Di-n-butylphthalate	n/a	=	106	%	EPA 625m		65	145	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Di-n-butylphthalate	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Di-n-butylphthalate	n/a	<	0.075	µg/L	EPA 625m	0.075		0.075	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Di-n-octylphthalate	n/a	=	75	%	EPA 625m		50	165	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Di-n-octylphthalate	n/a	=	76	%	EPA 625m		50	165	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Di-n-octylphthalate	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Di-n-octylphthalate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Di-n-octylphthalate	n/a	=	0.1905	µg/L	EPA 625m	0.01		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Di-n-octylphthalate	n/a	=	108	%	EPA 625m		50	165	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Di-n-octylphthalate	n/a	=	103	%	EPA 625m		50	165	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Di-n-octylphthalate	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Di-n-octylphthalate	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Fluoranthene	n/a	=	88	%	EPA 625m		65	135	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Fluoranthene	n/a	=	99	%	EPA 625m		65	135	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Fluoranthene	n/a	=	12	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Fluoranthene	n/a	=	0.0839	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Fluoranthene	n/a	=	110	%	EPA 625m		65	135	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Fluoranthene	n/a	=	99	%	EPA 625m		65	135	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Fluoranthene	n/a	=	11	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Fluoranthene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Fluorene	n/a	=	99	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Fluorene	n/a	=	98	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Fluorene	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Fluorene	n/a	=	0.004	µg/L	EPA 625m	0.001		30	EST
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Fluorene	n/a	=	94	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Fluorene	n/a	=	98	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Fluorene	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Fluorene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Hexachlorobenzene	n/a	=	97	%	EPA 625m		65	135	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Hexachlorobenzene	n/a	=	95	%	EPA 625m		65	135	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Hexachlorobenzene	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Hexachlorobenzene	n/a	=	92	%	EPA 625m		65	135	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Hexachlorobenzene	n/a	=	89	%	EPA 625m		65	135	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Hexachlorobenzene	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Hexachlorobenzene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	method blank	2/23/2008	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Hexachlorobutadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	method blank	2/23/2008	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Hexachlorocyclopentadiene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	method blank	2/23/2008	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Hexachloroethane	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	100	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	97	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	0.0658	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	124	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	111	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	11	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Indeno(1,2,3-cd)pyrene	n/a	=	0.0204	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	method blank	2/23/2008	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Isophorone	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Naphthalene	n/a	=	90	%	EPA 625m		50	120	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Naphthalene	n/a	=	87	%	EPA 625m		50	120	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Naphthalene	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Naphthalene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Naphthalene	n/a	=	0.019	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Naphthalene	n/a	=	74	%	EPA 625m		50	120	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Naphthalene	n/a	=	77	%	EPA 625m		50	120	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Naphthalene	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Naphthalene	n/a	=	0.0163	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	srgt LCS dup, rec	2/23/2008	Organic	Naphthalene-d8	n/a	=	89	%	EPA 625m		40	120	
2007/08-3	Lab	srgt LCS, rec	2/23/2008	Organic	Naphthalene-d8	n/a	=	62	%	EPA 625m		40	120	
2007/08-3	Lab	srgt method blank, rec	2/23/2008	Organic	Naphthalene-d8	n/a	=	91	%	EPA 625m		40	120	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Naphthalene-d8	n/a	=	74	%	EPA 625m		40	120	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Naphthalene-d8	n/a	=	83	%	EPA 625m		40	120	
2007/08-3	ME-CC	srgt matrix spike dup, rec	2/23/2008	Organic	Naphthalene-d8	n/a	=	74	%	EPA 625m		40	120	
2007/08-3	ME-CC	srgt matrix spike, rec	2/23/2008	Organic	Naphthalene-d8	n/a	=	75	%	EPA 625m		40	120	
2007/08-3	ME-SCR	srgt environ, rec	2/23/2008	Organic	Naphthalene-d8	n/a	=	53	%	EPA 625m		40	120	
2007/08-3	ME-SCR	srgt field blank, rec	2/23/2008	Organic	Naphthalene-d8	n/a	=	77	%	EPA 625m		40	120	
2007/08-3	ME-VR2	srgt environ, rec	2/23/2008	Organic	Naphthalene-d8	n/a	=	57	%	EPA 625m		40	120	
2007/08-3	Lab	method blank	2/23/2008	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Nitrobenzene	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	method blank	2/23/2008	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	N-Nitrosodimethylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	N-Nitrosodi-N-propylamine	n/a	=	95	%	EPA 625m		55	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	N-Nitrosodi-N-propylamine	n/a	=	92	%	EPA 625m		55	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	N-Nitrosodi-N-propylamine	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	N-Nitrosodi-N-propylamine	n/a	=	91	%	EPA 625m		55	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	N-Nitrosodi-N-propylamine	n/a	=	98	%	EPA 625m		55	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	N-Nitrosodi-N-propylamine	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	N-Nitrosodi-N-propylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	method blank	2/23/2008	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	N-Nitrosodiphenylamine	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Pentachlorophenol	n/a	=	79	%	EPA 625m		10	160	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Pentachlorophenol	n/a	=	85	%	EPA 625m		10	160	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Pentachlorophenol	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Pentachlorophenol	n/a	=	82	%	EPA 625m		10	160	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Pentachlorophenol	n/a	=	79	%	EPA 625m		10	160	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Pentachlorophenol	n/a	=	2.5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Pentachlorophenol	n/a	<	0.05	µg/L	EPA 625m	0.05		0.05	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Perylene	n/a	=	88	%	EPA 625m		65	135	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Perylene	n/a	=	90	%	EPA 625m		65	135	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Perylene	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Perylene	n/a	=	0.0257	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Perylene	n/a	=	94	%	EPA 625m		65	135	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Perylene	n/a	=	92	%	EPA 625m		65	135	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Perylene	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Perylene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	srgt LCS dup, rec	2/23/2008	Organic	Perylene-d12	n/a	=	85	%	EPA 625m		60	140	
2007/08-3	Lab	srgt LCS, rec	2/23/2008	Organic	Perylene-d12	n/a	=	95	%	EPA 625m		60	140	
2007/08-3	Lab	srgt method blank, rec	2/23/2008	Organic	Perylene-d12	n/a	=	84	%	EPA 625m		60	140	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Perylene-d12	n/a	=	102	%	EPA 625m		60	140	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Perylene-d12	n/a	=	93	%	EPA 625m		60	140	
2007/08-3	ME-CC	srgt matrix spike dup, rec	2/23/2008	Organic	Perylene-d12	n/a	=	107	%	EPA 625m		60	140	
2007/08-3	ME-CC	srgt matrix spike, rec	2/23/2008	Organic	Perylene-d12	n/a	=	98	%	EPA 625m		60	140	
2007/08-3	ME-SCR	srgt environ, rec	2/23/2008	Organic	Perylene-d12	n/a	=	93	%	EPA 625m		60	140	
2007/08-3	ME-SCR	srgt field blank, rec	2/23/2008	Organic	Perylene-d12	n/a	=	97	%	EPA 625m		60	140	
2007/08-3	ME-VR2	srgt environ, rec	2/23/2008	Organic	Perylene-d12	n/a	=	98	%	EPA 625m		60	140	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Phenanthrene	n/a	=	93	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Phenanthrene	n/a	=	95	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Phenanthrene	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Phenanthrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Phenanthrene	n/a	=	0.0423	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Phenanthrene	n/a	=	96	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Phenanthrene	n/a	=	88	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Phenanthrene	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Phenanthrene	n/a	=	0.0044	µg/L	EPA 625m	0.001		0.001	EST
2007/08-3	Lab	srgt LCS dup, rec	2/23/2008	Organic	Phenanthrene-d10	n/a	=	99	%	EPA 625m		70	130	
2007/08-3	Lab	srgt LCS, rec	2/23/2008	Organic	Phenanthrene-d10	n/a	=	77	%	EPA 625m		70	130	
2007/08-3	Lab	srgt method blank, rec	2/23/2008	Organic	Phenanthrene-d10	n/a	=	101	%	EPA 625m		70	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Phenanthrene-d10	n/a	=	84	%	EPA 625m		70	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Phenanthrene-d10	n/a	=	97	%	EPA 625m		70	130	
2007/08-3	ME-CC	srgt matrix spike dup, rec	2/23/2008	Organic	Phenanthrene-d10	n/a	=	94	%	EPA 625m		70	130	
2007/08-3	ME-CC	srgt matrix spike, rec	2/23/2008	Organic	Phenanthrene-d10	n/a	=	94	%	EPA 625m		70	130	
2007/08-3	ME-SCR	srgt environ, rec	2/23/2008	Organic	Phenanthrene-d10	n/a	=	81	%	EPA 625m		70	130	
2007/08-3	ME-SCR	srgt field blank, rec	2/23/2008	Organic	Phenanthrene-d10	n/a	=	94	%	EPA 625m		70	130	
2007/08-3	ME-VR2	srgt environ, rec	2/23/2008	Organic	Phenanthrene-d10	n/a	=	68	%	EPA 625m		70	130	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Phenol	n/a	=	112	%	EPA 625m		0	115	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Phenol	n/a	=	107	%	EPA 625m		0	115	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Phenol	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Phenol	n/a	=	47	%	EPA 625m		0	115	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Phenol	n/a	=	45	%	EPA 625m		0	115	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Phenol	n/a	=	1.8	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Phenol	n/a	<	0.1	µg/L	EPA 625m	0.1		0.1	
2007/08-3	Lab	srgt LCS dup, rec	2/23/2008	Organic	Phenol-d5	n/a	=	98	%	EPA 625m		10	110	
2007/08-3	Lab	srgt LCS, rec	2/23/2008	Organic	Phenol-d5	n/a	=	100	%	EPA 625m		10	110	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	srgt method blank, rec	2/23/2008	Organic	Phenol-d5	n/a	=	107	%	EPA 625m		10	110	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Phenol-d5	n/a	=	50	%	EPA 625m		10	110	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Phenol-d5	n/a	=	33	%	EPA 625m		10	110	
2007/08-3	ME-CC	srgt matrix spike dup, rec	2/23/2008	Organic	Phenol-d5	n/a	=	57	%	EPA 625m		10	110	
2007/08-3	ME-CC	srgt matrix spike, rec	2/23/2008	Organic	Phenol-d5	n/a	=	56	%	EPA 625m		10	110	
2007/08-3	ME-SCR	srgt environ, rec	2/23/2008	Organic	Phenol-d5	n/a	=	26	%	EPA 625m		10	110	
2007/08-3	ME-SCR	srgt field blank, rec	2/23/2008	Organic	Phenol-d5	n/a	=	22	%	EPA 625m		10	110	
2007/08-3	ME-VR2	srgt environ, rec	2/23/2008	Organic	Phenol-d5	n/a	=	29	%	EPA 625m		10	110	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Organic	Pyrene	n/a	=	111	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Organic	Pyrene	n/a	=	114	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Organic	Pyrene	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Pyrene	n/a	=	0.0818	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Organic	Pyrene	n/a	=	127	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Organic	Pyrene	n/a	=	118	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Organic	Pyrene	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Pyrene	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	srgt LCS dup, rec	2/23/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	87	%	EPA 625m		40	130	
2007/08-3	Lab	srgt LCS, rec	2/23/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	71	%	EPA 625m		40	130	
2007/08-3	Lab	srgt method blank, rec	2/23/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	107	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	83	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	70	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt matrix spike dup, rec	2/23/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	82	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt matrix spike, rec	2/23/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	85	%	EPA 625m		40	130	
2007/08-3	ME-SCR	srgt environ, rec	2/23/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	75	%	EPA 625m		40	130	
2007/08-3	ME-SCR	srgt field blank, rec	2/23/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	85	%	EPA 625m		40	130	
2007/08-3	ME-VR2	srgt environ, rec	2/23/2008	Organic	Tetrachloro-m-xylene (TCMX)	n/a	=	76	%	EPA 625m		40	130	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Organic	Total Detectable PAHs	n/a	=	0.771	µg/L	EPA 625m			30	
2007/08-3	ME-SCR	field blank	2/23/2008	Organic	Total Detectable PAHs	n/a	=	0.094	µg/L	EPA 625m				
2007/08-3	Lab	method blank	2/23/2008	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	Aroclor 1016	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	Lab	method blank	2/23/2008	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	Aroclor 1221	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	Lab	method blank	2/23/2008	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	Aroclor 1232	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	Lab	method blank	2/23/2008	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	Aroclor 1242	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	Lab	method blank	2/23/2008	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	Aroclor 1248	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	Lab	method blank	2/23/2008	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	Aroclor 1254	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	Lab	method blank	2/23/2008	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	Aroclor 1260	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 003	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 003	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 003	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 008	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 008	n/a	=	107	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 008	n/a	=	6	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 008	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 008	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 008	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 008	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 008	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 008	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 018	n/a	=	96	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 018	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 018	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 018	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 018	n/a	=	97	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 018	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 018	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 028	n/a	=	94	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 028	n/a	=	93	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 028	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 028	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 028	n/a	=	108	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 028	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 028	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	srgt LCS dup, rec	2/23/2008	PCB	PCB 030	n/a	=	92	%	EPA 625m		40	130	
2007/08-3	Lab	srgt LCS, rec	2/23/2008	PCB	PCB 030	n/a	=	72	%	EPA 625m		40	130	
2007/08-3	Lab	srgt method blank, rec	2/23/2008	PCB	PCB 030	n/a	=	109	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	PCB	PCB 030	n/a	=	87	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	PCB	PCB 030	n/a	=	72	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt matrix spike dup, rec	2/23/2008	PCB	PCB 030	n/a	=	92	%	EPA 625m		40	130	
2007/08-3	ME-CC	srgt matrix spike, rec	2/23/2008	PCB	PCB 030	n/a	=	86	%	EPA 625m		40	130	
2007/08-3	ME-SCR	srgt environ, rec	2/23/2008	PCB	PCB 030	n/a	=	80	%	EPA 625m		40	130	
2007/08-3	ME-SCR	srgt field blank, rec	2/23/2008	PCB	PCB 030	n/a	=	96	%	EPA 625m		40	130	
2007/08-3	ME-VR2	srgt environ, rec	2/23/2008	PCB	PCB 030	n/a	=	75	%	EPA 625m		40	130	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 031	n/a	=	109	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 031	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 031	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 031	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 031	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 031	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 031	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 033	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 033	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 033	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 033	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 033	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 033	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 033	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 037	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 037	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 037	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 037	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 037	n/a	=	108	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 037	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 037	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 044	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 044	n/a	=	109	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 044	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 044	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 044	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 044	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 044	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 049	n/a	=	107	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 049	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 049	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 049	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 049	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 049	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 049	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 052	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 052	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 052	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 052	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 052	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 052	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 052	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 056/060	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 056/060	n/a	=	97	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 056/060	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 056/060	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 056/060	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 056/060	n/a	=	107	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 056/060	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 056/060	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 056/060	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 066	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 066	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 066	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 066	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 066	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 066	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 066	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 070	n/a	=	92	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 070	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 070	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 070	n/a	=	99	%	EPA 625m		60	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 070	n/a	=	96	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 070	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 070	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 074	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 074	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 074	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 074	n/a	=	94	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 074	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 074	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 074	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 077	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 077	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 077	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 077	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 077	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 077	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 077	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 081	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 081	n/a	=	107	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 081	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 081	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 081	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 081	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 081	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 087	n/a	=	97	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 087	n/a	=	107	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 087	n/a	=	10	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 087	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 087	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 087	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 087	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 095	n/a	=	95	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 095	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 095	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 095	n/a	=	91	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 095	n/a	=	96	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 095	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 095	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 097	n/a	=	107	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 097	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 097	n/a	=	8	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 097	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 097	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 097	n/a	=	5	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 097	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 099	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 099	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 099	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 099	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 099	n/a	=	97	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 099	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 099	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 101	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 101	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 101	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 101	n/a	=	97	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 101	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 101	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 101	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 105	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 105	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 105	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 105	n/a	=	109	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 105	n/a	=	97	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 105	n/a	=	12	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 105	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 110	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 110	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 110	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 110	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 110	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 110	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 110	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	srgt LCS dup, rec	2/23/2008	PCB	PCB 112	n/a	=	92	%	EPA 625m		60	120	
2007/08-3	Lab	srgt LCS, rec	2/23/2008	PCB	PCB 112	n/a	=	92	%	EPA 625m		60	120	
2007/08-3	Lab	srgt method blank, rec	2/23/2008	PCB	PCB 112	n/a	=	107	%	EPA 625m		60	120	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	PCB	PCB 112	n/a	=	90	%	EPA 625m		60	120	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	PCB	PCB 112	n/a	=	80	%	EPA 625m		60	120	
2007/08-3	ME-CC	srgt matrix spike dup, rec	2/23/2008	PCB	PCB 112	n/a	=	98	%	EPA 625m		60	120	
2007/08-3	ME-CC	srgt matrix spike, rec	2/23/2008	PCB	PCB 112	n/a	=	93	%	EPA 625m		60	120	
2007/08-3	ME-SCR	srgt environ, rec	2/23/2008	PCB	PCB 112	n/a	=	87	%	EPA 625m		60	120	
2007/08-3	ME-SCR	srgt field blank, rec	2/23/2008	PCB	PCB 112	n/a	=	109	%	EPA 625m		60	120	
2007/08-3	ME-VR2	srgt environ, rec	2/23/2008	PCB	PCB 112	n/a	=	79	%	EPA 625m		60	120	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 114	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 114	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 114	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 114	n/a	=	109	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 114	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 114	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 114	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 118	n/a	=	96	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 118	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 118	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 118	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 118	n/a	=	96	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 118	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 118	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 119	n/a	=	95	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 119	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 119	n/a	=	10	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 119	n/a	=	96	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 119	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 119	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 119	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 123	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 123	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 123	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 123	n/a	=	95	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 123	n/a	=	97	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 123	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 123	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 126	n/a	=	96	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 126	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 126	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 126	n/a	=	108	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 126	n/a	=	108	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 126	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 126	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 128	n/a	=	95	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 128	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 128	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 128	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 128	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 128	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 128	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 128	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 128	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 138	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 138	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 138	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 138	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 138	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 138	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 138	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 141	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 141	n/a	=	108	%	EPA 625m		60	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 141	n/a	=	8	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 141	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 141	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 141	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 141	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 149	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 149	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 149	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 149	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 149	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 149	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 149	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 151	n/a	=	107	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 151	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 151	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 151	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 151	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 151	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 151	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 153	n/a	=	107	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 153	n/a	=	109	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 153	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 153	n/a	=	92	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 153	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 153	n/a	=	8	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 153	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 156	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 156	n/a	=	108	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 156	n/a	=	8	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 156	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 156	n/a	=	108	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 156	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 156	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 157	n/a	=	96	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 157	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 157	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 157	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 157	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 157	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 157	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 158	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 158	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 158	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 158	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 158	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 158	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 158	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 167	n/a	=	93	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 167	n/a	=	94	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 167	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 167	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 167	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 167	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 167	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 167	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 168 + 132	n/a	=	97	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 168 + 132	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 168 + 132	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 168 + 132	n/a	=	97	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 168 + 132	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 168 + 132	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 168 + 132	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 169	n/a	=	96	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 169	n/a	=	111	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 169	n/a	=	14	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 169	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 169	n/a	=	111	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 169	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 169	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 170	n/a	=	96	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 170	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 170	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 170	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 170	n/a	=	107	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 170	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 170	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 174	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 174	n/a	=	116	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 174	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 174	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 174	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 174	n/a	=	110	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 174	n/a	=	114	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 174	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 174	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 177	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 177	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 177	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 177	n/a	=	101	%	EPA 625m		60	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 177	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 177	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 177	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 180	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 180	n/a	=	111	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 180	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 180	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 180	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 180	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 180	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 183	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 183	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 183	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 183	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 183	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 183	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 183	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 187	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 187	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 187	n/a	=	8	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 187	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 187	n/a	=	109	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 187	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 187	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 189	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 189	n/a	=	95	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 189	n/a	=	10	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 189	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 189	n/a	=	112	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 189	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 189	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 194	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 194	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 194	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 194	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 194	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 194	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 194	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 195	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 195	n/a	=	109	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 195	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 195	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 195	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 195	n/a	=	108	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 195	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 195	n/a	=	8	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 195	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	srgt LCS dup, rec	2/23/2008	PCB	PCB 198	n/a	=	98	%	EPA 625m		60	120	
2007/08-3	Lab	srgt LCS, rec	2/23/2008	PCB	PCB 198	n/a	=	99	%	EPA 625m		60	120	
2007/08-3	Lab	srgt method blank, rec	2/23/2008	PCB	PCB 198	n/a	=	107	%	EPA 625m		60	120	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	PCB	PCB 198	n/a	=	81	%	EPA 625m		60	120	
2007/08-3	ME-CC	srgt environ, rec	2/23/2008	PCB	PCB 198	n/a	=	89	%	EPA 625m		60	120	
2007/08-3	ME-CC	srgt matrix spike dup, rec	2/23/2008	PCB	PCB 198	n/a	=	100	%	EPA 625m		60	120	
2007/08-3	ME-CC	srgt matrix spike, rec	2/23/2008	PCB	PCB 198	n/a	=	93	%	EPA 625m		60	120	
2007/08-3	ME-SCR	srgt environ, rec	2/23/2008	PCB	PCB 198	n/a	=	97	%	EPA 625m		60	120	
2007/08-3	ME-SCR	srgt field blank, rec	2/23/2008	PCB	PCB 198	n/a	=	105	%	EPA 625m		60	120	
2007/08-3	ME-VR2	srgt environ, rec	2/23/2008	PCB	PCB 198	n/a	=	79	%	EPA 625m		60	120	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 200	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 200	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 200	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 200	n/a	=	94	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 200	n/a	=	100	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 200	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 200	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 201	n/a	=	109	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 201	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 201	n/a	=	10	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 201	n/a	=	97	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 201	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 201	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 201	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 206	n/a	=	107	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 206	n/a	=	110	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 206	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 206	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 206	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 206	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 206	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	PCB	PCB 209	n/a	=	109	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	PCB	PCB 209	n/a	=	102	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	PCB	PCB 209	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	PCB	PCB 209	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	PCB 209	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	PCB	PCB 209	n/a	=	104	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	PCB	PCB 209	n/a	=	95	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	PCB	PCB 209	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	PCB 209	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m			30	
2007/08-3	ME-SCR	field blank	2/23/2008	PCB	Total Detectable PCBs	n/a	=	0	µg/L	EPA 625m				
2007/08-3	Lab	method blank	1/31/2007	Pesticide	2,4,5-T	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2007/08-3	Lab	LCS dup, rec	1/31/2008	Pesticide	2,4,5-T	n/a	=	96	%	EPA 8151A		30	130	
2007/08-3	Lab	LCS, rec	1/31/2008	Pesticide	2,4,5-T	n/a	=	103	%	EPA 8151A		30	130	
2007/08-3	Lab	LCS, RPD	1/31/2008	Pesticide	2,4,5-T	n/a	=	7	%	EPA 8151A		0	30	
2007/08-3	ME-CC	matrix spike dup, rec	1/31/2008	Pesticide	2,4,5-T	n/a	=	120	%	EPA 8151A		30	130	
2007/08-3	ME-CC	matrix spike, rec	1/31/2008	Pesticide	2,4,5-T	n/a	=	111	%	EPA 8151A		30	130	
2007/08-3	ME-CC	matrix spike, RPD	1/31/2008	Pesticide	2,4,5-T	n/a	=	8	%	EPA 8151A		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	method blank	1/31/2007	Pesticide	2,4,5-TP (Silvex)	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2007/08-3	Lab	method blank	1/31/2007	Pesticide	2,4-D	n/a	<	5	µg/L	EPA 8151A	5		5	
2007/08-3	Lab	LCS dup, rec	1/31/2008	Pesticide	2,4-D	n/a	=	74	%	EPA 8151A		30	130	
2007/08-3	Lab	LCS, rec	1/31/2008	Pesticide	2,4-D	n/a	=	80	%	EPA 8151A		30	130	
2007/08-3	Lab	LCS, RPD	1/31/2008	Pesticide	2,4-D	n/a	=	7	%	EPA 8151A		0	30	
2007/08-3	ME-CC	matrix spike dup, rec	1/31/2008	Pesticide	2,4-D	n/a	=	120	%	EPA 8151A		30	130	
2007/08-3	ME-CC	matrix spike, rec	1/31/2008	Pesticide	2,4-D	n/a	=	109	%	EPA 8151A		30	130	
2007/08-3	ME-CC	matrix spike, RPD	1/31/2008	Pesticide	2,4-D	n/a	=	10	%	EPA 8151A		0	30	
2007/08-3	Lab	method blank	1/31/2007	Pesticide	2,4-DB	n/a	<	5	µg/L	EPA 8151A	5		5	
2007/08-3	Lab	LCS dup, rec	1/31/2008	Pesticide	2,4-DB	n/a	=	84	%	EPA 8151A		30	130	
2007/08-3	Lab	LCS, rec	1/31/2008	Pesticide	2,4-DB	n/a	=	89	%	EPA 8151A		30	130	
2007/08-3	Lab	LCS, RPD	1/31/2008	Pesticide	2,4-DB	n/a	=	7	%	EPA 8151A		0	30	
2007/08-3	ME-CC	matrix spike dup, rec	1/31/2008	Pesticide	2,4-DB	n/a	=	94	%	EPA 8151A		30	130	
2007/08-3	ME-CC	matrix spike, rec	1/31/2008	Pesticide	2,4-DB	n/a	=	89	%	EPA 8151A		30	130	
2007/08-3	ME-CC	matrix spike, RPD	1/31/2008	Pesticide	2,4-DB	n/a	=	6	%	EPA 8151A		0	30	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	2,4'-DDD	n/a	=	110	%	EPA 625m		50	140	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	2,4'-DDD	n/a	=	114	%	EPA 625m		50	140	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	2,4'-DDD	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	2,4'-DDD	n/a	=	0.0088	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	2,4'-DDD	n/a	=	113	%	EPA 625m		50	140	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	2,4'-DDD	n/a	=	105	%	EPA 625m		50	140	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	2,4'-DDD	n/a	=	8	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	2,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	2,4'-DDE	n/a	=	107	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	2,4'-DDE	n/a	=	108	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	2,4'-DDE	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	2,4'-DDE	n/a	=	103	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	2,4'-DDE	n/a	=	109	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	2,4'-DDE	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	2,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	2,4'-DDT	n/a	=	87	%	EPA 625m		40	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	2,4'-DDT	n/a	=	90	%	EPA 625m		40	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	2,4'-DDT	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	2,4'-DDT	n/a	=	0.0081	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	2,4'-DDT	n/a	=	52	%	EPA 625m		40	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	2,4'-DDT	n/a	=	49	%	EPA 625m		40	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	2,4'-DDT	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	2,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	4,4'-DDD	n/a	=	105	%	EPA 625m		60	140	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	4,4'-DDD	n/a	=	104	%	EPA 625m		60	140	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	4,4'-DDD	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	4,4'-DDD	n/a	=	0.0264	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	4,4'-DDD	n/a	=	125	%	EPA 625m		60	140	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	4,4'-DDD	n/a	=	131	%	EPA 625m		60	140	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	4,4'-DDD	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	4,4'-DDD	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	4,4'-DDE	n/a	=	105	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	4,4'-DDE	n/a	=	108	%	EPA 625m		70	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	4,4'-DDE	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	4,4'-DDE	n/a	=	0.2797	µg/L	EPA 625m	0.001		30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	4,4'-DDE	n/a	=	100	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	4,4'-DDE	n/a	=	98	%	EPA 625m		70	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	4,4'-DDE	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	4,4'-DDE	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	4,4'-DDT	n/a	=	87	%	EPA 625m		0	150	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	4,4'-DDT	n/a	=	85	%	EPA 625m		0	150	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	4,4'-DDT	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	4,4'-DDT	n/a	=	0.053	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	4,4'-DDT	n/a	=	42	%	EPA 625m		0	150	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	4,4'-DDT	n/a	=	40	%	EPA 625m		0	150	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	4,4'-DDT	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	4,4'-DDT	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Aldrin	n/a	=	103	%	EPA 625m		50	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Aldrin	n/a	=	103	%	EPA 625m		50	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Aldrin	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Aldrin	n/a	=	108	%	EPA 625m		50	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Aldrin	n/a	=	100	%	EPA 625m		50	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Aldrin	n/a	=	8	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Aldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	BHC-alpha	n/a	=	102	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	BHC-alpha	n/a	=	99	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	BHC-alpha	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	BHC-alpha	n/a	=	94	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	BHC-alpha	n/a	=	101	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	BHC-alpha	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	BHC-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	BHC-beta	n/a	=	102	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	BHC-beta	n/a	=	90	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	BHC-beta	n/a	=	12	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	BHC-beta	n/a	=	97	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	BHC-beta	n/a	=	98	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	BHC-beta	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	BHC-beta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	BHC-delta	n/a	=	104	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	BHC-delta	n/a	=	102	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	BHC-delta	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	BHC-delta	n/a	=	106	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	BHC-delta	n/a	=	101	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	BHC-delta	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	BHC-delta	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	98	%	EPA 625m		50	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	105	%	EPA 625m		50	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	91	%	EPA 625m		50	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	94	%	EPA 625m		50	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	BHC-gamma (Lindane)	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	BHC-gamma (Lindane)	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Bolstar	n/a	=	80	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Bolstar	n/a	=	79	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Bolstar	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Bolstar	n/a	=	109	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Bolstar	n/a	=	101	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Bolstar	n/a	=	8	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Bolstar	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Chlordane-alpha	n/a	=	109	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Chlordane-alpha	n/a	=	103	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Chlordane-alpha	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Chlordane-alpha	n/a	=	0.0095	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Chlordane-alpha	n/a	=	104	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Chlordane-alpha	n/a	=	102	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Chlordane-alpha	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Chlordane-alpha	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Chlordane-gamma	n/a	=	103	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Chlordane-gamma	n/a	=	106	%	EPA 625m		60	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Chlordane-gamma	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Chlordane-gamma	n/a	=	0.0074	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Chlordane-gamma	n/a	=	112	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Chlordane-gamma	n/a	=	102	%	EPA 625m		60	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Chlordane-gamma	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Chlordane-gamma	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Chlorpyrifos	n/a	=	86	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Chlorpyrifos	n/a	=	99	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Chlorpyrifos	n/a	=	14	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Chlorpyrifos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Chlorpyrifos	n/a	=	0.0995	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Chlorpyrifos	n/a	=	102	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Chlorpyrifos	n/a	=	101	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Chlorpyrifos	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Chlorpyrifos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	cis-Nonachlor	n/a	=	94	%	EPA 625m		60	120	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	cis-Nonachlor	n/a	=	94	%	EPA 625m		60	120	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	cis-Nonachlor	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	cis-Nonachlor	n/a	=	0.005	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	cis-Nonachlor	n/a	=	91	%	EPA 625m		60	120	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	cis-Nonachlor	n/a	=	88	%	EPA 625m		60	120	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	cis-Nonachlor	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	cis-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	method blank	1/31/2007	Pesticide	Dalapon	n/a	<	13	µg/L	EPA 8151A	13		13	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Demeton-O	n/a	=	79	%	EPA 625m		45	105	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Demeton-O	n/a	=	82	%	EPA 625m		45	105	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Demeton-O	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Demeton-O	n/a	=	61	%	EPA 625m		45	105	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Demeton-O	n/a	=	63	%	EPA 625m		45	105	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Demeton-O	n/a	=	3	%	EPA 625m		0	30	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Demeton-O	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Diazinon	n/a	=	86	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Diazinon	n/a	=	80	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Diazinon	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Diazinon	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Diazinon	n/a	=	0.0374	µg/L	EPA 625m	0.002		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Diazinon	n/a	=	95	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Diazinon	n/a	=	95	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Diazinon	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Diazinon	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-3	Lab	method blank	1/31/2007	Pesticide	Dicamba	n/a	<	0.5	µg/L	EPA 8151A	0.5		0.5	
2007/08-3	Lab	method blank	1/31/2007	Pesticide	Dichlorprop	n/a	<	5	µg/L	EPA 8151A	5		5	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Dichlorvos	n/a	=	78	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Dichlorvos	n/a	=	78	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Dichlorvos	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Dichlorvos	n/a	=	86	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Dichlorvos	n/a	=	92	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Dichlorvos	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Dichlorvos	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Dieldrin	n/a	=	106	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Dieldrin	n/a	=	107	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Dieldrin	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Dieldrin	n/a	=	93	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Dieldrin	n/a	=	104	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Dieldrin	n/a	=	11	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Dieldrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Dimethoate	n/a	=	80	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Dimethoate	n/a	=	85	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Dimethoate	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Dimethoate	n/a	=	104	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Dimethoate	n/a	=	94	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Dimethoate	n/a	=	10	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Dimethoate	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-3	Lab	method blank	1/31/2007	Pesticide	Dinoseb	n/a	<	2.5	µg/L	EPA 8151A	2.5		2.5	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Disulfoton	n/a	=	80	%	EPA 625m		45	105	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Disulfoton	n/a	=	84	%	EPA 625m		45	105	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Disulfoton	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Disulfoton	n/a	=	52	%	EPA 625m		45	105	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Disulfoton	n/a	=	45	%	EPA 625m		45	105	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Disulfoton	n/a	=	14	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Disulfoton	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Endosulfan sulfate	n/a	=	108	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Endosulfan sulfate	n/a	=	106	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Endosulfan sulfate	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Endosulfan sulfate	n/a	=	99	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Endosulfan sulfate	n/a	=	102	%	EPA 625m		60	125	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Endosulfan sulfate	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Endosulfan sulfate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Endosulfan-I	n/a	=	98	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Endosulfan-I	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Endosulfan-I	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Endosulfan-I	n/a	=	96	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Endosulfan-I	n/a	=	101	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Endosulfan-I	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Endosulfan-I	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Endosulfan-II	n/a	=	103	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Endosulfan-II	n/a	=	105	%	EPA 625m		60	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Endosulfan-II	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Endosulfan-II	n/a	=	91	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Endosulfan-II	n/a	=	82	%	EPA 625m		60	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Endosulfan-II	n/a	=	10	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Endosulfan-II	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Endrin	n/a	=	108	%	EPA 625m		65	135	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Endrin	n/a	=	102	%	EPA 625m		65	135	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Endrin	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Endrin	n/a	=	108	%	EPA 625m		65	135	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Endrin	n/a	=	103	%	EPA 625m		65	135	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Endrin	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Endrin	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Endrin aldehyde	n/a	=	95	%	EPA 625m		0	149	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Endrin aldehyde	n/a	=	94	%	EPA 625m		0	149	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Endrin aldehyde	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Endrin aldehyde	n/a	=	87	%	EPA 625m		0	149	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Endrin aldehyde	n/a	=	75	%	EPA 625m		0	149	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Endrin aldehyde	n/a	=	15	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Endrin aldehyde	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Endrin ketone	n/a	=	97	%	EPA 625m		40	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Endrin ketone	n/a	=	97	%	EPA 625m		40	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Endrin ketone	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Endrin ketone	n/a	=	88	%	EPA 625m		40	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Endrin ketone	n/a	=	98	%	EPA 625m		40	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Endrin ketone	n/a	=	11	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Endrin ketone	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Ethoprop	n/a	=	89	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Ethoprop	n/a	=	77	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Ethoprop	n/a	=	14	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Ethoprop	n/a	=	102	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Ethoprop	n/a	=	97	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Ethoprop	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Ethoprop	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Fenclorophos (Ronnel)	n/a	=	88	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Fenclorophos (Ronnel)	n/a	=	90	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Fenclorophos (Ronnel)	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Fenclorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Fenclorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Fenclorophos (Ronnel)	n/a	=	108	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Fenclorophos (Ronnel)	n/a	=	103	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Fenclorophos (Ronnel)	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Fenclorophos (Ronnel)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Fensulfothion	n/a	=	92	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Fensulfothion	n/a	=	104	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Fensulfothion	n/a	=	12	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Fensulfothion	n/a	=	110	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Fensulfothion	n/a	=	111	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Fensulfothion	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Fensulfothion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Fenthion	n/a	=	80	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Fenthion	n/a	=	79	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Fenthion	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Fenthion	n/a	=	94	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Fenthion	n/a	=	99	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Fenthion	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Fenthion	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-3	Lab	LCS, rec	2/4/2008	Pesticide	Glyphosate	n/a	=	91	%	EPA 547		71	137	
2007/08-3	Lab	method blank	2/4/2008	Pesticide	Glyphosate	n/a	<	5	µg/L	EPA 547	5		5	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Heptachlor	n/a	=	107	%	EPA 625m		45	135	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Heptachlor	n/a	=	109	%	EPA 625m		45	135	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Heptachlor	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Heptachlor	n/a	=	105	%	EPA 625m		45	135	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Heptachlor	n/a	=	105	%	EPA 625m		45	135	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Heptachlor	n/a	=	0	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Heptachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Heptachlor epoxide	n/a	=	102	%	EPA 625m		65	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Heptachlor epoxide	n/a	=	103	%	EPA 625m		65	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Heptachlor epoxide	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Heptachlor epoxide	n/a	=	112	%	EPA 625m		65	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Heptachlor epoxide	n/a	=	104	%	EPA 625m		65	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Heptachlor epoxide	n/a	=	7	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Heptachlor epoxide	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Malathion	n/a	=	78	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Malathion	n/a	=	82	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Malathion	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Malathion	n/a	=	116	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Malathion	n/a	=	110	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Malathion	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Malathion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	method blank	1/31/2007	Pesticide	MCPA	n/a	<	500	µg/L	EPA 8151A	500		500	
2007/08-3	Lab	method blank	1/31/2007	Pesticide	MCPP	n/a	<	500	µg/L	EPA 8151A	500		500	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Merphos	n/a	=	90	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Merphos	n/a	=	85	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Merphos	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Merphos	n/a	=	105	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Merphos	n/a	=	108	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Merphos	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Merphos	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Methoxychlor	n/a	=	88	%	EPA 625m		0	155	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Methoxychlor	n/a	=	96	%	EPA 625m		0	155	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Methoxychlor	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Methoxychlor	n/a	=	47	%	EPA 625m		0	155	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Methoxychlor	n/a	=	46	%	EPA 625m		0	155	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Methoxychlor	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Methoxychlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Methyl parathion	n/a	=	95	%	EPA 625m		60	120	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Methyl parathion	n/a	=	97	%	EPA 625m		60	120	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Methyl parathion	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Methyl parathion	n/a	=	106	%	EPA 625m		60	120	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Methyl parathion	n/a	=	109	%	EPA 625m		60	120	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Methyl parathion	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Methyl parathion	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Mevinphos	n/a	=	75	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Mevinphos	n/a	=	81	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Mevinphos	n/a	=	8	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		0.008	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Mevinphos	n/a	=	102	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Mevinphos	n/a	=	98	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Mevinphos	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Mevinphos	n/a	<	0.008	µg/L	EPA 625m	0.008		0.008	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Mirex	n/a	=	99	%	EPA 625m		50	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Mirex	n/a	=	101	%	EPA 625m		50	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Mirex	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Mirex	n/a	=	88	%	EPA 625m		50	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Mirex	n/a	=	83	%	EPA 625m		50	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Mirex	n/a	=	6	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Mirex	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Oxychlorthane	n/a	=	102	%	EPA 625m		50	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Oxychlorthane	n/a	=	98	%	EPA 625m		50	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Oxychlorthane	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Oxychlorthane	n/a	=	99	%	EPA 625m		50	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Oxychlorthane	n/a	=	102	%	EPA 625m		50	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Oxychlorthane	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Oxychlorthane	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	

Appendix G
2007/08 QA/QC Analysis Results

Event ID	Site ID	QA/QC Sample Type	Analysis Date	Classification	Constituent	Fraction	Sign	Result	Units	Method	DL	QA Limit Min	QA Limit Max	DQO Compliance
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Phorate	n/a	=	105	%	EPA 625m		45	105	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Phorate	n/a	=	102	%	EPA 625m		45	105	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Phorate	n/a	=	3	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006		0.006	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Phorate	n/a	=	90	%	EPA 625m		45	105	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Phorate	n/a	=	104	%	EPA 625m		45	105	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Phorate	n/a	=	14	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Phorate	n/a	<	0.006	µg/L	EPA 625m	0.006		0.006	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	93	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	92	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	119	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	108	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	=	10	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Tetrachlorovinphos (Stirofos)	n/a	<	0.002	µg/L	EPA 625m	0.002		0.002	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Tokuthion	n/a	=	83	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Tokuthion	n/a	=	80	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Tokuthion	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Tokuthion	n/a	=	114	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Tokuthion	n/a	=	104	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Tokuthion	n/a	=	9	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Tokuthion	n/a	<	0.003	µg/L	EPA 625m	0.003		0.003	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Total Detectable DDTs	n/a	=	0.376	µg/L	EPA 625m			30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Total Detectable DDTs	n/a	=	0	µg/L	EPA 625m				
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Toxaphene	n/a	<	0.01	µg/L	EPA 625m	0.01		0.01	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	trans-Nonachlor	n/a	=	106	%	EPA 625m		55	130	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	trans-Nonachlor	n/a	=	108	%	EPA 625m		55	130	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	trans-Nonachlor	n/a	=	2	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	trans-Nonachlor	n/a	=	0.0052	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	trans-Nonachlor	n/a	=	108	%	EPA 625m		55	130	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	trans-Nonachlor	n/a	=	112	%	EPA 625m		55	130	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	trans-Nonachlor	n/a	=	4	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	trans-Nonachlor	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	Lab	LCS dup, rec	2/23/2008	Pesticide	Trichloronate	n/a	=	83	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, rec	2/23/2008	Pesticide	Trichloronate	n/a	=	79	%	EPA 625m		65	125	
2007/08-3	Lab	LCS, RPD	2/23/2008	Pesticide	Trichloronate	n/a	=	5	%	EPA 625m		0	30	
2007/08-3	Lab	method blank	2/23/2008	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	
2007/08-3	ME-CC	lab duplicate	2/23/2008	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		30	
2007/08-3	ME-CC	matrix spike dup, rec	2/23/2008	Pesticide	Trichloronate	n/a	=	101	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, rec	2/23/2008	Pesticide	Trichloronate	n/a	=	100	%	EPA 625m		65	125	
2007/08-3	ME-CC	matrix spike, RPD	2/23/2008	Pesticide	Trichloronate	n/a	=	1	%	EPA 625m		0	30	
2007/08-3	ME-SCR	field blank	2/23/2008	Pesticide	Trichloronate	n/a	<	0.001	µg/L	EPA 625m	0.001		0.001	